

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091922\
 Data File : BM036579.D
 Acq On : 19 Sep 2022 14:30
 Operator : CG/JU
 Sample : N4604-13
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Sep 19 22:31:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 02:39:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.028	152	328698	20.000	ng/u1	0.00
20) Naphthalene-d8	10.839	136	1445646	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	980146	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	2048482	20.000	ng/u1	0.00
79) Chrysene-d12	21.550	240	1930582	20.000	ng/u1	0.00
88) Perylene-d12	23.991	264	1844675	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	31801	3.627	ng/uL	0.00
4) Pyridine-d5	3.810	84	390	0.015	ng/u1	0.00
7) Phenol-d5	7.175	99	555044	18.896	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.351	67	358859	20.117	ng/u1	0.00
11) 2-Chlorophenol-d4	7.551	132	433540	20.500	ng/u1	0.00
15) 4-Methylphenol-d8	8.728	113	441403	20.337	ng/u1	0.00
21) Nitrobenzene-d5	9.192	128	207941	21.251	ng/u1	0.00
24) 2-Nitrophenol-d4	9.916	143	190145	21.695	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.451	165	401102	19.747	ng/u1	0.00
31) 4-Chloroaniline-d4	10.969	131	389488	11.533	ng/u1	0.00
46) Dimethylphthalate-d6	14.063	166	1438631	21.763	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	1643402	20.958	ng/u1	0.00
54) 4-Nitrophenol-d4	14.821	143	200816	15.662	ng/u1	0.00
60) Fluorene-d10	15.639	176	1289513	21.767	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	113136	13.959	ng/u1	0.00
73) Anthracene-d10	17.486	188	1973396	21.069	ng/u1	0.00
81) Pyrene-d10	19.762	212	2216928	23.342	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.833	264	1942633	21.555	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091922\
Data File : BM036579.D
Acq On : 19 Sep 2022 14:30
Operator : CG/JU
Sample : N4604-13
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Sep 19 22:31:20 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Sep 16 02:39:54 2022
Response via : Initial Calibration

