

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091922\
 Data File : BM036581.D
 Acq On : 19 Sep 2022 15:44
 Operator : CG/JU
 Sample : N4604-12
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Sep 19 22:31:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 02:39:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.028	152	617519	20.000	ng/u1	0.00
20) Naphthalene-d8	10.839	136	2768365	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	1825110	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	3707579	20.000	ng/u1	0.00
79) Chrysene-d12	21.551	240	3550797	20.000	ng/u1	0.00
88) Perylene-d12	23.997	264	3309693	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	51035	3.098	ng/uL	0.00
4) Pyridine-d5	3.811	84	71	0.001	ng/u1	0.00
7) Phenol-d5	7.175	99	949379	17.204	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.351	67	621817	18.554	ng/u1	0.00
11) 2-Chlorophenol-d4	7.551	132	746621	18.792	ng/u1	0.00
15) 4-Methylphenol-d8	8.728	113	668277	16.389	ng/u1	0.00
21) Nitrobenzene-d5	9.193	128	372120	19.859	ng/u1	0.00
24) 2-Nitrophenol-d4	9.916	143	351391	20.936	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.451	165	699500	17.984	ng/u1	0.00
31) 4-Chloroaniline-d4	10.969	131	682603	10.555	ng/u1	0.00
46) Dimethylphthalate-d6	14.063	166	2477973	20.131	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	2893556	19.818	ng/u1	0.00
54) 4-Nitrophenol-d4	14.827	143	370205	15.506	ng/u1	0.00
60) Fluorene-d10	15.639	176	2236804	20.277	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	239944	16.357	ng/u1	0.00
73) Anthracene-d10	17.486	188	3318143	19.574	ng/u1	0.00
81) Pyrene-d10	19.768	212	3993561	22.861	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.839	264	3435150	21.244	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091922\
Data File : BM036581.D
Acq On : 19 Sep 2022 15:44
Operator : CG/JU
Sample : N4604-12
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Sep 19 22:31:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Sep 16 02:39:54 2022
Response via : Initial Calibration

