

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091922\
 Data File : BM036597.D
 Acq On : 20 Sep 2022 01:24
 Operator : CG/JU
 Sample : N4576-02
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 CB8G3

Quant Time: Sep 20 02:02:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Sep 16 02:39:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.028	152	468617	20.000	ng/u1	0.00
20) Naphthalene-d8	10.839	136	2104145	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	1350040	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	2697860	20.000	ng/u1	0.00
79) Chrysene-d12	21.550	240	2496904	20.000	ng/u1	0.00
88) Perylene-d12	23.997	264	2158151	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	45046	3.604	ng/uL	0.00
4) Pyridine-d5	3.822	84	184032	5.048	ng/u1	0.00
7) Phenol-d5	7.175	99	198420	4.738	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.351	67	614907	24.178	ng/u1	0.00
11) 2-Chlorophenol-d4	7.551	132	612149	20.303	ng/u1	0.00
15) 4-Methylphenol-d8	8.728	113	381798	12.339	ng/u1	0.00
21) Nitrobenzene-d5	9.192	128	411238	28.875	ng/u1	0.00
24) 2-Nitrophenol-d4	9.916	143	390603	30.619	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.451	165	669304	22.639	ng/u1	0.00
31) 4-Chloroaniline-d4	10.969	131	872891	17.759	ng/u1	0.00
46) Dimethylphthalate-d6	14.062	166	2374722	26.081	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	2833070	26.231	ng/u1	0.00
54) 4-Nitrophenol-d4	14.827	143	67332	3.813	ng/u1	0.00
60) Fluorene-d10	15.639	176	2117048	25.945	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	216534	20.286	ng/u1	0.00
73) Anthracene-d10	17.486	188	3409477	27.640	ng/u1	0.00
81) Pyrene-d10	19.768	212	3844908	31.301	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.838	264	3042119	28.852	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091922\
Data File : BM036597.D
Acq On : 20 Sep 2022 01:24
Operator : CG/JU
Sample : N4576-02
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
CB8G3

Quant Time: Sep 20 02:02:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Sep 16 02:39:54 2022
Response via : Initial Calibration

