


Data Path : Z:\HPCHEM1\BNA M\DATA\BM092116\
 Data File : BM007497.D
 Acq On : 21 Sep 2016 18:24
 Operator : UM/SJ
 Sample : DFTPP62
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP62

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092116.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Sep 22 01:26:49 2016



AutoFind: Scans 2551, 2552, 2553; Background Corrected with Scan 2545

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.1	72069	PASS
68	69	0.00	2	1.9	1539	PASS
69	198	0.00	100	34.0	81296	PASS
70	69	0.00	2	0.8	660	PASS
127	198	10	80	41.6	99698	PASS
197	198	0.00	2	0.3	795	PASS
198	198	100	100	100.0	239402	PASS
199	198	5	9	6.8	16265	PASS
275	198	10	60	29.3	70045	PASS
365	198	1	100	4.3	10398	PASS
441	442	0.01	100	14.7	28848	PASS
442	198	50	100	82.0	196288	PASS
443	442	15	24	19.3	37957	PASS

Scan 2545 (17.557 min): BM007497.D
DFTPP62

m/z	Abundance
39.90	746.0
44.00	875.0
73.10	464.0
207.00	1383.0
281.00	379.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

m/z	Abundance
37.00	473.0
38.10	1060.0
39.10	5200.0
40.00	581.0
40.90	308.0
44.00	837.0
49.10	573.0
50.00	18536.0
51.10	72272.0
52.00	3724.0
53.00	313.0
55.10	558.0
56.00	2390.0
57.00	4965.0
61.00	1084.0
62.00	1394.0
63.10	3473.0
64.00	437.0
65.00	1832.0
68.10	1742.0
69.00	81112.0
70.00	670.0
73.00	1239.0
74.10	10115.0
75.00	14863.0
76.10	4815.0
77.10	89544.0
78.10	6064.0
79.00	7193.0
80.00	5740.0
81.00	7784.0
82.00	1499.0
83.00	1622.0
85.00	1386.0
86.00	2436.0
87.00	1195.0
88.00	357.0
91.00	1656.0
92.00	2157.0
93.00	12861.0
94.00	975.0
96.00	528.0
97.10	456.0
98.00	10243.0
99.00	8944.0
99.90	911.0
101.00	3583.0
102.90	1292.0
104.00	2387.0
105.00	1819.0
106.00	802.0
107.00	31816.0
108.00	4722.0
109.20	694.0
110.00	48688.0
111.00	8744.0
112.00	1002.0
113.00	404.0
116.00	1526.0
117.00	29896.0
118.00	2463.0
122.00	2291.0
123.00	3248.0
124.00	1413.0
124.90	1222.0
126.10	663.0
127.00	94144.0
128.00	7785.0
129.00	39760.0
130.00	3458.0
130.90	615.0
132.00	416.0
134.00	1241.0
135.00	3618.0
135.90	1308.0
137.10	2233.0
138.00	397.0
139.00	350.0
140.10	466.0
141.00	5563.0
142.00	1721.0
143.00	1321.0
145.90	962.0
147.00	2985.0
148.10	7150.0
149.10	1623.0
150.00	509.0
151.20	567.0
153.00	1772.0
154.10	1387.0
155.00	2970.0
156.00	4894.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

157.00	789.0
158.00	1039.0
158.90	955.0
160.00	2328.0
161.00	3189.0
162.00	640.0
165.00	1792.0
166.00	1624.0
167.00	12754.0
168.00	6850.0
169.00	1265.0
169.90	444.0
171.00	491.0
172.00	1104.0
173.00	1241.0
174.00	2193.0
175.10	4576.0
176.00	1483.0
177.00	2087.0
178.00	936.0
179.00	8776.0
180.00	6610.0
181.00	2527.0
182.10	433.0
184.00	728.0
185.00	4442.0
186.00	32824.0
187.00	9670.0
188.00	1293.0
189.00	2618.0
191.00	1420.0
192.00	2914.0
193.00	3477.0
194.00	724.0
195.00	618.0
196.00	5425.0
196.80	924.0
198.00	207872.0
199.00	13704.0
200.10	1113.0
201.50	1126.0
203.10	2019.0
204.00	9272.0
205.10	13868.0
206.00	59040.0
207.00	9685.0
208.00	2648.0
208.90	875.0
209.90	933.0
210.20	906.0
211.00	2185.0
215.00	640.0
216.00	1426.0
217.00	16261.0
218.00	1959.0
221.00	8118.0
221.80	2027.0
223.00	4154.0
224.00	34512.0
225.00	8043.0
226.00	909.0
227.00	16536.0
228.00	1948.0
229.00	3452.0
231.00	1100.0
233.00	378.0
234.00	876.0
234.90	1100.0
236.00	721.0
237.00	2062.0
238.90	540.0
240.00	440.0
241.10	796.0
242.00	2207.0
243.00	1494.0
244.00	25192.0
245.10	3130.0
246.00	6459.0
246.90	1006.0
249.00	1165.0
249.90	303.0
250.90	359.0
253.10	617.0
254.10	914.0
255.00	130904.0
256.00	19400.0
256.90	1827.0
258.00	7917.0
259.00	1082.0
265.00	3206.0
265.90	519.0
271.10	339.0
272.00	456.0
273.00	3747.0
274.00	10019.0
275.00	57024.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

276.00	7880.0
277.00	6924.0
278.00	1220.0
281.00	581.0
283.00	664.0
283.90	429.0
284.90	997.0
293.00	1214.0
293.90	531.0
295.00	364.0
296.00	18600.0
297.00	2629.0
297.90	320.0
301.00	343.0
303.10	1957.0
304.00	517.0
307.90	439.0
310.00	454.0
314.00	761.0
315.00	1993.0
316.10	828.0
321.10	650.0
323.00	5156.0
323.90	720.0
327.00	1264.0
327.90	453.0
332.90	366.0
334.00	3965.0
335.00	999.0
341.10	967.0
345.90	1199.0
346.90	351.0
352.00	1744.0
352.90	824.0
354.10	1546.0
355.00	457.0
365.00	7538.0
366.00	1092.0
371.10	630.0
372.10	2457.0
373.00	452.0
382.90	520.0
389.90	372.0
402.00	672.0
402.90	1152.0
404.00	670.0
421.00	1272.0
422.00	846.0
423.00	7748.0
424.00	2718.0
425.00	420.0
441.00	19816.0
442.10	132608.0
443.00	26192.0
444.00	2242.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

m/z	Abundance
37.00	509.0
38.00	981.0
39.10	6103.0
40.00	742.0
44.00	981.0
49.00	734.0
50.10	21680.0
51.10	84912.0
52.10	4010.0
55.00	616.0
56.00	2663.0
57.00	6176.0
58.10	367.0
61.00	1121.0
62.00	1761.0
63.10	3581.0
64.10	499.0
65.00	1890.0
68.00	1803.0
69.00	97280.0
70.00	597.0
73.00	995.0
74.00	11111.0
75.00	16164.0
76.10	5052.0
77.10	107360.0
78.10	6718.0
79.00	8154.0
80.00	6345.0
81.00	8328.0
82.00	2418.0
83.00	1597.0
84.00	302.0
85.00	1749.0
86.00	2715.0
87.00	1479.0
87.90	466.0
89.00	301.0
91.00	1933.0
92.00	2673.0
93.00	14631.0
94.00	1277.0
96.10	643.0
98.00	12345.0
99.00	9729.0
100.10	636.0
101.00	4097.0
103.00	1632.0
104.00	3039.0
105.00	3212.0
106.00	1231.0
107.00	38560.0
108.00	5894.0
109.00	1070.0
110.00	64160.0
111.00	10856.0
112.00	1301.0
113.10	305.0
114.90	340.0
116.10	2107.0
117.00	37360.0
118.00	2914.0
119.90	393.0
122.00	2586.0
123.00	3480.0
124.00	2241.0
125.00	1710.0
126.10	508.0
127.00	119352.0
128.00	8583.0
129.00	48768.0
130.00	4280.0
130.90	829.0
132.10	586.0
134.00	1444.0
135.00	4013.0
136.00	1836.0
137.00	2286.0
137.80	434.0
138.90	310.0
140.10	547.0
141.00	7137.0
142.00	2014.0
143.00	2030.0
144.00	350.0
145.00	591.0
146.00	1110.0
147.00	3241.0
148.00	8370.0
149.00	2187.0
150.00	634.0
150.90	1002.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

151.60	617.0
153.00	2523.0
154.00	1807.0
155.00	4259.0
156.00	6096.0
156.90	954.0
158.00	1579.0
159.00	871.0
160.00	2818.0
161.00	4069.0
162.10	733.0
164.00	395.0
165.00	2660.0
166.00	2825.0
167.00	15748.0
168.00	9075.0
169.00	1527.0
170.10	506.0
170.90	645.0
172.00	1566.0
173.00	1477.0
174.00	2899.0
175.00	6178.0
176.00	1624.0
177.00	2355.0
178.00	880.0
179.00	11853.0
180.00	7713.0
181.00	3584.0
182.00	487.0
183.90	1057.0
185.00	5658.0
186.00	43312.0
187.00	12774.0
188.00	1589.0
189.00	2634.0
190.00	390.0
191.00	1335.0
191.90	3733.0
193.00	3745.0
194.00	1146.0
196.00	7212.0
197.00	1235.0
198.00	288000.0
199.00	18840.0
200.00	1473.0
201.20	768.0
201.40	772.0
202.90	2582.0
204.00	12837.0
205.00	19376.0
206.00	81408.0
207.10	13030.0
208.00	3300.0
208.90	1408.0
210.00	875.0
211.00	3291.0
212.00	766.0
214.90	1389.0
216.10	2011.0
217.00	23664.0
218.00	3004.0
219.00	333.0
221.00	8614.0
221.90	2448.0
223.00	4939.0
224.00	47736.0
225.00	12289.0
226.10	1291.0
227.00	22672.0
228.00	3282.0
229.00	4378.0
230.10	431.0
231.00	1323.0
232.90	527.0
234.00	1207.0
235.00	1651.0
236.00	954.0
237.00	2368.0
238.00	372.0
238.90	988.0
239.90	573.0
241.00	1254.0
242.00	2127.0
243.00	2497.0
244.10	34392.0
245.10	4535.0
246.00	7837.0
246.90	2110.0
248.00	419.0
249.00	960.0
251.00	483.0
252.00	394.0
252.90	592.0
254.10	1343.0
255.00	199424.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

256.00	29304.0
256.90	2863.0
258.00	12625.0
258.90	1662.0
259.90	390.0
263.90	388.0
265.00	5076.0
266.00	784.0
269.90	303.0
271.00	466.0
272.00	312.0
273.00	5087.0
274.00	15588.0
275.00	82672.0
276.00	9954.0
277.00	8695.0
278.00	1403.0
278.90	519.0
281.00	365.0
282.00	441.0
283.10	978.0
284.00	626.0
285.00	1577.0
290.10	396.0
292.00	337.0
293.00	1388.0
293.90	684.0
294.90	486.0
296.00	27920.0
297.00	4076.0
297.90	341.0
300.90	405.0
301.90	340.0
303.00	3037.0
304.00	841.0
308.00	535.0
314.00	1253.0
315.00	2733.0
316.00	2144.0
317.00	601.0
321.10	813.0
322.00	371.0
323.10	7585.0
324.00	1540.0
327.00	1700.0
328.00	738.0
332.00	726.0
332.90	654.0
334.10	5487.0
335.00	1572.0
341.00	953.0
346.00	1354.0
347.00	316.0
352.00	2423.0
353.00	1553.0
354.10	2566.0
355.10	742.0
365.00	12157.0
366.00	1664.0
370.90	674.0
372.10	4665.0
373.00	1109.0
383.10	1134.0
383.90	396.0
390.00	400.0
390.90	336.0
391.90	397.0
402.00	1562.0
402.90	2157.0
404.00	787.0
421.00	1506.0
422.00	1806.0
423.00	12524.0
424.00	3452.0
425.10	453.0
441.00	32376.0
442.00	224896.0
443.00	44392.0
444.00	4439.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

m/z	Abundance
37.10	399.0
38.00	1000.0
39.00	3902.0
40.10	708.0
44.00	820.0
49.00	679.0
50.00	14730.0
51.10	59024.0
52.10	2785.0
55.10	437.0
56.00	2086.0
57.00	4208.0
60.90	804.0
62.00	1005.0
63.10	2292.0
65.00	1302.0
68.00	1074.0
69.00	65496.0
70.00	714.0
73.10	1083.0
74.00	8056.0
75.00	11209.0
76.00	3958.0
77.00	74712.0
78.10	5281.0
79.00	6205.0
80.00	4244.0
81.00	5938.0
81.90	1414.0
83.00	1498.0
85.10	1060.0
86.00	1896.0
87.00	838.0
87.90	420.0
91.00	1600.0
92.00	1612.0
93.00	11095.0
93.90	749.0
96.10	521.0
97.00	309.0
98.00	9279.0
99.00	7454.0
100.10	627.0
101.00	3231.0
103.00	1535.0
104.00	2126.0
105.00	2168.0
106.10	1161.0
107.00	27600.0
108.00	4053.0
109.00	902.0
110.00	45056.0
111.00	7113.0
112.00	884.0
113.00	349.0
116.00	1398.0
117.00	27856.0
118.00	2030.0
119.00	303.0
122.00	1745.0
123.00	3383.0
124.00	1149.0
125.10	1239.0
126.10	557.0
127.00	85600.0
128.00	6358.0
129.00	37712.0
130.00	3425.0
130.90	591.0
132.00	311.0
133.10	356.0
134.00	1037.0
135.00	3257.0
136.10	1587.0
137.00	1858.0
138.00	394.0
139.00	385.0
139.90	434.0
141.00	5508.0
142.00	1651.0
142.90	992.0
144.10	392.0
146.00	825.0
147.00	2706.0
148.00	6348.0
148.90	1369.0
150.10	441.0
151.00	719.0
152.00	384.0
153.00	2054.0
154.00	1390.0
155.00	2572.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

156.00	4455.0
157.10	651.0
157.90	1369.0
158.90	900.0
160.00	1809.0
161.00	2135.0
162.10	580.0
163.00	338.0
164.10	425.0
165.00	1992.0
166.00	1983.0
167.00	11014.0
168.00	7012.0
169.00	940.0
170.00	493.0
170.90	524.0
172.00	1415.0
172.90	1145.0
174.00	2301.0
175.00	5044.0
176.10	1134.0
177.10	2209.0
178.10	611.0
179.00	8983.0
180.00	6356.0
181.00	3075.0
183.90	638.0
185.00	4098.0
186.00	32320.0
187.00	10043.0
188.10	1038.0
188.90	2193.0
190.00	450.0
191.00	1118.0
192.00	2254.0
193.00	3415.0
194.00	478.0
195.20	379.0
196.00	5735.0
196.90	1150.0
198.00	222336.0
199.00	16253.0
200.00	1242.0
201.50	631.0
203.00	1785.0
204.00	9575.0
205.00	15203.0
206.00	63192.0
207.00	9877.0
208.00	2562.0
209.00	1101.0
210.00	1004.0
211.00	2690.0
212.90	329.0
215.00	772.0
216.10	1806.0
217.00	18360.0
218.00	2456.0
218.90	412.0
221.00	8270.0
221.90	2328.0
223.00	4288.0
224.00	38224.0
225.00	8418.0
226.00	1018.0
227.00	18480.0
228.00	1810.0
229.00	3351.0
230.00	400.0
231.10	1490.0
231.90	330.0
233.10	413.0
234.00	1290.0
234.90	876.0
236.00	1283.0
237.00	1678.0
238.00	331.0
239.00	749.0
239.80	543.0
241.00	1049.0
242.10	2842.0
243.00	2221.0
244.00	28760.0
245.00	3758.0
246.00	7136.0
247.10	1165.0
248.00	445.0
249.00	1102.0
251.90	451.0
253.00	715.0
254.20	1343.0
255.00	162304.0
256.00	24384.0
257.00	2360.0
258.00	10529.0
259.00	1886.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

263.90	319.0
265.00	4515.0
265.90	695.0
270.90	404.0
271.90	569.0
273.00	4843.0
274.00	13467.0
275.00	70440.0
276.00	8451.0
277.00	8136.0
278.00	1331.0
281.00	388.0
283.10	755.0
284.00	720.0
285.10	703.0
289.90	401.0
293.00	1372.0
294.00	485.0
295.10	542.0
296.00	23432.0
297.00	3529.0
301.00	399.0
301.90	356.0
303.00	2491.0
304.00	759.0
308.00	311.0
314.00	1069.0
315.00	2991.0
316.10	1264.0
321.00	763.0
322.00	377.0
323.10	7091.0
324.00	1041.0
327.00	1311.0
328.00	525.0
332.00	646.0
332.90	695.0
334.00	5283.0
335.00	1543.0
341.00	779.0
346.00	1410.0
346.90	354.0
352.10	1944.0
352.90	1427.0
354.00	2559.0
355.00	424.0
365.00	11501.0
366.00	2037.0
370.20	347.0
371.00	411.0
372.00	4004.0
373.00	836.0
383.00	1173.0
384.00	326.0
390.00	413.0
401.00	352.0
402.00	1907.0
403.10	2285.0
404.00	727.0
421.00	1792.0
422.10	1639.0
423.10	12975.0
424.10	2833.0
441.10	34352.0
442.10	231360.0
443.10	43288.0
444.00	4624.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62