

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092122\
 Data File : BM036660.D
 Acq On : 22 Sep 2022 13:29
 Operator : CG/JU
 Sample : N4708-05
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 CB8K5

Quant Time: Sep 23 01:00:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 21 22:42:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.022	152	378381	20.000	ng/u1	0.00
20) Naphthalene-d8	10.834	136	1718725	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.645	164	1085796	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.380	188	2152650	20.000	ng/u1	0.00
79) Chrysene-d12	21.545	240	1619300	20.000	ng/u1	0.00
88) Perylene-d12	23.991	264	1390253	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	44247	4.384	ng/uL	0.00
4) Pyridine-d5	3.828	84	107327	3.646	ng/u1	0.01
7) Phenol-d5	7.175	99	224217	6.631	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.345	67	579176	28.204	ng/u1	0.00
11) 2-Chlorophenol-d4	7.545	132	596711	24.511	ng/u1	0.00
15) 4-Methylphenol-d8	8.728	113	415264	16.620	ng/u1	0.00
21) Nitrobenzene-d5	9.187	128	376247	32.342	ng/u1	0.00
24) 2-Nitrophenol-d4	9.910	143	356792	34.241	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.445	165	686475	28.427	ng/u1	0.00
31) 4-Chloroaniline-d4	10.969	131	402135	10.016	ng/u1	0.00
46) Dimethylphthalate-d6	14.057	166	2411490	32.930	ng/u1	0.00
49) Acenaphthylene-d8	14.339	160	2738321	31.524	ng/u1	0.00
54) 4-Nitrophenol-d4	14.827	143	81361	5.728	ng/u1	0.00
60) Fluorene-d10	15.633	176	2156634	32.862	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	259484	30.467	ng/u1	0.00
73) Anthracene-d10	17.480	188	3475708	35.313	ng/u1	0.00
81) Pyrene-d10	19.762	212	3842972	48.240	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.833	264	2629657	38.716	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092122\
Data File : BM036660.D
Acq On : 22 Sep 2022 13:29
Operator : CG/JU
Sample : N4708-05
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
CB8K5

Quant Time: Sep 23 01:00:22 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Sep 21 22:42:41 2022
Response via : Initial Calibration

