

Data Path : Z:\HPCHEM1\BNA M\DATA\BM092215\
 Data File : BM002700.D
 Acq On : 22 Sep 2015 17:28
 Operator : UM/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.475

Quant Time: Sep 23 03:41:43 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM90315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 23 03:40:59 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	150	0.00
2 SURR	1,4-Dioxane-d8	0.403	0.364	9.7	143	0.00
3	1,4-Dioxane	0.465	0.409	12.0	143	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	128	0.00
5	Naphthalene	1.037	0.981	5.4	121	0.00
6 SURR	2-Methylnaphthalene-d10	0.537	0.472	12.1	112	0.00
7	2-Methylnaphthalene	0.691	0.606	12.3	112	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
9	Acenaphthylene	2.132	1.925	9.7	104	0.00
10 C	Acenaphthene	1.493	1.381	7.5	106	0.00
11	Fluorene	1.588	1.377	13.3	100	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
13	Pentachlorophenol	0.039	0.032	17.9	116	0.00
14	Phenanthrene	1.197	1.118	6.6	111	0.00
15	Anthracene	1.191	1.076	9.7	107	0.00
16 SURR	Fluoranthene-d10	0.913	0.910	0.3	123	0.00
17 C	Fluoranthene	1.252	1.223	2.3	123	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	129	0.00
19	Pyrene	1.512	1.483	1.9	125	0.00
20	Benzo(a)anthracene	1.326	1.169	11.8	114	0.00
21	Chrysene	1.283	1.236	3.7	125	0.00
22 I	Perylene-d12	1.000	1.000	0.0	124	0.00
23	Benzo(b)fluoranthene	1.413	1.296	8.3	118	0.00
24	Benzo(k)fluoranthene	1.326	1.271	4.1	117	0.00
25 C	Benzo(a)pyrene	1.320	1.169	11.4	110	0.00
26	Indeno(1,2,3-cd)pyrene	1.563	1.097	29.8#	87	0.00
27	Dibenzo(a,h)anthracene	1.280	0.787	38.5#	76	0.00
28	Benzo(q,h,i)perylene	1.339	1.066	20.4#	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0