

Data Path : Z:\HPCHEM1\BNA M\DATA\BM092215\
 Data File : BM002706.D
 Acq On : 22 Sep 2015 22:24
 Operator : UM/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.476

Quant Time: Sep 23 03:42:42 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM90315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 23 03:40:59 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	164#	0.00
2 SURR	1,4-Dioxane-d8	0.403	0.359	10.9	154#	0.00
3	1,4-Dioxane	0.465	0.406	12.7	155#	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	144	0.00
5	Naphthalene	1.037	0.980	5.5	136	0.00
6 SURR	2-Methylnaphthalene-d10	0.537	0.476	11.4	128	0.00
7	2-Methylnaphthalene	0.691	0.611	11.6	127	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	130	-0.02
9	Acenaphthylene	2.132	1.881	11.8	115	0.00
10 C	Acenaphthene	1.493	1.405	5.9	122	0.00
11	Fluorene	1.588	1.409	11.3	116	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	129	0.00
13	Pentachlorophenol	0.039	0.039	0.0	156#	0.00
14	Phenanthrene	1.197	1.122	6.3	121	0.00
15	Anthracene	1.191	1.093	8.2	119	0.00
16 SURR	Fluoranthene-d10	0.913	0.902	1.2	133	0.00
17 C	Fluoranthene	1.252	1.224	2.2	135	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	148	-0.01
19	Pyrene	1.512	1.394	7.8	135	0.00
20	Benzo(a)anthracene	1.326	1.185	10.6	134	0.00
21	Chrysene	1.283	1.197	6.7	140	0.00
22 I	Perylene-d12	1.000	1.000	0.0	148	-0.01
23	Benzo(b)fluoranthene	1.413	1.342	5.0	146	-0.01
24	Benzo(k)fluoranthene	1.326	1.179	11.1	130	-0.01
25 C	Benzo(a)pyrene	1.320	1.178	10.8	132	0.00
26	Indeno(1,2,3-cd)pyrene	1.563	1.136	27.3#	107	-0.01
27	Dibenzo(a,h)anthracene	1.280	0.869	32.1#	100	0.00
28	Benzo(q,h,i)perylene	1.339	1.022	23.7#	114	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0