

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM092216\
 Data File : BM007520.D
 Acq On : 22 Sep 2016 20:28
 Operator : UM/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.442

Quant Time: Sep 23 12:15:13 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 23 04:00:23 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	116	0.00
3	Naphthalene	0.931	0.934	-0.3	117	0.00
4 SURR	2-Methylnaphthalene-d10	0.466	0.464	0.4	115	0.00
5	2-Methylnaphthalene	0.630	0.630	0.0	116	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
7	Acenaphthylene	1.693	1.638	3.2	112	0.00
8 C	Acenaphthene	1.320	1.302	1.4	114	0.00
9	Fluorene	1.618	1.581	2.3	110	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
11	Pentachlorophenol	0.109	0.108	0.9	118	0.00
12	Phenanthrene	1.082	1.090	-0.7	110	0.02
13	Anthracene	0.981	0.978	0.3	109	0.00
14 SURR	Fluoranthene-d10	0.932	0.933	-0.1	109	0.00
15 C	Fluoranthene	1.364	1.377	-1.0	111	-0.01
16 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
17	Pyrene	1.322	1.374	-3.9	115	0.00
18	Benzo(a)anthracene	1.182	1.219	-3.1	116	0.00
19	Chrysene	1.262	1.276	-1.1	109	0.00
20 I	Perylene-d12	1.000	1.000	0.0	112	0.00
21	Benzo(b)fluoranthene	1.335	1.410	-5.6	117	0.00
22	Benzo(k)fluoranthene	1.349	1.340	0.7	112	0.00
23 C	Benzo(a)pyrene	1.241	1.290	-3.9	117	0.00
24	Indeno(1,2,3-cd)pyrene	1.278	1.369	-7.1	122	0.00
25	Dibenzo(a,h)anthracene	0.976	1.079	-10.6	126	0.00
26	Benzo(g,h,i)perylene	1.148	1.212	-5.6	119	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0