


Data Path : Z:\HPCHEM1\BNA M\DATA\BM092316\
 Data File : BM007521.D
 Acq On : 23 Sep 2016 12:38
 Operator : UM/SJ
 Sample : DFTPP64
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP64

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM092316.M
 Title : SVOA CALIBRATION
 Last Update : Mon Sep 26 03:32:16 2016



AutoFind: Scans 2551, 2552, 2553; Background Corrected with Scan 2545

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.5	71285	PASS
68	69	0.00	2	1.7	1440	PASS
69	198	0.00	100	34.4	83234	PASS
70	69	0.00	2	0.7	543	PASS
127	198	10	80	41.5	100514	PASS
197	198	0.00	2	0.4	884	PASS
198	198	100	100	100.0	241984	PASS
199	198	5	9	6.9	16696	PASS
275	198	10	60	30.1	72938	PASS
365	198	1	100	4.7	11387	PASS
441	443	0.01	100	82.9	33069	PASS
442	198	50	100	88.2	213525	PASS
443	442	15	24	18.7	39893	PASS

m/z	Abundance
35.90	752.0
39.90	422.0
44.00	1546.0
72.90	761.0
147.10	371.0
191.00	321.0
193.10	375.0
207.00	1871.0
208.00	464.0
209.20	418.0
281.10	521.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

m/z	Abundance
35.90	511.0
36.90	578.0
38.00	1166.0
39.10	5963.0
40.00	727.0
44.00	920.0
49.00	808.0
50.10	22488.0
51.10	88976.0
52.10	4533.0
53.10	362.0
54.90	594.0
55.90	2188.0
57.00	6118.0
61.00	1386.0
62.00	1534.0
63.10	3762.0
64.00	805.0
65.00	1472.0
66.10	309.0
68.10	1700.0
69.00	100928.0
70.00	971.0
73.10	1591.0
74.00	11457.0
75.00	17128.0
76.10	6261.0
77.10	110064.0
78.10	8736.0
79.00	8582.0
80.00	6428.0
81.00	8525.0
82.00	1975.0
83.00	1970.0
84.00	435.0
85.00	1538.0
86.10	2637.0
87.00	1094.0
88.00	306.0
90.90	1960.0
92.10	2082.0
93.00	14693.0
94.00	921.0
96.20	738.0
96.90	451.0
98.00	12759.0
99.00	10279.0
100.10	603.0
101.00	4480.0
103.00	1655.0
104.00	3582.0
105.00	3041.0
105.90	960.0
107.00	36808.0
108.00	5593.0
109.00	1421.0
110.00	64072.0
111.00	10194.0
112.00	1274.0
113.00	412.0
115.90	1968.0
117.00	39056.0
118.00	3277.0
118.90	399.0
119.90	581.0
120.90	455.0
122.00	2631.0
123.00	3318.0
124.00	1597.0
124.90	1290.0
126.10	634.0
127.00	115520.0
128.00	8743.0
129.00	48408.0
130.00	4527.0
131.00	881.0
132.20	331.0
134.00	1718.0
135.00	4486.0
136.00	2225.0
137.00	2306.0
138.10	314.0
140.00	711.0
141.00	6666.0
142.10	1787.0
143.10	1403.0
143.90	632.0
145.90	1554.0
147.00	3964.0
148.00	9009.0
149.00	1841.0
150.10	492.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

151.20	788.0
151.90	687.0
153.00	2214.0
154.00	1535.0
155.00	4324.0
156.00	5729.0
157.00	1124.0
158.00	1320.0
158.90	1030.0
160.00	1945.0
161.00	4048.0
161.90	964.0
163.10	376.0
165.00	2841.0
166.00	2561.0
167.00	14523.0
168.00	8784.0
169.00	1206.0
170.00	476.0
170.90	919.0
172.00	1085.0
173.00	1897.0
174.00	2841.0
175.00	6004.0
175.90	1500.0
177.00	3315.0
178.00	767.0
179.00	10568.0
180.00	8532.0
181.00	3168.0
182.00	591.0
184.00	903.0
185.00	4678.0
186.00	38680.0
187.00	12336.0
188.00	1937.0
189.10	3121.0
190.00	348.0
191.00	1274.0
192.00	3351.0
193.00	3949.0
194.10	805.0
194.90	347.0
196.00	6459.0
197.10	1517.0
198.00	260352.0
199.00	18240.0
200.00	1463.0
201.50	1072.0
203.00	2168.0
204.00	11113.0
205.10	16968.0
206.10	74928.0
207.00	12147.0
207.90	3184.0
209.00	1139.0
209.90	1058.0
211.00	2549.0
211.90	470.0
215.00	1020.0
216.00	1997.0
217.00	21616.0
218.00	2602.0
219.00	323.0
221.00	8725.0
222.00	3068.0
223.00	5425.0
224.10	43256.0
225.00	12126.0
226.10	1304.0
227.00	20240.0
228.00	3104.0
229.00	4466.0
229.80	389.0
231.10	1649.0
232.00	353.0
234.10	1034.0
235.00	1204.0
235.90	1015.0
237.00	2355.0
237.90	401.0
239.00	639.0
240.00	682.0
240.90	1319.0
242.00	2095.0
243.10	1925.0
244.00	32184.0
245.10	4808.0
246.00	6349.0
247.00	1886.0
247.90	371.0
249.00	1280.0
250.80	348.0
253.00	836.0
254.10	1192.0
255.00	172224.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

256.00	26264.0
257.00	2752.0
258.00	10755.0
259.00	2270.0
265.00	5061.0
266.00	801.0
267.00	338.0
271.00	606.0
272.00	470.0
273.00	4815.0
274.00	13677.0
275.00	74128.0
276.00	9695.0
277.00	7639.0
277.90	1335.0
279.00	309.0
281.00	616.0
282.00	484.0
283.00	901.0
284.00	861.0
285.00	1123.0
289.00	340.0
292.00	476.0
292.90	1307.0
294.10	588.0
296.00	23080.0
297.00	3152.0
302.10	544.0
303.00	2748.0
303.90	687.0
310.00	476.0
314.00	916.0
315.00	2543.0
316.00	1212.0
321.00	686.0
323.10	7599.0
324.20	1212.0
326.90	1470.0
328.00	727.0
331.90	578.0
333.00	628.0
334.10	5002.0
335.00	1462.0
341.10	741.0
346.00	1519.0
352.00	2077.0
353.00	1033.0
354.10	2818.0
355.10	515.0
365.00	9900.0
366.00	1370.0
371.00	655.0
372.00	3812.0
373.00	796.0
382.90	920.0
384.10	328.0
390.00	386.0
391.00	356.0
402.00	1256.0
403.00	1616.0
403.90	643.0
421.10	1432.0
422.00	1655.0
423.10	10577.0
424.00	3458.0
425.20	537.0
441.00	28168.0
442.10	178048.0
443.00	33744.0
444.00	3402.0
445.00	304.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

m/z	Abundance
36.00	695.0
37.10	339.0
38.00	1280.0
39.00	5210.0
40.00	553.0
40.90	364.0
44.00	1232.0
49.10	786.0
50.10	22264.0
51.10	80688.0
52.00	3756.0
54.90	688.0
56.00	2420.0
57.00	5394.0
58.00	303.0
61.10	1516.0
61.90	1553.0
63.00	4393.0
64.10	501.0
65.10	1849.0
68.10	1864.0
69.00	96144.0
70.00	332.0
73.00	1607.0
74.10	11048.0
75.00	15400.0
76.10	6016.0
77.10	108416.0
78.10	7336.0
79.00	8730.0
80.00	5865.0
81.00	8661.0
82.00	2016.0
83.00	2140.0
85.00	1439.0
86.00	2598.0
87.00	1062.0
88.10	303.0
91.00	1768.0
92.00	2750.0
93.00	14752.0
94.10	1146.0
96.10	677.0
97.10	478.0
98.00	12249.0
99.00	9486.0
100.00	1111.0
101.10	4988.0
103.00	1793.0
104.00	3156.0
105.10	2602.0
106.10	1138.0
107.00	37984.0
108.00	5204.0
109.10	1028.0
110.00	61816.0
111.00	9666.0
111.90	1470.0
113.00	694.0
115.00	329.0
116.00	2203.0
117.00	36816.0
118.00	2836.0
119.10	413.0
120.00	485.0
122.00	2611.0
123.00	4054.0
123.90	1435.0
125.00	1640.0
126.20	516.0
127.00	116352.0
128.00	9240.0
129.00	50904.0
130.00	4429.0
131.00	712.0
131.90	380.0
133.00	414.0
134.10	1607.0
135.00	4428.0
136.10	1867.0
137.00	2057.0
137.90	480.0
139.90	1021.0
141.00	7898.0
142.10	2042.0
143.00	1873.0
144.10	393.0
145.00	379.0
146.00	1693.0
147.10	3676.0
148.00	9604.0
149.10	1600.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

150.00	471.0
150.90	707.0
151.90	447.0
153.00	2187.0
154.00	1646.0
155.00	3903.0
156.00	5408.0
157.00	1008.0
157.80	1157.0
159.00	1118.0
160.00	2397.0
161.00	3364.0
162.00	687.0
163.00	616.0
165.00	3134.0
166.00	2217.0
167.00	14015.0
168.00	8157.0
169.00	1222.0
169.90	626.0
171.00	942.0
171.90	1316.0
173.00	1525.0
174.00	3191.0
175.00	6011.0
176.10	1044.0
176.90	2639.0
177.90	1456.0
179.00	11259.0
180.00	8308.0
181.00	4029.0
181.90	469.0
184.00	922.0
185.00	5687.0
186.00	43464.0
187.00	12791.0
188.00	1695.0
189.00	2755.0
190.00	586.0
191.00	1709.0
192.00	3344.0
193.00	3832.0
194.00	635.0
195.00	493.0
196.00	8139.0
196.80	1478.0
198.00	286208.0
199.00	19944.0
200.00	1781.0
201.60	1013.0
203.00	2606.0
204.00	11449.0
205.10	20760.0
206.00	83536.0
207.00	12354.0
208.00	2963.0
209.00	1559.0
209.90	1464.0
211.10	3207.0
211.80	605.0
214.90	1023.0
216.00	2076.0
217.00	24264.0
218.00	2857.0
219.00	307.0
220.30	705.0
221.00	9811.0
221.90	2733.0
223.00	4716.0
224.00	47216.0
225.10	13134.0
226.00	1343.0
227.00	21888.0
228.00	3847.0
229.00	4682.0
229.90	546.0
231.00	1528.0
233.00	362.0
234.00	1233.0
235.00	1431.0
236.00	1104.0
237.00	2005.0
238.90	1083.0
239.90	735.0
241.00	1429.0
242.10	2326.0
243.00	2635.0
244.10	34552.0
245.10	4224.0
246.00	8599.0
247.00	1967.0
248.20	341.0
248.90	1405.0
251.10	467.0
252.00	391.0
253.10	932.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

254.10	1524.0
255.00	203392.0
256.00	27776.0
257.00	2782.0
258.00	12415.0
259.00	1845.0
260.00	410.0
263.90	305.0
265.00	5419.0
266.00	1095.0
270.90	389.0
272.00	525.0
273.00	5295.0
274.00	17040.0
275.00	85952.0
276.00	11437.0
277.00	9522.0
278.00	1589.0
281.00	698.0
283.00	1099.0
284.00	879.0
285.10	1632.0
288.90	467.0
290.10	311.0
292.20	490.0
293.00	1489.0
294.00	538.0
294.80	712.0
295.20	708.0
296.00	28104.0
297.00	4743.0
300.90	419.0
302.10	464.0
303.00	3013.0
304.00	800.0
308.20	385.0
309.00	378.0
310.10	671.0
314.10	1817.0
315.00	3039.0
316.00	2229.0
321.00	1080.0
322.00	328.0
323.10	8091.0
324.00	1704.0
327.00	1598.0
327.90	1056.0
329.00	359.0
331.90	536.0
333.00	654.0
334.00	5033.0
335.00	1428.0
341.10	1077.0
346.10	2046.0
346.90	502.0
352.00	2270.0
353.10	1782.0
354.10	2513.0
355.00	433.0
365.00	14129.0
366.00	1898.0
370.90	703.0
372.00	4429.0
373.00	1079.0
383.00	1087.0
384.10	394.0
390.10	367.0
391.10	387.0
401.10	315.0
402.00	1722.0
403.10	2294.0
404.10	640.0
405.10	427.0
421.00	2512.0
422.00	1834.0
423.10	14166.0
424.10	4404.0
425.00	730.0
441.10	39752.0
442.10	257920.0
443.10	49080.0
444.00	4881.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

m/z	Abundance
35.90	680.0
36.90	301.0
38.10	754.0
39.00	3351.0
40.10	507.0
44.00	1387.0
49.20	502.0
50.10	11209.0
51.10	44192.0
52.00	2430.0
54.90	463.0
56.00	1503.0
57.00	3175.0
61.10	789.0
62.00	496.0
63.10	2408.0
64.00	565.0
65.10	979.0
68.00	757.0
69.00	52632.0
70.00	328.0
73.00	1729.0
74.00	6553.0
75.00	8877.0
76.10	3106.0
77.10	61008.0
78.10	3662.0
79.00	4368.0
80.00	3174.0
81.00	4318.0
82.10	1034.0
82.90	1161.0
85.10	857.0
86.00	1444.0
86.90	420.0
91.00	1128.0
92.00	1418.0
93.00	8386.0
93.90	568.0
96.10	461.0
98.00	6921.0
99.00	5047.0
100.00	748.0
101.00	2992.0
103.00	776.0
104.00	2013.0
105.00	2110.0
105.90	730.0
107.00	21952.0
108.00	2855.0
109.10	814.0
110.00	36448.0
111.00	5224.0
111.90	828.0
113.10	344.0
116.00	992.0
117.00	21144.0
118.00	1315.0
122.00	1747.0
122.90	2479.0
124.00	1082.0
125.10	1100.0
126.20	394.0
127.00	69672.0
128.00	4909.0
129.00	29208.0
130.00	2534.0
131.10	458.0
134.00	1159.0
135.00	2604.0
136.00	1223.0
136.90	1414.0
138.00	393.0
141.00	3981.0
142.10	1403.0
143.00	908.0
146.10	808.0
147.00	2298.0
148.00	6144.0
149.00	1100.0
150.00	463.0
151.00	784.0
151.90	337.0
152.90	1577.0
154.00	1085.0
155.00	2532.0
156.00	3672.0
157.10	730.0
157.90	860.0
158.90	648.0
159.90	1333.0
160.90	2116.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

162.00	804.0
162.90	314.0
165.00	1543.0
165.90	1164.0
167.00	9580.0
168.00	4912.0
169.00	1141.0
169.80	382.0
172.10	950.0
173.10	1098.0
174.00	1726.0
175.00	4318.0
176.00	1407.0
177.00	1780.0
178.00	723.0
179.00	7131.0
180.00	5422.0
181.00	1713.0
184.00	590.0
185.00	3034.0
186.00	27840.0
187.00	7936.0
187.90	885.0
189.00	2051.0
191.00	1380.0
192.10	2053.0
193.00	3023.0
194.10	653.0
196.00	4435.0
196.90	1175.0
198.00	179392.0
199.00	11905.0
200.00	1326.0
201.60	625.0
203.00	1671.0
204.00	7887.0
205.10	12303.0
206.00	53640.0
207.00	8662.0
208.00	2758.0
209.00	1079.0
210.10	687.0
211.00	2512.0
215.10	825.0
216.00	1381.0
217.00	15002.0
218.00	2286.0
221.00	5710.0
222.00	2101.0
223.00	3395.0
224.00	32400.0
225.00	7751.0
226.00	1161.0
227.00	15293.0
228.10	2168.0
229.00	2787.0
230.00	357.0
231.00	1374.0
233.90	888.0
235.00	1118.0
236.00	653.0
237.00	1181.0
238.90	518.0
240.00	420.0
240.90	910.0
242.00	1808.0
243.10	1699.0
244.10	21848.0
245.00	2967.0
246.00	5970.0
247.10	1020.0
248.00	355.0
248.90	1051.0
253.00	703.0
254.00	794.0
255.00	135936.0
256.00	18448.0
257.00	1394.0
258.00	9113.0
258.90	1242.0
260.90	321.0
264.00	302.0
265.00	3769.0
266.10	418.0
266.90	339.0
272.10	454.0
273.00	3968.0
274.00	10360.0
275.00	58736.0
276.00	7613.0
277.00	5779.0
278.00	1085.0
281.00	617.0
283.00	914.0
284.00	712.0
285.10	984.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64

293.00	1633.0
293.90	330.0
295.00	567.0
296.00	19816.0
297.00	2899.0
301.90	307.0
303.00	2635.0
303.90	570.0
314.00	1099.0
315.00	2530.0
316.00	1343.0
320.90	844.0
323.00	6437.0
324.00	1129.0
327.00	1181.0
327.90	816.0
331.90	528.0
333.00	643.0
334.00	4160.0
334.90	1128.0
341.10	649.0
346.00	1388.0
352.00	1696.0
352.90	1337.0
354.10	2236.0
355.10	416.0
365.00	10132.0
366.00	1199.0
371.00	543.0
372.00	2577.0
373.10	587.0
383.00	1065.0
390.00	505.0
392.10	359.0
402.00	1966.0
403.00	1795.0
404.10	820.0
421.00	1374.0
422.00	1391.0
423.10	10850.0
424.10	2984.0
425.00	375.0
441.00	31288.0
442.10	204608.0
443.10	36856.0
444.00	3642.0
445.00	343.0

Instrument :
BNA_M
ClientSampleId :
DFTPP64