

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092322\
 Data File : BM036670.D
 Acq On : 23 Sep 2022 10:47
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020184

Quant Time: Sep 23 23:34:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 21 22:42:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.010	152	419185	20.000	ng/ul	-0.01
20) Naphthalene-d8	10.828	136	1833741	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.639	164	1108447	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.380	188	2198063	20.000	ng/ul	0.00
79) Chrysene-d12	21.545	240	1675225	20.000	ng/ul	0.00
88) Perylene-d12	23.986	264	1420793	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	84374	7.546	ng/uL	0.00
4) Pyridine-d5	3.816	84	657805	20.171	ng/ul	0.00
7) Phenol-d5	7.169	99	761497	20.328	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.340	67	456043	20.046	ng/ul	0.00
11) 2-Chlorophenol-d4	7.540	132	579941	21.503	ng/ul	-0.01
15) 4-Methylphenol-d8	8.722	113	582746	21.053	ng/ul	0.00
21) Nitrobenzene-d5	9.181	128	281848	22.708	ng/ul	0.00
24) 2-Nitrophenol-d4	9.904	143	275405	24.772	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.439	165	559050	21.698	ng/ul	-0.01
31) 4-Chloroaniline-d4	10.963	131	827567	19.319	ng/ul	0.00
46) Dimethylphthalate-d6	14.051	166	1490856	19.942	ng/ul	0.00
49) Acenaphthylene-d8	14.333	160	1811702	20.430	ng/ul	0.00
54) 4-Nitrophenol-d4	14.827	143	270048	18.624	ng/ul	0.00
60) Fluorene-d10	15.627	176	1355399	20.231	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.739	200	164149	18.875	ng/ul	0.00
73) Anthracene-d10	17.474	188	2031147	20.210	ng/ul	0.00
81) Pyrene-d10	19.756	212	2050404	24.879	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.827	264	1435322	20.678	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.422	88	92135	7.649	ng/uL	97
5) Pyridine	3.834	79	658798	20.107	ng/ul	97
6) Benzaldehyde	7.145	77	461784	25.457	ng/ul	96
8) Phenol	7.192	94	815835	20.292	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.434	93	636080	20.334	ng/ul	99
12) 2-Chlorophenol	7.575	128	629473	21.576	ng/ul	97
13) 2-Methylphenol	8.457	108	599592	20.290	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.545	45	733340	19.780	ng/ul	99
16) Acetophenone	8.839	105	974612	20.300	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.828	70	477514	20.612	ng/ul	96
18) 4-Methylphenol	8.787	108	660540	20.670	ng/ul	98
19) Hexachloroethane	9.098	117	240418	20.345	ng/ul	98
22) Nitrobenzene	9.222	77	710796	21.570	ng/ul	97
23) Isophorone	9.751	82	1280728	19.061	ng/ul	100
25) 2-Nitrophenol	9.934	139	320938	23.372	ng/ul	99
26) 2,4-Dimethylphenol	9.992	107	708280	20.913	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.233	93	812171	20.280	ng/ul	100
29) 2,4-Dichlorophenol	10.469	162	582218	21.495	ng/ul	99
30) Naphthalene	10.875	128	2044496	20.466	ng/ul	100
32) 4-Chloroaniline	10.986	127	854885	19.330	ng/ul	99
33) Hexachlorobutadiene	11.163	225	329033	19.862	ng/ul	99
34) Caprolactam	11.769	113	159492	18.153	ng/ul	98
35) 4-Chloro-3-methylphenol	12.104	107	620806	20.763	ng/ul	100
36) 2-Methylnaphthalene	12.480	142	1353116	20.330	ng/ul	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092322\
 Data File : BM036670.D
 Acq On : 23 Sep 2022 10:47
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020184

Quant Time: Sep 23 23:34:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 21 22:42:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.698	142	1353164	20.224	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.839	216	650790	20.743	ng/ul	99
40) Hexachlorocyclopentadiene	12.822	237	236129	14.442	ng/ul	98
41) 2,4,6-Trichlorophenol	13.080	196	419474	21.837	ng/ul	100
42) 2,4,5-Trichlorophenol	13.151	196	454912	22.088	ng/ul	99
43) 1,1'-Biphenyl	13.480	154	1721536	21.069	ng/ul	100
44) 2-Chloronaphthalene	13.522	162	1347950	21.023	ng/ul	99
45) 2-Nitroaniline	13.722	65	367559	23.290	ng/ul	98
47) Dimethylphthalate	14.098	163	1581748	19.946	ng/ul	100
48) 2,6-Dinitrotoluene	14.216	165	295908	22.951	ng/ul	100
50) Acenaphthylene	14.363	152	2070286	20.420	ng/ul	100
51) 3-Nitroaniline	14.545	138	348100	22.512	ng/ul#	95
52) Acenaphthene	14.704	153	1438026	20.272	ng/ul	100
53) 2,4-Dinitrophenol	14.745	184	100542	17.575	ng/ul	96
55) 4-Nitrophenol	14.845	109	231392	18.668	ng/ul	96
56) Dibenzofuran	15.033	168	1961633	20.591	ng/ul	98
57) 2,4-Dinitrotoluene	14.998	165	430437	22.964	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.257	232	355282	20.608	ng/ul	96
59) Diethylphthalate	15.457	149	1561954	19.676	ng/ul	99
61) Fluorene	15.680	166	1654961	20.397	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.674	204	772693	19.913	ng/ul	98
63) 4-Nitroaniline	15.698	138	355526	23.664	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.757	198	181697	18.515	ng/ul	98
67) N-Nitrosodiphenylamine	15.886	169	1341933	20.728	ng/ul	99
68) 4-Bromophenyl-phenylether	16.568	248	411297	18.984	ng/ul	97
69) Hexachlorobenzene	16.680	284	487082	19.008	ng/ul	95
70) Atrazine	16.839	200	424748	18.593	ng/ul	97
71) Pentachlorophenol	17.021	266	260021	17.083	ng/ul	98
72) Phenanthrene	17.421	178	2465620	20.365	ng/ul	99
74) Anthracene	17.509	178	2488537	20.222	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.445	216	641899	21.351	ng/ul	100
76) Pentachlorobenzene	14.957	250	668670	20.564	ng/ul	99
77) Carbazole	17.780	167	2221713	19.500	ng/ul	100
78) Di-n-butylphthalate	18.345	149	2565903	19.869	ng/ul	100
80) Fluoranthene	19.427	202	2589418	25.640	ng/ul	99
82) Pyrene	19.786	202	2686542	25.131	ng/ul	99
83) Butylbenzylphthalate	20.686	149	992030	24.490	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.462	252	658604	20.121	ng/ul	99
85) Benzo(a)anthracene	21.533	228	2204612	21.004	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.462	149	1471296	24.865	ng/ul	99
87) Chrysene	21.586	228	2046847	20.545	ng/ul	100
89) Di-n-octyl phthalate	22.403	149	2190016	25.568	ng/ul	100
90) Benzo(b)fluoranthene	23.244	252	1961342	21.769	ng/ul	99
91) Benzo(k)fluoranthene	23.291	252	1868004	20.749	ng/ul	99
93) Benzo(a)pyrene	23.880	252	1644083	20.757	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.532	276	1839233	19.885	ng/ul	96
95) Dibenzo(a,h)anthracene	26.550	278	1686676	20.108	ng/ul	98
96) Benzo(g,h,i)perylene	27.309	276	1587641	19.800	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.649 ng/uL
RT: 3.422 min Scan# 51
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion: 88 Resp: 92135
Ion Ratio Lower Upper
88 100
43 24.5 23.1 34.7
58 71.3 56.6 85.0



#5
Pyridine
Concen: 20.107 ng/uL
RT: 3.834 min Scan# 127
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion: 79 Resp: 658798
Ion Ratio Lower Upper
79 100
52 69.1 57.2 85.8
51 31.8 26.8 40.2









#6
Benzaldehyde
Concen: 25.457 ng/ul
RT: 7.145 min Scan# 691
Delta R.T. -0.012 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion: 77 Resp: 461784
Ion Ratio Lower Upper
77 100
105 98.4 75.1 112.7
106 95.0 73.0 109.6



#8
Phenol
Concen: 20.292 ng/ul
RT: 7.192 min Scan# 698
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion: 94 Resp: 815835
Ion Ratio Lower Upper
94 100
65 27.7 22.2 33.4
66 35.1 28.9 43.3









#10
Bis(2-Chloroethyl)ether
Concen: 20.334 ng/ul
RT: 7.434 min Scan# 71
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion: 93 Resp: 636080
Ion Ratio Lower Upper
93 100
63 73.1 59.1 88.7
95 32.0 26.1 39.1



#12
2-Chlorophenol
Concen: 21.576 ng/ul
RT: 7.575 min Scan# 763
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion: 128 Resp: 629473
Ion Ratio Lower Upper
128 100
64 44.3 37.8 56.8
130 33.0 26.0 39.0





#13

2-Methylphenol

Concen: 20.290 ng/ul

RT: 8.457 min Scan# 91

Delta R.T. -0.006 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Instrument :

BNA_M

ClientSampleId :





SSTD020184

Tgt Ion:108 Resp: 599592

Ion Ratio Lower Upper

108 100

107 90.5 70.7 106.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.780 ng/ul

RT: 8.545 min Scan# 928

Delta R.T. -0.018 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47




Tgt Ion: 45 Resp: 733340


Ion Ratio Lower Upper

45 100

77 16.8 13.1 19.7

79 12.7 9.5 14.3









#16
Acetophenone
Concen: 20.300 ng/ul
RT: 8.839 min Scan# 978
Delta R.T. -0.012 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:105 Resp: 974612
Ion Ratio Lower Upper
105 100
77 76.7 61.6 92.4
51 24.6 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 20.612 ng/ul
RT: 8.828 min Scan# 976
Delta R.T. -0.012 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion: 70 Resp: 477514
Ion Ratio Lower Upper
70 100
42 50.3 43.0 64.4
101 9.8 7.7 11.5
130 20.8 15.1 22.7





#18

4-Methylphenol

Concen: 20.670 ng/ul

RT: 8.787 min Scan# 9

Delta R.T. -0.006 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Instrument :

BNA_M

ClientSampleId :





SSTD020184

Tgt Ion:108 Resp: 660540

Ion Ratio Lower Upper

108 100

107 109.3 89.4 134.0



#19

Hexachloroethane

Concen: 20.345 ng/ul

RT: 9.098 min Scan# 1022

Delta R.T. -0.012 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47




Tgt Ion:117 Resp: 240418


Ion Ratio Lower Upper

117 100

201 87.8 68.7 103.1

199 55.8 43.0 64.6









#22
Nitrobenzene
Concen: 21.570 ng/ul
RT: 9.222 min Scan# 1044
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion: 77 Resp: 710796
Ion Ratio Lower Upper
77 100
123 46.2 34.9 52.3
65 12.8 10.7 16.1



#23
Isophorone
Concen: 19.061 ng/ul
RT: 9.751 min Scan# 1133
Delta R.T. -0.012 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion: 82 Resp: 1280728
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.2
138 16.8 13.5 20.3









#25
2-Nitrophenol
Concen: 23.372 ng/ul
RT: 9.934 min Scan# 1164
Delta R.T. -0.012 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:139 Resp: 320938
Ion Ratio Lower Upper
139 100
65 52.0 40.9 61.3
109 30.8 23.9 35.9



#26
2,4-Dimethylphenol
Concen: 20.913 ng/ul
RT: 9.992 min Scan# 1174
Delta R.T. -0.012 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:107 Resp: 708280
Ion Ratio Lower Upper
107 100
121 49.5 39.4 59.2
122 85.2 69.1 103.7





#27

Bis(2-Chloroethoxy)methane

Concen: 20.280 ng/ul

RT: 10.233 min Scan# 1215

Delta R.T. -0.012 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Instrument :

BNA_M

ClientSampleId :

SSTD020184





Tgt Ion: 93 Resp: 812171

Ion Ratio Lower Upper

93 100

95 32.8 26.2 39.4

123 11.9 9.2 13.8



#29

2,4-Dichlorophenol

Concen: 21.495 ng/ul

RT: 10.469 min Scan# 1255

Delta R.T. -0.006 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47




Tgt Ion:162 Resp: 582218


Ion Ratio Lower Upper

162 100

164 65.4 52.1 78.1

98 37.2 29.1 43.7









#30
Naphthalene
Concen: 20.466 ng/ul
RT: 10.875 min Scan# 1
Delta R.T. -0.012 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:128 Resp: 2044496
Ion Ratio Lower Upper
128 100
129 11.2 8.7 13.1
127 12.9 10.3 15.5



#32
4-Chloroaniline
Concen: 19.330 ng/ul
RT: 10.986 min Scan# 1343
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:127 Resp: 854885
Ion Ratio Lower Upper
127 100
129 32.4 25.4 38.2









#33
Hexachlorobutadiene
Concen: 19.862 ng/ul
RT: 11.163 min Scan# 11
Delta R.T. -0.012 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:225 Resp: 329033
Ion Ratio Lower Upper
225 100
223 62.8 49.8 74.8
227 64.1 51.0 76.6



#34
Caprolactam
Concen: 18.153 ng/ul
RT: 11.769 min Scan# 1476
Delta R.T. 0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:113 Resp: 159492
Ion Ratio Lower Upper
113 100
55 160.2 127.5 191.3
56 122.7 94.6 142.0









#35
4-Chloro-3-methylphenol
Concen: 20.763 ng/ul
RT: 12.104 min Scan# 1533
Delta R.T. -0.000 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:107 Resp: 620806
Ion Ratio Lower Upper
107 100
144 26.1 21.5 32.3
142 79.9 63.9 95.9



#36
2-Methylnaphthalene
Concen: 20.330 ng/ul
RT: 12.480 min Scan# 1597
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:142 Resp: 1353116
Ion Ratio Lower Upper
142 100
141 86.9 70.7 106.1





#37

1-Methylnaphthalene

Concen: 20.224 ng/ul

RT: 12.698 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Instrument :

BNA_M

ClientSampleId :

SSTD020184





Tgt Ion:142 Resp: 1353164

Ion Ratio Lower Upper

142 100

141 92.1 73.4 110.2

116 3.8 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.743 ng/ul

RT: 12.839 min Scan# 1658

Delta R.T. -0.012 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Tgt Ion:216 Resp: 650790




Ion Ratio Lower Upper


216 100

214 79.4 62.4 93.6

179 21.2 16.7 25.1

108 20.1 15.4 23.0





#40

Hexachlorocyclopentadiene

Concen: 14.442 ng/ul

RT: 12.822 min Scan# 1

Delta R.T. -0.012 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Instrument :

BNA_M

ClientSampleId :

SSTD020184





Tgt Ion:237 Resp: 236129

Ion Ratio Lower Upper

237 100

235 62.0 50.9 76.3

272 12.7 10.3 15.5



#41

2,4,6-Trichlorophenol

Concen: 21.837 ng/ul

RT: 13.080 min Scan# 1699

Delta R.T. -0.006 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47




Tgt Ion:196 Resp: 419474


Ion Ratio Lower Upper

196 100

198 96.7 77.3 115.9

200 31.4 25.2 37.8









#42
2,4,5-Trichlorophenol
Concen: 22.088 ng/ul
RT: 13.151 min Scan# 1711
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:196 Resp: 454912
Ion Ratio Lower Upper
196 100
198 95.8 75.6 113.4
200 31.4 24.7 37.1



#43
1,1'-Biphenyl
Concen: 21.069 ng/ul
RT: 13.480 min Scan# 1767
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:154 Resp: 1721536
Ion Ratio Lower Upper
154 100
153 39.6 31.4 47.2
76 12.9 10.4 15.6









#44
2-Chloronaphthalene
Concen: 21.023 ng/ul
RT: 13.522 min Scan# 1775
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:162 Resp: 1347950
Ion Ratio Lower Upper
162 100
164 32.5 26.2 39.2
127 39.6 32.6 49.0



#45
2-Nitroaniline
Concen: 23.290 ng/ul
RT: 13.722 min Scan# 1808
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion: 65 Resp: 367559
Ion Ratio Lower Upper
65 100
92 71.3 54.4 81.6
138 106.2 86.1 129.1









#47
Dimethylphthalate
Concen: 19.946 ng/ul
RT: 14.098 min Scan# 1872
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:163 Resp: 1581748
Ion Ratio Lower Upper
163 100
194 3.8 3.0 4.6
164 10.2 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 22.951 ng/ul
RT: 14.216 min Scan# 1892
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:165 Resp: 295908
Ion Ratio Lower Upper
165 100
89 66.0 52.8 79.2
121 23.6 19.4 29.0









#50
Acenaphthylene
Concen: 20.420 ng/ul
RT: 14.363 min Scan# 1917
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:152 Resp: 2070286
Ion Ratio Lower Upper
152 100
151 20.0 15.7 23.5
153 13.1 10.5 15.7



#51
3-Nitroaniline
Concen: 22.512 ng/ul
RT: 14.545 min Scan# 1948
Delta R.T. -0.000 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:138 Resp: 348100
Ion Ratio Lower Upper
138 100
108 11.4 12.5 18.7#
92 114.9 95.4 143.2









#52
Acenaphthene
Concen: 20.272 ng/ul
RT: 14.704 min Scan# 1975
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:153 Resp: 1438026
Ion Ratio Lower Upper
153 100
152 48.2 38.2 57.2
154 89.1 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 17.575 ng/ul
RT: 14.745 min Scan# 1982
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:184 Resp: 100542
Ion Ratio Lower Upper
184 100
63 71.0 58.8 88.2
154 63.9 55.0 82.4









#55
4-Nitrophenol
Concen: 18.668 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. -0.000 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:109 Resp: 231392
Ion Ratio Lower Upper
109 100
139 132.2 102.2 153.4
65 127.3 105.7 158.5



#56
Dibenzofuran
Concen: 20.591 ng/ul
RT: 15.033 min Scan# 2031
Delta R.T. -0.012 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:168 Resp: 1961633
Ion Ratio Lower Upper
168 100
139 38.4 29.7 44.5





#57

2,4-Dinitrotoluene

Concen: 22.964 ng/ul

RT: 14.998 min Scan# 20

Delta R.T. -0.006 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Instrument :

BNA_M

ClientSampleId :

SSTD020184





Tgt Ion:165 Resp: 430437

Ion Ratio Lower Upper

165 100

63 45.4 39.7 59.5

182 2.6 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 20.608 ng/ul

RT: 15.257 min Scan# 2069

Delta R.T. -0.006 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Tgt Ion:232 Resp: 355282




Ion Ratio Lower Upper


232 100

131 52.4 39.1 58.7

130 2.6 1.8 2.6

166 33.5 25.4 38.0









#59
Diethylphthalate
Concen: 19.676 ng/ul
RT: 15.457 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:149 Resp: 1561954
Ion Ratio Lower Upper
149 100
177 21.1 17.3 25.9
150 12.3 10.1 15.1



#61
Fluorene
Concen: 20.397 ng/ul
RT: 15.680 min Scan# 2141
Delta R.T. -0.012 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:166 Resp: 1654961
Ion Ratio Lower Upper
166 100
165 97.9 78.5 117.7
167 13.6 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 19.913 ng/ul
RT: 15.674 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:204 Resp: 772693
Ion Ratio Lower Upper
204 100
206 32.4 25.7 38.5
141 63.8 49.4 74.0



#63
4-Nitroaniline
Concen: 23.664 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:138 Resp: 355526
Ion Ratio Lower Upper
138 100
92 52.0 42.1 63.1
108 81.0 66.2 99.2









#66
4,6-Dinitro-2-methylphenol
Concen: 18.515 ng/ul
RT: 15.757 min Scan# 2154
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:198 Resp: 181697
Ion Ratio Lower Upper
198 100
182 4.8 4.2 6.2
77 28.7 24.0 36.0



#67
N-Nitrosodiphenylamine
Concen: 20.728 ng/ul
RT: 15.886 min Scan# 2176
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:169 Resp: 1341933
Ion Ratio Lower Upper
169 100
168 67.2 54.6 81.8
167 37.7 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 18.984 ng/ul
RT: 16.568 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:248 Resp: 411297
Ion Ratio Lower Upper
248 100
250 97.3 78.9 118.3
141 85.8 65.7 98.5



#69
Hexachlorobenzene
Concen: 19.008 ng/ul
RT: 16.680 min Scan# 2311
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:284 Resp: 487082
Ion Ratio Lower Upper
284 100
142 42.9 30.8 46.2
249 31.9 24.2 36.2









#70
Atrazine
Concen: 18.593 ng/ul
RT: 16.839 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:200 Resp: 424748
Ion Ratio Lower Upper
200 100
173 27.4 21.8 32.6
215 46.2 39.4 59.2



#71
Pentachlorophenol
Concen: 17.083 ng/ul
RT: 17.021 min Scan# 2369
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:266 Resp: 260021
Ion Ratio Lower Upper
266 100
264 63.7 49.7 74.5
268 63.4 51.8 77.6





#72

Phenanthrene

Concen: 20.365 ng/ul

RT: 17.421 min Scan# 2437

Delta R.T. -0.006 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Instrument :

BNA_M

ClientSampleId :

SSTD020184





Tgt Ion:178 Resp: 2465620

Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.2

176 19.3 15.7 23.5



#74

Anthracene

Concen: 20.222 ng/ul

RT: 17.509 min Scan# 2452

Delta R.T. -0.006 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47




Tgt Ion:178 Resp: 2488537


Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 19.1 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.351 ng/uL
RT: 13.445 min Scan# 1761
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:216 Resp: 641899
Ion Ratio Lower Upper
216 100
214 78.1 62.6 94.0
218 48.2 38.2 57.2



#76
Pentachlorobenzene
Concen: 20.564 ng/uL
RT: 14.957 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:250 Resp: 668670
Ion Ratio Lower Upper
250 100
248 62.0 50.2 75.2
252 62.6 50.5 75.7









#77
 Carbazole
 Concen: 19.500 ng/ul
 RT: 17.780 min Scan# 2498
 Delta R.T. -0.000 min
 Lab File: BM036670.D
 Acq: 23 Sep 2022 10:47




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020184


Tgt Ion:167 Resp: 2221713
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.2 25.8
 139 13.1 10.6 15.8



#78
 Di-n-butylphthalate
 Concen: 19.869 ng/ul
 RT: 18.345 min Scan# 2594
 Delta R.T. -0.006 min
 Lab File: BM036670.D
 Acq: 23 Sep 2022 10:47

Tgt Ion:149 Resp: 2565903
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.2 10.8
 104 5.4 4.3 6.5









#80
Fluoranthene
Concen: 25.640 ng/ul
RT: 19.427 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47




Instrument :
BNA_M
ClientSampleId :
SSTD020184


Tgt Ion:202 Resp: 2589418
Ion Ratio Lower Upper
202 100
101 11.7 9.8 14.6
100 8.8 7.4 11.2



#82
Pyrene
Concen: 25.131 ng/ul
RT: 19.786 min Scan# 2839
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:202 Resp: 2686542
Ion Ratio Lower Upper
202 100
101 13.8 10.9 16.3
100 11.6 8.7 13.1





#83

Butylbenzylphthalate

Concen: 24.490 ng/ul

RT: 20.686 min Scan# 2992

Delta R.T. -0.000 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Instrument :

BNA_M

ClientSampleId :

SSTD020184





Tgt Ion:149 Resp: 992030

Ion Ratio Lower Upper

149 100

91 73.7 55.2 82.8

206 19.8 16.2 24.2



#84

3,3'-Dichlorobenzidine

Concen: 20.121 ng/ul

RT: 21.462 min Scan# 3124

Delta R.T. -0.000 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47




Tgt Ion:252 Resp: 658604


Ion Ratio Lower Upper

252 100

254 65.0 53.0 79.4

126 13.0 10.4 15.6





#85

Benzo(a)anthracene

Concen: 21.004 ng/ul

RT: 21.533 min Scan# 3136

Delta R.T. -0.000 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Instrument :

BNA_M

ClientSampleId :

SSTD020184





Tgt Ion:228 Resp: 2204612

Ion Ratio Lower Upper

228 100

229 19.6 15.8 23.8

226 27.2 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 24.865 ng/ul

RT: 21.462 min Scan# 3124

Delta R.T. -0.006 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47




Tgt Ion:149 Resp: 1471296


Ion Ratio Lower Upper

149 100

167 26.8 22.0 33.0

279 3.8 3.3 4.9





#87

Chrysene

Concen: 20.545 ng/ul

RT: 21.586 min Scan# 3145

Delta R.T. -0.000 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Instrument :

BNA_M

ClientSampleId :

SSTD020184





Tgt Ion:228 Resp: 2046847

Ion Ratio Lower Upper

228 100

226 30.5 24.4 36.6

229 19.9 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 25.568 ng/ul




RT: 22.403 min Scan# 3284


Delta R.T. -0.006 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Tgt Ion:149 Resp: 2190016





#90

Benzo(b)fluoranthene

Concen: 21.769 ng/ul

RT: 23.244 min Scan# 3427

Delta R.T. -0.000 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Instrument :

BNA_M

ClientSampleId :

SSTD020184





Tgt Ion:252 Resp: 1961342

Ion Ratio Lower Upper

252 100

253 21.5 17.6 26.4

125 10.8 9.1 13.7



#91

Benzo(k)fluoranthene

Concen: 20.749 ng/ul

RT: 23.291 min Scan# 3435

Delta R.T. -0.006 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47




Tgt Ion:252 Resp: 1868004


Ion Ratio Lower Upper

252 100

253 21.8 17.8 26.8

125 11.2 8.2 12.4





#93

Benzo(a)pyrene

Concen: 20.757 ng/ul

RT: 23.880 min Scan# 3

Delta R.T. -0.006 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47

Instrument :

BNA_M

ClientSampleId :

SSTD020184





Tgt Ion:252 Resp: 1644083

Ion Ratio Lower Upper

252 100

253 22.2 17.5 26.3

125 12.5 9.5 14.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.885 ng/ul

RT: 26.532 min Scan# 3986

Delta R.T. -0.000 min

Lab File: BM036670.D

Acq: 23 Sep 2022 10:47




Tgt Ion:276 Resp: 1839233


Ion Ratio Lower Upper

276 100

138 28.3 20.2 30.4

277 25.7 19.7 29.5









#95
Dibenzo(a,h)anthracene
Concen: 20.108 ng/ul
RT: 26.550 min Scan# 3990
Delta R.T. -0.006 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Instrument :
BNA_M
ClientSampleId :
SSTD020184

Tgt Ion:278 Resp: 1686676
Ion Ratio Lower Upper
278 100
139 18.9 14.5 21.7
279 23.9 19.7 29.5



#96
Benzo(g,h,i)perylene
Concen: 19.800 ng/ul
RT: 27.309 min Scan# 4118
Delta R.T. -0.012 min
Lab File: BM036670.D
Acq: 23 Sep 2022 10:47

Tgt Ion:276 Resp: 1587641
Ion Ratio Lower Upper
276 100
138 24.8 18.7 28.1
277 23.6 19.2 28.8

