

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092322\
 Data File : BM036692.D
 Acq On : 24 Sep 2022 00:47
 Operator : CG/JU
 Sample : N4729-03
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EXFJ7

Quant Time: Sep 24 01:24:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 24 00:41:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.010	152	370105	20.000	ng/ul	0.00
20) Naphthalene-d8	10.822	136	1636676	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.633	164	1004684	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.374	188	1843175	20.000	ng/ul	0.00
79) Chrysene-d12	21.538	240	1347284	20.000	ng/ul	0.00
88) Perylene-d12	23.974	264	1214730	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	66289	6.714	ng/uL	0.00
4) Pyridine-d5	3.822	84	313955	10.904	ng/ul	0.00
7) Phenol-d5	7.169	99	470033	14.211	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.339	67	732137	36.450	ng/ul	0.00
11) 2-Chlorophenol-d4	7.539	132	848760	35.644	ng/ul	0.00
15) 4-Methylphenol-d8	8.722	113	672379	27.513	ng/ul	0.00
21) Nitrobenzene-d5	9.175	128	481802	43.492	ng/ul	0.00
24) 2-Nitrophenol-d4	9.898	143	471293	47.496	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.439	165	832987	36.224	ng/ul	0.00
31) 4-Chloroaniline-d4	10.957	131	1036859	27.119	ng/ul	0.00
46) Dimethylphthalate-d6	14.051	166	2674041	39.463	ng/ul	0.00
49) Acenaphthylene-d8	14.333	160	3170356	39.444	ng/ul	0.00
54) 4-Nitrophenol-d4	14.821	143	138168	10.513	ng/ul	-0.01
60) Fluorene-d10	15.621	176	2344799	38.614	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.733	200	234951	32.219	ng/ul	0.00
73) Anthracene-d10	17.468	188	3364394	39.922	ng/ul	0.00
81) Pyrene-d10	19.750	212	3390302	51.150	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.815	264	2535036	42.716	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092322\
Data File : BM036692.D
Acq On : 24 Sep 2022 00:47
Operator : CG/JU
Sample : N4729-03
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EXFJ7

Quant Time: Sep 24 01:24:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 24 00:41:16 2022
Response via : Initial Calibration

