



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092322\
 Data File : BM036669.D
 Acq On : 23 Sep 2022 09:54
 Operator : CG/JU
 Sample : DFTPP079
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP079

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Title : SVOA CALIBRATION
 Last Update : Mon Sep 26 23:10:22 2022



AutoFind: Scans 2503, 2504, 2505; Background Corrected with Scan 2497

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.0	239307	PASS
68	69	0.00	2	1.5	3848	PASS
69	198	0.00	100	38.2	261192	PASS
70	69	0.00	2	0.5	1393	PASS
127	198	10	80	49.8	340309	PASS
197	198	0.00	2	0.4	2585	PASS
198	198	100	100	100.0	683477	PASS
199	198	5	9	6.7	45675	PASS
275	198	10	60	23.4	159829	PASS
365	198	1	100	2.7	18618	PASS
441	198	0.01	100	9.8	66896	PASS
442	198	50	100	61.5	420395	PASS
443	442	15	24	19.4	81480	PASS

m/z	Abundance
35.80	272.0
37.90	287.0
38.90	473.0
39.90	1314.0
40.90	529.0
42.00	160.0
43.00	410.0
43.90	3144.0
44.90	360.0
46.90	197.0
48.90	999.0
49.90	228.0
50.90	416.0
52.90	176.0
54.90	511.0
57.00	220.0
68.00	240.0
68.90	184.0
73.00	1814.0
74.00	237.0
75.00	657.0
77.00	1228.0
77.80	396.0
79.00	301.0
80.90	452.0
83.90	730.0
85.80	515.0
87.00	236.0
88.90	304.0
90.90	287.0
91.90	182.0
96.00	1212.0
96.90	272.0
103.00	363.0
105.00	205.0
107.00	153.0
114.90	298.0
115.90	174.0
116.90	168.0
118.80	453.0
130.90	153.0
132.90	1228.0
134.00	230.0
135.00	351.0
146.90	578.0
160.90	197.0
162.80	275.0
165.00	278.0
166.90	213.0
176.90	433.0
177.80	216.0
190.90	1606.0
191.90	234.0
193.00	968.0
193.90	206.0
204.80	277.0
206.90	12698.0
207.90	2450.0
208.90	1610.0
210.00	349.0
221.10	205.0
250.80	163.0
264.90	223.0
266.90	329.0
280.90	1625.0
282.00	486.0
283.00	364.0
283.70	178.0
355.10	184.0

Instrument :
BNA_M
ClientSampleId :
DFTPP079

m/z	Abundance
36.00	581.0
37.00	1094.0
38.00	2418.0
39.00	19144.0
40.00	2290.0
41.00	672.0
42.00	195.0
43.00	358.0
43.90	3496.0
44.90	614.0
46.90	385.0
47.90	286.0
49.00	2470.0
50.00	67560.0
51.00	264384.0
52.00	14348.0
53.00	402.0
54.90	1887.0
56.00	7549.0
57.00	17920.0
58.10	705.0
58.90	546.0
60.10	233.0
61.00	3706.0
62.00	4370.0
63.00	11893.0
64.00	1974.0
65.00	5514.0
65.80	358.0
67.00	625.0
68.00	4637.0
68.90	286144.0
70.00	1513.0
71.00	331.0
73.00	4353.0
74.00	29760.0
75.00	48440.0
76.00	16576.0
77.00	357056.0
78.00	23992.0
79.00	20416.0
79.90	16187.0
81.00	22960.0
82.00	5398.0
82.90	5859.0
83.90	1469.0
84.90	4148.0
86.00	6583.0
87.00	3026.0
87.90	1469.0
89.00	777.0
90.00	151.0
91.00	5337.0
92.00	5864.0
92.90	37096.0
93.90	2441.0
95.00	660.0
96.00	3137.0
96.90	790.0
98.00	28712.0
98.90	25680.0
99.90	2481.0
100.90	13302.0
101.90	944.0
102.90	4177.0
103.90	8197.0
104.90	7903.0
105.90	2707.0
107.00	102312.0
108.00	16071.0
109.00	2521.0
109.90	182336.0
110.90	29504.0
112.00	3390.0
112.90	1086.0
114.00	190.0
114.90	562.0
116.00	5304.0
116.90	89104.0
118.00	6705.0
118.90	1258.0
120.00	1112.0
120.80	544.0
121.90	6118.0
123.00	10210.0
123.90	4482.0
125.00	4023.0
126.00	1309.0
127.00	352704.0
127.90	25904.0
128.90	129168.0
129.90	12296.0

Instrument :
BNA_M
ClientSampleId :
DFTPP079

130.90	2803.0
132.00	1454.0
132.90	1689.0
133.90	3562.0
135.00	11499.0
135.90	4105.0
137.00	7664.0
137.90	1467.0
138.90	676.0
139.90	1383.0
141.00	15640.0
142.00	4877.0
142.90	4432.0
144.00	913.0
144.90	884.0
145.90	2712.0
146.90	9038.0
147.90	20600.0
149.00	5153.0
150.00	1658.0
151.00	2347.0
151.80	1454.0
152.90	5257.0
153.90	4153.0
155.00	9390.0
156.00	12953.0
156.90	2401.0
157.90	2681.0
158.90	1877.0
159.90	5330.0
161.00	8477.0
161.90	2235.0
163.00	902.0
164.00	630.0
164.90	5638.0
165.90	4837.0
166.90	35360.0
167.90	26672.0
168.90	4093.0
170.00	1299.0
170.90	1383.0
171.90	3030.0
173.00	3996.0
174.00	6744.0
175.00	13678.0
175.90	3560.0
176.90	6630.0
177.90	2206.0
178.90	24632.0
179.90	20128.0
180.90	7850.0
181.90	1412.0
182.80	621.0
183.90	2082.0
184.90	11641.0
186.00	88536.0
186.90	25632.0
188.00	3936.0
188.90	4793.0
190.00	828.0
190.90	3388.0
191.90	8188.0
193.00	9287.0
194.00	2119.0
195.00	895.0
195.90	17064.0
196.70	2702.0
197.90	652480.0
198.90	44872.0
199.90	3299.0
201.30	2299.0
202.90	4141.0
204.00	24200.0
205.00	38240.0
206.00	164480.0
207.00	37072.0
208.00	8044.0
208.90	3360.0
210.10	3053.0
211.00	7140.0
213.00	522.0
213.80	180.0
214.90	1614.0
215.90	3753.0
216.90	42144.0
217.90	5404.0
218.90	556.0
221.00	25376.0
221.70	6452.0
222.90	9343.0
224.00	86976.0
225.00	23856.0
226.00	2781.0
226.90	35656.0
227.90	5848.0
228.90	8468.0

Instrument :
BNA_M
ClientSampleId :
DFTPP079

229.90	831.0
231.00	3565.0
232.00	533.0
233.00	532.0
234.00	2211.0
234.90	2426.0
235.90	1801.0
236.90	5290.0
237.90	802.0
238.90	1303.0
239.90	1548.0
240.90	2538.0
242.00	4737.0
243.00	4174.0
244.00	69416.0
245.00	9373.0
245.90	11848.0
246.90	2610.0
247.80	471.0
248.90	2856.0
249.80	526.0
250.90	684.0
252.00	544.0
252.90	1691.0
253.90	1593.0
255.00	331712.0
255.90	48096.0
257.00	4570.0
257.90	18040.0
259.00	2968.0
259.90	569.0
260.90	461.0
263.00	190.0
264.00	342.0
264.90	7521.0
265.90	1208.0
266.90	566.0
269.00	266.0
269.80	266.0
270.90	1011.0
272.00	821.0
272.90	8891.0
273.90	26048.0
275.00	139776.0
275.90	19712.0
276.90	12272.0
278.00	2897.0
278.90	568.0
281.00	2142.0
282.00	930.0
283.00	1876.0
283.90	1051.0
285.00	2207.0
285.90	503.0
288.90	447.0
290.00	382.0
290.90	410.0
291.90	553.0
292.90	3091.0
294.00	931.0
294.90	874.0
295.90	38600.0
296.90	5823.0
297.80	340.0
300.90	420.0
301.90	740.0
302.90	4417.0
303.90	1124.0
305.00	272.0
307.90	784.0
308.90	455.0
309.90	597.0
310.80	159.0
311.80	179.0
312.90	359.0
313.90	1776.0
314.90	4422.0
315.90	2101.0
316.90	582.0
320.90	1307.0
321.90	487.0
323.00	11897.0
324.00	2539.0
325.00	328.0
325.70	202.0
326.90	2337.0
327.90	921.0
328.80	176.0
332.00	1136.0
332.80	912.0
333.90	7673.0
334.90	1979.0
335.80	403.0
339.90	168.0
341.00	1472.0
342.00	422.0

Instrument :
BNA_M
ClientSampleId :
DFTPP079

345.90	2541.0
346.80	558.0
350.80	169.0
351.90	3258.0
353.00	2608.0
354.00	3353.0
355.00	987.0
359.00	292.0
364.90	14886.0
365.90	2357.0
366.90	274.0
370.00	372.0
371.00	802.0
372.00	5228.0
373.00	1451.0
373.90	168.0
382.90	1512.0
383.90	544.0
385.10	226.0
389.90	739.0
390.80	413.0
391.90	230.0
400.90	351.0
402.00	2010.0
403.00	3028.0
404.00	1233.0
405.00	344.0
415.00	150.0
420.90	2326.0
422.00	2657.0
423.00	16968.0
424.00	7795.0
425.00	1214.0
441.00	46808.0
442.00	297344.0
443.00	59280.0
444.00	4945.0
444.90	288.0

Instrument :
BNA_M
ClientSampleId :
DFTPP079

m/z	Abundance
36.00	597.0
36.90	1205.0
38.00	3071.0
39.00	18792.0
39.90	2419.0
40.90	994.0
42.00	290.0
42.90	610.0
43.90	3426.0
44.90	1155.0
46.90	337.0
48.00	269.0
48.90	2647.0
50.00	72680.0
51.00	285056.0
52.00	15185.0
52.90	810.0
55.00	1588.0
56.00	8903.0
57.00	19536.0
58.00	1075.0
58.90	490.0
59.90	345.0
60.90	3761.0
62.00	4473.0
63.00	11876.0
63.90	1712.0
65.00	6057.0
66.00	480.0
67.00	488.0
68.00	4868.0
68.90	314432.0
69.90	1673.0
71.00	400.0
73.00	4590.0
74.00	33640.0
75.00	52544.0
76.00	18704.0
77.00	393728.0
78.00	27104.0
79.00	22424.0
79.90	18136.0
81.00	26696.0
82.00	6455.0
83.00	6370.0
83.90	1174.0
85.00	4739.0
85.90	7897.0
86.90	3338.0
87.90	1689.0
89.00	854.0
89.80	160.0
91.00	5844.0
92.00	6863.0
93.00	42192.0
93.90	3052.0
94.90	902.0
96.00	2932.0
96.90	981.0
97.90	32024.0
99.00	28760.0
99.90	2435.0
101.00	16245.0
101.90	860.0
103.00	4765.0
103.90	9125.0
105.00	8018.0
105.90	2597.0
107.00	117824.0
108.00	18208.0
109.00	2769.0
109.90	209472.0
111.00	33632.0
112.00	4196.0
112.90	1290.0
113.80	198.0
114.80	821.0
115.90	5552.0
116.90	104256.0
117.90	8210.0
119.00	1241.0
119.90	1596.0
121.00	751.0
121.90	7844.0
123.00	12067.0
123.90	5411.0
125.00	4898.0
126.00	1203.0
127.00	412736.0
128.00	30776.0
128.90	156032.0
129.90	13973.0

Instrument :
BNA_M
ClientSampleId :
DFTPP079

130.90	2581.0
131.90	1166.0
132.90	2560.0
133.90	4581.0
134.90	12375.0
136.00	5424.0
136.90	8553.0
138.00	1408.0
138.90	830.0
139.90	1986.0
141.00	19728.0
142.00	6363.0
142.90	4682.0
144.00	1255.0
144.90	1185.0
146.00	3286.0
146.90	11302.0
147.90	25112.0
149.00	6089.0
149.90	1612.0
151.00	2412.0
151.80	1369.0
152.90	6099.0
153.90	4278.0
154.90	11549.0
156.00	17872.0
157.00	3283.0
157.90	3960.0
159.00	2645.0
160.00	5728.0
161.00	10193.0
161.90	2819.0
162.90	1209.0
163.90	912.0
164.90	7170.0
166.00	5524.0
166.90	41104.0
168.00	29648.0
169.00	4102.0
170.00	1374.0
170.90	1567.0
171.90	3957.0
172.90	4399.0
173.90	8562.0
175.00	16752.0
175.90	4426.0
177.00	7631.0
177.90	2189.0
178.90	29616.0
180.00	22872.0
180.90	10190.0
181.90	1529.0
182.90	855.0
183.90	2094.0
185.00	13914.0
186.00	112752.0
187.00	31720.0
187.90	3886.0
189.00	5952.0
189.90	1209.0
191.00	4640.0
191.90	9430.0
193.00	12281.0
194.00	2987.0
195.00	946.0
195.90	23640.0
196.80	2882.0
197.90	833024.0
198.90	56008.0
199.90	4525.0
201.40	3697.0
202.90	4959.0
204.00	29576.0
205.00	49464.0
206.00	214656.0
207.00	41888.0
208.00	9757.0
209.00	4323.0
210.00	3223.0
211.00	7917.0
212.90	497.0
213.90	353.0
214.90	2450.0
215.90	4025.0
216.90	55928.0
217.90	6526.0
219.00	578.0
221.00	30752.0
221.80	8094.0
222.90	11882.0
224.00	113904.0
225.00	28832.0
225.90	3742.0
226.90	45640.0
227.90	6616.0
228.90	9802.0

Instrument :
BNA_M
ClientSampleId :
DFTPP079

229.90	1244.0
230.90	4020.0
231.90	684.0
232.90	645.0
233.90	3261.0
235.00	3008.0
235.90	2249.0
237.00	5341.0
237.90	821.0
238.90	1946.0
239.90	1577.0
241.00	2430.0
242.00	6486.0
243.00	6573.0
244.00	92600.0
245.00	11232.0
246.00	16480.0
246.90	3430.0
248.00	860.0
248.90	3096.0
250.00	606.0
250.90	757.0
251.90	610.0
252.90	1910.0
254.00	2404.0
255.00	445632.0
256.00	65176.0
257.00	6168.0
257.90	25696.0
258.90	3929.0
259.90	647.0
260.90	655.0
262.90	366.0
263.90	837.0
264.90	10616.0
265.90	1547.0
266.90	391.0
267.90	162.0
269.80	488.0
271.00	944.0
271.90	1064.0
272.90	12975.0
273.90	35568.0
275.00	194624.0
276.00	25160.0
276.90	14975.0
278.00	3579.0
278.80	633.0
280.00	156.0
281.00	2271.0
282.00	962.0
282.90	2055.0
284.00	1375.0
285.00	2834.0
285.90	690.0
287.80	228.0
288.90	738.0
289.90	844.0
290.90	523.0
292.00	702.0
292.90	3786.0
293.90	1007.0
294.90	959.0
295.90	53288.0
297.00	7435.0
297.90	406.0
300.90	598.0
301.90	1067.0
303.00	6527.0
303.90	1534.0
304.90	391.0
308.00	574.0
309.00	858.0
309.90	618.0
311.90	160.0
312.90	451.0
314.00	2445.0
314.90	5885.0
315.90	3225.0
317.00	647.0
319.90	193.0
320.90	1508.0
322.00	764.0
323.00	17736.0
324.00	3092.0
324.90	361.0
326.00	500.0
326.90	3828.0
327.90	2090.0
328.90	407.0
331.90	1537.0
333.00	1239.0
334.00	11359.0
335.00	2519.0
336.00	471.0
338.90	354.0

Instrument :
BNA_M
ClientSampleId :
DFTPP079

340.00	277.0
340.90	2179.0
342.00	507.0
345.90	4006.0
346.90	1045.0
351.00	243.0
351.90	4930.0
353.00	3056.0
353.90	4943.0
355.00	1470.0
359.00	426.0
363.90	187.0
364.90	22416.0
365.90	3196.0
366.80	325.0
370.00	452.0
370.90	1465.0
372.00	8439.0
373.00	2201.0
373.80	291.0
376.90	230.0
383.00	2624.0
383.90	753.0
390.00	1195.0
390.90	857.0
392.00	486.0
396.90	174.0
400.90	490.0
401.90	3377.0
403.00	4964.0
403.90	1523.0
405.00	388.0
414.80	258.0
421.00	4059.0
422.00	3931.0
423.00	28320.0
424.00	9713.0
425.00	1334.0
441.00	80752.0
442.00	508416.0
443.00	100128.0
444.00	8689.0
444.90	630.0

Instrument :
BNA_M
ClientSampleId :
DFTPP079

m/z	Abundance
35.90	390.0
37.00	801.0
38.00	1978.0
39.00	11829.0
40.00	1922.0
41.10	751.0
42.00	168.0
42.90	571.0
43.90	3349.0
45.00	583.0
46.90	389.0
47.90	166.0
48.90	2006.0
50.00	42504.0
51.00	169728.0
52.00	8283.0
53.00	541.0
55.00	1367.0
56.00	5136.0
57.00	12048.0
57.90	500.0
59.00	503.0
61.00	2291.0
62.00	2392.0
63.00	7257.0
63.90	1103.0
65.00	3740.0
65.90	453.0
67.00	365.0
67.90	2759.0
68.90	183552.0
69.90	993.0
71.00	405.0
73.00	3681.0
74.00	19864.0
75.00	32248.0
76.00	11369.0
77.00	237184.0
78.00	16568.0
78.90	13210.0
80.00	10605.0
81.00	15437.0
81.90	4496.0
83.00	3379.0
83.90	1105.0
84.90	3002.0
85.90	5467.0
87.00	2480.0
88.00	841.0
88.90	698.0
91.00	3609.0
92.00	4129.0
93.00	26040.0
94.00	1995.0
94.90	592.0
96.00	2995.0
96.90	773.0
97.90	20480.0
99.00	18424.0
99.90	1552.0
101.00	9210.0
102.00	560.0
102.90	2996.0
103.90	5831.0
105.00	5530.0
105.90	1771.0
107.00	72480.0
108.00	11084.0
109.00	1664.0
110.00	126368.0
111.00	19288.0
111.90	2392.0
112.90	842.0
113.80	273.0
115.00	625.0
115.90	4131.0
116.90	65488.0
118.00	5239.0
119.00	973.0
119.90	1060.0
120.90	462.0
121.90	4679.0
122.90	7714.0
123.90	2901.0
124.90	3765.0
126.00	690.0
127.00	255488.0
127.90	18696.0
128.90	97976.0
129.90	7698.0
130.90	2031.0
131.90	741.0

Instrument :
BNA_M
ClientSampleId :
DFTPP079

132.90	1679.0
133.90	2441.0
134.90	8317.0
135.80	3519.0
136.90	5378.0
137.90	1034.0
139.00	549.0
139.90	1233.0
140.90	12040.0
142.00	4245.0
142.90	2845.0
144.00	848.0
144.80	665.0
145.90	2640.0
146.90	7718.0
147.90	16358.0
148.90	3953.0
149.90	1125.0
151.00	1554.0
151.80	848.0
153.00	4112.0
153.90	2664.0
155.00	6844.0
156.00	11074.0
157.00	1706.0
157.90	2422.0
158.80	2136.0
159.90	4151.0
160.90	6322.0
161.90	1674.0
162.90	941.0
163.90	650.0
164.90	4493.0
166.00	4088.0
167.00	28440.0
167.90	18984.0
168.90	2324.0
169.90	675.0
170.80	1446.0
171.90	2709.0
172.90	2879.0
173.90	4978.0
175.00	10812.0
175.90	2842.0
176.90	5150.0
177.90	1584.0
178.90	20376.0
179.90	13590.0
181.00	6198.0
181.90	902.0
182.80	531.0
183.90	1900.0
184.90	9620.0
186.00	73928.0
187.00	20776.0
188.00	2326.0
188.90	4638.0
189.90	870.0
191.00	4005.0
192.00	6714.0
192.90	8346.0
193.90	1315.0
194.90	870.0
195.90	14407.0
196.90	2172.0
197.90	564928.0
198.90	36144.0
199.90	2920.0
201.40	2362.0
202.90	3949.0
204.00	21056.0
205.00	34368.0
206.00	143552.0
207.00	33248.0
208.00	7923.0
209.00	3166.0
210.00	2439.0
211.00	4860.0
212.90	312.0
213.80	219.0
214.90	1428.0
215.90	3314.0
216.90	37328.0
217.90	5233.0
218.90	598.0
221.00	21312.0
221.80	4922.0
223.00	7582.0
224.00	81912.0
225.00	19424.0
226.00	2344.0
226.90	32928.0
228.00	3745.0
228.90	7036.0
229.90	972.0
231.00	2525.0

Instrument :
BNA_M
ClientSampleId :
DFTPP079

231.90	789.0
232.90	423.0
233.90	2326.0
235.00	2565.0
235.90	1345.0
237.00	3928.0
237.80	598.0
239.00	1265.0
240.00	1284.0
241.00	1813.0
241.90	4228.0
243.00	4127.0
244.00	68840.0
245.00	8701.0
245.90	11973.0
246.80	2380.0
247.80	537.0
249.00	2473.0
249.90	472.0
250.90	865.0
252.00	498.0
252.90	1277.0
253.90	1432.0
254.90	324672.0
255.90	46976.0
256.90	3805.0
257.90	17640.0
258.90	3345.0
259.80	608.0
260.90	616.0
262.80	218.0
263.90	347.0
264.90	7583.0
265.90	1126.0
266.90	424.0
267.90	175.0
269.90	433.0
270.90	801.0
271.80	834.0
272.90	9494.0
273.90	25776.0
275.00	145088.0
276.00	19328.0
277.00	11239.0
277.90	2276.0
279.00	519.0
280.90	2099.0
282.00	1089.0
283.00	1982.0
284.00	1142.0
285.00	2086.0
285.90	357.0
288.00	198.0
288.90	402.0
290.00	443.0
291.00	293.0
291.80	530.0
293.00	2914.0
294.00	857.0
294.90	891.0
295.90	42184.0
296.90	5507.0
297.80	415.0
299.00	155.0
300.80	770.0
301.90	738.0
303.00	4883.0
303.90	1386.0
305.00	206.0
308.00	687.0
308.90	405.0
309.90	408.0
310.90	151.0
312.90	400.0
313.90	1803.0
314.90	4555.0
315.90	2428.0
316.90	336.0
320.00	281.0
320.90	1269.0
321.90	591.0
323.00	13836.0
324.00	2302.0
325.00	263.0
326.00	274.0
326.90	2881.0
327.80	1201.0
328.80	326.0
331.90	885.0
332.90	1027.0
334.00	9024.0
334.90	2106.0
336.00	228.0
339.90	226.0
341.00	1603.0
342.00	445.0

Instrument :
BNA_M
ClientSampleId :
DFTPP079

343.00	156.0
345.90	3297.0
346.90	521.0
350.90	178.0
352.00	3860.0
353.00	3200.0
354.00	3897.0
355.00	1077.0
357.00	196.0
358.90	303.0
364.90	18552.0
365.90	2340.0
366.80	189.0
370.00	282.0
370.90	1333.0
372.00	7410.0
373.00	1787.0
373.90	185.0
376.90	296.0
382.90	2269.0
384.00	600.0
385.00	261.0
389.90	991.0
391.00	491.0
392.00	381.0
400.90	515.0
401.90	2968.0
403.00	5029.0
403.90	1449.0
404.90	331.0
421.00	3680.0
422.00	3168.0
423.00	24720.0
424.00	7883.0
424.90	1226.0
439.90	231.0
441.00	73128.0
442.00	455424.0
443.00	85032.0
444.00	7181.0
444.90	354.0

Instrument :
BNA_M
ClientSampleId :
DFTPP079