

Data Path : Z:\HPCHEM1\BNA_M\Data\BM092415\
 Data File : BM002711.D
 Acq On : 24 Sep 2015 14:15
 Operator : UM/IZ
 Sample : SSTD0.880
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.880

Quant Time: Sep 24 14:46:14 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM092415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 24 14:06:19 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.74	152	8366	0.40	ng/ul	0.00
4) Naphthalene-d8	11.60	136	31621	0.40	ng/ul	0.00
8) Acenaphthene-d10	15.33	164	15564	0.40	ng/ul	0.00
12) Phenanthrene-d10	18.03	188	32646	0.40	ng/ul	0.00
18) Chrysene-d12	22.12	240	30112	0.40	ng/ul	-0.01
22) Perylene-d12	24.72	264	26977	0.40	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) 1,4-Dioxane-d8	3.80	96	17050	2.02	ng/uL	0.00
6) 2-Methylnaphthalene-d10	13.13	152	37114	0.87	ng/ul	0.00
16) Fluoranthene-d10	20.01	212	65076	0.87	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) 1,4-Dioxane	3.83	88	19387	1.99	ng/ul	97
5) Naphthalene	11.65	128	68431	0.83	ng/ul	98
7) 2-Methylnaphthalene	13.20	142	44858	0.82	ng/ul	99
9) Acenaphthylene	15.06	152	65875	0.79	ng/ul	94
10) Acenaphthene	15.38	153	46998	0.81	ng/ul	91
11) Fluorene	16.36	166	52294	0.85	ng/ul	87
13) Pentachlorophenol	17.69	266	5221	1.64	ng/ul	96
14) Phenanthrene	18.08	178	79528	0.81	ng/ul	96
15) Anthracene	18.17	178	79603	0.82	ng/ul	95
17) Fluoranthene	20.03	202	89177	0.87	ng/ul	95
19) Pyrene	20.39	202	91867	0.81	ng/ul	98
20) Benzo(a)anthracene	22.09	228	80211	0.80	ng/ul	98
21) Chrysene	22.16	228	77203	0.80	ng/ul	95
23) Benzo(b)fluoranthene	23.92	252	79430	0.83	ng/ul	98
24) Benzo(k)fluoranthene	23.96	252	74033	0.83	ng/ul	94
25) Benzo(a)pyrene	24.61	252	73158	0.82	ng/ul	96
26) Indeno(1,2,3-cd)pyrene	27.46	276	81005	0.77	ng/ul	99
27) Dibenzo(a,h)anthracene	27.46	278	66345	0.77	ng/ul	93
28) Benzo(g,h,i)perylene	28.32	276	68217	0.76	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\Data\BM092415\
Data File : BM002711.D
Acq On : 24 Sep 2015 14:15
Operator : UM/IZ
Sample : SSTD0.880
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.880

Quant Time: Sep 24 14:46:14 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM092415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 24 14:06:19 2015
Response via : Initial Calibration

