

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM092415\
 Data File : BM002718.D
 Acq On : 24 Sep 2015 18:52
 Operator : UM/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.483

Quant Time: Sep 25 02:01:56 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM092415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 24 17:00:38 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.74	152	9006	0.40	ng/ul	0.00
4) Naphthalene-d8	11.60	136	34360	0.40	ng/ul	0.00
8) Acenaphthene-d10	15.33	164	16377	0.40	ng/ul	0.00
12) Phenanthrene-d10	18.03	188	33221	0.40	ng/ul	0.00
18) Chrysene-d12	22.13	240	30900	0.40	ng/ul	0.00
22) Perylene-d12	24.72	264	28483	0.40	ng/ul	-0.01

System Monitoring Compounds						
2) 1,4-Dioxane-d8	3.80	96	9265	1.00	ng/uL	0.00
6) 2-Methylnaphthalene-d10	13.13	152	18687	0.37	ng/ul	0.00
16) Fluoranthene-d10	20.01	212	32740	0.40	ng/ul	0.00

Target Compounds				Qvalue		
3) 1,4-Dioxane	3.83	88	10878	1.01	ng/ul	99
5) Naphthalene	11.65	128	36273	0.39	ng/ul	98
7) 2-Methylnaphthalene	13.20	142	23757	0.40	ng/ul	99
9) Acenaphthylene	15.06	152	32234	0.38	ng/ul	95
10) Acenaphthene	15.38	153	24483	0.40	ng/ul	92
11) Fluorene	16.36	166	26577	0.38	ng/ul	86
13) Pentachlorophenol	17.69	266	2591	0.40	ng/ul	96
14) Phenanthrene	18.08	178	40474	0.37	ng/ul	95
15) Anthracene	18.17	178	39366	0.39	ng/ul	95
17) Fluoranthene	20.03	202	44964	0.38	ng/ul	94
19) Pyrene	20.39	202	46439	0.39	ng/ul	99
20) Benzo(a)anthracene	22.09	228	40774	0.39	ng/ul	99
21) Chrysene	22.16	228	39662	0.40	ng/ul	95
23) Benzo(b)fluoranthene	23.92	252	40352	0.39	ng/ul	99
24) Benzo(k)fluoranthene	23.97	252	38675	0.41	ng/ul	96
25) Benzo(a)pyrene	24.61	252	37884	0.40	ng/ul	94
26) Indeno(1,2,3-cd)pyrene	27.45	276	42342	0.40	ng/ul	99
27) Dibenzo(a,h)anthracene	27.46	278	34687	0.40	ng/ul	96
28) Benzo(g,h,i)perylene	28.31	276	36028	0.40	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM092415\
Data File : BM002718.D
Acq On : 24 Sep 2015 18:52
Operator : UM/IZ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.483

Quant Time: Sep 25 02:01:56 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM092415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 24 17:00:38 2015
Response via : Initial Calibration

