

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092421\
 Data File : BM032139.D
 Acq On : 24 Sep 2021 20:11
 Operator : CG/JU
 Sample : SSTD04035
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD040042

Quant Time: Sep 25 00:20:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 25 00:18:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.666	152	435658	20.000	ng/u1	0.00
20) Naphthalene-d8	10.466	136	1847477	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.319	164	1125293	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.042	188	2110690	20.000	ng/u1	0.00
79) Chrysene-d12	21.200	240	1612962	20.000	ng/u1	0.00
88) Perylene-d12	23.371	264	1508101	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.990	96	166660	15.761	ng/uL	0.00
4) Pyridine-d5	3.413	84	1214018	36.213	ng/u1	0.00
7) Phenol-d5	6.854	99	1426737	35.835	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.995	67	861523	35.950	ng/u1	0.00
11) 2-Chlorophenol-d4	7.207	132	1123565	39.509	ng/u1	0.00
15) 4-Methylphenol-d8	8.401	113	1116564	34.694	ng/u1	0.00
21) Nitrobenzene-d5	8.825	128	495970	47.131	ng/u1	0.00
24) 2-Nitrophenol-d4	9.548	143	538305	53.944	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.095	165	1126996	41.793	ng/u1	0.00
31) 4-Chloroaniline-d4	10.595	131	1504103	37.700	ng/u1	0.00
46) Dimethylphthalate-d6	13.719	166	3002969	38.203	ng/u1	0.00
49) Acenaphthylene-d8	14.007	160	3884892	39.225	ng/u1	0.00
54) 4-Nitrophenol-d4	14.519	143	499817	39.428	ng/u1	0.00
60) Fluorene-d10	15.307	176	2538724	38.897	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.418	200	420314	54.475	ng/u1	0.00
73) Anthracene-d10	17.142	188	3628546	39.993	ng/u1	0.00
81) Pyrene-d10	19.424	212	3571349	45.603	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.236	264	3052199	41.215	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.025	88	181642	15.179	ng/uL	99
5) Pyridine	3.431	79	1191381	36.187	ng/u1	97
6) Benzaldehyde	6.790	77	887938	37.841	ng/u1	98
8) Phenol	6.878	94	1450229	35.770	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.084	93	1163952	36.830	ng/u1	99
12) 2-Chlorophenol	7.237	128	1162601	39.309	ng/u1	99
13) 2-Methylphenol	8.131	108	1107749	35.664	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.207	45	1591050	38.782	ng/u1	99
16) Acetophenone	8.489	105	1743960	35.003	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.484	70	872371	31.436	ng/u1	98
18) 4-Methylphenol	8.460	108	1182591	35.468	ng/u1	100
19) Hexachloroethane	8.760	117	474998	38.844	ng/u1	99
22) Nitrobenzene	8.866	77	1232070	41.330	ng/u1	99
23) Isophorone	9.389	82	2480687	33.789	ng/u1	100
25) 2-Nitrophenol	9.584	139	602284	52.223	ng/u1	97
26) 2,4-Dimethylphenol	9.654	107	1311957	37.342	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.872	93	1553062	37.025	ng/u1	100
29) 2,4-Dichlorophenol	10.119	162	1098077	41.590	ng/u1	99
30) Naphthalene	10.513	128	3658061	40.201	ng/u1	99
32) 4-Chloroaniline	10.619	127	1535743	37.614	ng/u1	100
33) Hexachlorobutadiene	10.825	225	730843	42.313	ng/u1	99
34) Caprolactam	11.354	113	330693	29.316	ng/u1	98
35) 4-Chloro-3-methylphenol	11.760	107	1172387	35.608	ng/u1	99
36) 2-Methylnaphthalene	12.130	142	2471135	38.597	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092421\
 Data File : BM032139.D
 Acq On : 24 Sep 2021 20:11
 Operator : CG/JU
 Sample : SST04035
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST040042

Quant Time: Sep 25 00:20:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 25 00:18:47 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.348	142	2508726	38.495	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.513	216	1320870	43.889	ng/ul	98
40) Hexachlorocyclopentadiene	12.507	237	743150	41.552	ng/ul	99
41) 2,4,6-Trichlorophenol	12.748	196	857476	43.982	ng/ul	99
42) 2,4,5-Trichlorophenol	12.819	196	910429	44.401	ng/ul	100
43) 1,1'-Biphenyl	13.148	154	3270082	41.638	ng/ul	99
44) 2-Chloronaphthalene	13.189	162	2486227	41.726	ng/ul	99
45) 2-Nitroaniline	13.383	65	637489	47.550	ng/ul	99
47) Dimethylphthalate	13.766	163	2966704	37.875	ng/ul	99
48) 2,6-Dinitrotoluene	13.877	165	569425	51.330	ng/ul	98
50) Acenaphthylene	14.036	152	4023638	39.730	ng/ul	99
51) 3-Nitroaniline	14.213	138	585548	45.284	ng/ul	96
52) Acenaphthene	14.383	153	2577113	39.060	ng/ul	100
53) 2,4-Dinitrophenol	14.413	184	271298	47.179	ng/ul	91
55) 4-Nitrophenol	14.530	109	405550	33.408	ng/ul	97
56) Dibenzofuran	14.713	168	3671275	39.521	ng/ul	99
57) 2,4-Dinitrotoluene	14.666	165	775491	48.001	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.948	232	736522	42.608	ng/ul	97
59) Diethylphthalate	15.136	149	2903366	35.579	ng/ul	99
61) Fluorene	15.360	166	2800923	37.396	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.354	204	1437217	39.173	ng/ul	98
63) 4-Nitroaniline	15.371	138	522225	38.984	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.436	198	419604	52.627	ng/ul	97
67) N-Nitrosodiphenylamine	15.566	169	2462766	42.164	ng/ul	99
68) 4-Bromophenyl-phenylether	16.236	248	868154	43.586	ng/ul	97
69) Hexachlorobenzene	16.371	284	1000369	44.800	ng/ul	99
70) Atrazine	16.507	200	876270	38.842	ng/ul	99
71) Pentachlorophenol	16.707	266	597816	46.677	ng/ul	99
72) Phenanthrene	17.083	178	4259204	40.004	ng/ul	99
74) Anthracene	17.177	178	4320214	39.754	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.119	216	1382630	49.927	ng/ul	99
76) Pentachlorobenzene	14.648	250	1319960	48.294	ng/ul	99
77) Carbazole	17.436	167	3771889	38.294	ng/ul	100
78) Di-n-butylphthalate	17.995	149	4458347	36.167	ng/ul	99
80) Fluoranthene	19.089	202	4474536	45.171	ng/ul	99
82) Pyrene	19.453	202	4491523	44.873	ng/ul	98
83) Butylbenzylphthalate	20.342	149	1480902	37.723	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.118	252	1220468	35.861	ng/ul	99
85) Benzo(a)anthracene	21.183	228	3767836	38.631	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.112	149	2205596	37.010	ng/ul	98
87) Chrysene	21.236	228	3701552	39.242	ng/ul	99
89) Di-n-octyl phthalate	21.959	149	3491000	41.109	ng/ul	100
90) Benzo(b)fluoranthene	22.724	252	3718616	41.891	ng/ul	100
91) Benzo(k)fluoranthene	22.765	252	3459157	42.462	ng/ul	100
93) Benzo(a)pyrene	23.277	252	3548538	41.317	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	25.530	276	3852043	38.883	ng/ul	99
95) Dibenzo(a,h)anthracene	25.530	278	3303987	39.328	ng/ul	100
96) Benzo(g,h,i)perylene	26.194	276	3369728	39.619	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092421\
 Data File : BM032139.D
 Acq On : 24 Sep 2021 20:11
 Operator : CG/JU
 Sample : SSTD04035
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD040042

Quant Time: Sep 25 00:20:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 25 00:18:47 2021
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 15.179 ng/uL
RT: 3.025 min Scan# 32
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion: 88 Resp: 181642
Ion Ratio Lower Upper
88 100
43 28.3 22.3 33.5
58 79.3 62.4 93.6



#5
Pyridine
Concen: 36.187 ng/uL
RT: 3.431 min Scan# 101
Delta R.T. -0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion: 79 Resp: 1191381
Ion Ratio Lower Upper
79 100
52 76.1 58.3 87.5
51 36.1 29.1 43.7









#6
Benzaldehyde
Concen: 37.841 ng/ul
RT: 6.790 min Scan# 672
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion: 77 Resp: 887938
Ion Ratio Lower Upper
77 100
105 99.1 77.0 115.4
106 93.9 74.2 111.2



#8
Phenol
Concen: 35.770 ng/ul
RT: 6.878 min Scan# 687
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion: 94 Resp: 1450229
Ion Ratio Lower Upper
94 100
65 28.0 22.4 33.6
66 37.5 29.9 44.9





#10

Bis(2-Chloroethyl)ether

Concen: 36.830 ng/ul

RT: 7.084 min Scan# 722

Delta R.T. -0.000 min

Lab File: BM032139.D

Acq: 24 Sep 2021 20:11

Instrument :





BNA_M

ClientSampleId :

SSTD040042

Tgt Ion: 93 Resp: 1163952

Ion	Ratio	Lower	Upper
93	100		
63	80.0	63.4	95.2
95	32.8	25.5	38.3



#12

2-Chlorophenol

Concen: 39.309 ng/ul

RT: 7.237 min Scan# 748




Delta R.T. -0.000 min


Lab File: BM032139.D

Acq: 24 Sep 2021 20:11

Tgt Ion:128 Resp: 1162601

Ion	Ratio	Lower	Upper
128	100		
64	47.6	38.3	57.5
130	31.7	25.8	38.8









#13
2-Methylphenol
Concen: 35.664 ng/ul
RT: 8.131 min Scan# 900
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:108 Resp: 1107749
Ion Ratio Lower Upper
108 100
107 88.9 71.6 107.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 38.782 ng/ul
RT: 8.207 min Scan# 913
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion: 45 Resp: 1591050
Ion Ratio Lower Upper
45 100
77 14.2 11.2 16.8
79 11.0 9.1 13.7









#16
Acetophenone
Concen: 35.003 ng/ul
RT: 8.489 min Scan# 961
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:105 Resp: 1743960
Ion Ratio Lower Upper
105 100
77 77.2 63.0 94.4
51 26.8 22.2 33.2



#17
N-Nitroso-di-n-propylamine
Concen: 31.436 ng/ul
RT: 8.484 min Scan# 960
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion: 70 Resp: 872371
Ion Ratio Lower Upper
70 100
42 56.1 46.2 69.4
101 9.6 7.5 11.3
130 20.5 16.2 24.2









#18
4-Methylphenol
Concen: 35.468 ng/ul
RT: 8.460 min Scan# 956
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:108 Resp: 1182591
Ion Ratio Lower Upper
108 100
107 111.3 89.1 133.7



#19
Hexachloroethane
Concen: 38.844 ng/ul
RT: 8.760 min Scan# 1007
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:117 Resp: 474998
Ion Ratio Lower Upper
117 100
201 100.4 79.5 119.3
199 62.9 49.1 73.7









#22
Nitrobenzene
Concen: 41.330 ng/ul
RT: 8.866 min Scan# 1025
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion: 77 Resp: 1232070
Ion Ratio Lower Upper
77 100
123 46.0 36.1 54.1
65 13.8 11.0 16.6



#23
Isophorone
Concen: 33.789 ng/ul
RT: 9.389 min Scan# 1114
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion: 82 Resp: 2480687
Ion Ratio Lower Upper
82 100
95 7.3 5.8 8.6
138 17.8 14.1 21.1









#25
2-Nitrophenol
Concen: 52.223 ng/ul
RT: 9.584 min Scan# 1147
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:139 Resp: 602284
Ion Ratio Lower Upper
139 100
65 50.9 42.0 63.0
109 30.4 25.8 38.8



#26
2,4-Dimethylphenol
Concen: 37.342 ng/ul
RT: 9.654 min Scan# 1159
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:107 Resp: 1311957
Ion Ratio Lower Upper
107 100
121 50.4 40.4 60.6
122 86.3 67.6 101.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 37.025 ng/ul
 RT: 9.872 min Scan# 1196
 Delta R.T. -0.000 min
 Lab File: BM032139.D
 Acq: 24 Sep 2021 20:11




Instrument :
 BNA_M
 ClientSampleId :
 SSTD040042


Tgt Ion: 93 Resp: 1553062
 Ion Ratio Lower Upper
 93 100
 95 32.6 26.2 39.4
 123 11.1 9.0 13.6



#29
 2,4-Dichlorophenol
 Concen: 41.590 ng/ul
 RT: 10.119 min Scan# 1238
 Delta R.T. -0.000 min
 Lab File: BM032139.D
 Acq: 24 Sep 2021 20:11

Tgt Ion:162 Resp: 1098077
 Ion Ratio Lower Upper
 162 100
 164 65.5 51.6 77.4
 98 36.5 28.6 42.8









#30
Naphthalene
Concen: 40.201 ng/ul
RT: 10.513 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:128 Resp: 3658061
Ion Ratio Lower Upper
128 100
129 11.2 8.7 13.1
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 37.614 ng/ul
RT: 10.619 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:127 Resp: 1535743
Ion Ratio Lower Upper
127 100
129 32.2 25.8 38.6









#33
Hexachlorobutadiene
Concen: 42.313 ng/ul
RT: 10.825 min Scan# 1358
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:	225	Resp:	730843
Ion Ratio	Lower	Upper	
225	100		
223	63.2	50.6	76.0
227	63.6	52.2	78.4



#34
Caprolactam
Concen: 29.316 ng/ul
RT: 11.354 min Scan# 1448
Delta R.T. 0.012 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:	113	Resp:	330693
Ion Ratio	Lower	Upper	
113	100		
55	168.8	138.1	207.1
56	130.5	106.0	159.0









#35
4-Chloro-3-methylphenol
Concen: 35.608 ng/ul
RT: 11.760 min Scan# 1517
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:107 Resp: 1172387
Ion Ratio Lower Upper
107 100
144 25.4 20.3 30.5
142 76.9 62.8 94.2



#36
2-Methylnaphthalene
Concen: 38.597 ng/ul
RT: 12.130 min Scan# 1580
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:142 Resp: 2471135
Ion Ratio Lower Upper
142 100
141 88.8 70.2 105.2









#37
1-Methylnaphthalene
Concen: 38.495 ng/ul
RT: 12.348 min Scan# 1617
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:142 Resp: 2508726
Ion Ratio Lower Upper
142 100
141 93.0 74.3 111.5
116 3.9 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 43.889 ng/ul
RT: 12.513 min Scan# 1645
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:216 Resp: 1320870
Ion Ratio Lower Upper
216 100
214 78.6 64.2 96.4
179 20.9 17.1 25.7
108 18.3 16.0 24.0









#40

Hexachlorocyclopentadiene
Concen: 41.552 ng/ul
RT: 12.507 min Scan# 1644
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Instrument :
BNA_M
ClientSampleId :
SSTD040042




Tgt Ion:237 Resp: 743150
Ion Ratio Lower Upper
237 100
235 62.3 49.4 74.0
272 13.0 10.1 15.1




#41

2,4,6-Trichlorophenol
Concen: 43.982 ng/ul
RT: 12.748 min Scan# 1685
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:196 Resp: 857476
Ion Ratio Lower Upper
196 100
198 95.5 77.8 116.6
200 30.6 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 44.401 ng/ul
RT: 12.819 min Scan# 1697
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:196 Resp: 910429
Ion Ratio Lower Upper
196 100
198 96.7 77.4 116.2
200 30.9 24.3 36.5



#43
1,1'-Biphenyl
Concen: 41.638 ng/ul
RT: 13.148 min Scan# 1753
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:154 Resp: 3270082
Ion Ratio Lower Upper
154 100
153 40.1 31.9 47.9
76 12.7 10.7 16.1









#44
2-Chloronaphthalene
Concen: 41.726 ng/ul
RT: 13.189 min Scan# 1760
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:162 Resp: 2486227
Ion Ratio Lower Upper
162 100
164 33.3 25.9 38.9
127 39.7 32.0 48.0



#45
2-Nitroaniline
Concen: 47.550 ng/ul
RT: 13.383 min Scan# 1793
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion: 65 Resp: 637489
Ion Ratio Lower Upper
65 100
92 68.6 54.5 81.7
138 103.3 84.1 126.1









#47
Dimethylphthalate
Concen: 37.875 ng/ul
RT: 13.766 min Scan# 1858
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:163 Resp: 2966704
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 10.5 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 51.330 ng/ul
RT: 13.877 min Scan# 1877
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:165 Resp: 569425
Ion Ratio Lower Upper
165 100
89 60.5 46.8 70.2
121 21.8 17.3 25.9









#50
Acenaphthylene
Concen: 39.730 ng/ul
RT: 14.036 min Scan# 1904
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:152 Resp: 4023638
Ion Ratio Lower Upper
152 100
151 20.0 15.7 23.5
153 13.4 10.6 15.8



#51
3-Nitroaniline
Concen: 45.284 ng/ul
RT: 14.213 min Scan# 1934
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:138 Resp: 585548
Ion Ratio Lower Upper
138 100
108 12.1 10.5 15.7
92 112.5 94.0 141.0









#52
Acenaphthene
Concen: 39.060 ng/ul
RT: 14.383 min Scan# 1963
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:153 Resp: 2577113
Ion Ratio Lower Upper
153 100
152 47.8 38.4 57.6
154 89.3 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 47.179 ng/ul
RT: 14.413 min Scan# 1968
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:184 Resp: 271298
Ion Ratio Lower Upper
184 100
63 64.1 56.7 85.1
154 63.5 57.5 86.3









#55
4-Nitrophenol
Concen: 33.408 ng/ul
RT: 14.530 min Scan# 1988
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:109 Resp: 405550
Ion Ratio Lower Upper
109 100
139 134.4 103.7 155.5
65 133.8 109.4 164.2



#56
Dibenzofuran
Concen: 39.521 ng/ul
RT: 14.713 min Scan# 2019
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:168 Resp: 3671275
Ion Ratio Lower Upper
168 100
139 38.0 29.8 44.6









#57
2,4-Dinitrotoluene
Concen: 48.001 ng/ul
RT: 14.666 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:165 Resp: 775491
Ion Ratio Lower Upper
165 100
63 45.1 34.2 51.2
182 3.2 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 42.608 ng/ul
RT: 14.948 min Scan# 2059
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:232 Resp: 736522
Ion Ratio Lower Upper
232 100
131 42.3 36.2 54.2
130 2.3 1.9 2.9
166 30.2 24.8 37.2









#59
Diethylphthalate
Concen: 35.579 ng/ul
RT: 15.136 min Scan# 2091
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:149 Resp: 2903366
Ion Ratio Lower Upper
149 100
177 23.0 17.9 26.9
150 12.4 10.0 15.0



#61
Fluorene
Concen: 37.396 ng/ul
RT: 15.360 min Scan# 2129
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:166 Resp: 2800923
Ion Ratio Lower Upper
166 100
165 97.3 77.7 116.5
167 13.7 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 39.173 ng/ul
RT: 15.354 min Scan# 2128
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:204 Resp: 1437217
Ion Ratio Lower Upper
204 100
206 32.5 25.8 38.6
141 57.5 47.9 71.9



#63
4-Nitroaniline
Concen: 38.984 ng/ul
RT: 15.371 min Scan# 2131
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:138 Resp: 522225
Ion Ratio Lower Upper
138 100
92 51.0 41.4 62.2
108 80.0 63.8 95.6









#66
4,6-Dinitro-2-methylphenol
Concen: 52.627 ng/ul
RT: 15.436 min Scan# 2142
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:198 Resp: 419604
Ion Ratio Lower Upper
198 100
182 4.1 3.8 5.8
77 23.5 19.8 29.8



#67
N-Nitrosodiphenylamine
Concen: 42.164 ng/ul
RT: 15.566 min Scan# 2164
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:169 Resp: 2462766
Ion Ratio Lower Upper
169 100
168 66.0 52.3 78.5
167 36.5 28.6 42.8









#68
4-Bromophenyl-phenylether
Concen: 43.586 ng/ul
RT: 16.236 min Scan# 2278
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:248 Resp: 868154
Ion Ratio Lower Upper
248 100
250 97.6 77.0 115.6
141 76.1 57.6 86.4



#69
Hexachlorobenzene
Concen: 44.800 ng/ul
RT: 16.371 min Scan# 2301
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:284 Resp: 1000369
Ion Ratio Lower Upper
284 100
142 36.3 28.4 42.6
249 31.7 24.7 37.1









#70
Atrazine
Concen: 38.842 ng/ul
RT: 16.507 min Scan# 2324
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:200 Resp: 876270
Ion Ratio Lower Upper
200 100
173 26.3 21.4 32.0
215 48.9 39.4 59.2



#71
Pentachlorophenol
Concen: 46.677 ng/ul
RT: 16.707 min Scan# 2358
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:266 Resp: 597816
Ion Ratio Lower Upper
266 100
264 62.2 50.2 75.2
268 63.1 51.0 76.6









#72
Phenanthrene
Concen: 40.004 ng/ul
RT: 17.083 min Scan# 2422
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:178 Resp: 4259204
Ion Ratio Lower Upper
178 100
179 15.9 12.3 18.5
176 19.6 15.4 23.0



#74
Anthracene
Concen: 39.754 ng/ul
RT: 17.177 min Scan# 2438
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:178 Resp: 4320214
Ion Ratio Lower Upper
178 100
179 15.6 12.6 19.0
176 18.9 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 49.927 ng/uL
RT: 13.119 min Scan# 1748
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:216 Resp: 1382630
Ion Ratio Lower Upper
216 100
214 79.1 62.7 94.1
218 48.0 38.7 58.1



#76
Pentachlorobenzene
Concen: 48.294 ng/uL
RT: 14.648 min Scan# 2008
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:250 Resp: 1319960
Ion Ratio Lower Upper
250 100
248 62.9 50.0 75.0
252 63.9 51.6 77.4









#77
 Carbazole
 Concen: 38.294 ng/ul
 RT: 17.436 min Scan# 2482
 Delta R.T. -0.000 min
 Lab File: BM032139.D
 Acq: 24 Sep 2021 20:11




Instrument :
 BNA_M
 ClientSampleId :
 SSTD040042


Tgt Ion:167 Resp: 3771889
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.3 25.9
 139 13.1 10.3 15.5



#78
 Di-n-butylphthalate
 Concen: 36.167 ng/ul
 RT: 17.995 min Scan# 2577
 Delta R.T. -0.000 min
 Lab File: BM032139.D
 Acq: 24 Sep 2021 20:11

Tgt Ion:149 Resp: 4458347
 Ion Ratio Lower Upper
 149 100
 150 9.6 7.4 11.0
 104 5.3 4.2 6.2









#80
Fluoranthene
Concen: 45.171 ng/ul
RT: 19.089 min Scan# 2763
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:202 Resp: 4474536
Ion Ratio Lower Upper
202 100
101 11.8 9.2 13.8
100 9.0 6.9 10.3



#82
Pyrene
Concen: 44.873 ng/ul
RT: 19.453 min Scan# 2825
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:202 Resp: 4491523
Ion Ratio Lower Upper
202 100
101 13.4 10.1 15.1
100 11.2 8.4 12.6









#83
Butylbenzylphthalate
Concen: 37.723 ng/ul
RT: 20.342 min Scan# 2976
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:149 Resp: 1480902
Ion Ratio Lower Upper
149 100
91 71.1 61.8 92.8
206 22.4 16.2 24.4



#84
3,3'-Dichlorobenzidine
Concen: 35.861 ng/ul
RT: 21.118 min Scan# 3108
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:252 Resp: 1220468
Ion Ratio Lower Upper
252 100
254 64.4 52.2 78.2
126 11.3 8.8 13.2









#85
Benzo(a)anthracene
Concen: 38.631 ng/ul
RT: 21.183 min Scan# 3119
Delta R.T. -0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:228 Resp: 3767836
Ion Ratio Lower Upper
228 100
229 20.6 15.9 23.9
226 28.0 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 37.010 ng/ul
RT: 21.112 min Scan# 3107
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:149 Resp: 2205596
Ion Ratio Lower Upper
149 100
167 28.1 21.6 32.4
279 4.4 3.3 4.9









#87
Chrysene
Concen: 39.242 ng/ul
RT: 21.236 min Scan# 3128
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:228 Resp: 3701552
Ion Ratio Lower Upper
228 100
226 31.1 24.4 36.6
229 20.5 15.9 23.9



#89
Di-n-octyl phthalate
Concen: 41.109 ng/ul
RT: 21.959 min Scan# 3251
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:149 Resp: 3491000









#90
Benzo(b)fluoranthene
Concen: 41.891 ng/ul
RT: 22.724 min Scan# 3381
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:252 Resp: 3718616
Ion Ratio Lower Upper
252 100
253 22.7 18.2 27.4
125 10.2 8.5 12.7



#91
Benzo(k)fluoranthene
Concen: 42.462 ng/ul
RT: 22.765 min Scan# 3388
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:252 Resp: 3459157
Ion Ratio Lower Upper
252 100
253 22.3 17.9 26.9
125 10.4 8.4 12.6









#93
Benzo(a)pyrene
Concen: 41.317 ng/ul
RT: 23.277 min Scan# 3475
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11




Instrument :
BNA_M
ClientSampleId :
SSTD040042


Tgt Ion:252 Resp: 3548538
Ion Ratio Lower Upper
252 100
253 22.4 17.8 26.6
125 11.9 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 38.883 ng/ul
RT: 25.530 min Scan# 3858
Delta R.T. 0.006 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:276 Resp: 3852043
Ion Ratio Lower Upper
276 100
138 26.3 21.5 32.3
277 25.6 21.0 31.6









#95
Dibenzo(a,h)anthracene
Concen: 39.328 ng/ul
RT: 25.530 min Scan# 3858
Delta R.T. -0.000 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Instrument :
BNA_M
ClientSampleId :
SSTD040042

Tgt Ion:278 Resp: 3303987
Ion Ratio Lower Upper
278 100
139 16.7 13.5 20.3
279 24.1 19.4 29.0



#96
Benzo(g,h,i)perylene
Concen: 39.619 ng/ul
RT: 26.194 min Scan# 3971
Delta R.T. 0.012 min
Lab File: BM032139.D
Acq: 24 Sep 2021 20:11

Tgt Ion:276 Resp: 3369728
Ion Ratio Lower Upper
276 100
138 22.4 17.9 26.9
277 23.9 18.6 28.0

