

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092421\
 Data File : BM032141.D
 Acq On : 24 Sep 2021 21:24
 Operator : CG/JU
 Sample : SSTD16037
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD160044

Quant Time: Sep 25 00:21:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 25 00:18:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.672	152	456000	20.000	ng/u1	0.00
20) Naphthalene-d8	10.472	136	1870194	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.319	164	1081427	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.048	188	1898994	20.000	ng/u1	0.00
79) Chrysene-d12	21.207	240	1404269	20.000	ng/u1	0.00
88) Perylene-d12	23.377	264	1568332	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.984	96	678184	61.274	ng/uL	0.00
4) Pyridine-d5	3.414	84	4654435	132.645	ng/u1	0.00
7) Phenol-d5	6.872	99	5588291	134.099	ng/u1	0.02
9) Bis-(2-Chloroethyl)eth...	7.002	67	3192898	127.292	ng/u1	0.01
11) 2-Chlorophenol-d4	7.219	132	4433500	148.944	ng/u1	0.01
15) 4-Methylphenol-d8	8.419	113	4392100	130.382	ng/u1	0.02
21) Nitrobenzene-d5	8.837	128	2125161	199.496	ng/u1	0.01
24) 2-Nitrophenol-d4	9.560	143	2404790	238.057	ng/u1	0.01
28) 2,4-Dichlorophenol-d3	10.107	165	4172234	152.840	ng/u1	0.02
31) 4-Chloroaniline-d4	10.607	131	5314615	131.590	ng/u1	0.01
46) Dimethylphthalate-d6	13.731	166	9527266	126.119	ng/u1	0.01
49) Acenaphthylene-d8	14.019	160	12050345	126.605	ng/u1	0.01
54) 4-Nitrophenol-d4	14.542	143	1864957	153.083	ng/u1	0.03
60) Fluorene-d10	15.313	176	7573096	120.738	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.436	200	1698670	244.699	ng/u1	0.02
73) Anthracene-d10	17.148	188	10250718	125.577	ng/u1	0.01
81) Pyrene-d10	19.430	212	10086484	147.938	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.248	264	10716935	139.156	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.025	88	743402	59.353	ng/uL	97
5) Pyridine	3.437	79	4584452	133.036	ng/u1	98
6) Benzaldehyde	6.796	77	1881054	76.588	ng/u1	96
8) Phenol	6.902	94	5496747	129.531	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.102	93	4300804	130.015	ng/u1	97
12) 2-Chlorophenol	7.249	128	4420684	142.801	ng/u1	99
13) 2-Methylphenol	8.143	108	4194579	129.019	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.213	45	5737440	133.613	ng/u1	98
16) Acetophenone	8.507	105	5595261	107.292	ng/u1	94
17) N-Nitroso-di-n-propyla...	8.531	70	3159541	108.777	ng/u1	94
18) 4-Methylphenol	8.490	108	4061373	116.373	ng/u1	98
19) Hexachloroethane	8.766	117	1830020	142.978	ng/u1	95
22) Nitrobenzene	8.884	77	4743966	157.206	ng/u1	96
23) Isophorone	9.413	82	8929520	120.149	ng/u1	97
25) 2-Nitrophenol	9.596	139	2486873	213.014	ng/u1	96
26) 2,4-Dimethylphenol	9.666	107	4815281	135.391	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.884	93	5593780	131.736	ng/u1	99
29) 2,4-Dichlorophenol	10.137	162	3916211	146.524	ng/u1	98
30) Naphthalene	10.525	128	12051749	130.837	ng/u1	95
32) 4-Chloroaniline	10.637	127	5200753	125.832	ng/u1	99
33) Hexachlorobutadiene	10.831	225	2641608	151.082	ng/u1	99
34) Caprolactam	11.419	113	722782	63.297	ng/u1	95
35) 4-Chloro-3-methylphenol	11.790	107	4276212	128.302	ng/u1	98
36) 2-Methylnaphthalene	12.137	142	8078725	124.651	ng/u1	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092421\
 Data File : BM032141.D
 Acq On : 24 Sep 2021 21:24
 Operator : CG/JU
 Sample : SSTD16037
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD160044

Quant Time: Sep 25 00:21:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 25 00:18:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.360	142	8061688	122.201	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.525	216	3883390	134.269	ng/ul	98
40) Hexachlorocyclopentadiene	12.507	237	2601273	151.346	ng/ul	97
41) 2,4,6-Trichlorophenol	12.760	196	3056745	163.148	ng/ul	99
42) 2,4,5-Trichlorophenol	12.837	196	3249620	164.909	ng/ul	99
43) 1,1'-Biphenyl	13.160	154	9555176	126.601	ng/ul	95
44) 2-Chloronaphthalene	13.195	162	8027829	140.193	ng/ul	96
45) 2-Nitroaniline	13.401	65	2509170	194.751	ng/ul	93
47) Dimethylphthalate	13.784	163	9372177	124.504	ng/ul	98
48) 2,6-Dinitrotoluene	13.901	165	2180205	204.505	ng/ul	94
50) Acenaphthylene	14.048	152	11412274	117.257	ng/ul	93
51) 3-Nitroaniline	14.225	138	1903925	153.214	ng/ul	99
52) Acenaphthene	14.389	153	7789189	122.846	ng/ul	98
53) 2,4-Dinitrophenol	14.425	184	1302544	235.701	ng/ul	91
55) 4-Nitrophenol	14.560	109	1422862	121.966	ng/ul	95
56) Dibenzofuran	14.725	168	10661551	119.425	ng/ul	89
57) 2,4-Dinitrotoluene	14.684	165	2838068	182.796	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	14.954	232	2467356	148.526	ng/ul	99
59) Diethylphthalate	15.148	149	9182942	117.098	ng/ul	96
61) Fluorene	15.372	166	7125608	98.996	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.354	204	3803486	107.873	ng/ul	95
63) 4-Nitroaniline	15.401	138	1433576	111.357	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.454	198	1622542	226.188	ng/ul#	97
67) N-Nitrosodiphenylamine	15.572	169	7199510	137.000	ng/ul	96
68) 4-Bromophenyl-phenylether	16.242	248	2686055	149.888	ng/ul	95
69) Hexachlorobenzene	16.383	284	3021177	150.381	ng/ul	98
70) Atrazine	16.525	200	2821538	139.012	ng/ul	99
71) Pentachlorophenol	16.713	266	2060425	178.810	ng/ul	99
72) Phenanthrene	17.095	178	11529995	120.366	ng/ul	94
74) Anthracene	17.183	178	11075739	113.279	ng/ul	93
75) 1,2,3,4-Tetrachloroben...	13.125	216	4444893	178.399	ng/ul	99
76) Pentachlorobenzene	14.654	250	4031334	163.939	ng/ul	98
77) Carbazole	17.448	167	10584632	119.441	ng/ul	94
78) Di-n-butylphthalate	18.001	149	12130539	109.376	ng/ul#	94
80) Fluoranthene	19.095	202	11809740	136.938	ng/ul	95
82) Pyrene	19.466	202	11275942	129.395	ng/ul	95
83) Butylbenzylphthalate	20.342	149	5369941	157.117	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.124	252	3451695	116.494	ng/ul	100
85) Benzo(a)anthracene	21.195	228	10694291	125.942	ng/ul	93
86) Bis(2-ethylhexyl)phtha...	21.113	149	6576504	126.754	ng/ul#	96
87) Chrysene	21.248	228	10132065	123.378	ng/ul	93
89) Di-n-octyl phthalate	21.960	149	11995335	135.828	ng/ul	100
90) Benzo(b)fluoranthene	22.736	252	11832206	128.174	ng/ul	97
91) Benzo(k)fluoranthene	22.783	252	11093715	130.948	ng/ul	97
93) Benzo(a)pyrene	23.295	252	11603626	129.918	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	25.559	276	10122570	98.255	ng/ul	98
95) Dibenzo(a,h)anthracene	25.565	278	11106231	127.122	ng/ul	99
96) Benzo(g,h,i)perylene	26.230	276	12863798	145.435	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 59.353 ng/uL
RT: 3.025 min Scan# 32
Delta R.T. 0.000 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion: 88 Resp: 743402
Ion Ratio Lower Upper
88 100
43 26.9 22.3 33.5
58 75.4 62.4 93.6



#5
Pyridine
Concen: 133.036 ng/uL
RT: 3.437 min Scan# 102
Delta R.T. 0.000 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion: 79 Resp: 4584452
Ion Ratio Lower Upper
79 100
52 71.8 58.3 87.5
51 34.7 29.1 43.7









#6
Benzaldehyde
Concen: 76.588 ng/ul
RT: 6.796 min Scan# 673
Delta R.T. 0.006 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion: 77 Resp: 1881054
Ion Ratio Lower Upper
77 100
105 100.0 77.0 115.4
106 95.7 74.2 111.2



#8
Phenol
Concen: 129.531 ng/ul
RT: 6.902 min Scan# 691
Delta R.T. 0.024 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion: 94 Resp: 5496747
Ion Ratio Lower Upper
94 100
65 27.8 22.4 33.6
66 37.7 29.9 44.9









#10
 Bis(2-Chloroethyl)ether
 Concen: 130.015 ng/ul
 RT: 7.102 min Scan# 725
 Delta R.T. 0.018 min
 Lab File: BM032141.D
 Acq: 24 Sep 2021 21:24

Instrument :
 BNA_M
 ClientSampleId :
 SSTD160044

Tgt Ion: 93 Resp: 4300804




Ion	Ratio	Lower	Upper
93	100		
63	76.2	63.4	95.2
95	33.2	25.5	38.3




#12
 2-Chlorophenol
 Concen: 142.801 ng/ul
 RT: 7.249 min Scan# 750
 Delta R.T. 0.012 min
 Lab File: BM032141.D
 Acq: 24 Sep 2021 21:24

Tgt Ion:128 Resp: 4420684

Ion	Ratio	Lower	Upper
128	100		
64	46.9	38.3	57.5
130	32.9	25.8	38.8









#13
2-Methylphenol
Concen: 129.019 ng/ul
RT: 8.143 min Scan# 902
Delta R.T. 0.018 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:108 Resp: 4194579
Ion Ratio Lower Upper
108 100
107 90.2 71.6 107.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 133.613 ng/ul
RT: 8.213 min Scan# 914
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion: 45 Resp: 5737440
Ion Ratio Lower Upper
45 100
77 15.2 11.2 16.8
79 12.1 9.1 13.7









#16
Acetophenone
Concen: 107.292 ng/ul
RT: 8.507 min Scan# 964
Delta R.T. 0.024 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:105 Resp: 5595261
Ion Ratio Lower Upper
105 100
77 84.7 63.0 94.4
51 28.5 22.2 33.2



#17
N-Nitroso-di-n-propylamine
Concen: 108.777 ng/ul
RT: 8.531 min Scan# 968
Delta R.T. 0.053 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion: 70 Resp: 3159541
Ion Ratio Lower Upper
70 100
42 52.9 46.2 69.4
101 10.2 7.5 11.3
130 22.0 16.2 24.2









#18
4-Methylphenol
Concen: 116.373 ng/ul
RT: 8.490 min Scan# 961
Delta R.T. 0.036 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:108 Resp: 4061373
Ion Ratio Lower Upper
108 100
107 113.3 89.1 133.7



#19
Hexachloroethane
Concen: 142.978 ng/ul
RT: 8.766 min Scan# 1008
Delta R.T. 0.006 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:117 Resp: 1830020
Ion Ratio Lower Upper
117 100
201 104.0 79.5 119.3
199 65.4 49.1 73.7









#22
Nitrobenzene
Concen: 157.206 ng/ul
RT: 8.884 min Scan# 1028
Delta R.T. 0.018 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion: 77 Resp: 4743966
Ion Ratio Lower Upper
77 100
123 48.4 36.1 54.1
65 13.9 11.0 16.6



#23
Isophorone
Concen: 120.149 ng/ul
RT: 9.413 min Scan# 1118
Delta R.T. 0.030 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion: 82 Resp: 8929520
Ion Ratio Lower Upper
82 100
95 7.7 5.8 8.6
138 19.0 14.1 21.1









#25
2-Nitrophenol
Concen: 213.014 ng/ul
RT: 9.596 min Scan# 1149
Delta R.T. 0.018 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:139 Resp: 2486873
Ion Ratio Lower Upper
139 100
65 48.8 42.0 63.0
109 31.5 25.8 38.8



#26
2,4-Dimethylphenol
Concen: 135.391 ng/ul
RT: 9.666 min Scan# 1161
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:107 Resp: 4815281
Ion Ratio Lower Upper
107 100
121 51.1 40.4 60.6
122 86.0 67.6 101.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 131.736 ng/ul
 RT: 9.884 min Scan# 1198
 Delta R.T. 0.012 min
 Lab File: BM032141.D
 Acq: 24 Sep 2021 21:24

Instrument :
 BNA_M
 ClientSampleId :
 SSTD160044

Tgt Ion: 93 Resp: 5593780




Ion	Ratio	Lower	Upper
93	100		
95	33.7	26.2	39.4
123	11.3	9.0	13.6




#29
 2,4-Dichlorophenol
 Concen: 146.524 ng/ul
 RT: 10.137 min Scan# 1241
 Delta R.T. 0.018 min
 Lab File: BM032141.D
 Acq: 24 Sep 2021 21:24

Tgt Ion: 162 Resp: 3916211

Ion	Ratio	Lower	Upper
162	100		
164	66.1	51.6	77.4
98	36.9	28.6	42.8









#30
Naphthalene
Concen: 130.837 ng/ul
RT: 10.525 min Scan# 1307
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:128 Resp:12051749
Ion Ratio Lower Upper
128 100
129 12.3 8.7 13.1
127 14.9 10.2 15.4



#32
4-Chloroaniline
Concen: 125.832 ng/ul
RT: 10.637 min Scan# 1326
Delta R.T. 0.018 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:127 Resp: 5200753
Ion Ratio Lower Upper
127 100
129 32.9 25.8 38.6









#33
Hexachlorobutadiene
Concen: 151.082 ng/u1
RT: 10.831 min Scan# 1359
Delta R.T. 0.006 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:225 Resp: 2641608
Ion Ratio Lower Upper
225 100
223 63.7 50.6 76.0
227 64.3 52.2 78.4



#34
Caprolactam
Concen: 63.297 ng/u1
RT: 11.419 min Scan# 1459
Delta R.T. 0.077 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:113 Resp: 722782
Ion Ratio Lower Upper
113 100
55 166.7 138.1 207.1
56 124.5 106.0 159.0









#35
4-Chloro-3-methylphenol
Concen: 128.302 ng/ul
RT: 11.790 min Scan# 1522
Delta R.T. 0.036 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:107 Resp: 4276212
Ion Ratio Lower Upper
107 100
144 26.2 20.3 30.5
142 80.3 62.8 94.2



#36
2-Methylnaphthalene
Concen: 124.651 ng/ul
RT: 12.137 min Scan# 1581
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:142 Resp: 8078725
Ion Ratio Lower Upper
142 100
141 89.8 70.2 105.2





#37

1-Methylnaphthalene

Concen: 122.201 ng/ul

RT: 12.360 min Scan# 1619

Delta R.T. 0.012 min

Lab File: BM032141.D

Acq: 24 Sep 2021 21:24

Instrument :

BNA_M

ClientSampleId :

SSTD160044





Tgt Ion:142 Resp: 8061688

Ion Ratio Lower Upper

142 100

141 93.1 74.3 111.5

116 4.3 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 134.269 ng/ul

RT: 12.525 min Scan# 1647

Delta R.T. 0.018 min

Lab File: BM032141.D

Acq: 24 Sep 2021 21:24

Tgt Ion:216 Resp: 3883390




Ion Ratio Lower Upper


216 100

214 78.5 64.2 96.4

179 21.9 17.1 25.7

108 18.2 16.0 24.0









#40

Hexachlorocyclopentadiene
Concen: 151.346 ng/ul
RT: 12.507 min Scan# 1644
Delta R.T. 0.006 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Instrument :
BNA_M
ClientSampleId :
SSTD160044




Tgt Ion:237 Resp: 2601273
Ion Ratio Lower Upper
237 100
235 64.5 49.4 74.0
272 12.6 10.1 15.1




#41

2,4,6-Trichlorophenol
Concen: 163.148 ng/ul
RT: 12.760 min Scan# 1687
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:196 Resp: 3056745
Ion Ratio Lower Upper
196 100
198 96.2 77.8 116.6
200 31.6 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 164.909 ng/ul
RT: 12.837 min Scan# 1700
Delta R.T. 0.018 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:196 Resp: 3249620
Ion Ratio Lower Upper
196 100
198 95.7 77.4 116.2
200 31.6 24.3 36.5



#43
1,1'-Biphenyl
Concen: 126.601 ng/ul
RT: 13.160 min Scan# 1755
Delta R.T. 0.018 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:154 Resp: 9555176
Ion Ratio Lower Upper
154 100
153 43.9 31.9 47.9
76 13.6 10.7 16.1









#44
2-Chloronaphthalene
Concen: 140.193 ng/ul
RT: 13.195 min Scan# 1761
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:162 Resp: 8027829
Ion Ratio Lower Upper
162 100
164 34.4 25.9 38.9
127 43.2 32.0 48.0



#45
2-Nitroaniline
Concen: 194.751 ng/ul
RT: 13.401 min Scan# 1796
Delta R.T. 0.018 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion: 65 Resp: 2509170
Ion Ratio Lower Upper
65 100
92 71.7 54.5 81.7
138 113.2 84.1 126.1









#47
Dimethylphthalate
Concen: 124.504 ng/ul
RT: 13.784 min Scan# 1861
Delta R.T. 0.018 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:163 Resp: 9372177
Ion Ratio Lower Upper
163 100
194 4.4 3.3 4.9
164 10.9 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 204.505 ng/ul
RT: 13.901 min Scan# 1881
Delta R.T. 0.024 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:165 Resp: 2180205
Ion Ratio Lower Upper
165 100
89 53.8 46.8 70.2
121 19.7 17.3 25.9









#50
Acenaphthylene
Concen: 117.257 ng/ul
RT: 14.048 min Scan# 1906
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:152 Resp:11412274
Ion Ratio Lower Upper
152 100
151 23.5 15.7 23.5
153 15.4 10.6 15.8



#51
3-Nitroaniline
Concen: 153.214 ng/ul
RT: 14.225 min Scan# 1936
Delta R.T. 0.018 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:138 Resp: 1903925
Ion Ratio Lower Upper
138 100
108 11.4 10.5 15.7
92 116.4 94.0 141.0









#52
Acenaphthene
Concen: 122.846 ng/ul
RT: 14.389 min Scan# 1964
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:153 Resp: 7789189
Ion Ratio Lower Upper
153 100
152 50.6 38.4 57.6
154 89.6 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 235.701 ng/ul
RT: 14.425 min Scan# 1970
Delta R.T. 0.018 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:184 Resp: 1302544
Ion Ratio Lower Upper
184 100
63 63.7 56.7 85.1
154 63.7 57.5 86.3









#55
4-Nitrophenol
Concen: 121.966 ng/ul
RT: 14.560 min Scan# 1993
Delta R.T. 0.036 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:109 Resp: 1422862
Ion Ratio Lower Upper
109 100
139 133.4 103.7 155.5
65 129.0 109.4 164.2



#56
Dibenzofuran
Concen: 119.425 ng/ul
RT: 14.725 min Scan# 2021
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:168 Resp:10661551
Ion Ratio Lower Upper
168 100
139 43.8 29.8 44.6









#57
2,4-Dinitrotoluene
Concen: 182.796 ng/ul
RT: 14.684 min Scan# 2014
Delta R.T. 0.018 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:165 Resp: 2838068
Ion Ratio Lower Upper
165 100
63 41.7 34.2 51.2
182 2.8 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 148.526 ng/ul
RT: 14.954 min Scan# 2060
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:232 Resp: 2467356
Ion Ratio Lower Upper
232 100
131 45.6 36.2 54.2
130 2.4 1.9 2.9
166 31.9 24.8 37.2









#59
Diethylphthalate
Concen: 117.098 ng/ul
RT: 15.148 min Scan# 2093
Delta R.T. 0.018 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:149 Resp: 9182942
Ion Ratio Lower Upper
149 100
177 24.3 17.9 26.9
150 13.6 10.0 15.0



#61
Fluorene
Concen: 98.996 ng/ul
RT: 15.372 min Scan# 2131
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:166 Resp: 7125608
Ion Ratio Lower Upper
166 100
165 99.2 77.7 116.5
167 14.2 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 107.873 ng/ul
RT: 15.354 min Scan# 2128
Delta R.T. 0.006 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:204 Resp: 3803486
Ion Ratio Lower Upper
204 100
206 33.4 25.8 38.6
141 65.3 47.9 71.9



#63
4-Nitroaniline
Concen: 111.357 ng/ul
RT: 15.401 min Scan# 2136
Delta R.T. 0.036 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:138 Resp: 1433576
Ion Ratio Lower Upper
138 100
92 49.8 41.4 62.2
108 78.8 63.8 95.6









#66
4,6-Dinitro-2-methylphenol
Concen: 226.188 ng/ul
RT: 15.454 min Scan# 2145
Delta R.T. 0.024 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:198 Resp: 1622542
Ion Ratio Lower Upper
198 100
182 3.3 3.8 5.8#
77 23.3 19.8 29.8



#67
N-Nitrosodiphenylamine
Concen: 137.000 ng/ul
RT: 15.572 min Scan# 2165
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:169 Resp: 7199510
Ion Ratio Lower Upper
169 100
168 68.6 52.3 78.5
167 38.4 28.6 42.8









#68
4-Bromophenyl-phenylether
Concen: 149.888 ng/ul
RT: 16.242 min Scan# 2279
Delta R.T. 0.006 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:248 Resp: 2686055
Ion Ratio Lower Upper
248 100
250 96.7 77.0 115.6
141 80.7 57.6 86.4



#69
Hexachlorobenzene
Concen: 150.381 ng/ul
RT: 16.383 min Scan# 2303
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:284 Resp: 3021177
Ion Ratio Lower Upper
284 100
142 34.4 28.4 42.6
249 31.7 24.7 37.1





#70

Atrazine

Concen: 139.012 ng/ul

RT: 16.525 min Scan# 2327





Delta R.T. 0.024 min

Lab File: BM032141.D

Acq: 24 Sep 2021 21:24

Tgt Ion:200 Resp: 2821538

Ion	Ratio	Lower	Upper
200	100		
173	27.1	21.4	32.0
215	48.6	39.4	59.2



#71

Pentachlorophenol

Concen: 178.810 ng/ul

RT: 16.713 min Scan# 2359




Delta R.T. 0.006 min


Lab File: BM032141.D

Acq: 24 Sep 2021 21:24

Tgt Ion:266 Resp: 2060425

Ion	Ratio	Lower	Upper
266	100		
264	63.4	50.2	75.2
268	63.3	51.0	76.6









#72
Phenanthrene
Concen: 120.366 ng/ul
RT: 17.095 min Scan# 2424
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:178 Resp:11529995
Ion Ratio Lower Upper
178 100
179 17.4 12.3 18.5
176 22.3 15.4 23.0



#74
Anthracene
Concen: 113.279 ng/ul
RT: 17.183 min Scan# 2439
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:178 Resp:11075739
Ion Ratio Lower Upper
178 100
179 18.2 12.6 19.0
176 22.1 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 178.399 ng/uL
RT: 13.125 min Scan# 1749
Delta R.T. 0.006 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:216 Resp: 4444893
Ion Ratio Lower Upper
216 100
214 79.3 62.7 94.1
218 48.6 38.7 58.1



#76
Pentachlorobenzene
Concen: 163.939 ng/uL
RT: 14.654 min Scan# 2009
Delta R.T. 0.006 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:250 Resp: 4031334
Ion Ratio Lower Upper
250 100
248 64.6 50.0 75.0
252 64.2 51.6 77.4









#77
 Carbazole
 Concen: 119.441 ng/ul
 RT: 17.448 min Scan# 2484
 Delta R.T. 0.012 min
 Lab File: BM032141.D
 Acq: 24 Sep 2021 21:24




Instrument :
 BNA_M
 ClientSampleId :
 SSTD160044


Tgt Ion:167 Resp:10584632
 Ion Ratio Lower Upper
 167 100
 166 24.3 17.3 25.9
 139 15.3 10.3 15.5



#78
 Di-n-butylphthalate
 Concen: 109.376 ng/ul
 RT: 18.001 min Scan# 2578
 Delta R.T. 0.006 min
 Lab File: BM032141.D
 Acq: 24 Sep 2021 21:24

Tgt Ion:149 Resp:12130539
 Ion Ratio Lower Upper
 149 100
 150 11.6 7.4 11.0#
 104 6.6 4.2 6.2#









#80
Fluoranthene
Concen: 136.938 ng/ul
RT: 19.095 min Scan# 2764
Delta R.T. 0.006 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:202 Resp:11809740
Ion Ratio Lower Upper
202 100
101 13.4 9.2 13.8
100 10.2 6.9 10.3



#82
Pyrene
Concen: 129.395 ng/ul
RT: 19.466 min Scan# 2827
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:202 Resp:11275942
Ion Ratio Lower Upper
202 100
101 15.0 10.1 15.1
100 12.2 8.4 12.6





#83

Butylbenzylphthalate

Concen: 157.117 ng/ul

RT: 20.342 min Scan# 2976

Delta R.T. 0.006 min

Lab File: BM032141.D

Acq: 24 Sep 2021 21:24

Instrument :

BNA_M

ClientSampleId :

SSTD160044





Tgt Ion:149 Resp: 5369941

Ion Ratio Lower Upper

149 100

91 71.8 61.8 92.8

206 22.1 16.2 24.4



#84

3,3'-Dichlorobenzidine

Concen: 116.494 ng/ul

RT: 21.124 min Scan# 3109

Delta R.T. 0.006 min

Lab File: BM032141.D

Acq: 24 Sep 2021 21:24




Tgt Ion:252 Resp: 3451695


Ion Ratio Lower Upper

252 100

254 65.2 52.2 78.2

126 11.8 8.8 13.2









#85
Benzo(a)anthracene
Concen: 125.942 ng/ul
RT: 21.195 min Scan# 3121
Delta R.T. 0.006 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:228 Resp:10694291
Ion Ratio Lower Upper
228 100
229 22.4 15.9 23.9
226 31.0 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 126.754 ng/ul
RT: 21.113 min Scan# 3107
Delta R.T. 0.000 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:149 Resp: 6576504
Ion Ratio Lower Upper
149 100
167 29.4 21.6 32.4
279 5.0 3.3 4.9#





#87

Chrysene

Concen: 123.378 ng/ul

RT: 21.248 min Scan# 3130

Delta R.T. 0.012 min

Lab File: BM032141.D

Acq: 24 Sep 2021 21:24

Instrument :

BNA_M

ClientSampleId :

SSTD160044





Tgt Ion:228 Resp:10132065

Ion Ratio Lower Upper

228 100

226 34.4 24.4 36.6

229 22.5 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 135.828 ng/ul




RT: 21.960 min Scan# 3251


Delta R.T. 0.000 min

Lab File: BM032141.D

Acq: 24 Sep 2021 21:24

Tgt Ion:149 Resp:11995335









#90
Benzo(b)fluoranthene
Concen: 128.174 ng/ul
RT: 22.736 min Scan# 3383
Delta R.T. 0.012 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:252 Resp:11832206
Ion Ratio Lower Upper
252 100
253 24.4 18.2 27.4
125 11.0 8.5 12.7



#91
Benzo(k)fluoranthene
Concen: 130.948 ng/ul
RT: 22.783 min Scan# 3391
Delta R.T. 0.018 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:252 Resp:11093715
Ion Ratio Lower Upper
252 100
253 24.3 17.9 26.9
125 10.8 8.4 12.6









#93
Benzo(a)pyrene
Concen: 129.918 ng/ul
RT: 23.295 min Scan# 3478
Delta R.T. 0.018 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24




Instrument :
BNA_M
ClientSampleId :
SSTD160044


Tgt Ion:252 Resp:11603626
Ion Ratio Lower Upper
252 100
253 24.1 17.8 26.6
125 12.4 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 98.255 ng/ul
RT: 25.559 min Scan# 3863
Delta R.T. 0.036 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:276 Resp:10122570
Ion Ratio Lower Upper
276 100
138 27.9 21.5 32.3
277 27.2 21.0 31.6









#95
Dibenzo(a,h)anthracene
Concen: 127.122 ng/ul
RT: 25.565 min Scan# 3864
Delta R.T. 0.036 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Instrument :
BNA_M
ClientSampleId :
SSTD160044

Tgt Ion:278 Resp:11106231
Ion Ratio Lower Upper
278 100
139 16.2 13.5 20.3
279 24.7 19.4 29.0



#96
Benzo(g,h,i)perylene
Concen: 145.435 ng/ul
RT: 26.230 min Scan# 3977
Delta R.T. 0.047 min
Lab File: BM032141.D
Acq: 24 Sep 2021 21:24

Tgt Ion:276 Resp:12863798
Ion Ratio Lower Upper
276 100
138 22.8 17.9 26.9
277 24.8 18.6 28.0

