

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092421\
 Data File : BM032142.D
 Acq On : 24 Sep 2021 22:00
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV045

Quant Time: Sep 25 01:17:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 25 01:15:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.666	152	399274	20.000	ng/u1	0.00
20) Naphthalene-d8	10.460	136	1697720	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.313	164	1098616	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.042	188	2216179	20.000	ng/u1	0.00
79) Chrysene-d12	21.200	240	2051672	20.000	ng/u1	0.00
88) Perylene-d12	23.365	264	1906707	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.990	96	66544	7.037	ng/uL	0.00
4) Pyridine-d5	3.413	84	508162	18.930	ng/u1	0.00
7) Phenol-d5	6.848	99	599638	18.920	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.990	67	395007	20.728	ng/u1	0.00
11) 2-Chlorophenol-d4	7.207	132	479768	19.498	ng/u1	0.00
15) 4-Methylphenol-d8	8.390	113	480643	19.394	ng/u1	0.00
21) Nitrobenzene-d5	8.819	128	212374	20.944	ng/u1	0.00
24) 2-Nitrophenol-d4	9.548	143	229582	21.412	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.089	165	493608	20.138	ng/u1	0.00
31) 4-Chloroaniline-d4	10.589	131	670496	20.239	ng/u1	0.00
46) Dimethylphthalate-d6	13.719	166	1444177	20.232	ng/u1	0.00
49) Acenaphthylene-d8	14.007	160	1851911	20.421	ng/u1	0.00
54) 4-Nitrophenol-d4	14.513	143	247017	21.273	ng/u1	0.00
60) Fluorene-d10	15.301	176	1217329	19.987	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.413	200	198981	18.999	ng/u1	0.00
73) Anthracene-d10	17.136	188	1855441	20.065	ng/u1	0.00
81) Pyrene-d10	19.418	212	2019404	18.707	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.230	264	1778289	19.873	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.025	88	67844	6.530	ng/uL	98
5) Pyridine	3.437	79	412184	15.486	ng/u1	98
6) Benzaldehyde	6.790	77	393424	22.033	ng/u1	97
8) Phenol	6.878	94	613165	19.084	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.084	93	495696	19.363	ng/u1	98
12) 2-Chlorophenol	7.237	128	492975	19.207	ng/u1	98
13) 2-Methylphenol	8.125	108	469383	19.302	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.201	45	680686	19.319	ng/u1	99
16) Acetophenone	8.484	105	781730	20.812	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.478	70	386204	20.159	ng/u1	96
18) 4-Methylphenol	8.454	108	508414	19.892	ng/u1	100
19) Hexachloroethane	8.760	117	198829	18.874	ng/u1	98
22) Nitrobenzene	8.860	77	535962	20.493	ng/u1	99
23) Isophorone	9.384	82	1014765	18.639	ng/u1	99
25) 2-Nitrophenol	9.578	139	262769	21.452	ng/u1	96
26) 2,4-Dimethylphenol	9.648	107	566356	19.415	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.866	93	677509	19.467	ng/u1	99
29) 2,4-Dichlorophenol	10.113	162	484042	20.074	ng/u1	98
30) Naphthalene	10.513	128	1629669	19.555	ng/u1	100
32) 4-Chloroaniline	10.613	127	695614	20.583	ng/u1	99
33) Hexachlorobutadiene	10.819	225	316035	19.065	ng/u1	99
34) Caprolactam	11.342	113	150639	20.481	ng/u1	95
35) 4-Chloro-3-methylphenol	11.754	107	522924	20.135	ng/u1	98
36) 2-Methylnaphthalene	12.125	142	1183704	21.159	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092421\
 Data File : BM032142.D
 Acq On : 24 Sep 2021 22:00
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SICV045


Quant Time: Sep 25 01:17:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 25 01:15:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.348	142	1101356	19.342	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.507	216	601280	19.174	ng/ul	98
40) Hexachlorocyclopentadiene	12.501	237	344375	20.491	ng/ul	99
41) 2,4,6-Trichlorophenol	12.748	196	387727	19.767	ng/ul	99
42) 2,4,5-Trichlorophenol	12.819	196	414710	19.916	ng/ul	99
43) 1,1'-Biphenyl	13.142	154	1429232	18.369	ng/ul	99
44) 2-Chloronaphthalene	13.183	162	1145080	19.507	ng/ul	99
45) 2-Nitroaniline	13.383	65	289185	21.796	ng/ul	97
47) Dimethylphthalate	13.766	163	1439880	20.219	ng/ul	100
48) 2,6-Dinitrotoluene	13.877	165	287881	24.347	ng/ul	96
50) Acenaphthylene	14.036	152	1862938	19.559	ng/ul	100
51) 3-Nitroaniline	14.207	138	303888	23.499	ng/ul	98
52) Acenaphthene	14.377	153	1244463	20.084	ng/ul	99
53) 2,4-Dinitrophenol	14.407	184	124616	19.279	ng/ul	94
55) 4-Nitrophenol	14.524	109	199391	21.052	ng/ul	95
56) Dibenzofuran	14.713	168	1763974	19.835	ng/ul	99
57) 2,4-Dinitrotoluene	14.660	165	379096	22.984	ng/ul#	97
58) 2,3,4,6-Tetrachlorophenol	14.942	232	364780	21.474	ng/ul	99
59) Diethylphthalate	15.130	149	1408263	20.231	ng/ul	99
61) Fluorene	15.354	166	1388724	20.125	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.348	204	703596	20.165	ng/ul	100
63) 4-Nitroaniline	15.366	138	292174	25.776	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.430	198	205800	19.906	ng/ul	98
67) N-Nitrosodiphenylamine	15.560	169	1228625	19.891	ng/ul	100
68) 4-Bromophenyl-phenylether	16.236	248	422232	19.953	ng/ul	98
69) Hexachlorobenzene	16.371	284	501371	19.489	ng/ul	98
70) Atrazine	16.501	200	433876	19.871	ng/ul	98
71) Pentachlorophenol	16.707	266	286482	19.386	ng/ul	99
72) Phenanthrene	17.083	178	2195358	19.842	ng/ul	99
74) Anthracene	17.171	178	2248034	20.272	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.119	216	597386	17.487	ng/ul	98
76) Pentachlorobenzene	14.642	250	582696	17.615	ng/ul	99
77) Carbazole	17.436	167	2003855	21.363	ng/ul	99
78) Di-n-butylphthalate	17.995	149	2317655	21.507	ng/ul	99
80) Fluoranthene	19.089	202	2451674	18.302	ng/ul	100
82) Pyrene	19.448	202	2587472	18.835	ng/ul	98
83) Butylbenzylphthalate	20.336	149	873539	22.114	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.118	252	823432	24.200	ng/ul	99
85) Benzo(a)anthracene	21.183	228	2306928	19.957	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.106	149	1331615	22.771	ng/ul	99
87) Chrysene	21.236	228	2287115	19.635	ng/ul	99
89) Di-n-octyl phthalate	21.953	149	2115040	21.831	ng/ul	100
90) Benzo(b)fluoranthene	22.718	252	2221882	19.731	ng/ul	99
91) Benzo(k)fluoranthene	22.759	252	2142511	20.076	ng/ul	99
93) Benzo(a)pyrene	23.271	252	2084013	19.599	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	25.518	276	2094862	19.330	ng/ul	99
95) Dibenzo(a,h)anthracene	25.524	278	1767508	19.664	ng/ul	100
96) Benzo(g,h,i)perylene	26.177	276	1745494	18.171	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 25 01:17:02 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 25 01:15:33 2021
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.530 ng/uL
RT: 3.025 min Scan# 32
Delta R.T. 0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion: 88 Resp: 67844
Ion Ratio Lower Upper
88 100
43 28.2 22.3 33.5
58 76.1 62.4 93.6



#5
Pyridine
Concen: 15.486 ng/uL
RT: 3.437 min Scan# 102
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion: 79 Resp: 412184
Ion Ratio Lower Upper
79 100
52 72.3 58.3 87.5
51 34.4 29.1 43.7









#6
Benzaldehyde
Concen: 22.033 ng/ul
RT: 6.790 min Scan# 672
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion: 77 Resp: 393424
Ion Ratio Lower Upper
77 100
105 98.3 77.0 115.4
106 96.7 74.2 111.2



#8
Phenol
Concen: 19.084 ng/ul
RT: 6.878 min Scan# 687
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion: 94 Resp: 613165
Ion Ratio Lower Upper
94 100
65 26.8 22.4 33.6
66 36.1 29.9 44.9









#10
Bis(2-Chloroethyl)ether
Concen: 19.363 ng/ul
RT: 7.084 min Scan# 722
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:	93	Resp:	495696
Ion	Ratio	Lower	Upper
93	100		
63	76.6	63.4	95.2
95	32.2	25.5	38.3



#12
2-Chlorophenol
Concen: 19.207 ng/ul
RT: 7.237 min Scan# 748
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:	128	Resp:	492975
Ion	Ratio	Lower	Upper
128	100		
64	45.8	38.3	57.5
130	31.8	25.8	38.8









#13
2-Methylphenol
Concen: 19.302 ng/ul
RT: 8.125 min Scan# 899
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:108 Resp: 469383
Ion Ratio Lower Upper
108 100
107 89.8 71.6 107.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.319 ng/ul
RT: 8.201 min Scan# 912
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion: 45 Resp: 680686
Ion Ratio Lower Upper
45 100
77 14.5 11.2 16.8
79 11.1 9.1 13.7









#16
Acetophenone
Concen: 20.812 ng/ul
RT: 8.484 min Scan# 960
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:105 Resp: 781730
Ion Ratio Lower Upper
105 100
77 75.6 63.0 94.4
51 26.2 22.2 33.2



#17
N-Nitroso-di-n-propylamine
Concen: 20.159 ng/ul
RT: 8.478 min Scan# 959
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion: 70 Resp: 386204
Ion Ratio Lower Upper
70 100
42 55.0 46.2 69.4
101 9.8 7.5 11.3
130 22.4 16.2 24.2









#18
4-Methylphenol
Concen: 19.892 ng/ul
RT: 8.454 min Scan# 955
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:108 Resp: 508414
Ion Ratio Lower Upper
108 100
107 111.6 89.1 133.7



#19
Hexachloroethane
Concen: 18.874 ng/ul
RT: 8.760 min Scan# 1007
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:117 Resp: 198829
Ion Ratio Lower Upper
117 100
201 100.6 79.5 119.3
199 63.6 49.1 73.7









#22
Nitrobenzene
Concen: 20.493 ng/ul
RT: 8.860 min Scan# 1024
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion: 77 Resp: 535962
Ion Ratio Lower Upper
77 100
123 44.9 36.1 54.1
65 13.1 11.0 16.6



#23
Isophorone
Concen: 18.639 ng/ul
RT: 9.384 min Scan# 1113
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion: 82 Resp: 1014765
Ion Ratio Lower Upper
82 100
95 7.4 5.8 8.6
138 18.3 14.1 21.1









#25
2-Nitrophenol
Concen: 21.452 ng/ul
RT: 9.578 min Scan# 1146
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:139 Resp: 262769
Ion Ratio Lower Upper
139 100
65 50.9 42.0 63.0
109 28.8 25.8 38.8



#26
2,4-Dimethylphenol
Concen: 19.415 ng/ul
RT: 9.648 min Scan# 1158
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:107 Resp: 566356
Ion Ratio Lower Upper
107 100
121 50.8 40.4 60.6
122 85.7 67.6 101.4









#27

Bis(2-Chloroethoxy)methane
Concen: 19.467 ng/ul
RT: 9.866 min Scan# 1195
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Instrument :
BNA_M
ClientSampleId :
SICV045




Tgt Ion: 93	Resp: 677509
Ion Ratio	Lower Upper
93 100	
95 32.7	26.2 39.4
123 10.6	9.0 13.6




#29

2,4-Dichlorophenol
Concen: 20.074 ng/ul
RT: 10.113 min Scan# 1237
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:162	Resp: 484042		
Ion Ratio	Lower	Upper	
162 100			
164 65.8	51.6	77.4	
98 37.0	28.6	42.8	









#30
Naphthalene
Concen: 19.555 ng/ul
RT: 10.513 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:128 Resp: 1629669
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 12.9 10.2 15.4



#32
4-Chloroaniline
Concen: 20.583 ng/ul
RT: 10.613 min Scan# 1322
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:127 Resp: 695614
Ion Ratio Lower Upper
127 100
129 31.8 25.8 38.6









#33
Hexachlorobutadiene
Concen: 19.065 ng/ul
RT: 10.819 min Scan# 1357
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:225 Resp: 316035
Ion Ratio Lower Upper
225 100
223 63.1 50.6 76.0
227 63.2 52.2 78.4



#34
Caprolactam
Concen: 20.481 ng/ul
RT: 11.342 min Scan# 1446
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:113 Resp: 150639
Ion Ratio Lower Upper
113 100
55 168.1 138.1 207.1
56 124.3 106.0 159.0









#35
4-Chloro-3-methylphenol
Concen: 20.135 ng/ul
RT: 11.754 min Scan# 1516
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:107 Resp: 522924
Ion Ratio Lower Upper
107 100
144 26.1 20.3 30.5
142 79.9 62.8 94.2



#36
2-Methylnaphthalene
Concen: 21.159 ng/ul
RT: 12.125 min Scan# 1579
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:142 Resp: 1183704
Ion Ratio Lower Upper
142 100
141 88.0 70.2 105.2









#37
1-Methylnaphthalene
Concen: 19.342 ng/ul
RT: 12.348 min Scan# 1617
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:142 Resp: 1101356
Ion Ratio Lower Upper
142 100
141 91.8 74.3 111.5
116 3.8 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.174 ng/ul
RT: 12.507 min Scan# 1644
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:216 Resp: 601280
Ion Ratio Lower Upper
216 100
214 78.3 64.2 96.4
179 21.3 17.1 25.7
108 19.1 16.0 24.0









#40
Hexachlorocyclopentadiene
Concen: 20.491 ng/ul
RT: 12.501 min Scan# 1643
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:237 Resp: 344375
Ion Ratio Lower Upper
237 100
235 62.9 49.4 74.0
272 12.7 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 19.767 ng/ul
RT: 12.748 min Scan# 1685
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:196 Resp: 387727
Ion Ratio Lower Upper
196 100
198 98.5 77.8 116.6
200 31.4 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 19.916 ng/ul
RT: 12.819 min Scan# 1697
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:196 Resp: 414710
Ion Ratio Lower Upper
196 100
198 96.0 77.4 116.2
200 30.8 24.3 36.5



#43
1,1'-Biphenyl
Concen: 18.369 ng/ul
RT: 13.142 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:154 Resp: 1429232
Ion Ratio Lower Upper
154 100
153 39.6 31.9 47.9
76 12.6 10.7 16.1









#44
2-Chloronaphthalene
Concen: 19.507 ng/ul
RT: 13.183 min Scan# 1759
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:162 Resp: 1145080
Ion Ratio Lower Upper
162 100
164 32.8 25.9 38.9
127 39.8 32.0 48.0



#45
2-Nitroaniline
Concen: 21.796 ng/ul
RT: 13.383 min Scan# 1793
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion: 65 Resp: 289185
Ion Ratio Lower Upper
65 100
92 69.8 54.5 81.7
138 109.7 84.1 126.1









#47
Dimethylphthalate
Concen: 20.219 ng/ul
RT: 13.766 min Scan# 1858
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:163 Resp: 1439880
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 10.3 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 24.347 ng/ul
RT: 13.877 min Scan# 1877
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:165 Resp: 287881
Ion Ratio Lower Upper
165 100
89 55.3 46.8 70.2
121 20.4 17.3 25.9









#50
Acenaphthylene
Concen: 19.559 ng/ul
RT: 14.036 min Scan# 1904
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:152 Resp: 1862938
Ion Ratio Lower Upper
152 100
151 19.8 15.7 23.5
153 13.1 10.6 15.8



#51
3-Nitroaniline
Concen: 23.499 ng/ul
RT: 14.207 min Scan# 1933
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:138 Resp: 303888
Ion Ratio Lower Upper
138 100
108 11.6 10.5 15.7
92 116.1 94.0 141.0









#52
Acenaphthene
Concen: 20.084 ng/ul
RT: 14.377 min Scan# 1962
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:153 Resp: 1244463
Ion Ratio Lower Upper
153 100
152 47.3 38.4 57.6
154 89.6 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 19.279 ng/ul
RT: 14.407 min Scan# 1967
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:184 Resp: 124616
Ion Ratio Lower Upper
184 100
63 64.8 56.7 85.1
154 67.6 57.5 86.3









#55
4-Nitrophenol
Concen: 21.052 ng/ul
RT: 14.524 min Scan# 1987
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:109 Resp: 199391
Ion Ratio Lower Upper
109 100
139 134.2 103.7 155.5
65 129.4 109.4 164.2



#56
Dibenzofuran
Concen: 19.835 ng/ul
RT: 14.713 min Scan# 2019
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:168 Resp: 1763974
Ion Ratio Lower Upper
168 100
139 36.4 29.8 44.6









#57
2,4-Dinitrotoluene
Concen: 22.984 ng/ul
RT: 14.660 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:165 Resp: 379096
Ion Ratio Lower Upper
165 100
63 44.7 34.2 51.2
182 3.6 2.3 3.5#



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.474 ng/ul
RT: 14.942 min Scan# 2058
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:232 Resp: 364780
Ion Ratio Lower Upper
232 100
131 44.6 36.2 54.2
130 2.3 1.9 2.9
166 30.7 24.8 37.2









#59
Diethylphthalate
Concen: 20.231 ng/ul
RT: 15.130 min Scan# 2090
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:149 Resp: 1408263
Ion Ratio Lower Upper
149 100
177 23.0 17.9 26.9
150 12.5 10.0 15.0



#61
Fluorene
Concen: 20.125 ng/ul
RT: 15.354 min Scan# 2128
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:166 Resp: 1388724
Ion Ratio Lower Upper
166 100
165 98.8 77.7 116.5
167 13.5 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 20.165 ng/ul
RT: 15.348 min Scan# 2127
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:204 Resp: 703596
Ion Ratio Lower Upper
204 100
206 32.5 25.8 38.6
141 60.0 47.9 71.9



#63
4-Nitroaniline
Concen: 25.776 ng/ul
RT: 15.366 min Scan# 2130
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:138 Resp: 292174
Ion Ratio Lower Upper
138 100
92 50.6 41.4 62.2
108 77.0 63.8 95.6









#66
4,6-Dinitro-2-methylphenol
Concen: 19.906 ng/ul
RT: 15.430 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:198 Resp: 205800
Ion Ratio Lower Upper
198 100
182 4.3 3.8 5.8
77 23.8 19.8 29.8



#67
N-Nitrosodiphenylamine
Concen: 19.891 ng/ul
RT: 15.560 min Scan# 2163
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:169 Resp: 1228625
Ion Ratio Lower Upper
169 100
168 65.4 52.3 78.5
167 35.7 28.6 42.8









#68
4-Bromophenyl-phenylether
Concen: 19.953 ng/ul
RT: 16.236 min Scan# 2278
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampled :
SICV045


Tgt Ion:248 Resp: 422232
Ion Ratio Lower Upper
248 100
250 97.7 77.0 115.6
141 73.5 57.6 86.4



#69
Hexachlorobenzene
Concen: 19.489 ng/ul
RT: 16.371 min Scan# 2301
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:284 Resp: 501371
Ion Ratio Lower Upper
284 100
142 33.3 28.4 42.6
249 30.6 24.7 37.1





#70

Atrazine

Concen: 19.871 ng/ul

RT: 16.501 min Scan# 2323

Delta R.T. -0.000 min

Lab File: BM032142.D

Acq: 24 Sep 2021 22:00

Tgt Ion:200 Resp: 433876

Ion Ratio Lower Upper

200 100

173 26.2 21.4 32.0





215 47.3 39.4 59.2

Instrument :

BNA_M

ClientSampleId :

SICV045



#71

Pentachlorophenol

Concen: 19.386 ng/ul

RT: 16.707 min Scan# 2358

Delta R.T. -0.000 min

Lab File: BM032142.D

Acq: 24 Sep 2021 22:00




Tgt Ion:266 Resp: 286482


Ion Ratio Lower Upper

266 100

264 61.1 50.2 75.2

268 64.2 51.0 76.6









#72
Phenanthrene
Concen: 19.842 ng/ul
RT: 17.083 min Scan# 2422
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:178 Resp: 2195358
Ion Ratio Lower Upper
178 100
179 15.1 12.3 18.5
176 19.0 15.4 23.0



#74
Anthracene
Concen: 20.272 ng/ul
RT: 17.171 min Scan# 2437
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:178 Resp: 2248034
Ion Ratio Lower Upper
178 100
179 15.4 12.6 19.0
176 18.7 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 17.487 ng/uL
RT: 13.119 min Scan# 1748
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampled :
SICV045


Tgt Ion:216 Resp: 597386
Ion Ratio Lower Upper
216 100
214 76.7 62.7 94.1
218 47.8 38.7 58.1



#76
Pentachlorobenzene
Concen: 17.615 ng/uL
RT: 14.642 min Scan# 2007
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:250 Resp: 582696
Ion Ratio Lower Upper
250 100
248 64.4 50.0 75.0
252 64.4 51.6 77.4









#77
 Carbazole
 Concen: 21.363 ng/ul
 RT: 17.436 min Scan# 2482
 Delta R.T. -0.000 min
 Lab File: BM032142.D
 Acq: 24 Sep 2021 22:00




Instrument :
 BNA_M
 ClientSampleId :
 SICV045


Tgt Ion:167 Resp: 2003855
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.3 25.9
 139 12.8 10.3 15.5



#78
 Di-n-butylphthalate
 Concen: 21.507 ng/ul
 RT: 17.995 min Scan# 2577
 Delta R.T. -0.000 min
 Lab File: BM032142.D
 Acq: 24 Sep 2021 22:00

Tgt Ion:149 Resp: 2317655
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.4 11.0
 104 4.9 4.2 6.2









#80
Fluoranthene
Concen: 18.302 ng/ul
RT: 19.089 min Scan# 2763
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:202 Resp: 2451674
Ion Ratio Lower Upper
202 100
101 11.3 9.2 13.8
100 8.6 6.9 10.3



#82
Pyrene
Concen: 18.835 ng/ul
RT: 19.448 min Scan# 2824
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:202 Resp: 2587472
Ion Ratio Lower Upper
202 100
101 13.7 10.1 15.1
100 11.1 8.4 12.6









#83
Butylbenzylphthalate
Concen: 22.114 ng/ul
RT: 20.336 min Scan# 2975
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:149 Resp: 873539
Ion Ratio Lower Upper
149 100
91 72.7 61.8 92.8
206 21.5 16.2 24.4



#84
3,3'-Dichlorobenzidine
Concen: 24.200 ng/ul
RT: 21.118 min Scan# 3108
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:252 Resp: 823432
Ion Ratio Lower Upper
252 100
254 64.6 52.2 78.2
126 10.7 8.8 13.2









#85
Benzo(a)anthracene
Concen: 19.957 ng/ul
RT: 21.183 min Scan# 3119
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:228 Resp: 2306928
Ion Ratio Lower Upper
228 100
229 20.2 15.9 23.9
226 27.4 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 22.771 ng/ul
RT: 21.106 min Scan# 3106
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:149 Resp: 1331615
Ion Ratio Lower Upper
149 100
167 27.7 21.6 32.4
279 3.8 3.3 4.9









#87
Chrysene
Concen: 19.635 ng/ul
RT: 21.236 min Scan# 3128
Delta R.T. -0.000 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Instrument :
BNA_M
ClientSampleId :
SICV045




Tgt Ion:228 Resp: 2287115


Ion	Ratio	Lower	Upper
228	100		
226	30.0	24.4	36.6
229	19.3	15.9	23.9



#89
Di-n-octyl phthalate
Concen: 21.831 ng/ul
RT: 21.953 min Scan# 3250
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:149 Resp: 2115040









#90
Benzo(b)fluoranthene
Concen: 19.731 ng/ul
RT: 22.718 min Scan# 3380
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:252 Resp: 2221882
Ion Ratio Lower Upper
252 100
253 22.2 18.2 27.4
125 10.4 8.5 12.7



#91
Benzo(k)fluoranthene
Concen: 20.076 ng/ul
RT: 22.759 min Scan# 3387
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:252 Resp: 2142511
Ion Ratio Lower Upper
252 100
253 21.6 17.9 26.9
125 10.8 8.4 12.6









#93
Benzo(a)pyrene
Concen: 19.599 ng/ul
RT: 23.271 min Scan# 3474
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00




Instrument :
BNA_M
ClientSampleId :
SICV045


Tgt Ion:252 Resp: 2084013
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.6
125 11.2 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 19.330 ng/ul
RT: 25.518 min Scan# 3856
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:276 Resp: 2094862
Ion Ratio Lower Upper
276 100
138 26.4 21.5 32.3
277 25.4 21.0 31.6









#95
Dibenzo(a,h)anthracene
Concen: 19.664 ng/ul
RT: 25.524 min Scan# 3857
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Instrument :
BNA_M
ClientSampleId :
SICV045

Tgt Ion:278 Resp: 1767508
Ion Ratio Lower Upper
278 100
139 16.7 13.5 20.3
279 24.0 19.4 29.0



#96
Benzo(g,h,i)perylene
Concen: 18.171 ng/ul
RT: 26.177 min Scan# 3968
Delta R.T. -0.006 min
Lab File: BM032142.D
Acq: 24 Sep 2021 22:00

Tgt Ion:276 Resp: 1745494
Ion Ratio Lower Upper
276 100
138 22.6 17.9 26.9
277 23.0 18.6 28.0

