

Data Path : Z:\HPCHEM1\BNA_M\Data\BM092515\
 Data File : BM002727.D
 Acq On : 25 Sep 2015 12:40
 Operator : UM/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.486

Quant Time: Sep 25 15:18:48 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM092415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 15:05:30 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.74	152	10311	0.40	ng/ul	0.00
4) Naphthalene-d8	11.59	136	38525	0.40	ng/ul	0.00
8) Acenaphthene-d10	15.33	164	17907	0.40	ng/ul	0.00
12) Phenanthrene-d10	18.03	188	36661	0.40	ng/ul	0.00
18) Chrysene-d12	22.13	240	34716	0.40	ng/ul	0.00
22) Perylene-d12	24.72	264	31682	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) 1,4-Dioxane-d8	3.80	96	10615	1.00	ng/uL	0.00
6) 2-Methylnaphthalene-d10	13.13	152	20127	0.36	ng/ul	0.00
16) Fluoranthene-d10	20.01	212	35941	0.39	ng/ul	0.01

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) 1,4-Dioxane	3.83	88	12317	1.00	ng/ul	97
5) Naphthalene	11.64	128	40612	0.39	ng/ul	98
7) 2-Methylnaphthalene	13.20	142	26274	0.39	ng/ul	100
9) Acenaphthylene	15.04	152	34029	0.37	ng/ul	96
10) Acenaphthene	15.38	153	26822	0.40	ng/ul	88
11) Fluorene	16.36	166	29266	0.38	ng/ul	84
13) Pentachlorophenol	17.69	266	1950	0.27	ng/ul	95
14) Phenanthrene	18.08	178	44314	0.37	ng/ul	94
15) Anthracene	18.16	178	42425	0.38	ng/ul	96
17) Fluoranthene	20.03	202	49466	0.38	ng/ul	93
19) Pyrene	20.39	202	51036	0.38	ng/ul	98
20) Benzo(a)anthracene	22.09	228	45052	0.39	ng/ul	99
21) Chrysene	22.16	228	44692	0.40	ng/ul	96
23) Benzo(b)fluoranthene	23.92	252	45371	0.40	ng/ul	98
24) Benzo(k)fluoranthene	23.97	252	43264	0.41	ng/ul	97
25) Benzo(a)pyrene	24.61	252	42235	0.40	ng/ul	96
26) Indeno(1,2,3-cd)pyrene	27.45	276	47301	0.40	ng/ul	99
27) Dibenzo(a,h)anthracene	27.45	278	38494	0.40	ng/ul	95
28) Benzo(g,h,i)perylene	28.31	276	40402	0.41	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\Data\BM092515\
Data File : BM002727.D
Acq On : 25 Sep 2015 12:40
Operator : UM/IZ
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.486

Quant Time: Sep 25 15:18:48 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM092415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 25 15:05:30 2015
Response via : Initial Calibration

