

Data Path : Z:\HPCHEM1\BNA_M\Data\BM092515\
 Data File : BM002725.D
 Acq On : 25 Sep 2015 10:14
 Operator : UM/IZ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.485

Quant Time: Sep 25 15:10:16 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM092415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 15:05:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	0.00
2 SURR	1,4-Dioxane-d8	0.413	0.410	0.7	121	0.00
3	1,4-Dioxane	0.479	0.476	0.6	121	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
5	Naphthalene	1.084	1.055	2.7	115	0.00
6 SURR	2-Methylnaphthalene-d10	0.585	0.533	8.9	114	0.00
7	2-Methylnaphthalene	0.698	0.690	1.1	117	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
9	Acenaphthylene	2.063	1.875	9.1	111	0.00
10 C	Acenaphthene	1.506	1.493	0.9	114	0.00
11	Fluorene	1.717	1.641	4.4	113	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
13	Pentachlorophenol	0.078	0.052	33.3#	76	0.00
14	Phenanthrene	1.302	1.202	7.7	111	0.00
15	Anthracene	1.203	1.164	3.2	111	0.00
16 SURR	Fluoranthene-d10	0.997	0.995	0.2	113	0.00
17 C	Fluoranthene	1.416	1.354	4.4	112	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	116	0.00
19	Pyrene	1.559	1.482	4.9	112	0.00
20	Benzo(a)anthracene	1.337	1.306	2.3	114	0.00
21	Chrysene	1.280	1.286	-0.5	115	0.00
22 I	Perylene-d12	1.000	1.000	0.0	114	0.00
23	Benzo(b)fluoranthene	1.442	1.443	-0.1	120	0.00
24	Benzo(k)fluoranthene	1.336	1.353	-1.3	111	0.00
25 C	Benzo(a)pyrene	1.328	1.327	0.1	114	0.00
26	Indeno(1,2,3-cd)pyrene	1.481	1.474	0.5	114	0.00
27	Dibenzo(a,h)anthracene	1.215	1.198	1.4	113	0.00
28	Benzo(g,h,i)perylene	1.256	1.256	0.0	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0