

Data Path : Z:\HPCHEM1\BNA_M\Data\BM092515\
 Data File : BM002727.D
 Acq On : 25 Sep 2015 12:40
 Operator : UM/IZ
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.486

Quant Time: Sep 25 15:18:48 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM092415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 15:05:30 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	0.00
2 SURR	1,4-Dioxane-d8	0.413	0.412	0.2	122	0.00
3	1,4-Dioxane	0.479	0.478	0.2	121	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
5	Naphthalene	1.084	1.054	2.8	114	0.00
6 SURR	2-Methylnaphthalene-d10	0.585	0.522	10.8	111	0.00
7	2-Methylnaphthalene	0.698	0.682	2.3	115	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
9	Acenaphthylene	2.063	1.900	7.9	111	0.00
10 C	Acenaphthene	1.506	1.498	0.5	112	0.00
11	Fluorene	1.717	1.634	4.8	110	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
13	Pentachlorophenol	0.078	0.053	32.1#	76	0.00
14	Phenanthrene	1.302	1.209	7.1	109	0.00
15	Anthracene	1.203	1.157	3.8	107	0.00
16 SURR	Fluoranthene-d10	0.997	0.980	1.7	109	0.01
17 C	Fluoranthene	1.416	1.349	4.7	109	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
19	Pyrene	1.559	1.470	5.7	109	0.00
20	Benzo(a)anthracene	1.337	1.298	2.9	111	0.00
21	Chrysene	1.280	1.287	-0.5	113	0.00
22 I	Perylene-d12	1.000	1.000	0.0	113	0.00
23	Benzo(b)fluoranthene	1.442	1.432	0.7	117	0.00
24	Benzo(k)fluoranthene	1.336	1.366	-2.2	110	0.00
25 C	Benzo(a)pyrene	1.328	1.333	-0.4	113	0.00
26	Indeno(1,2,3-cd)pyrene	1.481	1.493	-0.8	114	0.00
27	Dibenzo(a,h)anthracene	1.215	1.215	0.0	112	-0.01
28	Benzo(g,h,i)perylene	1.256	1.275	-1.5	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0