

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM092520\  
 Data File : BM027544.D  
 Acq On : 26 Sep 2020 00:54  
 Operator : JU/CG  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 26 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTD02012

Quant Time: Sep 26 02:47:57 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SOM-EPA-BM092420MA.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Thu Sep 24 20:21:51 2020  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc  | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|-------|--------|----------|
| -----                         |        |      |          |       |        |          |
| Internal Standards            |        |      |          |       |        |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.804  | 152  | 112185   | 20.00 | ng/ul  | # 0.00   |
| 18) Naphthalene-d8            | 10.592 | 136  | 434324   | 20.00 | ng/ul  | # 0.00   |
| 36) Acenaphthene-d10          | 14.439 | 164  | 348833   | 20.00 | ng/ul  | 0.00     |
| 62) Phenanthrene-d10          | 17.180 | 188  | 867643   | 20.00 | ng/ul  | # 0.00   |
| 78) Chrysene-d12              | 21.356 | 240  | 871748   | 20.00 | ng/ul  | # 0.00   |
| 86) Perylene-d12              | 23.621 | 264  | 770119   | 20.00 | ng/ul  | -0.02    |
| System Monitoring Compounds   |        |      |          |       |        |          |
| 3) 1,4-Dioxane-d8             | 3.310  | 96   | 17987    | 9.12  | ng/uL  | 0.00     |
| 5) Phenol-d5                  | 6.969  | 99   | 149930   | 20.50 | ng/ul  | 0.00     |
| 7) Bis-(2-Chloroethyl)eth...  | 7.140  | 67   | 80168    | 18.65 | ng/ul  | 0.00     |
| 9) 2-Chlorophenol-d4          | 7.340  | 132  | 127720   | 20.18 | ng/ul  | 0.00     |
| 13) 4-Methylphenol-d8         | 8.504  | 113  | 130119   | 21.22 | ng/ul  | 0.00     |
| 19) Nitrobenzene-d5           | 8.957  | 128  | 63247    | 19.76 | ng/ul  | 0.00     |
| 22) 2-Nitrophenol-d4          | 9.675  | 143  | 71294    | 21.25 | ng/ul  | 0.00     |
| 26) 2,4-Dichlorophenol-d3     | 10.210 | 165  | 181613   | 21.22 | ng/ul  | 0.00     |
| 29) 4-Chloroaniline-d4        | 10.722 | 131  | 181170   | 21.36 | ng/ul  | 0.00     |
| 44) Dimethylphthalate-d6      | 13.851 | 166  | 554837   | 20.56 | ng/ul  | 0.00     |
| 47) Acenaphthylene-d8         | 14.127 | 160  | 642793   | 19.88 | ng/ul  | 0.00     |
| 52) 4-Nitrophenol-d4          | 14.627 | 143  | 78621    | 20.79 | ng/ul  | 0.00     |
| 58) Fluorene-d10              | 15.427 | 176  | 501334   | 20.65 | ng/ul  | 0.00     |
| 63) 4,6-Dinitro-2-methylph... | 15.545 | 200  | 91000    | 19.37 | ng/ul  | 0.00     |
| 71) Anthracene-d10            | 17.274 | 188  | 816500   | 19.95 | ng/ul  | 0.00     |
| 79) Pyrene-d10                | 19.568 | 212  | 940692   | 20.69 | ng/ul  | 0.00     |
| 90) Benzo(a)pyrene-d12        | 23.480 | 264  | 773746   | 18.91 | ng/ul  | -0.01    |
| Target Compounds              |        |      |          |       |        |          |
|                               |        |      |          |       |        | Qvalue   |
| 2) 1,4-Dioxane                | 3.346  | 88   | 17961    | 7.72  | ng/uL  | 97       |
| 4) Benzaldehyde               | 6.946  | 77   | 101539   | 20.47 | ng/ul  | 90       |
| 6) Phenol                     | 6.998  | 94   | 158404   | 21.04 | ng/ul  | 90       |
| 8) Bis(2-Chloroethyl)ether    | 7.234  | 93   | 121555   | 20.23 | ng/ul  | 90       |
| 10) 2-Chlorophenol            | 7.369  | 128  | 135583   | 20.85 | ng/ul# | 87       |
| 11) 2-Methylphenol            | 8.240  | 108  | 123430   | 20.79 | ng/ul  | 98       |
| 12) 2,2'-oxybis(1-Chloropr... | 8.334  | 45   | 59431    | 12.73 | ng/ul  | 97       |
| 14) Acetophenone              | 8.622  | 105  | 215597   | 21.65 | ng/ul  | 90       |
| 15) N-Nitroso-di-n-propyla... | 8.610  | 70   | 107101   | 21.55 | ng/ul# | 92       |
| 16) 4-Methylphenol            | 8.569  | 108  | 136684   | 21.65 | ng/ul  | 98       |
| 17) Hexachloroethane          | 8.887  | 117  | 58578    | 18.86 | ng/ul# | 52       |
| 20) Nitrobenzene              | 8.998  | 77   | 161861   | 18.48 | ng/ul# | 87       |
| 21) Isophorone                | 9.522  | 82   | 302621   | 20.07 | ng/ul# | 93       |
| 23) 2-Nitrophenol             | 9.704  | 139  | 81955    | 22.25 | ng/ul# | 87       |
| 24) 2,4-Dimethylphenol        | 9.769  | 107  | 171144   | 20.93 | ng/ul  | 94       |
| 25) Bis(2-Chloroethoxy)met... | 10.004 | 93   | 172522   | 19.98 | ng/ul  | 99       |
| 27) 2,4-Dichlorophenol        | 10.239 | 162  | 178096   | 21.43 | ng/ul# | 91       |
| 28) Naphthalene               | 10.639 | 128  | 444149   | 20.03 | ng/ul  | 100      |
| 30) 4-Chloroaniline           | 10.745 | 127  | 185395   | 22.21 | ng/ul  | 99       |
| 31) Hexachlorobutadiene       | 10.934 | 225  | 162093   | 19.26 | ng/ul  | 99       |
| 32) Caprolactam               | 11.510 | 113  | 43266    | 24.92 | ng/ul# | 73       |
| 33) 4-Chloro-3-methylphenol   | 11.869 | 107  | 154713   | 22.59 | ng/ul  | 90       |
| 34) 2-Methylnaphthalene       | 12.251 | 142  | 359010   | 21.32 | ng/ul  | 99       |
| 35) 1-Methylnaphthalene       | 12.475 | 142  | 338373   | 20.54 | ng/ul  | 97       |
| 37) 1,2,4,5-Tetrachloroben... | 12.622 | 216  | 306759   | 19.36 | ng/ul# | 96       |
| 38) Hexachlorocyclopentadiene | 12.610 | 237  | 202382   | 19.10 | ng/ul  | 98       |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM092520\  
 Data File : BM027544.D  
 Acq On : 26 Sep 2020 00:54  
 Operator : JU/CG  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 26 Sample Multiplier: 1


Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTD02012

Quant Time: Sep 26 02:47:57 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SOM-EPA-BM092420MA.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Thu Sep 24 20:21:51 2020  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc  | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|-------|--------|----------|
| 39) 2,4,6-Trichlorophenol     | 12.863 | 196  | 182135   | 21.25 | ng/ul  | 98       |
| 40) 2,4,5-Trichlorophenol     | 12.933 | 196  | 198957   | 21.81 | ng/ul  | 98       |
| 41) 1,1'-Biphenyl             | 13.269 | 154  | 520231   | 19.55 | ng/ul# | 98       |
| 42) 2-Chloronaphthalene       | 13.310 | 162  | 428540   | 19.28 | ng/ul  | 96       |
| 43) 2-Nitroaniline            | 13.510 | 65   | 92855    | 20.81 | ng/ul# | 80       |
| 45) Dimethylphthalate         | 13.898 | 163  | 566961   | 20.59 | ng/ul  | 99       |
| 46) 2,6-Dinitrotoluene        | 14.010 | 165  | 114914   | 21.96 | ng/ul# | 80       |
| 48) Acenaphthylene            | 14.157 | 152  | 615692   | 20.04 | ng/ul  | 98       |
| 49) 3-Nitroaniline            | 14.333 | 138  | 96092    | 23.17 | ng/ul  | 83       |
| 50) Acenaphthene              | 14.504 | 153  | 433570   | 19.86 | ng/ul  | 97       |
| 51) 2,4-Dinitrophenol         | 14.539 | 184  | 57832    | 19.24 | ng/ul# | 66       |
| 53) 4-Nitrophenol             | 14.639 | 109  | 85497    | 20.43 | ng/ul# | 83       |
| 54) Dibenzofuran              | 14.833 | 168  | 674096   | 20.53 | ng/ul  | 94       |
| 55) 2,4-Dinitrotoluene        | 14.792 | 165  | 162639   | 22.76 | ng/ul# | 80       |
| 56) 2,3,4,6-Tetrachlorophenol | 15.063 | 232  | 182397   | 22.96 | ng/ul# | 82       |
| 57) Diethylphthalate          | 15.269 | 149  | 544662   | 21.55 | ng/ul  | 95       |
| 59) Fluorene                  | 15.486 | 166  | 537137   | 20.13 | ng/ul  | 100      |
| 60) 4-Chlorophenyl-phenyle... | 15.480 | 204  | 342173   | 20.39 | ng/ul  | 90       |
| 61) 4-Nitroaniline            | 15.498 | 138  | 95225    | 21.67 | ng/ul# | 92       |
| 64) 4,6-Dinitro-2-methylph... | 15.557 | 198  | 95894    | 18.64 | ng/ul# | 80       |
| 65) N-Nitrosodiphenylamine    | 15.692 | 169  | 484648   | 20.10 | ng/ul  | 99       |
| 66) 4-Bromophenyl-phenylether | 16.374 | 248  | 232279   | 20.09 | ng/ul  | 93       |
| 67) Hexachlorobenzene         | 16.486 | 284  | 253225   | 19.43 | ng/ul# | 91       |
| 68) Atrazine                  | 16.645 | 200  | 214512   | 20.37 | ng/ul  | 97       |
| 69) Pentachlorophenol         | 16.827 | 266  | 136052   | 19.11 | ng/ul  | 98       |
| 70) Phenanthrene              | 17.221 | 178  | 926203   | 19.75 | ng/ul  | 99       |
| 72) Anthracene                | 17.310 | 178  | 938945   | 19.70 | ng/ul  | 99       |
| 73) 1,2,3,4-Tetrachloroben... | 13.233 | 216  | 302358   | 18.90 | ng/uL  | 98       |
| 74) Pentachlorobenzene        | 14.757 | 250  | 293509   | 18.84 | ng/uL  | 96       |
| 75) Carbazole                 | 17.580 | 167  | 777514   | 20.55 | ng/ul  | 99       |
| 76) Di-n-butylphthalate       | 18.157 | 149  | 902767   | 22.07 | ng/ul  | 98       |
| 77) Fluoranthene              | 19.233 | 202  | 1189468  | 19.88 | ng/ul# | 92       |
| 80) Pyrene                    | 19.598 | 202  | 1207106  | 20.34 | ng/ul# | 90       |
| 81) Butylbenzylphthalate      | 20.503 | 149  | 354973   | 25.62 | ng/ul# | 86       |
| 82) 3,3'-Dichlorobenzidine    | 21.274 | 252  | 374463   | 20.06 | ng/ul# | 97       |
| 83) Benzo(a)anthracene        | 21.339 | 228  | 1154863  | 19.91 | ng/ul  | 100      |
| 84) Bis(2-ethylhexyl)phtha... | 21.286 | 149  | 515770   | 24.22 | ng/ul# | 94       |
| 85) Chrysene                  | 21.392 | 228  | 1090478  | 19.44 | ng/ul  | 99       |
| 87) Di-n-octyl phthalate      | 22.174 | 149  | 785223   | 23.30 | ng/ul  | 100      |
| 88) Benzo(b)fluoranthene      | 22.939 | 252  | 996288   | 19.53 | ng/ul# | 99       |
| 89) Benzo(k)fluoranthene      | 22.986 | 252  | 993721   | 19.51 | ng/ul# | 98       |
| 91) Benzo(a)pyrene            | 23.521 | 252  | 857685   | 18.19 | ng/ul# | 97       |
| 92) Indeno(1,2,3-cd)pyrene    | 25.915 | 276  | 1011615  | 17.08 | ng/ul# | 94       |
| 93) Dibenzo(a,h)anthracene    | 25.927 | 278  | 853548   | 17.13 | ng/ul# | 97       |
| 94) Benzo(g,h,i)perylene      | 26.609 | 276  | 829757   | 16.93 | ng/ul# | 96       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed









#2  
1,4-Dioxane  
Concen: 7.72 ng/uL  
RT: 3.346 min Scan# 61  
Delta R.T. 0.000 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion: 88 Resp: 17961  
Ion Ratio Lower Upper  
88 100  
43 23.0 21.7 32.5  
58 59.5 47.4 71.0



#4  
Benzaldehyde  
Concen: 20.47 ng/uL  
RT: 6.946 min Scan# 673  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion: 77 Resp: 101539  
Ion Ratio Lower Upper  
77 100  
105 91.7 64.9 97.3  
106 86.1 62.8 94.2









#6  
Phenol  
Concen: 21.04 ng/ul  
RT: 6.998 min Scan# 682  
Delta R.T. 0.000 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion: 94 Resp: 158404  
Ion Ratio Lower Upper  
94 100  
65 30.8 30.2 45.4  
66 46.7 42.4 63.6



#8  
Bis(2-Chloroethyl)ether  
Concen: 20.23 ng/ul  
RT: 7.234 min Scan# 722  
Delta R.T. 0.000 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion: 93 Resp: 121555  
Ion Ratio Lower Upper  
93 100  
63 64.3 60.6 90.8  
95 32.6 25.0 37.4









#10  
2-Chlorophenol  
Concen: 20.85 ng/ul  
RT: 7.369 min Scan# 745  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:128 Resp: 135583  
Ion Ratio Lower Upper  
128 100  
64 37.0 41.1 61.7#  
130 33.9 26.8 40.2



#11  
2-Methylphenol  
Concen: 20.79 ng/ul  
RT: 8.240 min Scan# 893  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:108 Resp: 123430  
Ion Ratio Lower Upper  
108 100  
107 91.4 74.4 111.6









#12  
2,2'-oxybis(1-chloropropane)  
Concen: 12.73 ng/ul  
RT: 8.334 min Scan# 909  
Delta R.T. 0.000 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion: 45 Resp: 59431  
Ion Ratio Lower Upper  
45 100  
77 29.4 24.6 36.8  
79 22.6 16.8 25.2



#14  
Acetophenone  
Concen: 21.65 ng/ul  
RT: 8.622 min Scan# 958  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:105 Resp: 215597  
Ion Ratio Lower Upper  
105 100  
77 82.1 74.4 111.6  
51 21.1 19.2 28.8









#15  
N-Nitroso-di-n-propylamine  
Concen: 21.55 ng/ul  
RT: 8.610 min Scan# 956  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleID :  
SSTD02012


Tgt Ion: 70 Resp: 107101  
Ion Ratio Lower Upper  
70 100  
42 42.2 31.0 46.4  
101 9.8 7.0 10.6  
130 24.5 13.6 20.4#



#16  
4-Methylphenol  
Concen: 21.65 ng/ul  
RT: 8.569 min Scan# 949  
Delta R.T. 0.000 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:108 Resp: 136684  
Ion Ratio Lower Upper  
108 100  
107 117.4 92.1 138.1









#17  
Hexachloroethane  
Concen: 18.86 ng/ul  
RT: 8.887 min Scan# 1003  
Delta R.T. 0.000 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:117 Resp: 58578  
Ion Ratio Lower Upper  
117 100  
201 170.7 90.6 135.8#  
199 103.4 56.3 84.5#



#20  
Nitrobenzene  
Concen: 18.48 ng/ul  
RT: 8.998 min Scan# 1022  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion: 77 Resp: 161861  
Ion Ratio Lower Upper  
77 100  
123 43.8 27.2 40.8#  
65 13.7 11.8 17.6









#21  
Isophorone  
Concen: 20.07 ng/ul  
RT: 9.522 min Scan# 1111  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion: 82 Resp: 302621  
Ion Ratio Lower Upper  
82 100  
95 8.8 6.5 9.7  
138 19.7 12.6 18.8#



#23  
2-Nitrophenol  
Concen: 22.25 ng/ul  
RT: 9.704 min Scan# 1142  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion: 139 Resp: 81955  
Ion Ratio Lower Upper  
139 100  
65 47.9 50.2 75.4#  
109 40.3 31.2 46.8









#24  
2,4-Dimethylphenol  
Concen: 20.93 ng/ul  
RT: 9.769 min Scan# 1153  
Delta R.T. 0.000 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:107 Resp: 171144  
Ion Ratio Lower Upper  
107 100  
121 48.0 36.6 54.8  
122 81.3 59.9 89.9



#25  
Bis(2-Chloroethoxy)methane  
Concen: 19.98 ng/ul  
RT: 10.004 min Scan# 1193  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion: 93 Resp: 172522  
Ion Ratio Lower Upper  
93 100  
95 32.2 26.2 39.4  
123 11.5 8.6 13.0









#27  
2,4-Dichlorophenol  
Concen: 21.43 ng/ul  
RT: 10.239 min Scan# 1233  
Delta R.T. 0.000 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:162 Resp: 178096  
Ion Ratio Lower Upper  
162 100  
164 68.7 51.1 76.7  
98 26.8 27.7 41.5#



#28  
Naphthalene  
Concen: 20.03 ng/ul  
RT: 10.639 min Scan# 1301  
Delta R.T. -0.012 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:128 Resp: 444149  
Ion Ratio Lower Upper  
128 100  
129 10.5 8.4 12.6  
127 12.8 10.3 15.5









#30  
4-Chloroaniline  
Concen: 22.21 ng/ul  
RT: 10.745 min Scan# 1319  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:127 Resp: 185395  
Ion Ratio Lower Upper  
127 100  
129 33.2 25.9 38.9



#31  
Hexachlorobutadiene  
Concen: 19.26 ng/ul  
RT: 10.934 min Scan# 1351  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:225 Resp: 162093  
Ion Ratio Lower Upper  
225 100  
223 60.8 49.8 74.6  
227 61.3 49.4 74.2









#32  
Caprolactam  
Concen: 24.92 ng/ul  
RT: 11.510 min Scan# 1449  
Delta R.T. 0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:113 Resp: 43266  
Ion Ratio Lower Upper  
113 100  
55 87.9 100.5 150.7#  
56 83.4 83.7 125.5#



#33  
4-Chloro-3-methylphenol  
Concen: 22.59 ng/ul  
RT: 11.869 min Scan# 1510  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:107 Resp: 154713  
Ion Ratio Lower Upper  
107 100  
144 28.5 19.7 29.5  
142 87.1 62.0 93.0









#34  
2-Methylnaphthalene  
Concen: 21.32 ng/ul  
RT: 12.251 min Scan# 1575  
Delta R.T. -0.012 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:142 Resp: 359010  
Ion Ratio Lower Upper  
142 100  
141 86.4 68.2 102.4



#35  
1-Methylnaphthalene  
Concen: 20.54 ng/ul  
RT: 12.475 min Scan# 1613  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:142 Resp: 338373  
Ion Ratio Lower Upper  
142 100  
141 90.3 74.6 111.8  
116 4.3 3.6 5.4









#37  
1,2,4,5-Tetrachlorobenzene  
Concen: 19.36 ng/ul  
RT: 12.622 min Scan# 1638  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:216 Resp: 306759  
Ion Ratio Lower Upper  
216 100  
214 77.4 62.2 93.2  
179 17.5 17.0 25.6  
108 11.4 14.2 21.4#



#38  
Hexachlorocyclopentadiene  
Concen: 19.10 ng/ul  
RT: 12.610 min Scan# 1636  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:237 Resp: 202382  
Ion Ratio Lower Upper  
237 100  
235 63.7 49.4 74.2  
272 12.1 10.1 15.1









#39  
2,4,6-Trichlorophenol  
Concen: 21.25 ng/ul  
RT: 12.863 min Scan# 1679  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleID :  
SSTD02012


Tgt Ion:196 Resp: 182135  
Ion Ratio Lower Upper  
196 100  
198 101.5 79.4 119.2  
200 31.7 24.6 37.0



#40  
2,4,5-Trichlorophenol  
Concen: 21.81 ng/ul  
RT: 12.933 min Scan# 1691  
Delta R.T. 0.000 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:196 Resp: 198957  
Ion Ratio Lower Upper  
196 100  
198 95.8 78.0 117.0  
200 31.4 25.8 38.6









#41  
1,1'-Biphenyl  
Concen: 19.55 ng/ul  
RT: 13.269 min Scan# 1748  
Delta R.T. -0.012 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:154 Resp: 520231  
Ion Ratio Lower Upper  
154 100  
153 40.0 32.0 48.0  
76 9.0 9.8 14.8#



#42  
2-Chloronaphthalene  
Concen: 19.28 ng/ul  
RT: 13.310 min Scan# 1755  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:162 Resp: 428540  
Ion Ratio Lower Upper  
162 100  
164 33.7 27.1 40.7  
127 33.0 29.6 44.4









#43  
2-Nitroaniline  
Concen: 20.81 ng/ul  
RT: 13.510 min Scan# 1789  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion: 65 Resp: 92855  
Ion Ratio Lower Upper  
65 100  
92 70.6 49.9 74.9  
138 113.1 70.1 105.1#



#45  
Dimethylphthalate  
Concen: 20.59 ng/ul  
RT: 13.898 min Scan# 1855  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion: 163 Resp: 566961  
Ion Ratio Lower Upper  
163 100  
194 4.5 3.3 4.9  
164 10.0 8.1 12.1









#46  
2,6-Dinitrotoluene  
Concen: 21.96 ng/ul  
RT: 14.010 min Scan# 1874  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:165 Resp: 114914  
Ion Ratio Lower Upper  
165 100  
89 51.8 56.6 84.8#  
121 17.3 18.0 27.0#



#48  
Acenaphthylene  
Concen: 20.04 ng/ul  
RT: 14.157 min Scan# 1899  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:152 Resp: 615692  
Ion Ratio Lower Upper  
152 100  
151 19.6 14.9 22.3  
153 13.3 9.8 14.8









#49  
3-Nitroaniline  
Concen: 23.17 ng/ul  
RT: 14.333 min Scan# 1929  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:138 Resp: 96092  
Ion Ratio Lower Upper  
138 100  
108 10.6 8.9 13.3  
92 108.6 104.4 156.6



#50  
Acenaphthene  
Concen: 19.86 ng/ul  
RT: 14.504 min Scan# 1958  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:153 Resp: 433570  
Ion Ratio Lower Upper  
153 100  
152 45.6 38.1 57.1  
154 88.4 72.5 108.7









#51  
2,4-Dinitrophenol  
Concen: 19.24 ng/ul  
RT: 14.539 min Scan# 1964  
Delta R.T. 0.000 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleID :  
SSTD02012


Tgt Ion:184 Resp: 57832  
Ion Ratio Lower Upper  
184 100  
63 39.5 66.4 99.6#  
154 56.1 56.8 85.2#



#53  
4-Nitrophenol  
Concen: 20.43 ng/ul  
RT: 14.639 min Scan# 1981  
Delta R.T. 0.000 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:109 Resp: 85497  
Ion Ratio Lower Upper  
109 100  
139 98.5 80.8 121.2  
65 90.2 98.6 147.8#









#54  
Dibenzofuran  
Concen: 20.53 ng/ul  
RT: 14.833 min Scan# 2014  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:168 Resp: 674096  
Ion Ratio Lower Upper  
168 100  
139 34.0 30.3 45.5



#55  
2,4-Dinitrotoluene  
Concen: 22.76 ng/ul  
RT: 14.792 min Scan# 2007  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:165 Resp: 162639  
Ion Ratio Lower Upper  
165 100  
63 36.6 41.0 61.6#  
182 2.8 2.3 3.5









#56  
2,3,4,6-Tetrachlorophenol  
Concen: 22.96 ng/ul  
RT: 15.063 min Scan# 2053  
Delta R.T. 0.000 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:232 Resp: 182397  
Ion Ratio Lower Upper  
232 100  
131 29.5 34.0 51.0#  
130 1.8 2.1 3.1#  
166 23.0 25.9 38.9#



#57  
Diethylphthalate  
Concen: 21.55 ng/ul  
RT: 15.269 min Scan# 2088  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:149 Resp: 544662  
Ion Ratio Lower Upper  
149 100  
177 25.4 17.6 26.4  
150 12.7 9.6 14.4









#59  
Fluorene  
Concen: 20.13 ng/ul  
RT: 15.486 min Scan# 2125  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:166 Resp: 537137  
Ion Ratio Lower Upper  
166 100  
165 97.0 77.8 116.6  
167 13.3 10.7 16.1



#60  
4-Chlorophenyl-phenylether  
Concen: 20.39 ng/ul  
RT: 15.480 min Scan# 2124  
Delta R.T. -0.012 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:204 Resp: 342173  
Ion Ratio Lower Upper  
204 100  
206 33.5 27.0 40.6  
141 44.8 44.6 66.8









#61  
4-Nitroaniline  
Concen: 21.67 ng/ul  
RT: 15.498 min Scan# 2127  
Delta R.T. 0.000 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:138 Resp: 95225  
Ion Ratio Lower Upper  
138 100  
92 48.6 48.6 73.0#  
108 111.1 86.3 129.5



#64  
4,6-Dinitro-2-methylphenol  
Concen: 18.64 ng/ul  
RT: 15.557 min Scan# 2137  
Delta R.T. 0.000 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:198 Resp: 95894  
Ion Ratio Lower Upper  
198 100  
182 3.6 3.4 5.0  
77 19.2 25.1 37.7#









#65  
N-Nitrosodiphenylamine  
Concen: 20.10 ng/ul  
RT: 15.692 min Scan# 2160  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:169 Resp: 484648  
Ion Ratio Lower Upper  
169 100  
168 66.3 54.2 81.4  
167 36.7 29.0 43.4



#66  
4-Bromophenyl-phenylether  
Concen: 20.09 ng/ul  
RT: 16.374 min Scan# 2276  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:248 Resp: 232279  
Ion Ratio Lower Upper  
248 100  
250 99.1 80.7 121.1  
141 57.2 55.8 83.8









#67  
Hexachlorobenzene  
Concen: 19.43 ng/u1  
RT: 16.486 min Scan# 2295  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012

Tgt Ion:284 Resp: 253225




| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 284 | 100   |       |       |
| 142 | 26.4  | 27.4  | 41.2# |
| 249 | 33.8  | 25.6  | 38.4  |




#68  
Atrazine  
Concen: 20.37 ng/u1  
RT: 16.645 min Scan# 2322  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:200 Resp: 214512

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 200 | 100   |       |       |
| 173 | 21.7  | 20.1  | 30.1  |
| 215 | 49.8  | 39.0  | 58.4  |









#69  
 Pentachlorophenol  
 Concen: 19.11 ng/ul  
 RT: 16.827 min Scan# 2353  
 Delta R.T. -0.006 min  
 Lab File: BM027544.D  
 Acq: 26 Sep 2020 00:54




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTD02012


Tgt Ion:266 Resp: 136052  
 Ion Ratio Lower Upper  
 266 100  
 264 61.9 50.0 75.0  
 268 64.5 49.4 74.0



#70  
 Phenanthrene  
 Concen: 19.75 ng/ul  
 RT: 17.221 min Scan# 2420  
 Delta R.T. -0.006 min  
 Lab File: BM027544.D  
 Acq: 26 Sep 2020 00:54

Tgt Ion:178 Resp: 926203  
 Ion Ratio Lower Upper  
 178 100  
 179 15.4 12.2 18.4  
 176 19.3 15.1 22.7









#72  
 Anthracene  
 Concen: 19.70 ng/uL  
 RT: 17.310 min Scan# 2435  
 Delta R.T. -0.006 min  
 Lab File: BM027544.D  
 Acq: 26 Sep 2020 00:54




Instrument :  
 BNA\_M  
 ClientSampleID :  
 SSTD02012


Tgt Ion:178 Resp: 938945  
 Ion Ratio Lower Upper  
 178 100  
 179 15.8 12.5 18.7  
 176 18.0 14.6 22.0



#73  
 1,2,3,4-Tetrachlorobenzene  
 Concen: 18.90 ng/uL  
 RT: 13.233 min Scan# 1742  
 Delta R.T. -0.006 min  
 Lab File: BM027544.D  
 Acq: 26 Sep 2020 00:54

Tgt Ion:216 Resp: 302358  
 Ion Ratio Lower Upper  
 216 100  
 214 77.5 63.5 95.3  
 218 49.1 38.4 57.6









#74  
 Pentachlorobenzene  
 Concen: 18.84 ng/uL  
 RT: 14.757 min Scan# 2001  
 Delta R.T. -0.006 min  
 Lab File: BM027544.D  
 Acq: 26 Sep 2020 00:54




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTD02012


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 250   | Resp: | 293509 |
| Ion      | Ratio | Lower | Upper  |
| 250      | 100   |       |        |
| 248      | 63.6  | 49.0  | 73.6   |
| 252      | 65.0  | 49.3  | 73.9   |



#75  
 Carbazole  
 Concen: 20.55 ng/uL  
 RT: 17.580 min Scan# 2481  
 Delta R.T. 0.000 min  
 Lab File: BM027544.D  
 Acq: 26 Sep 2020 00:54

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 167   | Resp: | 777514 |
| Ion      | Ratio | Lower | Upper  |
| 167      | 100   |       |        |
| 166      | 20.9  | 16.8  | 25.2   |
| 139      | 11.1  | 10.1  | 15.1   |









#76  
Di-n-butylphthalate  
Concen: 22.07 ng/ul  
RT: 18.157 min Scan# 2579  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:149 Resp: 902767  
Ion Ratio Lower Upper  
149 100  
150 9.0 7.5 11.3  
104 5.4 5.1 7.7



#77  
Fluoranthene  
Concen: 19.88 ng/ul  
RT: 19.233 min Scan# 2762  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:202 Resp: 1189468  
Ion Ratio Lower Upper  
202 100  
101 4.7 6.2 9.2#  
100 3.7 5.0 7.6#









#80  
Pyrene  
Concen: 20.34 ng/ul  
RT: 19.598 min Scan# 2824  
Delta R.T. -0.006 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:202 Resp: 1207106  
Ion Ratio Lower Upper  
202 100  
101 5.1 7.1 10.7#  
100 4.3 6.1 9.1#



#81  
Butylbenzylphthalate  
Concen: 25.62 ng/ul  
RT: 20.503 min Scan# 2978  
Delta R.T. -0.012 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:149 Resp: 354973  
Ion Ratio Lower Upper  
149 100  
91 61.8 59.4 89.2  
206 28.9 18.1 27.1#









#82  
3,3'-Dichlorobenzidine  
Concen: 20.06 ng/ul  
RT: 21.274 min Scan# 3109  
Delta R.T. -0.012 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:252 Resp: 374463  
Ion Ratio Lower Upper  
252 100  
254 64.1 52.6 78.8  
126 5.8 6.9 10.3#



#83  
Benzo(a)anthracene  
Concen: 19.91 ng/ul  
RT: 21.339 min Scan# 3120  
Delta R.T. -0.012 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:228 Resp: 1154863  
Ion Ratio Lower Upper  
228 100  
229 19.6 15.8 23.6  
226 26.7 21.5 32.3









#84  
Bis(2-ethylhexyl)phthalate  
Concen: 24.22 ng/ul  
RT: 21.286 min Scan# 3111  
Delta R.T. -0.012 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:149 Resp: 515770  
Ion Ratio Lower Upper  
149 100  
167 30.7 21.7 32.5  
279 6.2 3.9 5.9#



#85  
Chrysene  
Concen: 19.44 ng/ul  
RT: 21.392 min Scan# 3129  
Delta R.T. -0.012 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:228 Resp: 1090478  
Ion Ratio Lower Upper  
228 100  
226 29.8 23.4 35.2  
229 19.5 15.8 23.8









#87  
Di-n-octyl phthalate  
Concen: 23.30 ng/ul  
RT: 22.174 min Scan# 3262  
Delta R.T. -0.018 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:149 Resp: 785223



#88  
Benzo(b)fluoranthene  
Concen: 19.53 ng/ul  
RT: 22.939 min Scan# 3392  
Delta R.T. -0.018 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:252 Resp: 996288  
Ion Ratio Lower Upper  
252 100  
253 21.9 17.7 26.5  
125 4.7 5.1 7.7#









#89  
Benzo(k)fluoranthene  
Concen: 19.51 ng/ul  
RT: 22.986 min Scan# 3400  
Delta R.T. -0.018 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54




Instrument :  
BNA\_M  
ClientSampleId :  
SSTD02012


Tgt Ion:252 Resp: 993721  
Ion Ratio Lower Upper  
252 100  
253 21.9 17.4 26.0  
125 4.7 5.6 8.4#



#91  
Benzo(a)pyrene  
Concen: 18.19 ng/ul  
RT: 23.521 min Scan# 3491  
Delta R.T. -0.018 min  
Lab File: BM027544.D  
Acq: 26 Sep 2020 00:54

Tgt Ion:252 Resp: 857685  
Ion Ratio Lower Upper  
252 100  
253 22.1 17.5 26.3  
125 5.1 6.6 10.0#









#92  
 Indeno(1,2,3-cd)pyrene  
 Concen: 17.08 ng/ul  
 RT: 25.915 min Scan# 3898  
 Delta R.T. -0.018 min  
 Lab File: BM027544.D  
 Acq: 26 Sep 2020 00:54




Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTD02012


Tgt Ion:276 Resp: 1011615  
 Ion Ratio Lower Upper  
 276 100  
 138 12.4 14.0 21.0#  
 277 24.2 20.1 30.1



#93  
 Dibenzo(a,h)anthracene  
 Concen: 17.13 ng/ul  
 RT: 25.927 min Scan# 3900  
 Delta R.T. -0.023 min  
 Lab File: BM027544.D  
 Acq: 26 Sep 2020 00:54

Tgt Ion:278 Resp: 853548  
 Ion Ratio Lower Upper  
 278 100  
 139 8.0 9.2 13.8#  
 279 23.8 19.0 28.6





#94

Benzo(g,h,i)perylene

Concen: 16.93 ng/ul

RT: 26.609 min Scan# 4016

Delta R.T. -0.023 min

Lab File: BM027544.D


Acq: 26 Sep 2020 00:54

Instrument :

BNA\_M

ClientSampleId :

SSTD02012



Tgt Ion:276 Resp: 829757

Ion Ratio Lower Upper

276 100

138 10.7 12.2 18.2#

277 23.4 18.6 28.0

