

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092520\
 Data File : BM027552.D
 Acq On : 26 Sep 2020 05:42
 Operator : JU/CG
 Sample : L4139-01
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BG499

Quant Time: Sep 26 08:44:57 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM092420MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 24 20:21:51 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	101183	20.00	ng/ul	# 0.00
18) Naphthalene-d8	10.592	136	377210	20.00	ng/ul	# 0.00
36) Acenaphthene-d10	14.433	164	308092	20.00	ng/ul	-0.01
62) Phenanthrene-d10	17.174	188	761616	20.00	ng/ul	#-0.01
78) Chrysene-d12	21.350	240	740792	20.00	ng/ul	#-0.01
86) Perylene-d12	23.621	264	654566	20.00	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.310	96	11156	6.27	ng/uL	0.00
5) Phenol-d5	6.969	99	52154	7.91	ng/ul	0.00
7) Bis-(2-Chloroethyl)eth...	7.140	67	122410	31.57	ng/ul	0.00
9) 2-Chlorophenol-d4	7.340	132	160019	28.03	ng/ul	0.00
13) 4-Methylphenol-d8	8.504	113	104825	18.95	ng/ul	0.00
19) Nitrobenzene-d5	8.951	128	97879	35.20	ng/ul	-0.01
22) 2-Nitrophenol-d4	9.675	143	115690	39.71	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.210	165	257708	34.67	ng/ul	0.00
29) 4-Chloroaniline-d4	10.722	131	21367	2.90	ng/ul	0.00
44) Dimethylphthalate-d6	13.845	166	995672	41.77	ng/ul	-0.01
47) Acenaphthylene-d8	14.127	160	1055131	36.96	ng/ul	0.00
52) 4-Nitrophenol-d4	14.621	143	28207	8.45	ng/ul	0.00
58) Fluorene-d10	15.427	176	841595	39.25	ng/ul	0.00
63) 4,6-Dinitro-2-methylph...	15.539	200	169731	41.16	ng/ul	0.00
71) Anthracene-d10	17.274	188	1559795	43.42	ng/ul	0.00
79) Pyrene-d10	19.568	212	1886574	48.82	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.474	264	1541113	44.32	ng/ul	-0.02
Target Compounds						
						Qvalue
45) Dimethylphthalate	13.892	163	88569	3.64	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092520\
Data File : BM027552.D
Acq On : 26 Sep 2020 05:42
Operator : JU/CG
Sample : L4139-01
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BG499

Quant Time: Sep 26 08:44:57 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM092420MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 24 20:21:51 2020
Response via : Initial Calibration





#45
 Dimethylphthalate
 Concen: 3.64 ng/ul
 RT: 13.892 min Scan# 1854
 Delta R.T. -0.012 min
 Lab File: BM027552.D
 Acq: 26 Sep 2020 05:42

Instrument :
 BNA_M
 ClientSampleId :
 BG499

Tgt Ion:163 Resp: 88569
 Ion Ratio Lower Upper
 163 100
 194 3.8 3.3 4.9
 164 10.1 8.1 12.1

