

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092521\
 Data File : BM032147.D
 Acq On : 25 Sep 2021 10:55
 Operator : CG/JU
 Sample : PB139359BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS359

Quant Time: Sep 26 00:19:25 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 25 01:15:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.666	152	358561	20.000	ng/u1	0.00
20) Naphthalene-d8	10.460	136	1495994	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.312	164	934061	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.042	188	1821050	20.000	ng/u1	0.00
79) Chrysene-d12	21.200	240	1460315	20.000	ng/u1	0.00
88) Perylene-d12	23.371	264	1273435	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.990	96	54070	6.367	ng/uL	0.00
4) Pyridine-d5	3.413	84	758619	31.469	ng/u1	0.00
7) Phenol-d5	6.848	99	970011	34.081	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.989	67	571233	33.379	ng/u1	0.00
11) 2-Chlorophenol-d4	7.201	132	762158	34.492	ng/u1	0.00
15) 4-Methylphenol-d8	8.395	113	771719	34.675	ng/u1	0.00
21) Nitrobenzene-d5	8.819	128	344189	38.520	ng/u1	0.00
24) 2-Nitrophenol-d4	9.548	143	369146	39.071	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.089	165	773731	35.823	ng/u1	0.00
31) 4-Chloroaniline-d4	10.589	131	989597	33.900	ng/u1	0.00
46) Dimethylphthalate-d6	13.718	166	2204077	36.318	ng/u1	0.00
49) Acenaphthylene-d8	14.007	160	2772356	35.956	ng/u1	0.00
54) 4-Nitrophenol-d4	14.512	143	383895	38.886	ng/u1	0.00
60) Fluorene-d10	15.301	176	1893636	36.568	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.418	200	321101	37.311	ng/u1	0.00
73) Anthracene-d10	17.136	188	2828059	37.219	ng/u1	0.00
81) Pyrene-d10	19.424	212	2982439	38.817	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.235	264	2177297	36.432	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.025	88	122787	13.160	ng/uL	96
5) Pyridine	3.431	79	790469	33.071	ng/u1	98
6) Benzaldehyde	6.789	77	577860	36.037	ng/u1	96
8) Phenol	6.878	94	1046888	36.282	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.084	93	828993	36.059	ng/u1	97
12) 2-Chlorophenol	7.236	128	832049	36.099	ng/u1	98
13) 2-Methylphenol	8.125	108	795905	36.446	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.201	45	1090922	34.478	ng/u1	98
16) Acetophenone	8.483	105	1251556	37.103	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.478	70	616843	35.853	ng/u1	97
18) 4-Methylphenol	8.454	108	860938	37.509	ng/u1	99
19) Hexachloroethane	8.760	117	333807	35.285	ng/u1	97
22) Nitrobenzene	8.866	77	887197	38.497	ng/u1	97
23) Isophorone	9.383	82	1708518	35.613	ng/u1	99
25) 2-Nitrophenol	9.578	139	438877	40.661	ng/u1	98
26) 2,4-Dimethylphenol	9.648	107	948842	36.913	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.872	93	1110833	36.221	ng/u1	99
29) 2,4-Dichlorophenol	10.119	162	808598	38.056	ng/u1	99
30) Naphthalene	10.513	128	2658490	36.202	ng/u1	99
32) 4-Chloroaniline	10.619	127	1012554	34.002	ng/u1	99
33) Hexachlorobutadiene	10.819	225	523044	35.808	ng/u1	97
34) Caprolactam	11.348	113	240404	37.093	ng/u1	91
35) 4-Chloro-3-methylphenol	11.754	107	862250	37.678	ng/u1	99
36) 2-Methylnaphthalene	12.124	142	1907132	38.688	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092521\
 Data File : BM032147.D
 Acq On : 25 Sep 2021 10:55
 Operator : CG/JU
 Sample : PB139359BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SLCS359

Quant Time: Sep 26 00:19:25 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 25 01:15:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.348	142	1819597	36.265	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.507	216	982138	36.836	ng/ul	98
40) Hexachlorocyclopentadiene	12.501	237	485892	34.005	ng/ul	97
41) 2,4,6-Trichlorophenol	12.748	196	640651	38.415	ng/ul	99
42) 2,4,5-Trichlorophenol	12.818	196	689998	38.975	ng/ul	98
43) 1,1'-Biphenyl	13.142	154	2405749	36.367	ng/ul	99
44) 2-Chloronaphthalene	13.183	162	1843413	36.937	ng/ul	99
45) 2-Nitroaniline	13.383	65	471330	41.782	ng/ul	96
47) Dimethylphthalate	13.765	163	2303286	38.042	ng/ul	100
48) 2,6-Dinitrotoluene	13.877	165	432842	43.056	ng/ul	98
50) Acenaphthylene	14.030	152	2795887	34.525	ng/ul	99
51) 3-Nitroaniline	14.207	138	418071	38.024	ng/ul	98
52) Acenaphthene	14.377	153	1978482	37.556	ng/ul	100
53) 2,4-Dinitrophenol	14.407	184	214119	38.961	ng/ul	93
55) 4-Nitrophenol	14.530	109	322207	40.013	ng/ul	94
56) Dibenzofuran	14.712	168	2806388	37.116	ng/ul	100
57) 2,4-Dinitrotoluene	14.665	165	616639	43.972	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.942	232	581805	40.283	ng/ul	100
59) Diethylphthalate	15.130	149	2302459	38.903	ng/ul	100
61) Fluorene	15.359	166	2189933	37.328	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.348	204	1102080	37.150	ng/ul	99
63) 4-Nitroaniline	15.371	138	408923	42.431	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.430	198	343685	40.456	ng/ul	98
67) N-Nitrosodiphenylamine	15.559	169	1954538	38.510	ng/ul	100
68) 4-Bromophenyl-phenylether	16.236	248	666333	38.320	ng/ul	98
69) Hexachlorobenzene	16.371	284	797565	37.729	ng/ul	99
70) Atrazine	16.501	200	681068	37.961	ng/ul	98
71) Pentachlorophenol	16.706	266	473004	38.954	ng/ul	99
72) Phenanthrene	17.083	178	3457666	38.032	ng/ul	100
74) Anthracene	17.171	178	3508179	38.500	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.118	216	1021482	36.390	ng/ul	99
76) Pentachlorobenzene	14.642	250	950978	34.986	ng/ul	98
77) Carbazole	17.436	167	3114836	40.413	ng/ul	100
78) Di-n-butylphthalate	17.995	149	3657252	41.301	ng/ul	100
80) Fluoranthene	19.089	202	3773194	39.573	ng/ul	99
82) Pyrene	19.453	202	3844923	39.323	ng/ul	100
83) Butylbenzylphthalate	20.336	149	1285756	45.730	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.118	252	952998	39.350	ng/ul	99
85) Benzo(a)anthracene	21.183	228	3211706	39.035	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.112	149	1825549	43.859	ng/ul	98
87) Chrysene	21.236	228	3139880	37.873	ng/ul	100
89) Di-n-octyl phthalate	21.959	149	2849641	44.040	ng/ul	100
90) Benzo(b)fluoranthene	22.724	252	2899235	38.549	ng/ul	99
91) Benzo(k)fluoranthene	22.765	252	2987834	41.919	ng/ul	100
93) Benzo(a)pyrene	23.277	252	2587887	36.441	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	25.524	276	2982792	41.210	ng/ul	98
95) Dibenzo(a,h)anthracene	25.529	278	2504622	41.723	ng/ul	99
96) Benzo(g,h,i)perylene	26.188	276	2547022	39.701	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Abundance Scan 32 (3.025 min): BM032145.D\data.ms (-27)




m/z-->

Abundance Scan 32 (3.025 min): BM032147.D\data.ms



m/z-->

Abundance Scan 32 (3.025 min): BM032147.D\data.ms (-1)



m/z-->

#2

1,4-Dioxane

Concen: 13.160 ng/uL

RT: 3.025 min Scan# 32

Delta R.T. -0.000 min


Lab File: BM032147.D

Acq: 25 Sep 2021 10:55

Tgt Ion: 88 Resp: 122787


Ion	Ratio	Lower	Upper
88	100		
43	27.7	22.3	33.5
58	73.8	62.4	93.6

Abundance




Time-->

Abundance Scan 101 (3.431 min): BM032145.D\data.ms (-96)




m/z-->

Abundance Scan 101 (3.431 min): BM032147.D\data.ms



m/z-->

Abundance Scan 101 (3.431 min): BM032147.D\data.ms (-85)



m/z-->

#5

Pyridine

Concen: 33.071 ng/uL

RT: 3.431 min Scan# 101

Delta R.T. -0.006 min


Lab File: BM032147.D

Acq: 25 Sep 2021 10:55

Tgt Ion: 79 Resp: 790469

Ion	Ratio	Lower	Upper
79	100		
52	74.6	58.3	87.5
51	34.7	29.1	43.7

Abundance




Time-->

Instrument :

BNA_M

ClientSampleId :





SLCS359



#6
Benzaldehyde
Concen: 36.037 ng/ul
RT: 6.789 min Scan# 672
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion: 77 Resp: 577860
Ion Ratio Lower Upper
77 100
105 101.0 77.0 115.4
106 95.7 74.2 111.2



#8
Phenol
Concen: 36.282 ng/ul
RT: 6.878 min Scan# 687
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion: 94 Resp: 1046888
Ion Ratio Lower Upper
94 100
65 27.8 22.4 33.6
66 36.8 29.9 44.9









#10
Bis(2-Chloroethyl)ether
Concen: 36.059 ng/ul
RT: 7.084 min Scan# 722
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Instrument :
BNA_M
ClientSampleId :
SLCS359

Tgt Ion: 93 Resp: 828993




Ion	Ratio	Lower	Upper
93	100		
63	75.7	63.4	95.2
95	32.4	25.5	38.3




#12
2-Chlorophenol
Concen: 36.099 ng/ul
RT: 7.236 min Scan# 748
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:128 Resp: 832049

Ion	Ratio	Lower	Upper
128	100		
64	45.6	38.3	57.5
130	32.2	25.8	38.8









#13
2-Methylphenol
Concen: 36.446 ng/ul
RT: 8.125 min Scan# 899
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:108 Resp: 795905
Ion Ratio Lower Upper
108 100
107 87.5 71.6 107.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 34.478 ng/ul
RT: 8.201 min Scan# 912
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion: 45 Resp: 1090922
Ion Ratio Lower Upper
45 100
77 15.1 11.2 16.8
79 11.5 9.1 13.7









#16
Acetophenone
Concen: 37.103 ng/ul
RT: 8.483 min Scan# 960
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:105 Resp: 1251556
Ion Ratio Lower Upper
105 100
77 76.6 63.0 94.4
51 25.8 22.2 33.2



#17
N-Nitroso-di-n-propylamine
Concen: 35.853 ng/ul
RT: 8.478 min Scan# 959
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion: 70 Resp: 616843
Ion Ratio Lower Upper
70 100
42 55.2 46.2 69.4
101 9.8 7.5 11.3
130 20.9 16.2 24.2









#18
4-Methylphenol
Concen: 37.509 ng/ul
RT: 8.454 min Scan# 955
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:108 Resp: 860938
Ion Ratio Lower Upper
108 100
107 112.2 89.1 133.7



#19
Hexachloroethane
Concen: 35.285 ng/ul
RT: 8.760 min Scan# 1007
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:117 Resp: 333807
Ion Ratio Lower Upper
117 100
201 102.5 79.5 119.3
199 63.0 49.1 73.7









#22
Nitrobenzene
Concen: 38.497 ng/ul
RT: 8.866 min Scan# 1025
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampled :
SLCS359


Tgt Ion: 77 Resp: 887197
Ion Ratio Lower Upper
77 100
123 47.6 36.1 54.1
65 13.7 11.0 16.6



#23
Isophorone
Concen: 35.613 ng/ul
RT: 9.383 min Scan# 1113
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion: 82 Resp: 1708518
Ion Ratio Lower Upper
82 100
95 7.3 5.8 8.6
138 18.0 14.1 21.1









#25
2-Nitrophenol
Concen: 40.661 ng/ul
RT: 9.578 min Scan# 1146
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampled :
SLCS359


Tgt Ion:139 Resp: 438877
Ion Ratio Lower Upper
139 100
65 50.7 42.0 63.0
109 31.3 25.8 38.8



#26
2,4-Dimethylphenol
Concen: 36.913 ng/ul
RT: 9.648 min Scan# 1158
Delta R.T. -0.006 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:107 Resp: 948842
Ion Ratio Lower Upper
107 100
121 50.1 40.4 60.6
122 86.1 67.6 101.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 36.221 ng/ul
 RT: 9.872 min Scan# 1196
 Delta R.T. -0.000 min
 Lab File: BM032147.D
 Acq: 25 Sep 2021 10:55

Instrument :
 BNA_M
 ClientSampleId :
 SLCS359

Tgt Ion: 93 Resp: 1110833




Ion	Ratio	Lower	Upper
93	100		
95	32.4	26.2	39.4
123	11.0	9.0	13.6




#29
 2,4-Dichlorophenol
 Concen: 38.056 ng/ul
 RT: 10.119 min Scan# 1238
 Delta R.T. -0.000 min
 Lab File: BM032147.D
 Acq: 25 Sep 2021 10:55

Tgt Ion: 162 Resp: 808598

Ion	Ratio	Lower	Upper
162	100		
164	65.0	51.6	77.4
98	35.4	28.6	42.8









#30
Naphthalene
Concen: 36.202 ng/u1
RT: 10.513 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:128 Resp: 2658490
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.2 10.2 15.4



#32
4-Chloroaniline
Concen: 34.002 ng/u1
RT: 10.619 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:127 Resp: 1012554
Ion Ratio Lower Upper
127 100
129 32.6 25.8 38.6









#33
Hexachlorobutadiene
Concen: 35.808 ng/ul
RT: 10.819 min Scan# 1357
Delta R.T. -0.006 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:225 Resp: 523044
Ion Ratio Lower Upper
225 100
223 60.9 50.6 76.0
227 62.7 52.2 78.4



#34
Caprolactam
Concen: 37.093 ng/ul
RT: 11.348 min Scan# 1447
Delta R.T. 0.006 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:113 Resp: 240404
Ion Ratio Lower Upper
113 100
55 159.2 138.1 207.1
56 122.2 106.0 159.0









#35
4-Chloro-3-methylphenol
Concen: 37.678 ng/ul
RT: 11.754 min Scan# 1516
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:107 Resp: 862250
Ion Ratio Lower Upper
107 100
144 25.6 20.3 30.5
142 77.9 62.8 94.2



#36
2-Methylnaphthalene
Concen: 38.688 ng/ul
RT: 12.124 min Scan# 1579
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:142 Resp: 1907132
Ion Ratio Lower Upper
142 100
141 89.4 70.2 105.2









#37
1-Methylnaphthalene
Concen: 36.265 ng/ul
RT: 12.348 min Scan# 1617
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:142 Resp: 1819597
Ion Ratio Lower Upper
142 100
141 91.1 74.3 111.5
116 3.9 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 36.836 ng/ul
RT: 12.507 min Scan# 1644
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:216 Resp: 982138
Ion Ratio Lower Upper
216 100
214 78.2 64.2 96.4
179 21.2 17.1 25.7
108 19.3 16.0 24.0









#40
Hexachlorocyclopentadiene
Concen: 34.005 ng/ul
RT: 12.501 min Scan# 1643
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:237 Resp: 485892
Ion Ratio Lower Upper
237 100
235 64.1 49.4 74.0
272 12.7 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 38.415 ng/ul
RT: 12.748 min Scan# 1685
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:196 Resp: 640651
Ion Ratio Lower Upper
196 100
198 95.9 77.8 116.6
200 31.3 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 38.975 ng/ul
RT: 12.818 min Scan# 1697
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampled :
SLCS359


Tgt Ion:196 Resp: 689998
Ion Ratio Lower Upper
196 100
198 98.4 77.4 116.2
200 31.7 24.3 36.5



#43
1,1'-Biphenyl
Concen: 36.367 ng/ul
RT: 13.142 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:154 Resp: 2405749
Ion Ratio Lower Upper
154 100
153 40.6 31.9 47.9
76 12.8 10.7 16.1









#44
2-Chloronaphthalene
Concen: 36.937 ng/ul
RT: 13.183 min Scan# 1759
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Instrument :
BNA_M
ClientSampleId :
SLCS359

Tgt Ion:162 Resp: 1843413




Ion	Ratio	Lower	Upper
162	100		
164	32.8	25.9	38.9
127	39.6	32.0	48.0




#45
2-Nitroaniline
Concen: 41.782 ng/ul
RT: 13.383 min Scan# 1793
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion: 65 Resp: 471330

Ion	Ratio	Lower	Upper
65	100		
92	71.8	54.5	81.7
138	108.7	84.1	126.1









#47
Dimethylphthalate
Concen: 38.042 ng/ul
RT: 13.765 min Scan# 1858
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:163 Resp: 2303286
Ion Ratio Lower Upper
163 100
194 4.2 3.3 4.9
164 10.1 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 43.056 ng/ul
RT: 13.877 min Scan# 1877
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:165 Resp: 432842
Ion Ratio Lower Upper
165 100
89 56.5 46.8 70.2
121 20.7 17.3 25.9









#50
Acenaphthylene
Concen: 34.525 ng/ul
RT: 14.030 min Scan# 1903
Delta R.T. -0.006 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Instrument :
BNA_M
ClientSampleId :
SLCS359

Tgt Ion:152 Resp: 2795887




Ion	Ratio	Lower	Upper
152	100		
151	20.0	15.7	23.5
153	13.5	10.6	15.8




#51
3-Nitroaniline
Concen: 38.024 ng/ul
RT: 14.207 min Scan# 1933
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:138 Resp: 418071

Ion	Ratio	Lower	Upper
138	100		
108	11.9	10.5	15.7
92	115.5	94.0	141.0









#52
Acenaphthene
Concen: 37.556 ng/ul
RT: 14.377 min Scan# 1962
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:153 Resp: 1978482
Ion Ratio Lower Upper
153 100
152 47.6 38.4 57.6
154 88.6 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 38.961 ng/ul
RT: 14.407 min Scan# 1967
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:184 Resp: 214119
Ion Ratio Lower Upper
184 100
63 65.5 56.7 85.1
154 66.5 57.5 86.3









#55
4-Nitrophenol
Concen: 40.013 ng/ul
RT: 14.530 min Scan# 1988
Delta R.T. 0.006 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:109 Resp: 322207
Ion Ratio Lower Upper
109 100
139 136.4 103.7 155.5
65 129.5 109.4 164.2



#56
Dibenzofuran
Concen: 37.116 ng/ul
RT: 14.712 min Scan# 2019
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:168 Resp: 2806388
Ion Ratio Lower Upper
168 100
139 37.5 29.8 44.6









#57
2,4-Dinitrotoluene
Concen: 43.972 ng/ul
RT: 14.665 min Scan# 2011
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampled :
SLCS359


Tgt Ion:165 Resp: 616639
Ion Ratio Lower Upper
165 100
63 42.1 34.2 51.2
182 3.1 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 40.283 ng/ul
RT: 14.942 min Scan# 2058
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:232 Resp: 581805
Ion Ratio Lower Upper
232 100
131 45.4 36.2 54.2
130 2.4 1.9 2.9
166 31.0 24.8 37.2









#59
Diethylphthalate
Concen: 38.903 ng/ul
RT: 15.130 min Scan# 2090
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampled :
SLCS359


Tgt Ion:149 Resp: 2302459
Ion Ratio Lower Upper
149 100
177 22.5 17.9 26.9
150 12.3 10.0 15.0



#61
Fluorene
Concen: 37.328 ng/ul
RT: 15.359 min Scan# 2129
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:166 Resp: 2189933
Ion Ratio Lower Upper
166 100
165 97.1 77.7 116.5
167 13.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 37.150 ng/ul
RT: 15.348 min Scan# 2127
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:204 Resp: 1102080
Ion Ratio Lower Upper
204 100
206 32.7 25.8 38.6
141 60.5 47.9 71.9



#63
4-Nitroaniline
Concen: 42.431 ng/ul
RT: 15.371 min Scan# 2131
Delta R.T. 0.006 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:138 Resp: 408923
Ion Ratio Lower Upper
138 100
92 49.1 41.4 62.2
108 76.9 63.8 95.6









#66
4,6-Dinitro-2-methylphenol
Concen: 40.456 ng/ul
RT: 15.430 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:198 Resp: 343685
Ion Ratio Lower Upper
198 100
182 4.3 3.8 5.8
77 25.6 19.8 29.8



#67
N-Nitrosodiphenylamine
Concen: 38.510 ng/ul
RT: 15.559 min Scan# 2163
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:169 Resp: 1954538
Ion Ratio Lower Upper
169 100
168 65.6 52.3 78.5
167 36.0 28.6 42.8









#68
4-Bromophenyl-phenylether
Concen: 38.320 ng/ul
RT: 16.236 min Scan# 2278
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:248 Resp: 666333
Ion Ratio Lower Upper
248 100
250 97.2 77.0 115.6
141 74.1 57.6 86.4



#69
Hexachlorobenzene
Concen: 37.729 ng/ul
RT: 16.371 min Scan# 2301
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:284 Resp: 797565
Ion Ratio Lower Upper
284 100
142 34.6 28.4 42.6
249 31.2 24.7 37.1









#70
Atrazine
Concen: 37.961 ng/ul
RT: 16.501 min Scan# 2323
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:200 Resp: 681068
Ion Ratio Lower Upper
200 100
173 26.4 21.4 32.0
215 47.7 39.4 59.2



#71
Pentachlorophenol
Concen: 38.954 ng/ul
RT: 16.706 min Scan# 2358
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:266 Resp: 473004
Ion Ratio Lower Upper
266 100
264 62.6 50.2 75.2
268 64.6 51.0 76.6









#72
Phenanthrene
Concen: 38.032 ng/u1
RT: 17.083 min Scan# 2422
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:178 Resp: 3457666
Ion Ratio Lower Upper
178 100
179 15.3 12.3 18.5
176 19.3 15.4 23.0



#74
Anthracene
Concen: 38.500 ng/u1
RT: 17.171 min Scan# 2437
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:178 Resp: 3508179
Ion Ratio Lower Upper
178 100
179 15.7 12.6 19.0
176 19.0 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 36.390 ng/uL
RT: 13.118 min Scan# 1748
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:216 Resp: 1021482
Ion Ratio Lower Upper
216 100
214 77.4 62.7 94.1
218 47.9 38.7 58.1



#76
Pentachlorobenzene
Concen: 34.986 ng/uL
RT: 14.642 min Scan# 2007
Delta R.T. -0.006 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:250 Resp: 950978
Ion Ratio Lower Upper
250 100
248 64.0 50.0 75.0
252 63.5 51.6 77.4









#77
 Carbazole
 Concen: 40.413 ng/ul
 RT: 17.436 min Scan# 2482
 Delta R.T. -0.000 min
 Lab File: BM032147.D
 Acq: 25 Sep 2021 10:55




Instrument :
 BNA_M
 ClientSampled :
 SLCS359


Tgt Ion:167 Resp: 3114836
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.3 25.9
 139 12.9 10.3 15.5



#78
 Di-n-butylphthalate
 Concen: 41.301 ng/ul
 RT: 17.995 min Scan# 2577
 Delta R.T. -0.000 min
 Lab File: BM032147.D
 Acq: 25 Sep 2021 10:55

Tgt Ion:149 Resp: 3657252
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.0
 104 5.2 4.2 6.2









#80
Fluoranthene
Concen: 39.573 ng/u1
RT: 19.089 min Scan# 2763
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:202 Resp: 3773194
Ion Ratio Lower Upper
202 100
101 11.8 9.2 13.8
100 8.8 6.9 10.3



#82
Pyrene
Concen: 39.323 ng/u1
RT: 19.453 min Scan# 2825
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:202 Resp: 3844923
Ion Ratio Lower Upper
202 100
101 12.7 10.1 15.1
100 10.4 8.4 12.6









#83
Butylbenzylphthalate
Concen: 45.730 ng/ul
RT: 20.336 min Scan# 2975
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:149 Resp: 1285756
Ion Ratio Lower Upper
149 100
91 73.0 61.8 92.8
206 20.9 16.2 24.4



#84
3,3'-Dichlorobenzidine
Concen: 39.350 ng/ul
RT: 21.118 min Scan# 3108
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:252 Resp: 952998
Ion Ratio Lower Upper
252 100
254 64.2 52.2 78.2
126 11.4 8.8 13.2









#85
Benzo(a)anthracene
Concen: 39.035 ng/ul
RT: 21.183 min Scan# 3119
Delta R.T. -0.006 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:228 Resp: 3211706
Ion Ratio Lower Upper
228 100
229 20.3 15.9 23.9
226 27.5 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 43.859 ng/ul
RT: 21.112 min Scan# 3107
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:149 Resp: 1825549
Ion Ratio Lower Upper
149 100
167 28.0 21.6 32.4
279 4.5 3.3 4.9









#87
Chrysene
Concen: 37.873 ng/ul
RT: 21.236 min Scan# 3128
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:228 Resp: 3139880
Ion Ratio Lower Upper
228 100
226 30.7 24.4 36.6
229 20.0 15.9 23.9



#89
Di-n-octyl phthalate
Concen: 44.040 ng/ul
RT: 21.959 min Scan# 3251
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:149 Resp: 2849641









#90
Benzo(b)fluoranthene
Concen: 38.549 ng/ul
RT: 22.724 min Scan# 3381
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:252 Resp: 2899235
Ion Ratio Lower Upper
252 100
253 22.2 18.2 27.4
125 10.3 8.5 12.7



#91
Benzo(k)fluoranthene
Concen: 41.919 ng/ul
RT: 22.765 min Scan# 3388
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:252 Resp: 2987834
Ion Ratio Lower Upper
252 100
253 22.5 17.9 26.9
125 10.6 8.4 12.6









#93
Benzo(a)pyrene
Concen: 36.441 ng/ul
RT: 23.277 min Scan# 3475
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55




Instrument :
BNA_M
ClientSampleId :
SLCS359


Tgt Ion:252 Resp: 2587887
Ion Ratio Lower Upper
252 100
253 22.4 17.8 26.6
125 11.5 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 41.210 ng/ul
RT: 25.524 min Scan# 3857
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:276 Resp: 2982792
Ion Ratio Lower Upper
276 100
138 25.8 21.5 32.3
277 25.5 21.0 31.6









#95
Dibenzo(a,h)anthracene
Concen: 41.723 ng/ul
RT: 25.529 min Scan# 3858
Delta R.T. -0.000 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Instrument :
BNA_M
ClientSampleId :
SLCS359

Tgt Ion:278 Resp: 2504622
Ion Ratio Lower Upper
278 100
139 15.9 13.5 20.3
279 23.9 19.4 29.0



#96
Benzo(g,h,i)perylene
Concen: 39.701 ng/ul
RT: 26.188 min Scan# 3970
Delta R.T. 0.006 min
Lab File: BM032147.D
Acq: 25 Sep 2021 10:55

Tgt Ion:276 Resp: 2547022
Ion Ratio Lower Upper
276 100
138 21.7 17.9 26.9
277 23.9 18.6 28.0

