

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092521\
 Data File : BM032184.D
 Acq On : 26 Sep 2021 10:28
 Operator : CG/JU
 Sample : PB139298BS
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS298

Quant Time: Sep 27 01:08:35 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 25 01:15:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.660	152	350711	20.000	ng/u1	0.00
20) Naphthalene-d8	10.454	136	1357979	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.307	164	770437	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.036	188	1442790	20.000	ng/u1	0.00
79) Chrysene-d12	21.195	240	1321320	20.000	ng/u1	0.00
88) Perylene-d12	23.365	264	1326567	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.990	96	64002	7.706	ng/uL	0.00
4) Pyridine-d5	3.414	84	855792	36.294	ng/u1	0.00
7) Phenol-d5	6.849	99	1044212	37.510	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.984	67	608613	36.359	ng/u1	0.00
11) 2-Chlorophenol-d4	7.202	132	849618	39.311	ng/u1	0.00
15) 4-Methylphenol-d8	8.390	113	796907	36.609	ng/u1	0.00
21) Nitrobenzene-d5	8.819	128	383170	47.241	ng/u1	0.00
24) 2-Nitrophenol-d4	9.543	143	409307	47.725	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.084	165	787779	40.181	ng/u1	0.00
31) 4-Chloroaniline-d4	10.590	131	969727	36.595	ng/u1	0.00
46) Dimethylphthalate-d6	13.713	166	1979129	39.538	ng/u1	0.00
49) Acenaphthylene-d8	14.001	160	2604288	40.949	ng/u1	0.00
54) 4-Nitrophenol-d4	14.507	143	353891	43.460	ng/u1	0.00
60) Fluorene-d10	15.295	176	1718523	40.234	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.413	200	292573	42.909	ng/u1	0.00
73) Anthracene-d10	17.130	188	2484006	41.262	ng/u1	0.00
81) Pyrene-d10	19.418	212	2673462	38.456	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.230	264	2526933	40.589	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.025	88	126389	13.850	ng/uL	94
5) Pyridine	3.431	79	812798	34.766	ng/u1	98
6) Benzaldehyde	6.784	77	566706	36.132	ng/u1	96
8) Phenol	6.872	94	1024637	36.306	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.078	93	820725	36.498	ng/u1	96
12) 2-Chlorophenol	7.231	128	851542	37.771	ng/u1	96
13) 2-Methylphenol	8.125	108	763592	35.749	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.196	45	1007287	32.547	ng/u1	97
16) Acetophenone	8.478	105	1179154	35.739	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.472	70	561530	33.369	ng/u1	95
18) 4-Methylphenol	8.454	108	810454	36.100	ng/u1	98
19) Hexachloroethane	8.754	117	348287	37.639	ng/u1	98
22) Nitrobenzene	8.860	77	857978	41.013	ng/u1	96
23) Isophorone	9.378	82	1496793	34.371	ng/u1	99
25) 2-Nitrophenol	9.572	139	429572	43.844	ng/u1	96
26) 2,4-Dimethylphenol	9.648	107	869576	37.267	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.866	93	1033782	37.134	ng/u1	100
29) 2,4-Dichlorophenol	10.113	162	751272	38.951	ng/u1	99
30) Naphthalene	10.507	128	2521794	37.831	ng/u1	100
32) 4-Chloroaniline	10.613	127	914689	33.837	ng/u1	99
33) Hexachlorobutadiene	10.813	225	497223	37.500	ng/u1	99
34) Caprolactam	11.348	113	197522	33.574	ng/u1	87
35) 4-Chloro-3-methylphenol	11.754	107	746551	35.938	ng/u1	99
36) 2-Methylnaphthalene	12.119	142	1738174	38.844	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092521\
 Data File : BM032184.D
 Acq On : 26 Sep 2021 10:28
 Operator : CG/JU
 Sample : PB139298BS
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS298

Quant Time: Sep 27 01:08:35 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 25 01:15:33 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.342	142	1659577	36.437	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.501	216	879418	39.988	ng/ul	98
40) Hexachlorocyclopentadiene	12.495	237	457078	38.783	ng/ul	99
41) 2,4,6-Trichlorophenol	12.742	196	553450	40.234	ng/ul	100
42) 2,4,5-Trichlorophenol	12.813	196	590110	40.412	ng/ul	99
43) 1,1'-Biphenyl	13.136	154	2137212	39.169	ng/ul	99
44) 2-Chloronaphthalene	13.178	162	1631856	39.642	ng/ul	99
45) 2-Nitroaniline	13.378	65	405990	43.633	ng/ul	95
47) Dimethylphthalate	13.760	163	1888691	37.819	ng/ul	100
48) 2,6-Dinitrotoluene	13.872	165	362051	43.662	ng/ul	98
50) Acenaphthylene	14.031	152	2394742	35.852	ng/ul	99
51) 3-Nitroaniline	14.201	138	363707	40.105	ng/ul	95
52) Acenaphthene	14.372	153	1704712	39.232	ng/ul	99
53) 2,4-Dinitrophenol	14.401	184	179924	39.692	ng/ul	97
55) 4-Nitrophenol	14.525	109	268154	40.373	ng/ul	93
56) Dibenzofuran	14.707	168	2368529	37.977	ng/ul	100
57) 2,4-Dinitrotoluene	14.660	165	506699	43.806	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	14.936	232	471922	39.615	ng/ul	99
59) Diethylphthalate	15.125	149	1826188	37.409	ng/ul	99
61) Fluorene	15.354	166	1822506	37.662	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.342	204	920419	37.616	ng/ul	99
63) 4-Nitroaniline	15.366	138	340731	42.864	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.425	198	283046	42.053	ng/ul	99
67) N-Nitrosodiphenylamine	15.554	169	1592644	39.607	ng/ul	99
68) 4-Bromophenyl-phenylether	16.230	248	540709	39.248	ng/ul	97
69) Hexachlorobenzene	16.366	284	628105	37.502	ng/ul	99
70) Atrazine	16.495	200	528479	37.178	ng/ul	98
71) Pentachlorophenol	16.701	266	367118	38.160	ng/ul	99
72) Phenanthrene	17.077	178	2792671	38.771	ng/ul	99
74) Anthracene	17.166	178	2842861	39.378	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.113	216	898925	40.419	ng/uL	100
76) Pentachlorobenzene	14.642	250	794763	36.904	ng/uL	98
77) Carbazole	17.430	167	2509435	41.094	ng/ul	99
78) Di-n-butylphthalate	17.989	149	2855939	40.708	ng/ul	100
80) Fluoranthene	19.083	202	3080961	35.712	ng/ul	99
82) Pyrene	19.448	202	3165290	35.778	ng/ul	100
83) Butylbenzylphthalate	20.330	149	1144974	45.007	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.112	252	916687	41.832	ng/ul	99
85) Benzo(a)anthracene	21.177	228	2882737	38.723	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.107	149	1668206	44.295	ng/ul	99
87) Chrysene	21.230	228	2854490	38.052	ng/ul	100
89) Di-n-octyl phthalate	21.954	149	2876944	42.682	ng/ul	100
90) Benzo(b)fluoranthene	22.718	252	2982149	38.064	ng/ul	99
91) Benzo(k)fluoranthene	22.759	252	3019406	40.665	ng/ul	99
93) Benzo(a)pyrene	23.271	252	2736476	36.990	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	25.518	276	3395325	45.030	ng/ul	99
95) Dibenzo(a,h)anthracene	25.518	278	2858760	45.714	ng/ul	99
96) Benzo(g,h,i)perylene	26.177	276	2918579	43.670	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092521\
Data File : BM032184.D
Acq On : 26 Sep 2021 10:28
Operator : CG/JU
Sample : PB139298BS
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampled :
SLCS298

Quant Time: Sep 27 01:08:35 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 25 01:15:33 2021
Response via : Initial Calibration









1,4-Dioxane
Concen: 13.850 ng/uL
RT: 3.025 min Scan# 32
Delta R.T. 0.000 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion: 88 Resp: 126389
Ion Ratio Lower Upper
88 100
43 26.0 22.3 33.5
58 72.6 62.4 93.6



Pyridine
Concen: 34.766 ng/uL
RT: 3.431 min Scan# 101
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion: 79 Resp: 812798
Ion Ratio Lower Upper
79 100
52 72.1 58.3 87.5
51 34.0 29.1 43.7









#6
Benzaldehyde
Concen: 36.132 ng/ul
RT: 6.784 min Scan# 671
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion: 77 Resp: 566706
Ion Ratio Lower Upper
77 100
105 100.5 77.0 115.4
106 97.1 74.2 111.2



#8
Phenol
Concen: 36.306 ng/ul
RT: 6.872 min Scan# 686
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion: 94 Resp: 1024637
Ion Ratio Lower Upper
94 100
65 27.0 22.4 33.6
66 36.2 29.9 44.9









#10

Bis(2-Chloroethyl)ether
Concen: 36.498 ng/ul
RT: 7.078 min Scan# 721
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Instrument :
BNA_M
ClientSampleId :
SLCS298

Tgt Ion: 93 Resp: 820725

Ion	Ratio	Lower	Upper
93	100		
63	75.0	63.4	95.2
95	32.8	25.5	38.3







#12

2-Chlorophenol
Concen: 37.771 ng/ul
RT: 7.231 min Scan# 747
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:128 Resp: 851542

Ion	Ratio	Lower	Upper
128	100		
64	44.0	38.3	57.5
130	33.0	25.8	38.8









#13
2-Methylphenol
Concen: 35.749 ng/ul
RT: 8.125 min Scan# 899
Delta R.T. 0.000 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:108 Resp: 763592
Ion Ratio Lower Upper
108 100
107 87.7 71.6 107.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 32.547 ng/ul
RT: 8.196 min Scan# 911
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion: 45 Resp: 1007287
Ion Ratio Lower Upper
45 100
77 15.4 11.2 16.8
79 12.3 9.1 13.7









#16
Acetophenone
Concen: 35.739 ng/ul
RT: 8.478 min Scan# 959
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:105 Resp: 1179154
Ion Ratio Lower Upper
105 100
77 75.3 63.0 94.4
51 25.2 22.2 33.2



#17
N-Nitroso-di-n-propylamine
Concen: 33.369 ng/ul
RT: 8.472 min Scan# 958
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion: 70 Resp: 561530
Ion Ratio Lower Upper
70 100
42 53.3 46.2 69.4
101 9.6 7.5 11.3
130 21.8 16.2 24.2









#18
4-Methylphenol
Concen: 36.100 ng/ul
RT: 8.454 min Scan# 955
Delta R.T. 0.000 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:108 Resp: 810454
Ion Ratio Lower Upper
108 100
107 109.7 89.1 133.7



#19
Hexachloroethane
Concen: 37.639 ng/ul
RT: 8.754 min Scan# 1006
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:117 Resp: 348287
Ion Ratio Lower Upper
117 100
201 101.3 79.5 119.3
199 63.4 49.1 73.7









#22
Nitrobenzene
Concen: 41.013 ng/ul
RT: 8.860 min Scan# 1024
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion: 77 Resp: 857978
Ion Ratio Lower Upper
77 100
123 48.0 36.1 54.1
65 13.4 11.0 16.6



#23
Isophorone
Concen: 34.371 ng/ul
RT: 9.378 min Scan# 1112
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion: 82 Resp: 1496793
Ion Ratio Lower Upper
82 100
95 7.2 5.8 8.6
138 18.0 14.1 21.1









#25
2-Nitrophenol
Concen: 43.844 ng/ul
RT: 9.572 min Scan# 1145
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Instrument :
BNA_M
ClientSampleId :
SLCS298

Tgt Ion:139 Resp: 429572




Ion	Ratio	Lower	Upper
139	100		
65	49.4	42.0	63.0
109	30.2	25.8	38.8




#26
2,4-Dimethylphenol
Concen: 37.267 ng/ul
RT: 9.648 min Scan# 1158
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:107 Resp: 869576

Ion	Ratio	Lower	Upper
107	100		
121	50.5	40.4	60.6
122	85.9	67.6	101.4









#27

Bis(2-Chloroethoxy)methane
Concen: 37.134 ng/ul
RT: 9.866 min Scan# 1195
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Instrument :
BNA_M
ClientSampleId :
SLCS298




Tgt Ion: 93 Resp: 1033782
Ion Ratio Lower Upper
93 100
95 32.7 26.2 39.4
123 11.0 9.0 13.6




#29

2,4-Dichlorophenol
Concen: 38.951 ng/ul
RT: 10.113 min Scan# 1237
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:162 Resp: 751272
Ion Ratio Lower Upper
162 100
164 65.4 51.6 77.4
98 35.2 28.6 42.8









#30
Naphthalene
Concen: 37.831 ng/ul
RT: 10.507 min Scan# 1304
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:128 Resp: 2521794
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 33.837 ng/ul
RT: 10.613 min Scan# 1322
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:127 Resp: 914689
Ion Ratio Lower Upper
127 100
129 32.7 25.8 38.6









#33
Hexachlorobutadiene
Concen: 37.500 ng/ul
RT: 10.813 min Scan# 1356
Delta R.T. -0.012 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:225 Resp: 497223
Ion Ratio Lower Upper
225 100
223 64.2 50.6 76.0
227 64.7 52.2 78.4



#34
Caprolactam
Concen: 33.574 ng/ul
RT: 11.348 min Scan# 1447
Delta R.T. 0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:113 Resp: 197522
Ion Ratio Lower Upper
113 100
55 152.2 138.1 207.1
56 118.5 106.0 159.0









#35
4-Chloro-3-methylphenol
Concen: 35.938 ng/ul
RT: 11.754 min Scan# 1516
Delta R.T. 0.000 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampled :
SLCS298


Tgt Ion:107 Resp: 746551
Ion Ratio Lower Upper
107 100
144 25.7 20.3 30.5
142 79.5 62.8 94.2



#36
2-Methylnaphthalene
Concen: 38.844 ng/ul
RT: 12.119 min Scan# 1578
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:142 Resp: 1738174
Ion Ratio Lower Upper
142 100
141 88.6 70.2 105.2









#37
1-Methylnaphthalene
Concen: 36.437 ng/ul
RT: 12.342 min Scan# 1616
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:142 Resp: 1659577
Ion Ratio Lower Upper
142 100
141 91.6 74.3 111.5
116 3.7 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 39.988 ng/ul
RT: 12.501 min Scan# 1643
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:216 Resp: 879418
Ion Ratio Lower Upper
216 100
214 78.4 64.2 96.4
179 20.6 17.1 25.7
108 18.8 16.0 24.0









#40
Hexachlorocyclopentadiene
Concen: 38.783 ng/ul
RT: 12.495 min Scan# 1642
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:237 Resp: 457078
Ion Ratio Lower Upper
237 100
235 62.9 49.4 74.0
272 12.7 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 40.234 ng/ul
RT: 12.742 min Scan# 1684
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:196 Resp: 553450
Ion Ratio Lower Upper
196 100
198 96.9 77.8 116.6
200 30.8 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 40.412 ng/ul
RT: 12.813 min Scan# 1696
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:196 Resp: 590110
Ion Ratio Lower Upper
196 100
198 95.6 77.4 116.2
200 31.2 24.3 36.5



#43
1,1'-Biphenyl
Concen: 39.169 ng/ul
RT: 13.136 min Scan# 1751
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:154 Resp: 2137212
Ion Ratio Lower Upper
154 100
153 39.9 31.9 47.9
76 12.6 10.7 16.1









#44
2-Chloronaphthalene
Concen: 39.642 ng/ul
RT: 13.178 min Scan# 1758
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampled :
SLCS298


Tgt Ion:162 Resp: 1631856
Ion Ratio Lower Upper
162 100
164 33.1 25.9 38.9
127 39.9 32.0 48.0



#45
2-Nitroaniline
Concen: 43.633 ng/ul
RT: 13.378 min Scan# 1792
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion: 65 Resp: 405990
Ion Ratio Lower Upper
65 100
92 71.5 54.5 81.7
138 110.8 84.1 126.1









#47
Dimethylphthalate
Concen: 37.819 ng/ul
RT: 13.760 min Scan# 1857
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:163 Resp: 1888691
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 9.9 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 43.662 ng/ul
RT: 13.872 min Scan# 1876
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:165 Resp: 362051
Ion Ratio Lower Upper
165 100
89 57.0 46.8 70.2
121 20.8 17.3 25.9









#50
Acenaphthylene
Concen: 35.852 ng/ul
RT: 14.031 min Scan# 1903
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:152 Resp: 2394742
Ion Ratio Lower Upper
152 100
151 19.8 15.7 23.5
153 12.9 10.6 15.8



#51
3-Nitroaniline
Concen: 40.105 ng/ul
RT: 14.201 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:138 Resp: 363707
Ion Ratio Lower Upper
138 100
108 11.4 10.5 15.7
92 112.5 94.0 141.0









#52
Acenaphthene
Concen: 39.232 ng/ul
RT: 14.372 min Scan# 1961
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:153 Resp: 1704712
Ion Ratio Lower Upper
153 100
152 47.4 38.4 57.6
154 89.4 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 39.692 ng/ul
RT: 14.401 min Scan# 1966
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:184 Resp: 179924
Ion Ratio Lower Upper
184 100
63 67.2 56.7 85.1
154 70.9 57.5 86.3









#55
4-Nitrophenol
Concen: 40.373 ng/ul
RT: 14.525 min Scan# 1987
Delta R.T. 0.000 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampled :
SLCS298


Tgt Ion:109 Resp: 268154
Ion Ratio Lower Upper
109 100
139 134.7 103.7 155.5
65 125.8 109.4 164.2



#56
Dibenzofuran
Concen: 37.977 ng/ul
RT: 14.707 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:168 Resp: 2368529
Ion Ratio Lower Upper
168 100
139 37.4 29.8 44.6









#57
2,4-Dinitrotoluene
Concen: 43.806 ng/ul
RT: 14.660 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampled :
SLCS298


Tgt Ion:165 Resp: 506699
Ion Ratio Lower Upper
165 100
63 41.6 34.2 51.2
182 3.0 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 39.615 ng/ul
RT: 14.936 min Scan# 2057
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:232 Resp: 471922
Ion Ratio Lower Upper
232 100
131 46.3 36.2 54.2
130 2.5 1.9 2.9
166 31.2 24.8 37.2









#59
Diethylphthalate
Concen: 37.409 ng/ul
RT: 15.125 min Scan# 2089
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:149 Resp: 1826188
Ion Ratio Lower Upper
149 100
177 22.8 17.9 26.9
150 12.6 10.0 15.0



#61
Fluorene
Concen: 37.662 ng/ul
RT: 15.354 min Scan# 2128
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:166 Resp: 1822506
Ion Ratio Lower Upper
166 100
165 96.9 77.7 116.5
167 13.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 37.616 ng/ul
RT: 15.342 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:204 Resp: 920419
Ion Ratio Lower Upper
204 100
206 33.0 25.8 38.6
141 60.5 47.9 71.9



#63
4-Nitroaniline
Concen: 42.864 ng/ul
RT: 15.366 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:138 Resp: 340731
Ion Ratio Lower Upper
138 100
92 49.5 41.4 62.2
108 77.9 63.8 95.6









#66
4,6-Dinitro-2-methylphenol
Concen: 42.053 ng/ul
RT: 15.425 min Scan# 2140
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampled :
SLCS298


Tgt Ion:198 Resp: 283046
Ion Ratio Lower Upper
198 100
182 4.9 3.8 5.8
77 24.0 19.8 29.8



#67
N-Nitrosodiphenylamine
Concen: 39.607 ng/ul
RT: 15.554 min Scan# 2162
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:169 Resp: 1592644
Ion Ratio Lower Upper
169 100
168 66.1 52.3 78.5
167 36.7 28.6 42.8









#68
4-Bromophenyl-phenylether
Concen: 39.248 ng/ul
RT: 16.230 min Scan# 2277
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampled :
SLCS298


Tgt Ion:248	Resp:	540709
Ion	Ratio	Lower Upper
248	100	
250	97.9	77.0 115.6
141	75.4	57.6 86.4



#69
Hexachlorobenzene
Concen: 37.502 ng/ul
RT: 16.366 min Scan# 2300
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:284	Resp:	628105
Ion	Ratio	Lower Upper
284	100	
142	35.3	28.4 42.6
249	31.8	24.7 37.1









#70
Atrazine
Concen: 37.178 ng/ul
RT: 16.495 min Scan# 2322
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampled :
SLCS298


Tgt Ion:200 Resp: 528479
Ion Ratio Lower Upper
200 100
173 27.2 21.4 32.0
215 47.5 39.4 59.2



#71
Pentachlorophenol
Concen: 38.160 ng/ul
RT: 16.701 min Scan# 2357
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:266 Resp: 367118
Ion Ratio Lower Upper
266 100
264 61.2 50.2 75.2
268 63.6 51.0 76.6









#72
Phenanthrene
Concen: 38.771 ng/u1
RT: 17.077 min Scan# 2421
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:178 Resp: 2792671
Ion Ratio Lower Upper
178 100
179 15.2 12.3 18.5
176 19.7 15.4 23.0



#74
Anthracene
Concen: 39.378 ng/u1
RT: 17.166 min Scan# 2436
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:178 Resp: 2842861
Ion Ratio Lower Upper
178 100
179 15.5 12.6 19.0
176 18.7 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 40.419 ng/uL
RT: 13.113 min Scan# 1747
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:216 Resp: 898925
Ion Ratio Lower Upper
216 100
214 78.1 62.7 94.1
218 48.1 38.7 58.1



#76
Pentachlorobenzene
Concen: 36.904 ng/uL
RT: 14.642 min Scan# 2007
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:250 Resp: 794763
Ion Ratio Lower Upper
250 100
248 60.6 50.0 75.0
252 62.7 51.6 77.4









#77
 Carbazole
 Concen: 41.094 ng/ul
 RT: 17.430 min Scan# 2481
 Delta R.T. -0.006 min
 Lab File: BM032184.D
 Acq: 26 Sep 2021 10:28

Instrument :
 BNA_M
 ClientSampleId :
 SLCS298

Tgt Ion:167 Resp: 2509435




Ion	Ratio	Lower	Upper
167	100		
166	20.9	17.3	25.9
139	13.0	10.3	15.5




#78
 Di-n-butylphthalate
 Concen: 40.708 ng/ul
 RT: 17.989 min Scan# 2576
 Delta R.T. -0.006 min
 Lab File: BM032184.D
 Acq: 26 Sep 2021 10:28

Tgt Ion:149 Resp: 2855939

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.4	11.0
104	5.2	4.2	6.2









#80
Fluoranthene
Concen: 35.712 ng/ul
RT: 19.083 min Scan# 2762
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:202 Resp: 3080961
Ion Ratio Lower Upper
202 100
101 11.8 9.2 13.8
100 8.8 6.9 10.3



#82
Pyrene
Concen: 35.778 ng/ul
RT: 19.448 min Scan# 2824
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:202 Resp: 3165290
Ion Ratio Lower Upper
202 100
101 12.7 10.1 15.1
100 10.3 8.4 12.6









#83
Butylbenzylphthalate
Concen: 45.007 ng/ul
RT: 20.330 min Scan# 2974
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:149 Resp: 1144974
Ion Ratio Lower Upper
149 100
91 73.3 61.8 92.8
206 20.5 16.2 24.4



#84
3,3'-Dichlorobenzidine
Concen: 41.832 ng/ul
RT: 21.112 min Scan# 3107
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:252 Resp: 916687
Ion Ratio Lower Upper
252 100
254 64.5 52.2 78.2
126 11.7 8.8 13.2









#85
Benzo(a)anthracene
Concen: 38.723 ng/ul
RT: 21.177 min Scan# 3118
Delta R.T. -0.012 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampled :
SLCS298


Tgt Ion:228 Resp: 2882737
Ion Ratio Lower Upper
228 100
229 20.4 15.9 23.9
226 27.6 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 44.295 ng/ul
RT: 21.107 min Scan# 3106
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:149 Resp: 1668206
Ion Ratio Lower Upper
149 100
167 27.7 21.6 32.4
279 4.3 3.3 4.9









#87
Chrysene
Concen: 38.052 ng/ul
RT: 21.230 min Scan# 3127
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:228 Resp: 2854490
Ion Ratio Lower Upper
228 100
226 30.5 24.4 36.6
229 20.0 15.9 23.9



#89
Di-n-octyl phthalate
Concen: 42.682 ng/ul
RT: 21.954 min Scan# 3250
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:149 Resp: 2876944









#90
Benzo(b)fluoranthene
Concen: 38.064 ng/ul
RT: 22.718 min Scan# 3380
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampled :
SLCS298


Tgt Ion:252 Resp: 2982149
Ion Ratio Lower Upper
252 100
253 22.3 18.2 27.4
125 10.4 8.5 12.7



#91
Benzo(k)fluoranthene
Concen: 40.665 ng/ul
RT: 22.759 min Scan# 3387
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:252 Resp: 3019406
Ion Ratio Lower Upper
252 100
253 22.1 17.9 26.9
125 10.2 8.4 12.6









#93
Benzo(a)pyrene
Concen: 36.990 ng/ul
RT: 23.271 min Scan# 3474
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28




Instrument :
BNA_M
ClientSampleId :
SLCS298


Tgt Ion:252 Resp: 2736476
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.6
125 11.3 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 45.030 ng/ul
RT: 25.518 min Scan# 3856
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:276 Resp: 3395325
Ion Ratio Lower Upper
276 100
138 27.0 21.5 32.3
277 25.7 21.0 31.6









#95
Dibenzo(a,h)anthracene
Concen: 45.714 ng/ul
RT: 25.518 min Scan# 3856
Delta R.T. -0.012 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Instrument :
BNA_M
ClientSampleId :
SLCS298

Tgt Ion:278 Resp: 2858760
Ion Ratio Lower Upper
278 100
139 17.4 13.5 20.3
279 23.9 19.4 29.0



#96
Benzo(g,h,i)perylene
Concen: 43.670 ng/ul
RT: 26.177 min Scan# 3968
Delta R.T. -0.006 min
Lab File: BM032184.D
Acq: 26 Sep 2021 10:28

Tgt Ion:276 Resp: 2918579
Ion Ratio Lower Upper
276 100
138 23.1 17.9 26.9
277 23.7 18.6 28.0

