

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092521\
 Data File : BM032193.D
 Acq On : 26 Sep 2021 15:51
 Operator : CG/JU
 Sample : M3873-03
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EW7X3

Quant Time: Sep 27 01:10:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 25 01:15:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.660	152	415174	20.000	ng/u1	0.00
20) Naphthalene-d8	10.454	136	1619389	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.307	164	955658	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.036	188	1845665	20.000	ng/u1	0.00
79) Chrysene-d12	21.195	240	1710184	20.000	ng/u1	0.00
88) Perylene-d12	23.365	264	1787826	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.984	96	49981	5.083	ng/uL	0.00
4) Pyridine-d5	3.419	84	218517	7.828	ng/u1	0.00
7) Phenol-d5	6.843	99	264555	8.028	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.984	67	507298	25.601	ng/u1	0.00
11) 2-Chlorophenol-d4	7.201	132	597878	23.368	ng/u1	0.00
15) 4-Methylphenol-d8	8.384	113	436952	16.956	ng/u1	-0.01
21) Nitrobenzene-d5	8.813	128	315577	32.627	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.542	143	329010	32.169	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.084	165	616247	26.358	ng/u1	0.00
31) 4-Chloroaniline-d4	10.584	131	706962	22.372	ng/u1	-0.01
46) Dimethylphthalate-d6	13.707	166	2012059	32.405	ng/u1	-0.01
49) Acenaphthylene-d8	13.995	160	2330191	29.538	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.501	143	98160	9.718	ng/u1	-0.01
60) Fluorene-d10	15.295	176	1701239	32.110	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.407	200	288912	33.123	ng/u1	-0.01
73) Anthracene-d10	17.130	188	2840984	36.891	ng/u1	0.00
81) Pyrene-d10	19.418	212	3180468	35.346	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.230	264	3031736	36.134	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092521\
Data File : BM032193.D
Acq On : 26 Sep 2021 15:51
Operator : CG/JU
Sample : M3873-03
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EW7X3

Quant Time: Sep 27 01:10:45 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM092421.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 25 01:15:33 2021
Response via : Initial Calibration

