


Data Path : Z:\HPCHEM1\BNA M\DATA\BM092616\  
 Data File : BM007564.D  
 Acq On : 26 Sep 2016 09:16  
 Operator : UM/SJ  
 Sample : DFTPP65  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP65

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM02.2-EPA-BM092316.M  
 Title : SVOA CALIBRATION  
 Last Update : Tue Sep 27 03:37:16 2016



AutoFind: Scans 2550, 2551, 2552; Background Corrected with Scan 2544

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 28.9      | 96941   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.8       | 2032    | PASS             |
| 69          | 198          | 0.00         | 100          | 33.7      | 113165  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.5       | 550     | PASS             |
| 127         | 198          | 10           | 80           | 41.3      | 138581  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.5       | 1702    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 335765  | PASS             |
| 199         | 198          | 5            | 9            | 6.9       | 23184   | PASS             |
| 275         | 198          | 10           | 60           | 29.8      | 100112  | PASS             |
| 365         | 198          | 1            | 100          | 4.3       | 14307   | PASS             |
| 441         | 443          | 0.01         | 100          | 78.7      | 42970   | PASS             |
| 442         | 198          | 50           | 100          | 84.6      | 284053  | PASS             |
| 443         | 442          | 15           | 24           | 19.2      | 54632   | PASS             |

Scan 2544 (17.551 min): BM007564.D  
DFTPP65

| m/z    | Abundance |
|--------|-----------|
| 39.90  | 564.0     |
| 43.90  | 910.0     |
| 73.10  | 646.0     |
| 147.10 | 364.0     |
| 190.90 | 320.0     |
| 207.10 | 2365.0    |
| 207.90 | 443.0     |
| 208.90 | 395.0     |
| 281.00 | 731.0     |
| 282.00 | 376.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP65

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 498.0     |
| 38.00  | 1578.0    |
| 39.10  | 8098.0    |
| 40.00  | 817.0     |
| 43.90  | 745.0     |
| 49.00  | 1083.0    |
| 50.10  | 30408.0   |
| 51.10  | 111840.0  |
| 52.10  | 5883.0    |
| 55.20  | 537.0     |
| 56.10  | 3522.0    |
| 57.00  | 7812.0    |
| 61.00  | 1929.0    |
| 62.00  | 1888.0    |
| 63.00  | 4851.0    |
| 64.10  | 843.0     |
| 65.10  | 2215.0    |
| 66.00  | 353.0     |
| 68.00  | 2082.0    |
| 69.00  | 131712.0  |
| 70.00  | 509.0     |
| 73.00  | 2051.0    |
| 74.10  | 15313.0   |
| 75.00  | 21848.0   |
| 76.00  | 7329.0    |
| 77.10  | 147392.0  |
| 78.10  | 10069.0   |
| 79.00  | 11883.0   |
| 80.00  | 8008.0    |
| 81.00  | 11492.0   |
| 82.00  | 2745.0    |
| 83.00  | 2376.0    |
| 85.00  | 2087.0    |
| 86.00  | 4080.0    |
| 87.10  | 1791.0    |
| 87.90  | 596.0     |
| 91.00  | 2682.0    |
| 92.00  | 3251.0    |
| 93.00  | 20376.0   |
| 94.00  | 1389.0    |
| 95.10  | 345.0     |
| 96.10  | 1042.0    |
| 98.00  | 15982.0   |
| 99.00  | 13625.0   |
| 100.00 | 1122.0    |
| 101.00 | 5741.0    |
| 103.00 | 2082.0    |
| 104.00 | 3930.0    |
| 105.00 | 4477.0    |
| 106.10 | 1376.0    |
| 107.00 | 50312.0   |
| 108.00 | 7732.0    |
| 109.00 | 1285.0    |
| 110.00 | 85072.0   |
| 111.00 | 14185.0   |
| 112.00 | 1753.0    |
| 112.90 | 671.0     |
| 114.90 | 378.0     |
| 116.00 | 2738.0    |
| 117.00 | 48616.0   |
| 118.00 | 3199.0    |
| 119.00 | 489.0     |
| 120.00 | 501.0     |
| 120.90 | 350.0     |
| 122.00 | 3385.0    |
| 123.00 | 4933.0    |
| 124.10 | 2401.0    |
| 125.00 | 1964.0    |
| 126.10 | 708.0     |
| 127.00 | 152704.0  |
| 128.00 | 12071.0   |
| 129.00 | 64800.0   |
| 130.00 | 5675.0    |
| 131.10 | 1231.0    |
| 132.00 | 592.0     |
| 132.80 | 436.0     |
| 134.00 | 2528.0    |
| 135.00 | 6211.0    |
| 136.00 | 2250.0    |
| 137.10 | 3021.0    |
| 138.00 | 716.0     |
| 139.00 | 399.0     |
| 140.00 | 1157.0    |
| 141.00 | 8843.0    |
| 142.00 | 3316.0    |
| 143.00 | 2479.0    |
| 143.90 | 459.0     |
| 145.00 | 385.0     |
| 146.00 | 1981.0    |
| 147.00 | 5669.0    |
| 148.00 | 10429.0   |
| 149.00 | 2949.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP65

|        |          |
|--------|----------|
| 150.00 | 562.0    |
| 151.10 | 1301.0   |
| 151.70 | 699.0    |
| 153.00 | 3323.0   |
| 154.00 | 2012.0   |
| 155.00 | 4883.0   |
| 156.00 | 8296.0   |
| 157.00 | 1524.0   |
| 158.00 | 1789.0   |
| 159.00 | 1686.0   |
| 160.00 | 2697.0   |
| 161.10 | 4279.0   |
| 162.00 | 1222.0   |
| 162.90 | 384.0    |
| 163.80 | 638.0    |
| 165.00 | 3807.0   |
| 166.00 | 2577.0   |
| 167.00 | 20432.0  |
| 168.00 | 10346.0  |
| 169.10 | 2000.0   |
| 170.20 | 475.0    |
| 170.90 | 872.0    |
| 172.00 | 2106.0   |
| 173.00 | 2424.0   |
| 174.00 | 3664.0   |
| 175.00 | 7641.0   |
| 176.10 | 1989.0   |
| 177.00 | 3758.0   |
| 177.90 | 1350.0   |
| 179.00 | 14456.0  |
| 180.00 | 10962.0  |
| 181.00 | 4231.0   |
| 182.00 | 468.0    |
| 184.00 | 1470.0   |
| 185.00 | 7886.0   |
| 186.00 | 51440.0  |
| 187.00 | 16616.0  |
| 188.10 | 2375.0   |
| 189.00 | 3198.0   |
| 190.00 | 609.0    |
| 191.00 | 1660.0   |
| 192.00 | 4788.0   |
| 193.00 | 6199.0   |
| 193.90 | 677.0    |
| 195.00 | 660.0    |
| 196.00 | 8944.0   |
| 197.00 | 1707.0   |
| 198.00 | 343552.0 |
| 199.00 | 24568.0  |
| 200.00 | 1580.0   |
| 201.60 | 1511.0   |
| 203.00 | 2712.0   |
| 204.00 | 14386.0  |
| 205.00 | 23608.0  |
| 206.10 | 97240.0  |
| 207.10 | 14497.0  |
| 208.00 | 3997.0   |
| 209.10 | 1432.0   |
| 209.90 | 1419.0   |
| 211.10 | 4571.0   |
| 211.80 | 924.0    |
| 215.00 | 1172.0   |
| 216.00 | 1672.0   |
| 217.00 | 27216.0  |
| 218.00 | 3636.0   |
| 218.90 | 377.0    |
| 219.90 | 384.0    |
| 221.00 | 12731.0  |
| 221.90 | 3015.0   |
| 223.00 | 7004.0   |
| 224.00 | 57856.0  |
| 225.00 | 14855.0  |
| 226.00 | 2157.0   |
| 227.00 | 27008.0  |
| 228.00 | 4163.0   |
| 229.00 | 6121.0   |
| 229.90 | 661.0    |
| 231.00 | 1588.0   |
| 232.90 | 460.0    |
| 233.90 | 1579.0   |
| 235.00 | 1742.0   |
| 235.90 | 889.0    |
| 237.00 | 3390.0   |
| 237.90 | 637.0    |
| 239.10 | 849.0    |
| 240.10 | 1075.0   |
| 241.00 | 1141.0   |
| 242.00 | 2894.0   |
| 243.10 | 2560.0   |
| 244.10 | 41872.0  |
| 245.10 | 5019.0   |
| 246.00 | 10207.0  |
| 247.00 | 2410.0   |
| 247.90 | 339.0    |
| 249.00 | 1859.0   |
| 250.00 | 404.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP65

|        |          |
|--------|----------|
| 251.90 | 491.0    |
| 252.90 | 921.0    |
| 254.20 | 1233.0   |
| 255.00 | 232640.0 |
| 256.00 | 35032.0  |
| 257.00 | 3423.0   |
| 258.00 | 15282.0  |
| 259.00 | 2188.0   |
| 260.00 | 548.0    |
| 264.10 | 761.0    |
| 265.00 | 5748.0   |
| 266.00 | 1208.0   |
| 271.00 | 486.0    |
| 272.00 | 379.0    |
| 273.00 | 6909.0   |
| 274.00 | 17352.0  |
| 275.00 | 94728.0  |
| 276.00 | 12963.0  |
| 277.00 | 9752.0   |
| 278.00 | 2518.0   |
| 279.00 | 488.0    |
| 281.00 | 404.0    |
| 282.10 | 360.0    |
| 283.00 | 823.0    |
| 284.10 | 965.0    |
| 285.00 | 1793.0   |
| 286.10 | 419.0    |
| 288.90 | 473.0    |
| 289.90 | 306.0    |
| 292.20 | 355.0    |
| 293.00 | 1967.0   |
| 293.90 | 577.0    |
| 295.10 | 811.0    |
| 296.00 | 29576.0  |
| 297.00 | 3947.0   |
| 298.00 | 511.0    |
| 300.80 | 327.0    |
| 302.00 | 376.0    |
| 303.00 | 3482.0   |
| 304.00 | 1009.0   |
| 308.00 | 391.0    |
| 308.90 | 398.0    |
| 310.10 | 534.0    |
| 314.10 | 1385.0   |
| 315.00 | 2951.0   |
| 316.00 | 2145.0   |
| 317.10 | 349.0    |
| 320.90 | 735.0    |
| 322.00 | 507.0    |
| 323.10 | 8384.0   |
| 324.00 | 1617.0   |
| 327.00 | 2061.0   |
| 328.00 | 961.0    |
| 331.90 | 632.0    |
| 333.00 | 723.0    |
| 334.00 | 6198.0   |
| 335.10 | 1345.0   |
| 341.10 | 1216.0   |
| 342.00 | 450.0    |
| 346.00 | 1321.0   |
| 346.90 | 359.0    |
| 352.00 | 2468.0   |
| 353.00 | 1657.0   |
| 354.00 | 2820.0   |
| 355.00 | 890.0    |
| 365.00 | 11884.0  |
| 366.00 | 1831.0   |
| 371.00 | 428.0    |
| 372.00 | 4467.0   |
| 373.10 | 959.0    |
| 383.10 | 1088.0   |
| 383.90 | 339.0    |
| 389.90 | 345.0    |
| 391.00 | 416.0    |
| 402.10 | 1596.0   |
| 403.00 | 2166.0   |
| 404.10 | 915.0    |
| 405.00 | 381.0    |
| 421.00 | 1864.0   |
| 422.00 | 1745.0   |
| 423.10 | 11703.0  |
| 424.10 | 4620.0   |
| 425.00 | 661.0    |
| 441.00 | 35016.0  |
| 442.10 | 223872.0 |
| 443.10 | 43688.0  |
| 444.00 | 4020.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP65

| m/z    | Abundance |
|--------|-----------|
| 36.90  | 487.0     |
| 38.00  | 1250.0    |
| 39.10  | 7515.0    |
| 40.10  | 1044.0    |
| 44.00  | 930.0     |
| 49.10  | 1290.0    |
| 50.10  | 28200.0   |
| 51.10  | 113032.0  |
| 52.10  | 5962.0    |
| 53.00  | 356.0     |
| 55.00  | 646.0     |
| 56.00  | 2974.0    |
| 57.00  | 8192.0    |
| 58.00  | 508.0     |
| 61.00  | 1855.0    |
| 62.10  | 1467.0    |
| 63.00  | 4971.0    |
| 64.00  | 929.0     |
| 65.00  | 2336.0    |
| 68.10  | 2595.0    |
| 69.00  | 130608.0  |
| 69.90  | 736.0     |
| 73.10  | 1951.0    |
| 74.10  | 14731.0   |
| 75.00  | 22640.0   |
| 76.10  | 7006.0    |
| 77.10  | 147776.0  |
| 78.10  | 10747.0   |
| 79.00  | 11244.0   |
| 80.00  | 8144.0    |
| 81.00  | 11441.0   |
| 82.00  | 2311.0    |
| 83.00  | 2223.0    |
| 85.00  | 2273.0    |
| 86.00  | 4146.0    |
| 87.00  | 1807.0    |
| 88.00  | 663.0     |
| 89.10  | 315.0     |
| 91.10  | 2265.0    |
| 92.00  | 3201.0    |
| 93.00  | 20128.0   |
| 94.00  | 1108.0    |
| 94.90  | 361.0     |
| 96.00  | 1132.0    |
| 97.00  | 405.0     |
| 98.00  | 17072.0   |
| 99.00  | 12530.0   |
| 99.90  | 925.0     |
| 101.00 | 6071.0    |
| 101.90 | 310.0     |
| 103.00 | 1878.0    |
| 104.00 | 4155.0    |
| 105.00 | 3898.0    |
| 106.00 | 1154.0    |
| 107.00 | 54016.0   |
| 108.00 | 7469.0    |
| 109.10 | 1288.0    |
| 110.00 | 86872.0   |
| 111.00 | 13719.0   |
| 112.10 | 1963.0    |
| 113.10 | 582.0     |
| 116.00 | 2807.0    |
| 117.00 | 49096.0   |
| 118.00 | 3191.0    |
| 118.90 | 451.0     |
| 120.00 | 542.0     |
| 122.00 | 3195.0    |
| 123.00 | 5880.0    |
| 124.00 | 2652.0    |
| 125.00 | 1757.0    |
| 126.20 | 779.0     |
| 127.00 | 164224.0  |
| 128.10 | 12237.0   |
| 129.00 | 68696.0   |
| 130.00 | 5654.0    |
| 131.00 | 1227.0    |
| 131.80 | 341.0     |
| 133.10 | 424.0     |
| 134.00 | 2396.0    |
| 135.10 | 5728.0    |
| 135.90 | 2278.0    |
| 137.00 | 3401.0    |
| 138.00 | 703.0     |
| 139.00 | 420.0     |
| 140.00 | 897.0     |
| 141.00 | 10081.0   |
| 142.00 | 3179.0    |
| 142.90 | 2290.0    |
| 143.90 | 602.0     |
| 145.00 | 518.0     |
| 146.10 | 1891.0    |
| 147.00 | 6452.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP65

|        |          |
|--------|----------|
| 148.00 | 12608.0  |
| 149.00 | 3008.0   |
| 149.90 | 651.0    |
| 151.00 | 1367.0   |
| 151.90 | 862.0    |
| 152.90 | 2776.0   |
| 154.00 | 2288.0   |
| 155.10 | 5602.0   |
| 156.10 | 7703.0   |
| 156.90 | 1334.0   |
| 158.00 | 2059.0   |
| 159.00 | 1548.0   |
| 160.10 | 3769.0   |
| 161.00 | 4652.0   |
| 161.90 | 1163.0   |
| 163.00 | 343.0    |
| 163.90 | 494.0    |
| 165.00 | 3994.0   |
| 166.00 | 3314.0   |
| 167.00 | 20576.0  |
| 168.00 | 12710.0  |
| 169.00 | 1504.0   |
| 169.90 | 538.0    |
| 170.80 | 724.0    |
| 172.00 | 1920.0   |
| 173.00 | 2520.0   |
| 174.00 | 5043.0   |
| 175.10 | 8052.0   |
| 176.00 | 1624.0   |
| 177.00 | 4527.0   |
| 178.00 | 1541.0   |
| 179.00 | 15409.0  |
| 180.00 | 10938.0  |
| 181.00 | 4185.0   |
| 182.00 | 813.0    |
| 182.90 | 937.0    |
| 184.00 | 1287.0   |
| 185.00 | 7381.0   |
| 186.00 | 61368.0  |
| 187.00 | 17392.0  |
| 188.10 | 1752.0   |
| 189.00 | 3570.0   |
| 189.90 | 714.0    |
| 191.00 | 1863.0   |
| 192.00 | 5651.0   |
| 193.10 | 5676.0   |
| 194.00 | 1207.0   |
| 194.80 | 302.0    |
| 196.00 | 10579.0  |
| 197.10 | 2181.0   |
| 198.00 | 396480.0 |
| 199.00 | 27440.0  |
| 200.00 | 2226.0   |
| 201.50 | 1751.0   |
| 203.00 | 3248.0   |
| 204.00 | 16528.0  |
| 205.00 | 27664.0  |
| 206.00 | 114864.0 |
| 207.00 | 15082.0  |
| 208.00 | 4643.0   |
| 209.00 | 2046.0   |
| 210.10 | 1820.0   |
| 211.00 | 4773.0   |
| 215.00 | 1528.0   |
| 216.00 | 2374.0   |
| 217.00 | 33744.0  |
| 218.00 | 3971.0   |
| 218.90 | 536.0    |
| 221.00 | 14089.0  |
| 221.90 | 4561.0   |
| 223.00 | 7049.0   |
| 224.00 | 65880.0  |
| 225.00 | 15993.0  |
| 226.10 | 2274.0   |
| 227.00 | 30800.0  |
| 228.00 | 4251.0   |
| 229.00 | 6046.0   |
| 230.00 | 745.0    |
| 231.00 | 2685.0   |
| 232.00 | 313.0    |
| 232.90 | 554.0    |
| 233.90 | 1973.0   |
| 235.10 | 2087.0   |
| 236.00 | 1649.0   |
| 237.00 | 3601.0   |
| 239.00 | 1159.0   |
| 239.90 | 1008.0   |
| 240.90 | 1624.0   |
| 242.00 | 4357.0   |
| 243.10 | 3016.0   |
| 244.10 | 49432.0  |
| 245.00 | 7373.0   |
| 246.00 | 12170.0  |
| 247.00 | 2547.0   |
| 248.00 | 526.0    |
| 249.00 | 2250.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP65

|        |          |
|--------|----------|
| 249.90 | 442.0    |
| 251.10 | 554.0    |
| 252.10 | 532.0    |
| 253.00 | 1250.0   |
| 254.00 | 2016.0   |
| 255.00 | 282240.0 |
| 256.00 | 39472.0  |
| 257.00 | 3793.0   |
| 258.00 | 18128.0  |
| 259.00 | 2439.0   |
| 260.00 | 576.0    |
| 263.90 | 497.0    |
| 265.00 | 8322.0   |
| 266.00 | 855.0    |
| 269.90 | 422.0    |
| 271.00 | 557.0    |
| 271.90 | 787.0    |
| 273.00 | 8407.0   |
| 274.00 | 22384.0  |
| 275.00 | 121184.0 |
| 276.00 | 16260.0  |
| 277.00 | 11939.0  |
| 278.00 | 2576.0   |
| 279.10 | 405.0    |
| 281.10 | 685.0    |
| 282.00 | 502.0    |
| 283.00 | 1739.0   |
| 284.00 | 1041.0   |
| 285.00 | 1869.0   |
| 286.00 | 335.0    |
| 288.90 | 498.0    |
| 289.90 | 417.0    |
| 292.00 | 377.0    |
| 293.00 | 2713.0   |
| 294.00 | 682.0    |
| 295.00 | 917.0    |
| 296.00 | 39136.0  |
| 297.00 | 5568.0   |
| 298.00 | 377.0    |
| 301.00 | 462.0    |
| 302.00 | 707.0    |
| 303.10 | 4377.0   |
| 304.00 | 1323.0   |
| 307.90 | 498.0    |
| 308.80 | 314.0    |
| 309.90 | 587.0    |
| 312.90 | 322.0    |
| 314.10 | 2303.0   |
| 315.00 | 4282.0   |
| 316.00 | 2291.0   |
| 317.00 | 359.0    |
| 321.00 | 1427.0   |
| 322.00 | 633.0    |
| 323.00 | 11198.0  |
| 324.00 | 2183.0   |
| 327.00 | 2442.0   |
| 328.00 | 1280.0   |
| 329.00 | 389.0    |
| 332.00 | 1078.0   |
| 333.00 | 1149.0   |
| 334.00 | 7745.0   |
| 335.00 | 1593.0   |
| 341.10 | 1401.0   |
| 341.90 | 416.0    |
| 346.00 | 2734.0   |
| 346.90 | 532.0    |
| 352.10 | 3481.0   |
| 353.10 | 2398.0   |
| 354.10 | 3625.0   |
| 355.10 | 1097.0   |
| 365.00 | 16512.0  |
| 366.00 | 2446.0   |
| 370.00 | 345.0    |
| 371.00 | 919.0    |
| 372.10 | 6450.0   |
| 373.10 | 1122.0   |
| 383.00 | 1513.0   |
| 384.10 | 614.0    |
| 390.00 | 769.0    |
| 391.10 | 528.0    |
| 392.00 | 370.0    |
| 401.00 | 500.0    |
| 401.90 | 2113.0   |
| 403.00 | 3168.0   |
| 404.00 | 1328.0   |
| 405.20 | 316.0    |
| 421.00 | 2795.0   |
| 422.00 | 2433.0   |
| 423.00 | 17568.0  |
| 424.10 | 5155.0   |
| 425.00 | 691.0    |
| 439.80 | 365.0    |
| 441.00 | 51936.0  |
| 442.10 | 340800.0 |
| 443.10 | 65744.0  |
| 444.10 | 6954.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP65

445.10

330.0

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP65

| m/z    | Abundance |
|--------|-----------|
| 38.10  | 819.0     |
| 39.10  | 3846.0    |
| 40.10  | 605.0     |
| 43.90  | 1017.0    |
| 49.10  | 560.0     |
| 50.00  | 17288.0   |
| 51.10  | 65952.0   |
| 52.00  | 3891.0    |
| 55.00  | 467.0     |
| 56.00  | 1960.0    |
| 57.00  | 5310.0    |
| 61.00  | 999.0     |
| 62.10  | 1223.0    |
| 63.00  | 3146.0    |
| 64.10  | 367.0     |
| 65.00  | 1742.0    |
| 68.00  | 1420.0    |
| 69.00  | 77176.0   |
| 70.10  | 405.0     |
| 73.10  | 1020.0    |
| 74.10  | 8751.0    |
| 75.00  | 12516.0   |
| 76.10  | 4121.0    |
| 77.10  | 86944.0   |
| 78.10  | 5189.0    |
| 79.00  | 6407.0    |
| 80.00  | 4488.0    |
| 81.00  | 7575.0    |
| 82.00  | 1493.0    |
| 83.00  | 1471.0    |
| 84.10  | 310.0     |
| 85.00  | 1276.0    |
| 86.10  | 2328.0    |
| 87.00  | 1100.0    |
| 88.00  | 326.0     |
| 91.10  | 1751.0    |
| 92.00  | 2077.0    |
| 93.00  | 12945.0   |
| 94.00  | 677.0     |
| 94.90  | 356.0     |
| 96.00  | 784.0     |
| 96.90  | 315.0     |
| 98.00  | 10345.0   |
| 99.00  | 7687.0    |
| 99.90  | 912.0     |
| 101.00 | 4172.0    |
| 103.00 | 1535.0    |
| 104.10 | 2332.0    |
| 104.90 | 1965.0    |
| 106.10 | 522.0     |
| 107.00 | 32024.0   |
| 108.00 | 5002.0    |
| 108.90 | 831.0     |
| 110.00 | 51912.0   |
| 111.00 | 8238.0    |
| 112.00 | 1519.0    |
| 113.10 | 487.0     |
| 116.00 | 2279.0    |
| 117.00 | 31968.0   |
| 118.00 | 2980.0    |
| 119.10 | 315.0     |
| 120.00 | 396.0     |
| 122.00 | 2253.0    |
| 123.00 | 3242.0    |
| 124.00 | 1833.0    |
| 125.00 | 1467.0    |
| 126.20 | 705.0     |
| 127.00 | 98816.0   |
| 128.00 | 7788.0    |
| 129.00 | 41912.0   |
| 130.00 | 3524.0    |
| 131.00 | 887.0     |
| 131.90 | 427.0     |
| 134.00 | 1517.0    |
| 135.00 | 3335.0    |
| 136.00 | 1306.0    |
| 137.00 | 2257.0    |
| 138.00 | 463.0     |
| 139.00 | 378.0     |
| 140.00 | 573.0     |
| 141.00 | 7029.0    |
| 142.00 | 1945.0    |
| 142.90 | 1344.0    |
| 144.10 | 335.0     |
| 146.00 | 1599.0    |
| 147.00 | 3865.0    |
| 148.00 | 8260.0    |
| 149.00 | 1277.0    |
| 150.00 | 427.0     |
| 151.00 | 979.0     |
| 151.90 | 400.0     |
| 153.00 | 1907.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP65

|        |          |
|--------|----------|
| 154.00 | 1441.0   |
| 155.00 | 3994.0   |
| 156.10 | 4917.0   |
| 157.20 | 859.0    |
| 158.00 | 1392.0   |
| 158.90 | 946.0    |
| 159.90 | 2004.0   |
| 161.00 | 3651.0   |
| 161.80 | 965.0    |
| 165.00 | 2059.0   |
| 166.00 | 1825.0   |
| 167.00 | 13265.0  |
| 168.00 | 8582.0   |
| 168.90 | 953.0    |
| 169.90 | 315.0    |
| 171.00 | 769.0    |
| 172.10 | 948.0    |
| 173.00 | 1701.0   |
| 174.00 | 2434.0   |
| 175.00 | 5540.0   |
| 176.10 | 1010.0   |
| 177.00 | 2910.0   |
| 177.90 | 598.0    |
| 179.00 | 9975.0   |
| 180.00 | 6513.0   |
| 181.00 | 3201.0   |
| 184.00 | 889.0    |
| 185.10 | 4773.0   |
| 186.00 | 38328.0  |
| 187.00 | 10969.0  |
| 188.10 | 1043.0   |
| 189.00 | 2818.0   |
| 191.00 | 1567.0   |
| 192.00 | 2660.0   |
| 193.00 | 4380.0   |
| 194.10 | 606.0    |
| 195.00 | 413.0    |
| 196.00 | 6448.0   |
| 196.80 | 1081.0   |
| 197.10 | 1220.0   |
| 198.00 | 267264.0 |
| 199.00 | 17544.0  |
| 200.00 | 1345.0   |
| 201.40 | 1099.0   |
| 202.00 | 602.0    |
| 203.00 | 2302.0   |
| 204.00 | 11525.0  |
| 205.00 | 17536.0  |
| 206.00 | 77144.0  |
| 207.10 | 11735.0  |
| 208.00 | 3126.0   |
| 209.00 | 1162.0   |
| 210.10 | 1478.0   |
| 211.10 | 2971.0   |
| 211.90 | 789.0    |
| 215.00 | 1021.0   |
| 215.90 | 1726.0   |
| 217.00 | 23768.0  |
| 218.00 | 2606.0   |
| 220.30 | 483.0    |
| 221.00 | 10489.0  |
| 221.90 | 2936.0   |
| 223.00 | 5454.0   |
| 224.00 | 44240.0  |
| 225.00 | 12095.0  |
| 226.00 | 1217.0   |
| 227.00 | 20872.0  |
| 228.00 | 2761.0   |
| 229.00 | 3648.0   |
| 229.90 | 540.0    |
| 231.00 | 1584.0   |
| 232.90 | 301.0    |
| 233.90 | 941.0    |
| 235.00 | 1713.0   |
| 235.90 | 1204.0   |
| 237.00 | 1962.0   |
| 239.00 | 570.0    |
| 239.90 | 680.0    |
| 241.00 | 1057.0   |
| 242.00 | 2438.0   |
| 243.00 | 2761.0   |
| 244.10 | 36288.0  |
| 245.10 | 4095.0   |
| 246.00 | 7824.0   |
| 247.00 | 1661.0   |
| 248.00 | 404.0    |
| 248.90 | 1270.0   |
| 251.00 | 433.0    |
| 251.90 | 460.0    |
| 253.00 | 888.0    |
| 254.10 | 1477.0   |
| 255.00 | 195840.0 |
| 256.00 | 28160.0  |
| 257.00 | 2609.0   |
| 258.00 | 14013.0  |
| 259.00 | 1732.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP65

|        |          |
|--------|----------|
| 261.00 | 394.0    |
| 263.90 | 470.0    |
| 265.00 | 5638.0   |
| 265.90 | 800.0    |
| 271.00 | 554.0    |
| 271.90 | 540.0    |
| 273.00 | 5421.0   |
| 274.00 | 14481.0  |
| 275.00 | 84424.0  |
| 276.00 | 11313.0  |
| 277.00 | 7868.0   |
| 277.90 | 1502.0   |
| 279.00 | 485.0    |
| 281.10 | 696.0    |
| 282.00 | 407.0    |
| 283.00 | 1048.0   |
| 284.00 | 699.0    |
| 285.10 | 1693.0   |
| 289.00 | 462.0    |
| 290.00 | 312.0    |
| 292.00 | 493.0    |
| 293.00 | 2283.0   |
| 294.00 | 608.0    |
| 294.90 | 441.0    |
| 296.00 | 29936.0  |
| 297.00 | 4453.0   |
| 298.00 | 354.0    |
| 301.10 | 304.0    |
| 301.90 | 451.0    |
| 303.00 | 3208.0   |
| 304.20 | 774.0    |
| 309.90 | 324.0    |
| 312.90 | 304.0    |
| 314.10 | 1892.0   |
| 315.00 | 2818.0   |
| 316.00 | 1786.0   |
| 317.00 | 504.0    |
| 321.00 | 888.0    |
| 322.00 | 429.0    |
| 323.10 | 8976.0   |
| 324.00 | 1566.0   |
| 326.90 | 1846.0   |
| 328.00 | 778.0    |
| 331.90 | 827.0    |
| 332.90 | 679.0    |
| 334.00 | 5897.0   |
| 335.00 | 1498.0   |
| 340.90 | 949.0    |
| 346.00 | 1590.0   |
| 346.90 | 406.0    |
| 351.90 | 2236.0   |
| 353.10 | 1836.0   |
| 354.00 | 2286.0   |
| 354.90 | 499.0    |
| 365.00 | 14527.0  |
| 366.00 | 1777.0   |
| 370.90 | 531.0    |
| 372.00 | 4580.0   |
| 373.10 | 1058.0   |
| 383.10 | 1521.0   |
| 383.90 | 330.0    |
| 390.00 | 601.0    |
| 391.10 | 533.0    |
| 402.00 | 1810.0   |
| 403.00 | 2289.0   |
| 404.00 | 1270.0   |
| 421.00 | 2126.0   |
| 422.10 | 1932.0   |
| 423.10 | 15835.0  |
| 424.00 | 4245.0   |
| 425.10 | 484.0    |
| 441.10 | 41960.0  |
| 442.10 | 287488.0 |
| 443.10 | 54464.0  |
| 444.10 | 4285.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP65