

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092622\
 Data File : BM036745.D
 Acq On : 26 Sep 2022 14:57
 Operator : CG/JU
 Sample : PB147746BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS746

Quant Time: Sep 26 23:13:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 26 23:10:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.004	152	434789	20.000	ng/u1	0.00
20) Naphthalene-d8	10.816	136	1840759	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.627	164	1120182	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.368	188	2175529	20.000	ng/u1	0.00
79) Chrysene-d12	21.533	240	1725026	20.000	ng/u1	0.00
88) Perylene-d12	23.968	264	1407920	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	62423	5.382	ng/uL	0.00
4) Pyridine-d5	3.810	84	809179	23.922	ng/u1	0.00
7) Phenol-d5	7.163	99	1187118	30.553	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.334	67	711727	30.163	ng/u1	0.00
11) 2-Chlorophenol-d4	7.534	132	909892	32.527	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	911804	31.759	ng/u1	0.00
21) Nitrobenzene-d5	9.175	128	446043	35.800	ng/u1	0.00
24) 2-Nitrophenol-d4	9.898	143	436277	39.093	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.433	165	854089	33.023	ng/u1	0.00
31) 4-Chloroaniline-d4	10.951	131	1180214	27.446	ng/u1	0.00
46) Dimethylphthalate-d6	14.045	166	2402392	31.799	ng/u1	0.00
49) Acenaphthylene-d8	14.327	160	2869846	32.024	ng/u1	0.00
54) 4-Nitrophenol-d4	14.821	143	431536	29.449	ng/u1	0.00
60) Fluorene-d10	15.615	176	2100733	31.028	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.733	200	304906	35.424	ng/u1	0.00
73) Anthracene-d10	17.468	188	3118201	31.348	ng/u1	0.00
81) Pyrene-d10	19.750	212	3303859	38.931	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.809	264	2292205	33.324	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.422	88	131387	10.516	ng/uL	97
5) Pyridine	3.834	79	916026	26.955	ng/u1	95
6) Benzaldehyde	7.140	77	696105	36.997	ng/u1	96
8) Phenol	7.192	94	1219274	29.238	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.428	93	938324	28.920	ng/u1	99
12) 2-Chlorophenol	7.563	128	931777	30.792	ng/u1	100
13) 2-Methylphenol	8.445	108	894567	29.185	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.539	45	1111953	28.915	ng/u1	99
16) Acetophenone	8.834	105	1419273	28.501	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.828	70	711134	29.595	ng/u1	97
18) 4-Methylphenol	8.781	108	980890	29.593	ng/u1	99
19) Hexachloroethane	9.092	117	362575	29.582	ng/u1	93
22) Nitrobenzene	9.216	77	1066640	32.244	ng/u1	98
23) Isophorone	9.751	82	1957742	29.026	ng/u1	99
25) 2-Nitrophenol	9.928	139	475769	34.516	ng/u1	99
26) 2,4-Dimethylphenol	9.986	107	985664	28.993	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.228	93	1201812	29.894	ng/u1	99
29) 2,4-Dichlorophenol	10.463	162	842637	30.991	ng/u1	99
30) Naphthalene	10.869	128	2929650	29.215	ng/u1	100
32) 4-Chloroaniline	10.975	127	1195552	26.930	ng/u1	99
33) Hexachlorobutadiene	11.151	225	476404	28.648	ng/u1	99
34) Caprolactam	11.769	113	182224	20.661	ng/u1	99
35) 4-Chloro-3-methylphenol	12.092	107	919514	30.636	ng/u1	98
36) 2-Methylnaphthalene	12.469	142	1950347	29.192	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092622\
 Data File : BM036745.D
 Acq On : 26 Sep 2022 14:57
 Operator : CG/JU
 Sample : PB147746BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS746

Quant Time: Sep 26 23:13:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 26 23:10:22 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.686	142	1975316	29.409	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.833	216	936015	29.522	ng/ul	100
40) Hexachlorocyclopentadiene	12.810	237	397005	24.027	ng/ul	99
41) 2,4,6-Trichlorophenol	13.069	196	603892	31.108	ng/ul	100
42) 2,4,5-Trichlorophenol	13.139	196	662040	31.808	ng/ul	99
43) 1,1'-Biphenyl	13.469	154	2490249	30.157	ng/ul	99
44) 2-Chloronaphthalene	13.516	162	1921726	29.658	ng/ul	98
45) 2-Nitroaniline	13.716	65	578273	36.258	ng/ul	98
47) Dimethylphthalate	14.092	163	2385462	29.766	ng/ul	99
48) 2,6-Dinitrotoluene	14.210	165	460939	35.377	ng/ul	98
50) Acenaphthylene	14.357	152	2997532	29.256	ng/ul	99
51) 3-Nitroaniline	14.533	138	517972	33.146	ng/ul#	98
52) Acenaphthene	14.692	153	2105107	29.365	ng/ul	99
53) 2,4-Dinitrophenol	14.739	184	172550	29.846	ng/ul	97
55) 4-Nitrophenol	14.839	109	344223	27.480	ng/ul	98
56) Dibenzofuran	15.027	168	2811013	29.197	ng/ul	98
57) 2,4-Dinitrotoluene	14.992	165	662877	34.995	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.251	232	525548	30.164	ng/ul#	96
59) Diethylphthalate	15.451	149	2403558	29.960	ng/ul	100
61) Fluorene	15.674	166	2428047	29.611	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.668	204	1140581	29.085	ng/ul	100
63) 4-Nitroaniline	15.692	138	493433	32.500	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.751	198	311135	32.033	ng/ul	98
67) N-Nitrosodiphenylamine	15.880	169	1972322	30.781	ng/ul	99
68) 4-Bromophenyl-phenylether	16.557	248	641576	29.920	ng/ul	99
69) Hexachlorobenzene	16.668	284	726021	28.626	ng/ul	94
70) Atrazine	16.833	200	656939	29.054	ng/ul	98
71) Pentachlorophenol	17.015	266	412066	27.352	ng/ul	99
72) Phenanthrene	17.409	178	3635354	30.337	ng/ul	99
74) Anthracene	17.504	178	3659696	30.047	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.433	216	974608	32.754	ng/uL	100
76) Pentachlorobenzene	14.945	250	850447	26.426	ng/uL	99
77) Carbazole	17.768	167	3244037	28.767	ng/ul	100
78) Di-n-butylphthalate	18.333	149	4115255	32.197	ng/ul	100
80) Fluoranthene	19.415	202	3882303	37.332	ng/ul	99
82) Pyrene	19.780	202	4036984	36.673	ng/ul	98
83) Butylbenzylphthalate	20.674	149	1635797	39.217	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.450	252	1069017	31.716	ng/ul	98
85) Benzo(a)anthracene	21.521	228	3376755	31.242	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.444	149	2493219	40.918	ng/ul	99
87) Chrysene	21.574	228	3147878	30.684	ng/ul	99
89) Di-n-octyl phthalate	22.386	149	3657396	43.089	ng/ul	100
90) Benzo(b)fluoranthene	23.227	252	2932622	32.847	ng/ul	99
91) Benzo(k)fluoranthene	23.274	252	2882044	32.306	ng/ul	98
93) Benzo(a)pyrene	23.862	252	2481131	31.612	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.509	276	2676682	29.203	ng/ul	96
95) Dibenzo(a,h)anthracene	26.526	278	2413000	29.030	ng/ul	99
96) Benzo(g,h,i)perylene	27.285	276	2334493	29.380	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092622\
 Data File : BM036745.D
 Acq On : 26 Sep 2022 14:57
 Operator : CG/JU
 Sample : PB147746BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS746

Quant Time: Sep 26 23:13:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM091522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Sep 26 23:10:22 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 10.516 ng/uL
RT: 3.422 min Scan# 57
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion: 88 Resp: 131387
Ion Ratio Lower Upper
88 100
43 27.1 23.1 34.7
58 68.4 56.6 85.0



#5
Pyridine
Concen: 26.955 ng/uL
RT: 3.834 min Scan# 127
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion: 79 Resp: 916026
Ion Ratio Lower Upper
79 100
52 66.0 57.2 85.8
51 32.1 26.8 40.2









#6
Benzaldehyde
Concen: 36.997 ng/ul
RT: 7.140 min Scan# 689
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion: 77 Resp: 696105
Ion Ratio Lower Upper
77 100
105 96.7 75.1 112.7
106 95.2 73.0 109.6



#8
Phenol
Concen: 29.238 ng/ul
RT: 7.192 min Scan# 698
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion: 94 Resp: 1219274
Ion Ratio Lower Upper
94 100
65 27.4 22.2 33.4
66 35.0 28.9 43.3









#10
Bis(2-Chloroethyl)ether
Concen: 28.920 ng/ul
RT: 7.428 min Scan# 738
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion: 93 Resp: 938324
Ion Ratio Lower Upper
93 100
63 72.9 59.1 88.7
95 32.6 26.1 39.1



#12
2-Chlorophenol
Concen: 30.792 ng/ul
RT: 7.563 min Scan# 761
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion: 128 Resp: 931777
Ion Ratio Lower Upper
128 100
64 47.5 37.8 56.8
130 32.1 26.0 39.0





#13

2-Methylphenol

Concen: 29.185 ng/ul

RT: 8.445 min Scan# 911

Delta R.T. 0.000 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Instrument :

BNA_M

ClientSampleId :





SLCS746

Tgt Ion:108 Resp: 894567

Ion Ratio Lower Upper

108 100

107 87.9 70.7 106.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 28.915 ng/ul

RT: 8.539 min Scan# 927

Delta R.T. 0.000 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57




Tgt Ion: 45 Resp: 1111953


Ion Ratio Lower Upper

45 100

77 16.7 13.1 19.7

79 12.8 9.5 14.3









#16
Acetophenone
Concen: 28.501 ng/ul
RT: 8.834 min Scan# 91
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:105 Resp: 1419273
Ion Ratio Lower Upper
105 100
77 78.5 61.6 92.4
51 24.7 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 29.595 ng/ul
RT: 8.828 min Scan# 976
Delta R.T. 0.012 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion: 70 Resp: 711134
Ion Ratio Lower Upper
70 100
42 51.6 43.0 64.4
101 9.9 7.7 11.5
130 21.5 15.1 22.7









#18
4-Methylphenol
Concen: 29.593 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:108 Resp: 980890
Ion Ratio Lower Upper
108 100
107 110.5 89.4 134.0



#19
Hexachloroethane
Concen: 29.582 ng/ul
RT: 9.092 min Scan# 1021
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:117 Resp: 362575
Ion Ratio Lower Upper
117 100
201 93.8 68.7 103.1
199 57.7 43.0 64.6









#22
Nitrobenzene
Concen: 32.244 ng/ul
RT: 9.216 min Scan# 1042
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion: 77 Resp: 1066640
Ion Ratio Lower Upper
77 100
123 45.0 34.9 52.3
65 13.0 10.7 16.1



#23
Isophorone
Concen: 29.026 ng/ul
RT: 9.751 min Scan# 1133
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion: 82 Resp: 1957742
Ion Ratio Lower Upper
82 100
95 7.0 5.4 8.2
138 17.5 13.5 20.3









#25
2-Nitrophenol
Concen: 34.516 ng/ul
RT: 9.928 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:139 Resp: 475769
Ion Ratio Lower Upper
139 100
65 51.9 40.9 61.3
109 30.8 23.9 35.9



#26
2,4-Dimethylphenol
Concen: 28.993 ng/ul
RT: 9.986 min Scan# 1173
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:107 Resp: 985664
Ion Ratio Lower Upper
107 100
121 50.4 39.4 59.2
122 87.0 69.1 103.7





#27

Bis(2-Chloroethoxy)methane

Concen: 29.894 ng/ul

RT: 10.228 min Scan# 1213

Delta R.T. 0.006 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Instrument :





BNA_M

ClientSampleId :

SLCS746

Tgt Ion: 93 Resp: 1201812

Ion	Ratio	Lower	Upper
93	100		
95	32.4	26.2	39.4
123	11.7	9.2	13.8



#29

2,4-Dichlorophenol

Concen: 30.991 ng/ul

RT: 10.463 min Scan# 1254




Delta R.T. 0.006 min


Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Tgt Ion: 162 Resp: 842637

Ion	Ratio	Lower	Upper
162	100		
164	66.0	52.1	78.1
98	36.7	29.1	43.7









#30
Naphthalene
Concen: 29.215 ng/ul
RT: 10.869 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:128 Resp: 2929650
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.0 10.3 15.5



#32
4-Chloroaniline
Concen: 26.930 ng/ul
RT: 10.975 min Scan# 1341
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:127 Resp: 1195552
Ion Ratio Lower Upper
127 100
129 32.6 25.4 38.2









#33
Hexachlorobutadiene
Concen: 28.648 ng/ul
RT: 11.151 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:225 Resp: 476404
Ion Ratio Lower Upper
225 100
223 63.3 49.8 74.8
227 64.8 51.0 76.6



#34
Caprolactam
Concen: 20.661 ng/ul
RT: 11.769 min Scan# 1476
Delta R.T. 0.012 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:113 Resp: 182224
Ion Ratio Lower Upper
113 100
55 158.5 127.5 191.3
56 120.6 94.6 142.0









#35
4-Chloro-3-methylphenol
Concen: 30.636 ng/ul
RT: 12.092 min Scan# 1531
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:107 Resp: 919514
Ion Ratio Lower Upper
107 100
144 25.4 21.5 32.3
142 77.9 63.9 95.9



#36
2-Methylnaphthalene
Concen: 29.192 ng/ul
RT: 12.469 min Scan# 1595
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:142 Resp: 1950347
Ion Ratio Lower Upper
142 100
141 89.5 70.7 106.1





#37

1-Methylnaphthalene

Concen: 29.409 ng/ul

RT: 12.686 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Instrument :

BNA_M

ClientSampleId :

SLCS746





Tgt Ion:142 Resp: 1975316

Ion Ratio Lower Upper

142 100

141 91.9 73.4 110.2

116 4.0 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 29.522 ng/ul

RT: 12.833 min Scan# 1657

Delta R.T. 0.006 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Tgt Ion:216 Resp: 936015




Ion Ratio Lower Upper


216 100

214 77.9 62.4 93.6

179 21.1 16.7 25.1


108 19.2 15.4 23.0








#40
Hexachlorocyclopentadiene
Concen: 24.027 ng/ul
RT: 12.810 min Scan# 1653
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57


Instrument :
BNA_M
ClientSampleId :
SLCS746





Tgt Ion:237 Resp: 397005
Ion Ratio Lower Upper
237 100
235 62.5 50.9 76.3
272 13.2 10.3 15.5




#41
2,4,6-Trichlorophenol
Concen: 31.108 ng/ul
RT: 13.069 min Scan# 1697
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57



Tgt Ion:196 Resp: 603892
Ion Ratio Lower Upper
196 100
198 96.6 77.3 115.9
200 30.5 25.2 37.8









#42
 2,4,5-Trichlorophenol
 Concen: 31.808 ng/ul
 RT: 13.139 min Scan# 1709
 Delta R.T. 0.000 min
 Lab File: BM036745.D
 Acq: 26 Sep 2022 14:57




Instrument :
 BNA_M
 ClientSampleId :
 SLCS746


Tgt Ion:196 Resp: 662040
 Ion Ratio Lower Upper
 196 100
 198 93.8 75.6 113.4
 200 30.6 24.7 37.1



#43
 1,1'-Biphenyl
 Concen: 30.157 ng/ul
 RT: 13.469 min Scan# 1765
 Delta R.T. 0.000 min
 Lab File: BM036745.D
 Acq: 26 Sep 2022 14:57

Tgt Ion:154 Resp: 2490249
 Ion Ratio Lower Upper
 154 100
 153 39.8 31.4 47.2
 76 13.5 10.4 15.6





#44

2-Chloronaphthalene

Concen: 29.658 ng/ul

RT: 13.516 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Instrument :

BNA_M

ClientSampleId :

SLCS746





Tgt Ion:162 Resp: 1921726

Ion Ratio Lower Upper

162 100

164 32.7 26.2 39.2

127 39.1 32.6 49.0



#45

2-Nitroaniline

Concen: 36.258 ng/ul

RT: 13.716 min Scan# 1807

Delta R.T. 0.000 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57




Tgt Ion: 65 Resp: 578273


Ion Ratio Lower Upper

65 100

92 70.1 54.4 81.6

138 105.6 86.1 129.1









#47
Dimethylphthalate
Concen: 29.766 ng/ul
RT: 14.092 min Scan# 1871
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:163 Resp: 2385462
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 10.1 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 35.377 ng/ul
RT: 14.210 min Scan# 1891
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:165 Resp: 460939
Ion Ratio Lower Upper
165 100
89 64.8 52.8 79.2
121 22.7 19.4 29.0





#50

Acenaphthylene

Concen: 29.256 ng/ul

RT: 14.357 min Scan# 1915

Delta R.T. 0.006 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Instrument :

BNA_M

ClientSampleId :

SLCS746





Tgt Ion:152 Resp: 2997532

Ion Ratio Lower Upper

152 100

151 20.2 15.7 23.5

153 13.0 10.5 15.7



#51

3-Nitroaniline

Concen: 33.146 ng/ul

RT: 14.533 min Scan# 1946

Delta R.T. 0.000 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57




Tgt Ion:138 Resp: 517972


Ion Ratio Lower Upper

138 100

108 10.8 12.5 18.7#

92 119.9 95.4 143.2









#52
Acenaphthene
Concen: 29.365 ng/ul
RT: 14.692 min Scan# 1973
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:153 Resp: 2105107
Ion Ratio Lower Upper
153 100
152 48.4 38.2 57.2
154 88.1 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 29.846 ng/ul
RT: 14.739 min Scan# 1981
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:184 Resp: 172550
Ion Ratio Lower Upper
184 100
63 72.8 58.8 88.2
154 64.9 55.0 82.4









#55
4-Nitrophenol
Concen: 27.480 ng/ul
RT: 14.839 min Scan# 1997
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:109 Resp: 344223
Ion Ratio Lower Upper
109 100
139 130.6 102.2 153.4
65 129.7 105.7 158.5



#56
Dibenzofuran
Concen: 29.197 ng/ul
RT: 15.027 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:168 Resp: 2811013
Ion Ratio Lower Upper
168 100
139 38.1 29.7 44.5





#57

2,4-Dinitrotoluene

Concen: 34.995 ng/ul

RT: 14.992 min Scan# 2024

Delta R.T. 0.006 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Instrument :

BNA_M

ClientSampleId :

SLCS746





Tgt Ion:165 Resp: 662877

Ion Ratio Lower Upper

165 100

63 44.9 39.7 59.5

182 2.8 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 30.164 ng/ul

RT: 15.251 min Scan# 2068

Delta R.T. 0.006 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Tgt Ion:232 Resp: 525548




Ion Ratio Lower Upper


232 100

131 52.0 39.1 58.7

130 2.7 1.8 2.6#

166 33.0 25.4 38.0









#59
Diethylphthalate
Concen: 29.960 ng/ul
RT: 15.451 min Scan# 2101
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:149 Resp: 2403558
Ion Ratio Lower Upper
149 100
177 21.5 17.3 25.9
150 12.5 10.1 15.1



#61
Fluorene
Concen: 29.611 ng/ul
RT: 15.674 min Scan# 2140
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:166 Resp: 2428047
Ion Ratio Lower Upper
166 100
165 96.8 78.5 117.7
167 13.4 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 29.085 ng/ul
RT: 15.668 min Scan# 2139
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:204 Resp: 1140581
Ion Ratio Lower Upper
204 100
206 32.6 25.7 38.5
141 61.6 49.4 74.0



#63
4-Nitroaniline
Concen: 32.500 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:138 Resp: 493433
Ion Ratio Lower Upper
138 100
92 53.2 42.1 63.1
108 82.5 66.2 99.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 32.033 ng/ul
 RT: 15.751 min Scan# 2153
 Delta R.T. 0.006 min
 Lab File: BM036745.D
 Acq: 26 Sep 2022 14:57




Instrument :
 BNA_M
 ClientSampleId :
 SLCS746


Tgt Ion:198 Resp: 311135
 Ion Ratio Lower Upper
 198 100
 182 4.7 4.2 6.2
 77 28.9 24.0 36.0



#67
 N-Nitrosodiphenylamine
 Concen: 30.781 ng/ul
 RT: 15.880 min Scan# 2175
 Delta R.T. 0.000 min
 Lab File: BM036745.D
 Acq: 26 Sep 2022 14:57

Tgt Ion:169 Resp: 1972322
 Ion Ratio Lower Upper
 169 100
 168 66.9 54.6 81.8
 167 37.0 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 29.920 ng/ul
RT: 16.557 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:248 Resp: 641576
Ion Ratio Lower Upper
248 100
250 97.6 78.9 118.3
141 82.7 65.7 98.5



#69
Hexachlorobenzene
Concen: 28.626 ng/ul
RT: 16.668 min Scan# 2309
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:284 Resp: 726021
Ion Ratio Lower Upper
284 100
142 43.2 30.8 46.2
249 31.8 24.2 36.2





#70

Atrazine

Concen: 29.054 ng/ul

RT: 16.833 min Scan# 21

Delta R.T. 0.006 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Instrument :

BNA_M

ClientSampleId :

SLCS746





Tgt Ion:200 Resp: 656939

Ion Ratio Lower Upper

200 100

173 26.9 21.8 32.6

215 47.4 39.4 59.2



#71

Pentachlorophenol

Concen: 27.352 ng/ul

RT: 17.015 min Scan# 2368

Delta R.T. 0.000 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57




Tgt Ion:266 Resp: 412066


Ion Ratio Lower Upper

266 100

264 62.3 49.7 74.5

268 63.5 51.8 77.6









#72
Phenanthrene
Concen: 30.337 ng/ul
RT: 17.409 min Scan# 2435
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:178 Resp: 3635354
Ion Ratio Lower Upper
178 100
179 15.7 12.2 18.2
176 19.7 15.7 23.5



#74
Anthracene
Concen: 30.047 ng/ul
RT: 17.504 min Scan# 2451
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:178 Resp: 3659696
Ion Ratio Lower Upper
178 100
179 15.5 12.3 18.5
176 19.2 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 32.754 ng/uL
RT: 13.433 min Scan# 1759
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:216 Resp: 974608
Ion Ratio Lower Upper
216 100
214 77.9 62.6 94.0
218 47.5 38.2 57.2



#76
Pentachlorobenzene
Concen: 26.426 ng/uL
RT: 14.945 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:250 Resp: 850447
Ion Ratio Lower Upper
250 100
248 63.1 50.2 75.2
252 64.4 50.5 75.7





#77

Carbazole

Concen: 28.767 ng/ul

RT: 17.768 min Scan# 2496

Delta R.T. 0.000 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Instrument :

BNA_M

ClientSampleId :

SLCS746





Tgt Ion:167 Resp: 3244037

Ion Ratio Lower Upper

167 100

166 21.7 17.2 25.8

139 13.4 10.6 15.8



#78

Di-n-butylphthalate

Concen: 32.197 ng/ul

RT: 18.333 min Scan# 2592

Delta R.T. 0.000 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57




Tgt Ion:149 Resp: 4115255


Ion Ratio Lower Upper

149 100

150 9.2 7.2 10.8

104 5.3 4.3 6.5









#80
Fluoranthene
Concen: 37.332 ng/ul
RT: 19.415 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:202 Resp: 3882303
Ion Ratio Lower Upper
202 100
101 12.9 9.8 14.6
100 9.6 7.4 11.2



#82
Pyrene
Concen: 36.673 ng/ul
RT: 19.780 min Scan# 2838
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:202 Resp: 4036984
Ion Ratio Lower Upper
202 100
101 14.3 10.9 16.3
100 11.6 8.7 13.1





#83

Butylbenzylphthalate

Concen: 39.217 ng/ul

RT: 20.674 min Scan# 2989

Delta R.T. 0.006 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Instrument :

BNA_M

ClientSampleId :

SLCS746





Tgt Ion:149 Resp: 1635797

Ion Ratio Lower Upper

149 100

91 73.7 55.2 82.8

206 19.6 16.2 24.2



#84

3,3'-Dichlorobenzidine

Concen: 31.716 ng/ul

RT: 21.450 min Scan# 3122

Delta R.T. 0.000 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57




Tgt Ion:252 Resp: 1069017


Ion Ratio Lower Upper

252 100

254 64.2 53.0 79.4

126 12.8 10.4 15.6





#85

Benzo(a)anthracene

Concen: 31.242 ng/ul

RT: 21.521 min Scan# 3134

Delta R.T. 0.006 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Instrument :

BNA_M

ClientSampleId :

SLCS746





Tgt Ion:228 Resp: 3376755

Ion Ratio Lower Upper

228 100

229 19.7 15.8 23.8

226 27.7 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 40.918 ng/ul

RT: 21.444 min Scan# 3121

Delta R.T. 0.000 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57




Tgt Ion:149 Resp: 2493219


Ion Ratio Lower Upper

149 100

167 26.8 22.0 33.0

279 3.5 3.3 4.9









#87
Chrysene
Concen: 30.684 ng/ul
RT: 21.574 min Scan# 3143
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:228 Resp: 3147878
Ion Ratio Lower Upper
228 100
226 31.0 24.4 36.6
229 20.0 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 43.089 ng/ul
RT: 22.386 min Scan# 3281
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:149 Resp: 3657396









#90
Benzo(b)fluoranthene
Concen: 32.847 ng/ul
RT: 23.227 min Scan# 3423
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57




Instrument :
BNA_M
ClientSampleId :
SLCS746


Tgt Ion:252 Resp: 2932622
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 11.8 9.1 13.7



#91
Benzo(k)fluoranthene
Concen: 32.306 ng/ul
RT: 23.274 min Scan# 3432
Delta R.T. 0.006 min
Lab File: BM036745.D
Acq: 26 Sep 2022 14:57

Tgt Ion:252 Resp: 2882044
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.8
125 12.0 8.2 12.4





#93

Benzo(a)pyrene

Concen: 31.612 ng/ul

RT: 23.862 min Scan# 3

Delta R.T. 0.006 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Instrument :

BNA_M

ClientSampleId :

SLCS746





Tgt Ion:252 Resp: 2481131

Ion Ratio Lower Upper

252 100

253 22.0 17.5 26.3

125 12.7 9.5 14.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 29.203 ng/ul

RT: 26.509 min Scan# 3982

Delta R.T. 0.012 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57




Tgt Ion:276 Resp: 2676682


Ion Ratio Lower Upper

276 100

138 28.9 20.2 30.4

277 24.7 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 29.030 ng/ul

RT: 26.526 min Scan# 3985

Delta R.T. 0.012 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Instrument :

BNA_M

ClientSampleId :

SLCS746





Tgt Ion:278 Resp: 2413000

Ion Ratio Lower Upper

278 100

139 18.1 14.5 21.7

279 23.6 19.7 29.5



#96

Benzo(g,h,i)perylene

Concen: 29.380 ng/ul

RT: 27.285 min Scan# 4114

Delta R.T. 0.012 min

Lab File: BM036745.D

Acq: 26 Sep 2022 14:57

Tgt Ion:276 Resp: 2334493

Ion Ratio Lower Upper

276 100

138 25.3 18.7 28.1

277 23.5 19.2 28.8

