



2-Nitrophenol



$\text{Response} = 1.631\text{e-}001 * \text{Amt} - 2.056\text{e-}002$
 Coef of Det (r^2) = 0.999183 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM092722.M
 Calibration Table Last Updated: Wed Sep 28 01:39:14 2022

3-Nitroaniline

Response Ratio



$$\text{Response} = 3.667\text{e-}001 * \text{Amt} - 4.036\text{e-}002$$


Coef of Det (r^2) = 0.999532 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM092722.M

Calibration Table Last Updated: Wed Sep 28 01:39:14 2022

Benzoic acid


Response Ratio



$R = 2.639e-002 A^2 + 1.403e-001 A - 1.527e-002$
 Coef of Det (r^2) = 0.996115 Curve Fit: Quadratic
 Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM092722.M
 Calibration Table Last Updated: Wed Sep 28 01:39:14 2022

2,4-Dinitrotoluene

Response Ratio




$$\text{Response} = 3.938\text{e-}001 * \text{Amt} - 5.306\text{e-}002$$

Coef of Det (r^2) = 0.998403 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM092722.M

Calibration Table Last Updated: Wed Sep 28 01:39:14 2022


2,4-Dinitrophenol



$\text{Response} = 1.572\text{e-}001 * \text{Amt} - 4.408\text{e-}002$
 Coef of Det (r^2) = 0.992342 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM092722.M
 Calibration Table Last Updated: Wed Sep 28 01:39:14 2022

Bis(2-ethylhexyl)phthalate

Response Ratio



$$\text{Response} = 6.890\text{e-}001 * \text{Amt} - 8.483\text{e-}002$$


Coef of Det (r^2) = 0.998682 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM092722.M

Calibration Table Last Updated: Wed Sep 28 01:39:14 2022

4-Nitroaniline

Response Ratio




$$\text{Response} = 3.738\text{e-}001 * \text{Amt} - 4.044\text{e-}002$$

Coef of Det (r^2) = 0.999873 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM092722.M

Calibration Table Last Updated: Wed Sep 28 01:39:14 2022


4,6-Dinitro-2-methylphenol



Response = $1.177 \times 10^{-1} \times \text{Amt} - 3.234 \times 10^{-2}$
 Coef of Det (r^2) = 0.991842 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM092722.M
 Calibration Table Last Updated: Wed Sep 28 01:39:14 2022

Di-n-octyl phthalate

Response Ratio



$$\text{Response} = 1.181\text{e}+000 * \text{Amt} - 3.101\text{e}-001$$

Coef of Det (r^2) = 0.998051 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM092722.M

Calibration Table Last Updated: Wed Sep 28 01:39:14 2022