

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092722\
 Data File : BM036788.D
 Acq On : 27 Sep 2022 22:29
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM092722

Quant Time: Sep 28 01:51:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM092722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 01:39:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	115	0.00
2	1,4-Dioxane	40.000	38.846	2.9	112	0.00
3	Pyridine	40.000	41.308	-3.3	115	0.00
4	n-Nitrosodimethylamine	40.000	39.096	2.3	111	0.00
5 S	2-Fluorophenol	80.000	81.706	-2.1	115	0.00
6	Aniline	40.000	42.298	-5.7	117	0.00
7 S	Phenol-d6	80.000	84.015	-5.0	117	0.00
8	2-Chlorophenol	40.000	41.586	-4.0	116	0.00
9	Benzaldehyde	40.000	40.985	-2.5	118	0.00
10 C	Phenol	40.000	42.211	-5.5	117	0.00
11	bis(2-Chloroethyl)ether	40.000	41.131	-2.8	115	0.00
12	1,3-Dichlorobenzene	40.000	40.694	-1.7	114	0.00
13 C	1,4-Dichlorobenzene	40.000	40.469	-1.2	114	0.00
14	1,2-Dichlorobenzene	40.000	40.468	-1.2	114	0.00
15	Benzyl Alcohol	40.000	42.949	-7.4	119	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.631	0.9	112	0.00
17	2-Methylphenol	40.000	42.209	-5.5	118	0.00
18	Hexachloroethane	40.000	40.542	-1.4	114	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.958	-7.4	119	0.00
20	3+4-Methylphenols	40.000	42.884	-7.2	119	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	118	0.00
22	Acetophenone	40.000	41.689	-4.2	118	0.00
23 S	Nitrobenzene-d5	80.000	83.413	-4.3	117	0.00
24	Nitrobenzene	40.000	41.170	-2.9	116	0.00
25	Isophorone	40.000	43.394	-8.5	122	0.00
26 C	2-Nitrophenol	40.000	43.489	-8.7	127	0.00
27	2,4-Dimethylphenol	40.000	43.099	-7.7	121	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.481	-3.7	119	0.00
29 C	2,4-Dichlorophenol	40.000	43.733	-9.3	122	0.00
30	1,2,4-Trichlorobenzene	40.000	41.646	-4.1	120	0.00
31	Naphthalene	40.000	40.779	-1.9	117	0.00
32	Benzoic acid	40.000	45.885	-14.7	135	0.01
33	4-Chloroaniline	40.000	42.197	-5.5	121	0.00
34 C	Hexachlorobutadiene	40.000	41.488	-3.7	120	0.00
35	Caprolactam	40.000	43.341	-8.4	129	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.148	-7.9	123	0.00
37	2-Methylnaphthalene	40.000	42.326	-5.8	122	0.00
38	1-Methylnaphthalene	40.000	42.247	-5.6	121	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	124	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.788	-4.5	124	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.690	-6.7	122	0.00
42 S	2,4,6-Tribromophenol	80.000	85.618	-7.0	134	0.00
43 C	2,4,6-Trichlorophenol	40.000	44.479	-11.2	129	0.00
44	2,4,5-Trichlorophenol	40.000	44.148	-10.4	129	0.00
45 S	2-Fluorobiphenyl	80.000	83.459	-4.3	122	0.00
46	1,1'-Biphenyl	40.000	41.050	-2.6	123	0.00
47	2-Chloronaphthalene	40.000	41.198	-3.0	123	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092722\
 Data File : BM036788.D
 Acq On : 27 Sep 2022 22:29
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM092722

Quant Time: Sep 28 01:51:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM092722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 01:39:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.601	-9.0	125	0.00
49	Acenaphthylene	40.000	42.488	-6.2	124	0.00
50	Dimethylphthalate	40.000	41.364	-3.4	127	0.00
51	2,6-Dinitrotoluene	40.000	44.327	-10.8	130	0.00
52 C	Acenaphthene	40.000	41.203	-3.0	125	0.00
53	3-Nitroaniline	40.000	41.232	-3.1	128	0.00
54 P	2,4-Dinitrophenol	40.000	42.571	-6.4	139	0.00
55	Dibenzofuran	40.000	41.139	-2.8	125	0.00
56 P	4-Nitrophenol	40.000	41.610	-4.0	128	0.00
57	2,4-Dinitrotoluene	40.000	41.744	-4.4	132	0.00
58	Fluorene	40.000	41.279	-3.2	125	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.454	-8.6	130	0.00
60	Diethylphthalate	40.000	40.736	-1.8	126	0.00
61	4-Chlorophenyl-phenylether	40.000	41.649	-4.1	126	0.00
62	4-Nitroaniline	40.000	40.700	-1.8	127	0.00
63	Azobenzene	40.000	41.486	-3.7	122	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	126	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.036	-5.1	136	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.521	-6.3	125	0.00
67	4-Bromophenyl-phenylether	40.000	44.039	-10.1	130	0.00
68	Hexachlorobenzene	40.000	43.380	-8.5	131	0.00
69	Atrazine	40.000	43.377	-8.4	129	0.00
70 C	Pentachlorophenol	40.000	43.835	-9.6	134	0.00
71	Phenanthrene	40.000	40.687	-1.7	124	0.00
72	Anthracene	40.000	42.024	-5.1	124	0.00
73	Carbazole	40.000	40.388	-1.0	122	0.00
74	Di-n-butylphthalate	40.000	41.533	-3.8	126	0.00
75 C	Fluoranthene	40.000	39.942	0.1	122	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	118	0.00
77	Benzidine	40.000	40.705	-1.8	115	0.00
78	Pyrene	40.000	42.776	-6.9	122	0.00
79 S	Terphenyl-d14	80.000	86.095	-7.6	122	0.00
80	Butylbenzylphthalate	40.000	47.331	-18.3	122	0.00
81	Benzo(a)anthracene	40.000	41.821	-4.6	119	0.00
82	3,3'-Dichlorobenzidine	40.000	43.689	-9.2	120	0.00
83	Chrysene	40.000	40.748	-1.9	117	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.025	-2.6	120	0.00
85 c	Di-n-octyl phthalate	40.000	39.012	2.5	118	0.00
86 I	Perylene-d12	20.000	20.000	0.0	117	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.003	-5.0	115	0.00
88	Benzo(b)fluoranthene	40.000	41.215	-3.0	119	0.00
89	Benzo(k)fluoranthene	40.000	41.125	-2.8	112	0.00
90 C	Benzo(a)pyrene	40.000	41.817	-4.5	116	0.00
91	Dibenzo(a,h)anthracene	40.000	41.801	-4.5	115	0.00
92	Benzo(g,h,i)perylene	40.000	42.381	-6.0	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092722\
Data File : BM036788.D
Acq On : 27 Sep 2022 22:29
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM092722

Quant Time: Sep 28 01:51:20 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM092722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 28 01:39:14 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0