

Data Path : Z:\HPCHEM1\BNA M\DATA\BM092815\
 Data File : BM002753.D
 Acq On : 29 Sep 2015 02:59
 Operator : UM/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.489

Quant Time: Sep 29 04:42:43 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM092415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 29 02:52:19 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	132	0.02
2 SURR	1,4-Dioxane-d8	0.413	0.388	6.1	127	0.00
3	1,4-Dioxane	0.479	0.453	5.4	127	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	133	0.00
5	Naphthalene	1.084	1.052	3.0	130	0.00
6 SURR	2-Methylnaphthalene-d10	0.585	0.528	9.7	127	0.00
7	2-Methylnaphthalene	0.698	0.682	2.3	130	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	126	0.00
9	Acenaphthylene	2.063	2.058	0.2	135	0.00
10 C	Acenaphthene	1.506	1.529	-1.5	129	0.00
11	Fluorene	1.717	1.711	0.3	130	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	119	0.00
13	Pentachlorophenol	0.078	0.058	25.6#	89	0.00
14	Phenanthrene	1.302	1.216	6.6	118	0.00
15	Anthracene	1.203	1.238	-2.9	124	0.00
16 SURR	Fluoranthene-d10	0.997	1.005	-0.8	121	0.00
17 C	Fluoranthene	1.416	1.370	3.2	120	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	117	0.00
19	Pyrene	1.559	1.544	1.0	118	0.01
20	Benzo(a)anthracene	1.337	1.350	-1.0	119	0.00
21	Chrysene	1.280	1.280	0.0	116	0.00
22 I	Perylene-d12	1.000	1.000	0.0	120	0.00
23	Benzo(b)fluoranthene	1.442	1.490	-3.3	129	0.00
24	Benzo(k)fluoranthene	1.336	1.248	6.6	107	0.00
25 C	Benzo(a)pyrene	1.328	1.343	-1.1	121	0.00
26	Indeno(1,2,3-cd)pyrene	1.481	1.515	-2.3	123	0.00
27	Dibenzo(a,h)anthracene	1.215	1.248	-2.7	123	0.00
28	Benzo(q,h,i)perylene	1.256	1.288	-2.5	123	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0