


Data Path : Z:\HPCHEM1\BNA M\DATA\BM092815\
 Data File : BM002728.D
 Acq On : 28 Sep 2015 10:50
 Operator : UM/IZ
 Sample : DFTPP17
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP17

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM092415.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Sep 29 02:52:19 2015



AutoFind: Scans 2600, 2601, 2602; Background Corrected with Scan 2594

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.8	33501	PASS
68	69	0.00	2	2.0	683	PASS
69	198	0.00	100	34.1	34870	PASS
70	69	0.00	2	1.0	350	PASS
127	198	10	80	49.6	50650	PASS
197	198	0.00	1	0.4	415	PASS
198	198	100	100	100.0	102210	PASS
199	198	5	9	6.7	6809	PASS
275	198	10	60	26.1	26709	PASS
365	198	1	100	2.9	2946	PASS
441	443	0.01	100	77.0	14180	PASS
442	198	50	100	92.1	94146	PASS
443	442	15	24	19.6	18411	PASS

m/z	Abundance
39.90	516.0
41.00	565.0
43.00	761.0
44.00	1916.0
55.00	471.0
57.00	528.0
58.00	347.0
69.00	300.0
71.00	376.0
73.00	400.0
91.00	376.0
147.00	300.0
207.00	512.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

m/z	Abundance
38.00	503.0
39.00	3046.0
40.00	503.0
41.00	787.0
43.00	1167.0
44.00	1795.0
50.00	8584.0
51.00	33536.0
52.00	1766.0
55.00	717.0
56.00	1182.0
57.00	2813.0
58.00	694.0
59.10	478.0
60.90	367.0
62.00	446.0
63.00	1455.0
65.00	719.0
68.10	770.0
69.00	35456.0
70.00	331.0
71.00	553.0
73.00	886.0
74.00	3425.0
75.00	5476.0
76.00	1846.0
77.00	39992.0
78.10	3113.0
79.00	2713.0
80.00	1957.0
81.00	3002.0
82.00	1016.0
83.00	850.0
85.00	791.0
85.90	769.0
87.00	331.0
91.00	960.0
92.00	839.0
93.00	4195.0
94.00	438.0
95.00	318.0
96.00	514.0
97.00	427.0
98.00	2974.0
99.00	2712.0
101.00	1735.0
102.90	646.0
104.00	1230.0
105.00	1161.0
106.00	449.0
107.00	12189.0
108.00	2222.0
109.00	501.0
110.00	24128.0
111.00	3926.0
111.90	455.0
114.90	318.0
116.00	866.0
117.00	8233.0
118.00	708.0
122.00	839.0
123.00	1309.0
124.00	720.0
125.00	578.0
126.10	317.0
127.00	47720.0
128.00	3286.0
129.00	17560.0
129.90	1438.0
131.00	462.0
133.90	613.0
135.00	1465.0
136.00	585.0
137.00	792.0
141.00	2167.0
142.00	901.0
142.90	627.0
146.00	447.0
147.00	1616.0
148.00	2265.0
149.00	761.0
151.10	392.0
152.90	634.0
154.00	552.0
155.00	1301.0
156.00	1742.0
156.90	491.0
158.00	479.0
158.90	441.0
160.00	638.0
160.90	1061.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

161.80	389.0
164.90	899.0
166.00	669.0
167.00	4024.0
168.00	2045.0
169.00	440.0
171.90	407.0
172.90	553.0
174.00	965.0
175.00	1690.0
176.00	609.0
177.00	863.0
178.90	2983.0
180.00	2244.0
181.00	1103.0
185.00	1465.0
186.00	11550.0
187.00	3170.0
187.90	326.0
188.90	763.0
190.90	497.0
192.00	969.0
193.00	1323.0
194.00	405.0
196.00	2623.0
197.10	572.0
197.90	89544.0
199.00	6045.0
199.90	511.0
201.40	607.0
203.00	720.0
204.00	2739.0
205.00	5111.0
206.00	22544.0
207.00	3574.0
208.00	839.0
211.00	830.0
215.90	425.0
217.00	5503.0
218.00	812.0
221.00	4144.0
222.90	1263.0
224.00	11949.0
225.00	3005.0
225.90	375.0
227.00	4346.0
227.90	628.0
228.90	979.0
231.10	349.0
235.00	339.0
237.00	441.0
241.00	334.0
242.00	627.0
243.00	663.0
244.00	8899.0
245.00	1152.0
246.00	1543.0
246.90	313.0
248.90	300.0
254.10	435.0
255.00	47752.0
256.00	7170.0
257.00	614.0
258.00	2375.0
258.90	552.0
265.00	893.0
273.00	1401.0
274.00	3743.0
275.00	22016.0
276.00	2952.0
277.00	1596.0
278.00	308.0
285.00	309.0
292.90	339.0
296.00	5032.0
296.90	784.0
303.00	598.0
313.90	344.0
315.00	521.0
316.00	428.0
323.00	1863.0
334.00	1196.0
334.90	316.0
345.90	416.0
352.00	521.0
353.00	413.0
354.00	684.0
365.00	2106.0
365.90	450.0
372.00	1003.0
402.00	454.0
403.00	702.0
421.00	425.0
422.00	493.0
423.00	4008.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

423.90	891.0
441.00	9925.0
442.00	63488.0
443.00	12994.0
444.00	1257.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

m/z	Abundance
38.00	586.0
39.10	3510.0
40.00	544.0
41.00	738.0
43.10	953.0
44.00	1762.0
50.00	9750.0
51.00	39392.0
52.00	2116.0
55.10	678.0
56.00	1453.0
57.00	3300.0
58.00	893.0
59.00	408.0
61.00	537.0
62.00	558.0
63.00	1809.0
65.00	957.0
68.00	707.0
69.00	40864.0
70.00	360.0
71.10	566.0
73.10	964.0
74.00	4044.0
75.00	6674.0
76.00	2339.0
77.00	47984.0
78.00	3287.0
79.00	2951.0
80.00	2214.0
81.00	3580.0
82.00	1068.0
83.00	1143.0
85.00	842.0
86.00	965.0
87.00	521.0
91.00	1079.0
92.00	979.0
93.00	4929.0
94.00	455.0
95.10	371.0
95.90	488.0
96.90	435.0
98.00	3618.0
99.00	2895.0
100.00	392.0
100.90	1978.0
103.00	719.0
104.00	1458.0
105.00	1244.0
106.00	389.0
107.00	15236.0
108.00	2448.0
109.00	502.0
110.00	30104.0
111.00	3987.0
111.90	584.0
116.00	874.0
117.00	10142.0
117.90	844.0
120.00	307.0
122.00	1198.0
123.00	1751.0
124.00	682.0
125.00	669.0
126.20	408.0
127.00	59168.0
128.00	4375.0
129.00	21496.0
130.00	1969.0
131.00	390.0
132.00	307.0
134.00	719.0
135.00	1830.0
136.00	778.0
137.00	1002.0
140.00	310.0
141.00	2674.0
141.90	1054.0
142.90	784.0
145.90	492.0
147.00	2135.0
148.00	2789.0
148.90	833.0
150.90	367.0
152.00	333.0
153.00	968.0
154.00	678.0
155.00	1652.0
156.00	2563.0
157.00	585.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

158.00	610.0
159.00	411.0
160.00	840.0
161.00	1147.0
162.00	419.0
165.00	1003.0
166.00	947.0
167.00	5351.0
168.00	2619.0
169.00	532.0
171.00	359.0
172.00	536.0
173.00	701.0
174.00	1191.0
175.00	2387.0
175.90	745.0
177.00	1274.0
178.00	389.0
179.00	4077.0
180.00	2827.0
181.00	1241.0
184.00	370.0
184.90	2056.0
186.00	15748.0
187.00	4536.0
187.90	394.0
189.00	1019.0
191.00	522.0
192.00	1460.0
193.00	1633.0
194.00	504.0
196.00	3840.0
197.00	675.0
198.00	120560.0
199.00	7947.0
200.00	609.0
201.40	640.0
202.90	791.0
204.00	4164.0
205.00	6917.0
206.00	30960.0
207.00	4213.0
208.00	1142.0
209.00	396.0
210.20	492.0
211.00	1301.0
215.90	614.0
217.00	7158.0
218.00	983.0
220.10	304.0
221.00	5559.0
223.00	1653.0
224.00	15745.0
225.00	4131.0
226.00	452.0
227.00	6281.0
227.90	834.0
228.90	1351.0
231.00	546.0
234.00	506.0
235.00	474.0
235.90	307.0
237.00	594.0
242.00	791.0
243.00	922.0
244.00	13243.0
245.00	1808.0
246.00	2169.0
246.90	412.0
249.00	430.0
254.10	489.0
255.00	67040.0
256.00	10134.0
257.00	736.0
258.00	3345.0
259.00	593.0
265.00	1421.0
273.00	1687.0
274.00	5180.0
275.00	31256.0
276.00	3938.0
277.00	2362.0
278.00	362.0
283.00	328.0
285.00	419.0
293.00	465.0
296.00	7222.0
297.00	1202.0
303.00	942.0
303.90	306.0
314.10	370.0
314.90	753.0
316.00	518.0
323.00	2758.0
324.00	528.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

326.90	558.0
333.90	1682.0
335.00	468.0
340.90	419.0
345.90	695.0
352.00	721.0
353.00	571.0
354.00	856.0
365.00	3478.0
365.90	541.0
372.00	1541.0
373.00	408.0
382.90	422.0
402.00	583.0
403.00	1094.0
403.90	310.0
421.00	716.0
422.00	723.0
423.00	5457.0
424.10	1261.0
441.00	16009.0
442.00	108888.0
443.00	20720.0
444.10	2122.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

m/z	Abundance
38.00	391.0
39.10	2268.0
40.00	600.0
41.00	706.0
43.00	1093.0
44.00	1821.0
50.00	6846.0
51.00	27576.0
52.00	1334.0
55.00	660.0
56.00	927.0
57.00	2404.0
58.00	654.0
59.00	447.0
61.00	395.0
62.00	394.0
63.00	1120.0
65.00	701.0
68.10	572.0
69.00	29192.0
70.00	359.0
71.10	420.0
73.10	1377.0
74.00	2948.0
75.00	4708.0
76.10	1590.0
77.00	34768.0
78.00	2403.0
79.00	2399.0
80.00	1724.0
81.00	2550.0
82.00	736.0
83.00	700.0
85.00	665.0
86.00	595.0
86.90	352.0
91.00	801.0
92.00	802.0
93.00	3540.0
94.00	387.0
95.00	377.0
96.00	404.0
97.10	379.0
98.00	2665.0
99.00	2496.0
101.00	1483.0
103.00	552.0
104.00	1073.0
105.00	974.0
106.00	332.0
107.00	11002.0
108.00	1581.0
109.10	456.0
110.00	22080.0
111.00	3357.0
112.00	391.0
116.00	791.0
117.00	7599.0
118.00	592.0
121.90	876.0
123.00	1112.0
124.00	786.0
125.00	590.0
127.00	45064.0
128.00	3438.0
129.00	16130.0
129.90	1544.0
131.00	433.0
133.90	563.0
135.00	1438.0
136.00	653.0
136.90	685.0
141.00	2072.0
142.00	795.0
142.90	559.0
145.00	318.0
146.00	362.0
147.00	1662.0
148.00	2226.0
149.00	646.0
151.00	349.0
151.70	312.0
153.00	724.0
154.00	683.0
155.00	1290.0
156.00	1908.0
157.00	521.0
158.00	474.0
158.90	363.0
159.90	573.0
160.90	1051.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

162.00	330.0
165.00	955.0
166.00	673.0
167.00	4212.0
168.00	2049.0
169.00	355.0
171.90	377.0
172.90	588.0
174.00	1114.0
175.10	1714.0
176.00	666.0
177.00	854.0
178.00	353.0
179.00	3075.0
180.00	2265.0
180.90	1111.0
184.00	304.0
184.90	1438.0
186.00	12724.0
187.00	3542.0
187.90	452.0
188.90	810.0
191.10	458.0
192.00	1059.0
193.00	1544.0
194.10	315.0
196.00	2887.0
196.80	588.0
197.90	96528.0
199.00	6437.0
199.90	500.0
201.40	517.0
202.90	740.0
204.00	3101.0
205.00	5925.0
206.00	24272.0
207.00	3790.0
208.00	822.0
209.00	342.0
210.20	480.0
211.00	824.0
216.00	506.0
217.00	5885.0
218.00	804.0
221.00	4713.0
221.70	1056.0
223.00	1484.0
224.00	12753.0
225.00	3577.0
226.00	379.0
227.00	5288.0
228.00	670.0
228.90	1081.0
231.00	541.0
234.00	388.0
234.90	380.0
236.90	387.0
241.90	664.0
243.10	803.0
244.10	10821.0
245.00	1331.0
246.00	2003.0
246.90	384.0
248.90	463.0
254.10	494.0
255.00	56216.0
256.00	8009.0
257.00	599.0
258.00	2799.0
259.00	372.0
265.00	1199.0
273.00	1667.0
274.00	4322.0
275.00	26856.0
276.00	3490.0
277.00	1970.0
281.00	424.0
284.90	364.0
293.00	511.0
296.00	6234.0
297.00	859.0
303.00	965.0
314.00	310.0
315.00	739.0
316.00	451.0
323.00	2436.0
324.00	404.0
326.90	380.0
334.00	1733.0
335.00	472.0
341.00	450.0
346.00	625.0
352.10	807.0
353.00	577.0
354.00	777.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

365.00	3254.0
365.90	472.0
372.00	1505.0
373.00	446.0
383.00	481.0
402.00	612.0
403.00	957.0
404.00	381.0
421.00	768.0
422.00	684.0
423.00	5740.0
424.00	1244.0
441.00	16608.0
442.00	110064.0
443.00	21520.0
444.00	1903.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17