


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM092816\
 Data File : BM007593.D
 Acq On : 28 Sep 2016 08:44
 Operator : UM/SJ
 Sample : DFTPP66
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP66

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM020615.M
 Title : SVOA CALIBRATION
 Last Update : Sat Feb 07 02:55:14 2015



AutoFind: Scans 2549, 2550, 2551; Background Corrected with Scan 2544

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.4	118997	PASS
68	69	0.00	2	1.9	2605	PASS
69	198	0.00	100	35.7	135282	PASS
70	69	0.00	2	0.5	678	PASS
127	198	10	80	42.6	161496	PASS
197	198	0.00	2	0.5	1883	PASS
198	198	100	100	100.0	378794	PASS
199	198	5	9	6.9	26250	PASS
275	198	10	60	29.1	110397	PASS
365	198	1	100	4.1	15607	PASS
441	443	0.01	100	79.8	42642	PASS
442	198	50	100	74.0	280384	PASS
443	442	15	24	19.1	53434	PASS

m/z	Abundance
37.00	460.0
38.10	1247.0
39.10	7321.0
40.00	887.0
44.00	808.0
49.00	756.0
50.10	24808.0
51.10	96352.0
52.10	5147.0
55.00	676.0
56.00	2935.0
57.10	6105.0
58.00	456.0
61.00	1660.0
62.00	1534.0
63.00	4216.0
64.10	861.0
65.10	2500.0
68.10	1937.0
69.00	107928.0
70.10	731.0
73.00	1584.0
74.00	12317.0
75.10	19088.0
76.10	6297.0
77.10	120904.0
78.10	8597.0
79.00	9453.0
80.00	6196.0
81.00	9606.0
81.90	2591.0
83.00	2078.0
84.00	419.0
85.10	1826.0
86.00	2390.0
86.90	1014.0
88.00	345.0
88.90	330.0
91.00	2287.0
92.00	2875.0
93.00	17472.0
94.00	1067.0
94.90	354.0
96.00	817.0
97.10	399.0
98.00	12560.0
99.00	12459.0
100.00	948.0
101.00	5468.0
103.00	1753.0
104.00	3712.0
105.00	3145.0
106.00	1095.0
107.00	41360.0
108.00	6292.0
109.00	1395.0
110.00	67048.0
111.00	11593.0
112.00	1305.0
112.90	486.0
116.10	2268.0
117.00	38320.0
118.00	3203.0
118.90	355.0
120.00	306.0
122.00	2071.0
123.00	4693.0
123.90	1628.0
125.00	1945.0
126.10	574.0
127.00	119624.0
128.00	9333.0
129.00	52656.0
130.00	3990.0
131.00	836.0
132.90	591.0
133.90	1651.0
134.90	3944.0
136.00	1782.0
137.10	2537.0
138.00	472.0
139.90	630.0
141.00	7206.0
142.10	2186.0
143.00	1779.0
144.10	332.0
146.00	1164.0
147.00	4211.0
148.00	9566.0
149.00	2071.0
150.10	727.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

151.10	937.0
151.70	498.0
153.10	2331.0
153.90	1550.0
155.00	4124.0
156.00	5841.0
157.00	1179.0
157.90	1686.0
159.00	997.0
159.90	2141.0
161.00	3477.0
162.10	860.0
163.10	456.0
164.20	328.0
165.00	2720.0
166.00	2282.0
167.00	14100.0
168.00	9282.0
169.00	1364.0
170.00	410.0
171.00	946.0
172.00	1662.0
173.10	1299.0
174.00	3228.0
175.00	5613.0
176.00	1582.0
176.90	2502.0
177.90	1021.0
179.00	12093.0
180.00	8293.0
181.00	3524.0
182.20	445.0
184.00	1254.0
185.00	4922.0
186.00	40440.0
187.00	12210.0
188.00	1332.0
189.00	3199.0
189.90	386.0
191.00	1491.0
191.90	3527.0
193.00	4141.0
194.10	727.0
194.80	362.0
195.30	350.0
196.00	7150.0
196.80	1617.0
198.00	262528.0
199.00	18784.0
200.00	1424.0
200.90	316.0
201.50	907.0
202.90	2254.0
204.00	11062.0
205.00	18248.0
206.10	75896.0
207.10	13211.0
208.00	3264.0
209.00	1633.0
209.90	1206.0
211.00	3062.0
212.00	386.0
214.90	974.0
216.00	1509.0
217.00	20288.0
218.00	2960.0
218.90	300.0
221.00	8522.0
221.90	2360.0
223.10	4454.0
224.00	42928.0
225.00	10831.0
226.10	1487.0
227.00	20760.0
228.00	2361.0
229.00	3954.0
229.90	448.0
231.00	1481.0
233.00	483.0
234.00	1087.0
235.00	1386.0
236.00	598.0
237.00	2678.0
237.90	482.0
239.00	793.0
239.90	715.0
241.00	655.0
242.00	1821.0
243.10	1935.0
244.00	31320.0
245.00	4339.0
246.00	6623.0
247.00	1800.0
247.90	333.0
248.90	915.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

250.80	314.0
251.90	328.0
253.00	717.0
254.00	1251.0
255.00	167424.0
256.00	25152.0
257.00	2482.0
258.00	10601.0
259.10	1650.0
259.90	311.0
265.10	4171.0
265.90	663.0
267.10	356.0
270.90	446.0
271.90	450.0
273.00	4423.0
274.00	12314.0
275.00	70936.0
276.00	9028.0
277.00	6372.0
278.00	1839.0
279.10	322.0
281.00	922.0
282.00	323.0
282.90	727.0
284.00	820.0
285.00	939.0
293.00	1633.0
294.10	356.0
295.00	304.0
296.00	21528.0
297.00	3133.0
302.00	445.0
303.00	2178.0
304.10	580.0
308.00	307.0
309.90	317.0
314.00	763.0
315.00	1950.0
316.00	1572.0
320.90	420.0
323.10	5879.0
324.10	1187.0
326.90	1612.0
327.90	624.0
332.00	412.0
332.90	481.0
334.10	4572.0
335.00	870.0
341.00	747.0
346.00	1190.0
346.90	305.0
352.00	1969.0
353.00	1153.0
354.10	1966.0
355.00	537.0
365.00	9390.0
366.00	1195.0
371.00	525.0
372.00	2588.0
373.00	1020.0
383.00	664.0
389.90	323.0
402.00	1403.0
403.10	1572.0
404.10	494.0
421.00	1295.0
422.00	1144.0
423.10	8381.0
424.00	3176.0
425.10	577.0
441.10	23328.0
442.10	148736.0
443.00	29864.0
444.00	2677.0
444.90	318.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

m/z	Abundance
37.10	496.0
38.00	1486.0
39.10	10243.0
40.00	957.0
41.00	404.0
44.10	816.0
45.00	333.0
48.90	861.0
50.10	36352.0
51.10	137408.0
52.10	7248.0
53.20	441.0
55.00	731.0
56.00	4355.0
57.00	9489.0
58.00	517.0
61.00	1605.0
62.00	2572.0
63.00	6612.0
64.10	856.0
65.10	2864.0
68.00	2993.0
69.00	158528.0
70.00	579.0
73.10	2120.0
74.00	17600.0
75.00	27688.0
76.00	8471.0
77.10	180224.0
78.10	12130.0
79.00	12712.0
80.00	10351.0
81.00	13279.0
82.00	3159.0
83.00	2779.0
84.00	469.0
85.00	2784.0
86.00	3897.0
87.00	2142.0
88.00	693.0
88.90	300.0
91.00	2583.0
92.00	4031.0
93.00	25232.0
94.00	1465.0
95.00	422.0
96.00	890.0
97.10	333.0
98.00	19064.0
99.00	16248.0
100.10	1287.0
101.00	7752.0
101.90	503.0
103.00	2313.0
104.00	4995.0
105.00	4728.0
106.10	1361.0
107.00	63304.0
108.00	9611.0
109.00	1483.0
110.00	101952.0
111.00	17072.0
112.00	2240.0
113.10	613.0
114.90	321.0
116.00	3270.0
117.00	58272.0
118.00	4173.0
118.90	500.0
120.10	910.0
121.10	301.0
122.00	4533.0
123.00	6480.0
124.00	2731.0
125.00	2199.0
126.20	944.0
127.00	188288.0
128.00	15708.0
129.00	78608.0
130.00	7171.0
131.10	1353.0
132.00	799.0
132.80	583.0
134.00	2140.0
135.00	6326.0
136.00	2886.0
137.00	3852.0
138.10	511.0
139.10	417.0
140.10	1030.0
141.00	10275.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

142.00	3434.0
143.00	2542.0
144.00	643.0
145.00	696.0
146.00	1886.0
147.00	6352.0
148.00	13203.0
149.00	3159.0
150.00	1020.0
151.20	1275.0
151.90	609.0
153.00	3026.0
154.00	2576.0
155.10	5909.0
156.00	9243.0
157.10	1955.0
158.00	2264.0
159.00	1658.0
160.00	3174.0
161.00	4558.0
162.00	1663.0
163.10	501.0
164.00	622.0
165.00	4532.0
166.00	3490.0
167.00	21648.0
168.00	13682.0
169.00	2585.0
169.90	870.0
170.90	1165.0
171.90	2205.0
173.00	2514.0
174.00	5335.0
175.00	9338.0
176.00	1809.0
177.00	4603.0
178.00	1569.0
179.00	17624.0
180.00	13324.0
181.00	5938.0
182.00	1048.0
182.70	449.0
184.00	1519.0
185.10	7584.0
186.00	64968.0
187.00	18368.0
188.00	2661.0
189.00	4926.0
189.90	763.0
191.00	2262.0
192.00	5717.0
193.00	6934.0
193.90	1489.0
194.90	757.0
196.00	10937.0
196.90	2236.0
198.00	448128.0
199.00	29816.0
200.00	2526.0
201.40	1588.0
203.00	3187.0
204.00	17352.0
205.00	30376.0
206.00	123856.0
207.10	18032.0
208.00	5041.0
209.00	1829.0
210.10	1936.0
211.00	4752.0
211.90	968.0
213.00	386.0
215.00	1578.0
216.10	2776.0
217.00	34096.0
218.00	4459.0
219.10	578.0
220.20	717.0
221.10	15059.0
221.90	4131.0
223.00	7441.0
224.10	70424.0
225.00	18264.0
226.10	2397.0
227.00	33368.0
228.00	4532.0
229.00	6690.0
229.90	632.0
231.00	3300.0
232.80	428.0
234.00	1663.0
235.00	2759.0
236.00	1648.0
237.00	4642.0
238.00	327.0
239.10	1502.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

240.10	1132.0
241.00	1777.0
242.00	4025.0
243.10	4332.0
244.10	52616.0
245.00	7034.0
246.00	12516.0
247.00	2550.0
248.10	539.0
249.00	2373.0
249.90	359.0
250.90	372.0
251.80	412.0
252.90	950.0
254.00	1590.0
255.00	292928.0
256.00	44496.0
257.00	4056.0
258.00	19624.0
259.00	2707.0
260.00	544.0
260.90	698.0
263.00	336.0
264.00	624.0
265.00	7322.0
266.00	1352.0
271.00	542.0
272.00	1013.0
273.00	8948.0
274.00	21760.0
275.00	125344.0
276.00	16648.0
277.00	12683.0
278.00	2455.0
278.90	363.0
281.10	889.0
282.10	507.0
283.00	1625.0
284.00	1189.0
285.10	2037.0
285.90	519.0
288.90	426.0
289.90	591.0
292.00	593.0
292.90	2641.0
294.10	991.0
295.10	765.0
296.00	41024.0
297.00	6121.0
298.10	590.0
301.00	583.0
302.00	595.0
303.00	3896.0
304.00	1192.0
305.00	419.0
307.90	489.0
309.00	412.0
310.10	430.0
312.90	352.0
314.00	1927.0
315.00	3717.0
316.00	2285.0
316.90	366.0
321.00	1384.0
321.90	553.0
323.10	11377.0
324.00	2574.0
327.00	1809.0
328.10	1155.0
329.00	324.0
331.90	679.0
333.00	1214.0
334.00	7426.0
335.00	2017.0
341.00	1345.0
346.00	2356.0
347.00	452.0
352.10	3483.0
353.00	2624.0
354.10	3121.0
355.10	788.0
365.00	16592.0
365.90	2312.0
370.00	398.0
370.90	793.0
372.10	5566.0
373.10	1486.0
383.00	1363.0
384.10	406.0
389.90	511.0
391.00	422.0
391.90	388.0
402.00	2336.0
403.00	3160.0
404.00	1303.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

404.90	333.0
421.10	2469.0
422.00	2377.0
423.00	17056.0
424.00	5487.0
425.00	675.0
441.10	46896.0
442.10	313472.0
443.10	58384.0
444.00	5148.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

m/z	Abundance
36.90	337.0
38.00	1509.0
39.10	8734.0
40.10	863.0
44.00	898.0
49.00	1046.0
50.10	31776.0
51.10	123232.0
52.10	5758.0
54.90	665.0
56.00	3463.0
57.00	8135.0
57.90	372.0
61.10	1453.0
62.10	2210.0
63.00	5799.0
64.00	828.0
65.00	2519.0
65.80	400.0
68.10	2887.0
69.00	139392.0
70.10	724.0
73.10	1803.0
74.00	15473.0
75.00	23824.0
76.10	7974.0
77.10	159808.0
78.10	11207.0
79.00	12301.0
80.00	9284.0
81.00	11654.0
82.00	3508.0
83.00	2999.0
85.00	2093.0
86.00	3621.0
87.00	1713.0
88.00	597.0
89.00	307.0
91.00	2786.0
92.00	3003.0
93.00	21472.0
94.00	1820.0
95.90	980.0
97.30	603.0
98.00	17896.0
99.00	13425.0
100.00	940.0
101.00	7406.0
101.80	473.0
103.00	2612.0
104.00	4482.0
105.10	3388.0
106.00	1517.0
107.00	57368.0
108.10	7981.0
109.00	1539.0
110.00	90816.0
111.00	14886.0
112.00	1675.0
112.90	473.0
116.10	2511.0
117.00	55008.0
118.00	4237.0
118.90	470.0
119.90	569.0
122.00	3441.0
123.00	5282.0
124.00	2699.0
125.10	2460.0
126.20	592.0
127.00	176576.0
128.00	13160.0
129.00	71064.0
130.00	5909.0
131.00	1073.0
132.00	747.0
132.80	554.0
134.00	2743.0
135.00	5667.0
136.00	2707.0
137.10	3676.0
138.00	742.0
138.80	475.0
140.00	1287.0
141.00	10399.0
142.00	3063.0
143.10	2075.0
144.00	525.0
145.10	589.0
146.00	2058.0
147.00	5513.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

148.00	13547.0
149.10	3125.0
150.00	859.0
151.20	1490.0
151.90	682.0
153.00	3523.0
154.00	2507.0
155.00	6061.0
156.00	8458.0
157.00	2126.0
157.90	1960.0
158.90	1251.0
160.00	3571.0
161.00	4839.0
162.00	1259.0
163.00	315.0
164.00	761.0
165.00	4821.0
166.00	2937.0
167.00	24024.0
168.00	12478.0
169.10	2112.0
170.10	891.0
171.00	911.0
172.10	2030.0
173.00	2770.0
174.00	4795.0
175.00	9070.0
176.00	2433.0
177.00	4048.0
178.00	1288.0
179.00	16960.0
180.00	12457.0
181.00	5043.0
182.00	671.0
183.00	522.0
184.00	1666.0
185.00	7808.0
186.00	64304.0
187.00	18112.0
188.00	2031.0
189.00	4484.0
190.00	691.0
191.00	1828.0
192.00	5553.0
193.00	6813.0
194.00	1101.0
195.00	612.0
196.00	11279.0
197.00	1798.0
198.00	425728.0
199.00	30152.0
200.00	2193.0
201.40	2104.0
203.00	2915.0
204.00	16936.0
205.00	28488.0
206.00	121616.0
207.10	16552.0
208.00	5175.0
209.10	1730.0
210.00	1675.0
211.00	5532.0
211.90	1009.0
214.90	1546.0
216.10	2868.0
217.00	35408.0
218.00	4766.0
220.20	772.0
221.00	15176.0
222.00	4465.0
223.00	7507.0
224.00	69840.0
225.00	17032.0
226.10	1963.0
227.00	34144.0
228.00	5249.0
229.00	6249.0
229.90	719.0
231.10	2541.0
232.90	517.0
234.00	2052.0
235.00	2286.0
236.00	1980.0
237.10	3105.0
237.90	570.0
239.00	1278.0
240.00	1370.0
241.00	2004.0
242.10	4070.0
243.00	3665.0
244.10	54032.0
245.10	6710.0
246.00	12100.0
247.00	2391.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

247.90	599.0
249.00	1925.0
250.10	400.0
250.90	467.0
252.00	537.0
253.00	1372.0
254.10	1794.0
255.00	296768.0
256.00	41768.0
257.00	3657.0
258.00	20160.0
259.00	3430.0
259.90	498.0
261.00	458.0
264.00	551.0
265.00	7569.0
266.00	1447.0
267.10	394.0
271.00	798.0
271.90	918.0
273.00	8782.0
274.00	24224.0
275.00	134912.0
276.00	17744.0
277.00	12767.0
278.00	2450.0
278.90	612.0
281.00	625.0
281.90	560.0
283.00	1433.0
284.00	961.0
285.00	2460.0
286.00	509.0
288.90	471.0
289.90	393.0
292.00	464.0
293.00	2504.0
294.00	824.0
295.00	959.0
296.00	44688.0
297.00	6488.0
298.10	552.0
301.10	530.0
302.00	753.0
303.00	4594.0
304.00	1129.0
305.00	319.0
307.80	545.0
309.10	553.0
310.00	712.0
313.00	318.0
314.00	2197.0
315.00	5300.0
316.00	2838.0
321.00	1416.0
322.00	499.0
323.10	11965.0
324.00	2295.0
326.00	317.0
327.00	3015.0
328.00	1098.0
331.90	808.0
333.10	1292.0
334.00	8354.0
335.00	2065.0
339.00	426.0
341.10	1440.0
346.00	2301.0
347.00	744.0
352.00	3838.0
353.00	2662.0
354.10	4254.0
355.00	892.0
359.10	333.0
365.00	20840.0
366.00	2243.0
371.10	1013.0
372.10	7124.0
373.10	1650.0
383.00	1590.0
390.00	1177.0
391.00	516.0
392.10	506.0
401.10	303.0
402.00	2934.0
403.00	3381.0
404.00	1678.0
421.00	2659.0
422.10	2797.0
423.10	19568.0
424.10	5658.0
425.00	966.0
441.10	57704.0
442.10	378944.0
443.10	72056.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

444.006538.0

444.90523.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66

Scan 2544 (17.551 min): BM007593.D
DFTPP66

m/z	Abundance
39.90	625.0
44.00	986.0
73.10	717.0
147.00	407.0
207.00	2449.0
208.10	366.0
281.00	723.0
282.10	381.0

Instrument :
BNA_M
ClientSampleId :
DFTPP66