

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092820\
 Data File : BM027561.D
 Acq On : 28 Sep 2020 12:45
 Operator : JU/CG
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02015

Quant Time: Sep 28 13:25:16 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM092420MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 24 20:21:51 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	41692	20.00	ng/ul	0.00
18) Naphthalene-d8	10.592	136	144634	20.00	ng/ul	# 0.00
36) Acenaphthene-d10	14.433	164	122464	20.00	ng/ul	-0.01
62) Phenanthrene-d10	17.180	188	307045	20.00	ng/ul	# 0.00
78) Chrysene-d12	21.362	240	339691	20.00	ng/ul	# 0.00
86) Perylene-d12	23.633	264	356106	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.311	96	5168	7.05	ng/uL	0.00
5) Phenol-d5	6.969	99	42763	15.74	ng/ul	0.00
7) Bis-(2-Chloroethyl)eth...	7.140	67	21310	13.34	ng/ul	0.00
9) 2-Chlorophenol-d4	7.340	132	43140	18.34	ng/ul	0.00
13) 4-Methylphenol-d8	8.504	113	39773	17.45	ng/ul	0.00
19) Nitrobenzene-d5	8.957	128	20594	19.32	ng/ul	0.00
22) 2-Nitrophenol-d4	9.675	143	26330	23.57	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.210	165	67321	23.62	ng/ul	0.00
29) 4-Chloroaniline-d4	10.722	131	55356	19.60	ng/ul	0.00
44) Dimethylphthalate-d6	13.851	166	193489	20.42	ng/ul	0.00
47) Acenaphthylene-d8	14.128	160	220824	19.46	ng/ul	0.00
52) 4-Nitrophenol-d4	14.628	143	23356	17.59	ng/ul	0.00
58) Fluorene-d10	15.427	176	176009	20.65	ng/ul	0.00
63) 4,6-Dinitro-2-methylph...	15.539	200	36016	21.66	ng/ul	0.00
71) Anthracene-d10	17.274	188	285195	19.69	ng/ul	0.00
79) Pyrene-d10	19.568	212	338436	19.10	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.486	264	360435	19.05	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.346	88	5439	6.29	ng/uL#	83
4) Benzaldehyde	6.946	77	30684	16.65	ng/ul	86
6) Phenol	6.999	94	45466	16.25	ng/ul	95
8) Bis(2-Chloroethyl)ether	7.234	93	31520	14.12	ng/ul	89
10) 2-Chlorophenol	7.369	128	44392	18.37	ng/ul#	85
11) 2-Methylphenol	8.240	108	37446	16.97	ng/ul	98
12) 2,2'-oxybis(1-Chloropr...	8.340	45	17008	9.80	ng/ul#	88
14) Acetophenone	8.622	105	67868	18.34	ng/ul	84
15) N-Nitroso-di-n-propyla...	8.610	70	29497	15.97	ng/ul#	83
16) 4-Methylphenol	8.569	108	40406	17.22	ng/ul	98
17) Hexachloroethane	8.881	117	21405	18.55	ng/ul#	25
20) Nitrobenzene	8.998	77	49127	16.84	ng/ul#	82
21) Isophorone	9.528	82	82711	16.47	ng/ul#	86
23) 2-Nitrophenol	9.704	139	28581	23.30	ng/ul#	84
24) 2,4-Dimethylphenol	9.769	107	56806	20.86	ng/ul	98
25) Bis(2-Chloroethoxy)met...	10.010	93	44187	15.37	ng/ul#	96
27) 2,4-Dichlorophenol	10.240	162	65819	23.78	ng/ul#	88
28) Naphthalene	10.640	128	141902	19.21	ng/ul	98
30) 4-Chloroaniline	10.745	127	56228	20.23	ng/ul	96
31) Hexachlorobutadiene	10.934	225	73490	26.22	ng/ul	99
32) Caprolactam	11.504	113	10088	17.44	ng/ul#	53
33) 4-Chloro-3-methylphenol	11.869	107	45556	19.97	ng/ul#	83
34) 2-Methylnaphthalene	12.257	142	121679	21.69	ng/ul	97
35) 1-Methylnaphthalene	12.475	142	112443	20.49	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM092820\
 Data File : BM027561.D
 Acq On : 28 Sep 2020 12:45
 Operator : JU/CG
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02015


Quant Time: Sep 28 13:25:16 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM092420MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 24 20:21:51 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1,2,4,5-Tetrachloroben...	12.622	216	130963	23.55	ng/ul#	95
38) Hexachlorocyclopentadiene	12.610	237	92750	24.94	ng/ul	99
39) 2,4,6-Trichlorophenol	12.863	196	69947	23.25	ng/ul	99
40) 2,4,5-Trichlorophenol	12.934	196	76989	24.04	ng/ul	98
41) 1,1'-Biphenyl	13.269	154	182897	19.58	ng/ul#	96
42) 2-Chloronaphthalene	13.310	162	155192	19.89	ng/ul	93
43) 2-Nitroaniline	13.510	65	26004	16.60	ng/ul#	71
45) Dimethylphthalate	13.898	163	195651	20.23	ng/ul#	98
46) 2,6-Dinitrotoluene	14.010	165	38692	21.06	ng/ul#	80
48) Acenaphthylene	14.157	152	204215	18.93	ng/ul	96
49) 3-Nitroaniline	14.333	138	27751	19.06	ng/ul#	71
50) Acenaphthene	14.504	153	146158	19.07	ng/ul	98
51) 2,4-Dinitrophenol	14.539	184	24037	22.78	ng/ul#	61
53) 4-Nitrophenol	14.639	109	32394	22.05	ng/ul#	59
54) Dibenzofuran	14.833	168	240201	20.84	ng/ul	94
55) 2,4-Dinitrotoluene	14.792	165	53588	21.36	ng/ul#	82
56) 2,3,4,6-Tetrachlorophenol	15.057	232	71695	25.71	ng/ul#	76
57) Diethylphthalate	15.263	149	171837	19.37	ng/ul#	92
59) Fluorene	15.486	166	191279	20.42	ng/ul	99
60) 4-Chlorophenyl-phenyle...	15.480	204	138035	23.44	ng/ul#	87
61) 4-Nitroaniline	15.498	138	28902	18.73	ng/ul#	69
64) 4,6-Dinitro-2-methylph...	15.557	198	37481	20.59	ng/ul#	71
65) N-Nitrosodiphenylamine	15.692	169	169664	19.88	ng/ul	97
66) 4-Bromophenyl-phenylether	16.374	248	92686	22.65	ng/ul#	88
67) Hexachlorobenzene	16.486	284	98848	21.43	ng/ul#	89
68) Atrazine	16.645	200	79274	21.27	ng/ul#	94
69) Pentachlorophenol	16.827	266	53048	21.06	ng/ul	98
70) Phenanthrene	17.221	178	317596	19.14	ng/ul	99
72) Anthracene	17.310	178	325020	19.27	ng/ul	99
73) 1,2,3,4-Tetrachloroben...	13.233	216	127602	22.54	ng/uL	97
74) Pentachlorobenzene	14.757	250	123854	22.47	ng/uL	96
75) Carbazole	17.580	167	246887	18.44	ng/ul	97
76) Di-n-butylphthalate	18.151	149	261567	18.07	ng/ul	100
77) Fluoranthene	19.233	202	419718	19.83	ng/ul#	86
80) Pyrene	19.598	202	422441	18.27	ng/ul#	84
81) Butylbenzylphthalate	20.504	149	102911	19.06	ng/ul#	80
82) 3,3'-Dichlorobenzidine	21.280	252	155418	21.36	ng/ul#	97
83) Benzo(a)anthracene	21.345	228	442500	19.58	ng/ul	99
84) Bis(2-ethylhexyl)phtha...	21.286	149	151918	18.30	ng/ul#	91
85) Chrysene	21.398	228	424055	19.40	ng/ul	99
87) Di-n-octyl phthalate	22.180	149	247112	15.86	ng/ul	100
88) Benzo(b)fluoranthene	22.945	252	430750	18.26	ng/ul#	97
89) Benzo(k)fluoranthene	22.992	252	444584	18.88	ng/ul#	96
91) Benzo(a)pyrene	23.533	252	394838	18.11	ng/ul#	95
92) Indeno(1,2,3-cd)pyrene	25.927	276	514777	18.80	ng/ul#	90
93) Dibenzo(a,h)anthracene	25.939	278	433447	18.81	ng/ul#	95
94) Benzo(g,h,i)perylene	26.621	276	423829	18.70	ng/ul#	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 28 13:25:16 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM092420MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 24 20:21:51 2020
Response via : Initial Calibration









1,4-Dioxane
 Concen: 6.29 ng/uL
 RT: 3.346 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BM027561.D
 Acq: 28 Sep 2020 12:45




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02015


Tgt Ion:	88	Resp:	5439
Ion	Ratio	Lower	Upper
88	100		
43	19.9	21.7	32.5#
58	45.0	47.4	71.0#



Benzaldehyde
 Concen: 16.65 ng/uL
 RT: 6.946 min Scan# 673
 Delta R.T. -0.006 min
 Lab File: BM027561.D
 Acq: 28 Sep 2020 12:45

Tgt Ion:	77	Resp:	30684
Ion	Ratio	Lower	Upper
77	100		
105	94.1	64.9	97.3
106	90.3	62.8	94.2









Phenol
 Concen: 16.25 ng/ul
 RT: 6.999 min Scan# 682
 Delta R.T. 0.000 min
 Lab File: BM027561.D
 Acq: 28 Sep 2020 12:45




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02015


Tgt Ion:	94	Resp:	45466
Ion	Ratio	Lower	Upper
94	100		
65	32.7	30.2	45.4
66	54.6	42.4	63.6



Bis(2-Chloroethyl)ether
 Concen: 14.12 ng/ul
 RT: 7.234 min Scan# 722
 Delta R.T. 0.000 min
 Lab File: BM027561.D
 Acq: 28 Sep 2020 12:45

Tgt Ion:	93	Resp:	31520
Ion	Ratio	Lower	Upper
93	100		
63	66.3	60.6	90.8
95	36.9	25.0	37.4









#10
2-Chlorophenol
Concen: 18.37 ng/ul
RT: 7.369 min Scan# 745
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45




Instrument :
BNA_M
ClientSampleId :
SSTD02015

Tgt Ion:	128	Resp:	44392
Ion	Ratio	Lower	Upper
128	100		
64	35.2	41.1	61.7#
130	35.4	26.8	40.2




#11
2-Methylphenol
Concen: 16.97 ng/ul
RT: 8.240 min Scan# 893
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:	108	Resp:	37446
Ion	Ratio	Lower	Upper
108	100		
107	91.5	74.4	111.6






Abundance Scan 909 (8.334 min): BM027544.D\data.ms (-9027) (-)




2,2'-oxybis(1-chloropropane)
Concen: 9.80 ng/ul
RT: 8.340 min Scan# 910
Delta R.T. 0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Instrument :
BNA_M
ClientSampleID :
SSTD02015

Tgt Ion:	45	Resp:	17008
Ion	Ratio	Lower	Upper
45	100		
77	37.7	24.6	36.8#
79	26.1	16.8	25.2#







Abundance Scan 958 (8.622 min): BM027544.D\data.ms (-9527) (-)



Acetophenone
Concen: 18.34 ng/ul
RT: 8.622 min Scan# 958
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:	105	Resp:	67868
Ion	Ratio	Lower	Upper
105	100		
77	75.1	74.4	111.6
51	21.9	19.2	28.8









N-Nitroso-di-n-propylamine
Concen: 15.97 ng/ul
RT: 8.610 min Scan# 956
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45




Instrument :
BNA_M
ClientSampleId :
SSTD02015

Tgt Ion:	70	Resp:	29497
Ion	Ratio	Lower	Upper
70	100		
42	50.2	31.0	46.4#
101	9.0	7.0	10.6
130	27.0	13.6	20.4#




4-Methylphenol
Concen: 17.22 ng/ul
RT: 8.569 min Scan# 949
Delta R.T. 0.000 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:	108	Resp:	40406
Ion	Ratio	Lower	Upper
108	100		
107	113.3	92.1	138.1






Abundance Scan 1003 (8.887 min): BM027544.D\data.ms (-996) (-)




Hexachloroethane
Concen: 18.55 ng/ul
RT: 8.881 min Scan# 1002
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Instrument :
BNA_M
ClientSampleId :
SSTD02015

Tgt Ion:	117	Resp:	21405
Ion	Ratio	Lower	Upper
117	100		
201	205.0	90.6	135.8#
199	117.4	56.3	84.5#






Abundance Scan 1022 (8.998 min): BM027544.D\data.ms (-1012) (-)




Nitrobenzene
Concen: 16.84 ng/ul
RT: 8.998 min Scan# 1022
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:	77	Resp:	49127
Ion	Ratio	Lower	Upper
77	100		
123	47.8	27.2	40.8#
65	16.8	11.8	17.6






Abundance Scan 1111 (9.522 min): BM027544.D\data.ms (-1111) (-)




Isophorone
Concen: 16.47 ng/ul
RT: 9.528 min Scan# 1112
Delta R.T. 0.000 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Instrument :
BNA_M
ClientSampleId :
SSTD02015

Tgt Ion:	82	Resp:	82711
Ion	Ratio	Lower	Upper
82	100		
95	12.6	6.5	9.7#
138	21.6	12.6	18.8#






Abundance Scan 1142 (9.704 min): BM027544.D\data.ms (-1142) (-)



2-Nitrophenol
Concen: 23.30 ng/ul
RT: 9.704 min Scan# 1142
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45


Tgt Ion:	139	Resp:	28581
Ion	Ratio	Lower	Upper
139	100		
65	46.8	50.2	75.4#
109	43.5	31.2	46.8





Abundance Scan 1153 (9.769 min): BM027544.D\data.ms (-1184) (-)

2,4-Dimethylphenol
Concen: 20.86 ng/ul
RT: 9.769 min Scan# 1153
Delta R.T. 0.000 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45


Instrument :
BNA_M
ClientSampleId :
SSTD02015



Tgt Ion:	107	Resp:	56806
Ion	Ratio	Lower	Upper
107	100		
121	43.2	36.6	54.8
122	75.2	59.9	89.9





Abundance Scan 1153 (9.769 min): BM027561.D\data.ms (-1119) (-)




Abundance Scan 1193 (10.004 min): BM027544.D\data.ms (-1185) (-)

Bis(2-Chloroethoxy)methane
Concen: 15.37 ng/ul
RT: 10.010 min Scan# 1194
Delta R.T. 0.000 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:	93	Resp:	44187
Ion	Ratio	Lower	Upper
93	100		
95	34.3	26.2	39.4
123	14.1	8.6	13.0#




Abundance Scan 1194 (10.010 min): BM027561.D\data.ms (-1160) (-)



Abundance Scan 1233 (10.239 min): BM027544.D\data.ms (-125) (-)


2,4-Dichlorophenol
Concen: 23.78 ng/ul
RT: 10.240 min Scan# 1233
Delta R.T. 0.000 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Instrument :
BNA_M
ClientSampleId :
SSTD02015




m/z-->

Tgt Ion:	162	Resp:	65819
Ion	Ratio	Lower	Upper
162	100		
164	69.3	51.1	76.7
98	21.8	27.7	41.5#




m/z-->

Abundance Scan 1233 (10.240 min): BM027561.D\data.ms (-1199) (-)




m/z-->




Abundance Scan 1301 (10.639 min): BM027544.D\data.ms (-129) (-)

Naphthalene
Concen: 19.21 ng/ul
RT: 10.640 min Scan# 1301
Delta R.T. -0.012 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45




m/z-->

Tgt Ion:	128	Resp:	141902
Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.4	12.6
127	13.7	10.3	15.5




m/z-->

Abundance Scan 1301 (10.640 min): BM027561.D\data.ms (-1269) (-)




m/z-->






Abundance Scan 1319 (10.745 min): BM027544.D\data.ms (-1319) (-)

4-Chloroaniline
Concen: 20.23 ng/ul
RT: 10.745 min Scan# 1319
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Instrument :
BNA_M
ClientSampleId :
SSTD02015




Tgt Ion:127 Resp: 56228
Ion Ratio Lower Upper
127 100
129 30.0 25.9 38.9







Abundance Scan 1351 (10.934 min): BM027544.D\data.ms (-1351) (-)

Hexachlorobutadiene
Concen: 26.22 ng/ul
RT: 10.934 min Scan# 1351
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45



Tgt Ion:225 Resp: 73490
Ion Ratio Lower Upper
225 100
223 61.6 49.8 74.6
227 63.2 49.4 74.2









Caprolactam
Concen: 17.44 ng/ul
RT: 11.504 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45




Instrument :
BNA_M
ClientSampleId :
SSTD02015

Tgt Ion:	113	Resp:	10088
Ion	Ratio	Lower	Upper
113	100		
55	64.3	100.5	150.7#
56	65.2	83.7	125.5#



4-Chloro-3-methylphenol
Concen: 19.97 ng/ul
RT: 11.869 min Scan# 1510
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:	107	Resp:	45556
Ion	Ratio	Lower	Upper
107	100		
144	30.2	19.7	29.5#
142	93.5	62.0	93.0#



Abundance Scan 1575 (12.251 min): BM027544.D\data.ms (-1534) (-)

2-Methylnaphthalene

Concen: 21.69 ng/ul

RT: 12.257 min Scan# 1576

Delta R.T. -0.006 min

Lab File: BM027561.D



Acq: 28 Sep 2020 12:45

Tgt Ion:142 Resp: 121679



Ion Ratio Lower Upper

142 100

141 88.2 68.2 102.4



Abundance Scan 1576 (12.257 min): BM027561.D\data.ms (-1543) (-)



Abundance Scan 1613 (12.475 min): BM027544.D\data.ms (-1535) (-)

1-Methylnaphthalene

Concen: 20.49 ng/ul

RT: 12.475 min Scan# 1613

Delta R.T. -0.006 min

Lab File: BM027561.D

Acq: 28 Sep 2020 12:45



Tgt Ion:142 Resp: 112443

Ion Ratio Lower Upper



142 100

141 91.1 74.6 111.8

116 5.2 3.6 5.4




Abundance Scan 1613 (12.475 min): BM027561.D\data.ms (-1580) (-)






Abundance Scan 1638 (12.622 min): BM027544.D\data.ms (-1631) (-)

1,2,4,5-Tetrachlorobenzene
Concen: 23.55 ng/ul
RT: 12.622 min Scan# 1638
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Instrument :
BNA_M
ClientSampleID :
SSTD02015




Tgt Ion:	216	Resp:	130963
Ion Ratio		Lower	Upper
216	100		
214	79.4	62.2	93.2
179	17.9	17.0	25.6
108	10.2	14.2	21.4#






Abundance Scan 1636 (12.610 min): BM027544.D\data.ms (-1636) (-)

Hexachlorocyclopentadiene
Concen: 24.94 ng/ul
RT: 12.610 min Scan# 1636
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45




Tgt Ion:	237	Resp:	92750
Ion Ratio		Lower	Upper
237	100		
235	61.5	49.4	74.2
272	13.8	10.1	15.1





Abundance Scan 1679 (12.863 min): BM027544.D\data.ms (-1639) (-)

2,4,6-Trichlorophenol
Concen: 23.25 ng/ul
RT: 12.863 min Scan# 1679
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45


Instrument :
BNA_M
ClientSampleId :
SSTD02015



Tgt Ion:	196	Resp:	69947
Ion	Ratio	Lower	Upper
196	100		
198	99.0	79.4	119.2
200	32.3	24.6	37.0



Abundance Scan 1679 (12.863 min): BM027561.D\data.ms (-1646) (-)






Abundance Scan 1691 (12.933 min): BM027544.D\data.ms (-1646) (-)

2,4,5-Trichlorophenol
Concen: 24.04 ng/ul
RT: 12.934 min Scan# 1691
Delta R.T. 0.000 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:	196	Resp:	76989
Ion	Ratio	Lower	Upper
196	100		
198	100.4	78.0	117.0
200	32.6	25.8	38.6




Abundance Scan 1691 (12.934 min): BM027561.D\data.ms (-1674) (-)




Abundance Scan 1748 (13.269 min): BM027544.D\data.ms (-1744) (-)

1,1'-Biphenyl
Concen: 19.58 ng/ul
RT: 13.269 min Scan# 1748
Delta R.T. -0.012 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Instrument :
BNA_M
ClientSampleId :
SSTD02015






Tgt Ion:	154	Resp:	182897
Ion	Ratio	Lower	Upper
154	100		
153	40.0	32.0	48.0
76	6.3	9.8	14.8#




Abundance Scan 1755 (13.310 min): BM027544.D\data.ms (-1748) (-)

2-Chloronaphthalene
Concen: 19.89 ng/ul
RT: 13.310 min Scan# 1755
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45




Tgt Ion:	162	Resp:	155192
Ion	Ratio	Lower	Upper
162	100		
164	33.0	27.1	40.7
127	30.1	29.6	44.4





Abundance Scan 1789 (13.510 min): BM027544.D\data.ms (-1742) (-)

2-Nitroaniline
Concen: 16.60 ng/ul
RT: 13.510 min Scan# 1789
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45


Instrument :
BNA_M
ClientSampleId :
SSTD02015



Tgt Ion:	65	Resp:	26004
Ion	Ratio	Lower	Upper
65	100		
92	72.2	49.9	74.9
138	126.0	70.1	105.1#




Abundance Scan 1789 (13.510 min): BM027561.D\data.ms (-1756) (-)





Abundance Scan 1855 (13.898 min): BM027544.D\data.ms (-1845) (-)


Dimethylphthalate
Concen: 20.23 ng/ul
RT: 13.898 min Scan# 1855
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45




Tgt Ion:	163	Resp:	195651
Ion	Ratio	Lower	Upper
163	100		
194	5.4	3.3	4.9#
164	9.5	8.1	12.1



Abundance Scan 1855 (13.898 min): BM027561.D\data.ms (-1822) (-)









2,6-Dinitrotoluene
Concen: 21.06 ng/ul
RT: 14.010 min Scan# 1874
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45




Instrument :
BNA_M
ClientSampleId :
SSTD02015

Tgt Ion:	165	Resp:	38692
Ion	Ratio	Lower	Upper
165	100		
89	53.1	56.6	84.8#
121	15.2	18.0	27.0#




Acenaphthylene
Concen: 18.93 ng/ul
RT: 14.157 min Scan# 1899
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:	152	Resp:	204215
Ion	Ratio	Lower	Upper
152	100		
151	20.6	14.9	22.3
153	13.3	9.8	14.8






Abundance Scan 1929 (14.333 min): BM027544.D\data.ms (-1929) (-)




3-Nitroaniline
Concen: 19.06 ng/ul
RT: 14.333 min Scan# 1929
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Instrument :
BNA_M
ClientSampleId :
SSTD02015

Tgt Ion:	138	Resp:	27751
Ion	Ratio	Lower	Upper
138	100		
108	11.1	8.9	13.3
92	93.3	104.4	156.6#







Abundance Scan 1958 (14.504 min): BM027544.D\data.ms (-1958) (-)



Acenaphthene
Concen: 19.07 ng/ul
RT: 14.504 min Scan# 1958
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:	153	Resp:	146158
Ion	Ratio	Lower	Upper
153	100		
152	46.6	38.1	57.1
154	89.0	72.5	108.7









2,4-Dinitrophenol
Concen: 22.78 ng/ul
RT: 14.539 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45




Instrument :
BNA_M
ClientSampleId :
SSTD02015

Tgt Ion:	184	Resp:	24037
Ion	Ratio	Lower	Upper
184	100		
63	33.8	66.4	99.6#
154	54.3	56.8	85.2#



4-Nitrophenol
Concen: 22.05 ng/ul
RT: 14.639 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45


Tgt Ion:	109	Resp:	32394
Ion	Ratio	Lower	Upper
109	100		
139	74.8	80.8	121.2#
65	63.1	98.6	147.8#






Abundance Scan 2014 (14.833 min): BM027544.D\data.ms (-2054) (-)

Dibenzofuran
Concen: 20.84 ng/ul
RT: 14.833 min Scan# 2014
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Instrument :
BNA_M
ClientSampleId :
SSTD02015




Tgt Ion:168 Resp: 240201
Ion Ratio Lower Upper
168 100
139 34.1 30.3 45.5






Abundance Scan 2007 (14.792 min): BM027544.D\data.ms (-2055) (-)

2,4-Dinitrotoluene
Concen: 21.36 ng/ul
RT: 14.792 min Scan# 2007
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45




Tgt Ion:165 Resp: 53588
Ion Ratio Lower Upper
165 100
63 37.8 41.0 61.6#
182 3.2 2.3 3.5






Abundance Scan 2053 (15.063 min): BM027544.D\data.ms (-2055) (-)

2,3,4,6-Tetrachlorophenol
Concen: 25.71 ng/ul
RT: 15.057 min Scan# 2052
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Instrument :
BNA_M
ClientSampleId :
SSTD02015




Tgt Ion:	232	Resp:	71695
Ion Ratio	Lower	Upper	
232	100		
131	25.0	34.0	51.0#
130	2.2	2.1	3.1
166	20.6	25.9	38.9#






Abundance Scan 2088 (15.269 min): BM027544.D\data.ms (-2057) (-)

Diethylphthalate
Concen: 19.37 ng/ul
RT: 15.263 min Scan# 2087
Delta R.T. -0.012 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45



Tgt Ion:	149	Resp:	171837
Ion Ratio	Lower	Upper	
149	100		
177	27.0	17.6	26.4#
150	13.5	9.6	14.4



Abundance Scan 2125 (15.486 min): BM027544.D\data.ms (-2159) (-)

Fluorene

Concen: 20.42 ng/ul

RT: 15.486 min Scan# 2125

Delta R.T. -0.006 min

Lab File: BM027561.D

Acq: 28 Sep 2020 12:45

Tgt Ion:166 Resp: 191279

Ion Ratio Lower Upper

166 100

165 98.0 77.8 116.6

167 13.2 10.7 16.1

Instrument :

BNA_M

ClientSampleId :

SSTD02015

m/z-->

Abundance Scan 2125 (15.486 min): BM027561.D\data.ms

Raw

m/z-->

Abundance Scan 2125 (15.486 min): BM027561.D\data.ms (-2092) (-)

Sub

m/z-->

Abundance

15.486

100000

50000

0

15.45 15.50

Time-->

Abundance Scan 2124 (15.480 min): BM027544.D\data.ms (-2160) (-)

4-Chlorophenyl-phenylether

Concen: 23.44 ng/ul

RT: 15.480 min Scan# 2124

Delta R.T. -0.012 min

Lab File: BM027561.D

Acq: 28 Sep 2020 12:45

Tgt Ion:204 Resp: 138035

Ion Ratio Lower Upper

204 100

206 32.1 27.0 40.6

141 41.7 44.6 66.8#

m/z-->

Abundance Scan 2124 (15.480 min): BM027561.D\data.ms

Raw

m/z-->

Abundance Scan 2124 (15.480 min): BM027561.D\data.ms (-2092) (-)

Sub

m/z-->

Abundance

15.480

100000

80000

60000

40000

20000

0


15.45 15.50

Time-->



Abundance Scan 2127 (15.498 min): BM027544.D\data.ms (-2100) (-)

4-Nitroaniline
Concen: 18.73 ng/ul
RT: 15.498 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45


Instrument :
BNA_M
ClientSampleId :
SSTD02015



Tgt Ion:138 Resp: 28902
Ion Ratio Lower Upper
138 100
92 49.0 48.6 73.0
108 149.4 86.3 129.5#






Abundance Scan 2127 (15.498 min): BM027561.D\data.ms (-2093) (-)



Abundance Scan 2137 (15.557 min): BM027544.D\data.ms (-2101) (-)

4,6-Dinitro-2-methylphenol
Concen: 20.59 ng/ul
RT: 15.557 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:198 Resp: 37481
Ion Ratio Lower Upper
198 100
182 2.9 3.4 5.0#
77 13.6 25.1 37.7#




Abundance Scan 2137 (15.557 min): BM027561.D\data.ms (-2103) (-)





Abundance Scan 2160 (15.692 min): BM027544.D\data.ms (-2160) (-)

N-Nitrosodiphenylamine
Concen: 19.88 ng/ul
RT: 15.692 min Scan# 2160
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45


Instrument :
BNA_M
ClientSampleId :
SSTD02015



Tgt Ion:	169	Resp:	169664
Ion	Ratio	Lower	Upper
169	100		
168	64.6	54.2	81.4
167	35.1	29.0	43.4




Abundance Scan 2160 (15.692 min): BM027561.D\data.ms (-2127) (-)





Abundance Scan 2276 (16.374 min): BM027544.D\data.ms (-2276) (-)


4-Bromophenyl-phenylether
Concen: 22.65 ng/ul
RT: 16.374 min Scan# 2276
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45



Tgt Ion:	248	Resp:	92686
Ion	Ratio	Lower	Upper
248	100		
250	95.4	80.7	121.1
141	52.4	55.8	83.8#



Abundance Scan 2276 (16.374 min): BM027561.D\data.ms (-2243) (-)




Abundance Scan 2295 (16.486 min): BM027544.D\data.ms (-2262) (-)

Hexachlorobenzene
Concen: 21.43 ng/ul
RT: 16.486 min Scan# 2295
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45



Instrument :
BNA_M
ClientSampleId :
SSTD02015

Ref 50
m/z-->



Tgt Ion:	284	Resp:	98848
Ion Ratio	Lower	Upper	
284	100		
142	22.7	27.4	41.2#
249	31.7	25.6	38.4


Abundance Scan 2295 (16.486 min): BM027561.D\data.ms (-2262) (-)



Abundance Scan 2322 (16.645 min): BM027544.D\data.ms (-2289) (-)



Atrazine
Concen: 21.27 ng/ul
RT: 16.645 min Scan# 2322
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Ref 50
m/z-->



Tgt Ion:	200	Resp:	79274
Ion Ratio	Lower	Upper	
200	100		
173	19.1	20.1	30.1#
215	50.5	39.0	58.4


Abundance Scan 2322 (16.645 min): BM027561.D\data.ms (-2289) (-)




Abundance Scan 2353 (16.827 min): BM027544.D\data.ms (-2353) (-)

Pentachlorophenol
Concen: 21.06 ng/ul
RT: 16.827 min Scan# 2353
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45


Instrument :
BNA_M
ClientSampleId :
SSTD02015



m/z-->




m/z-->




m/z-->

Tgt Ion:	266	Resp:	53048
Ion	Ratio	Lower	Upper
266	100		
264	60.0	50.0	75.0
268	63.1	49.4	74.0




Abundance Scan 2420 (17.221 min): BM027544.D\data.ms (-2420) (-)


Phenanthrene
Concen: 19.14 ng/ul
RT: 17.221 min Scan# 2420
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45



m/z-->





m/z-->



m/z-->

Tgt Ion:	178	Resp:	317596
Ion	Ratio	Lower	Upper
178	100		
179	15.4	12.2	18.4
176	18.3	15.1	22.7









Anthracene
Concen: 19.27 ng/uL
RT: 17.310 min Scan# 2435
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45




Instrument :
BNA_M
ClientSampleID :
SSTD02015

Tgt Ion:	178	Resp:	325020
Ion	Ratio	Lower	Upper
178	100		
179	15.2	12.5	18.7
176	18.7	14.6	22.0



1,2,3,4-Tetrachlorobenzene
Concen: 22.54 ng/uL
RT: 13.233 min Scan# 1742
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45


Tgt Ion:	216	Resp:	127602
Ion	Ratio	Lower	Upper
216	100		
214	74.6	63.5	95.3
218	48.0	38.4	57.6






Abundance Scan 2001 (14.757 min): BM027544.D\data.ms (-1974) (-)

Pentachlorobenzene
Concen: 22.47 ng/uL
RT: 14.757 min Scan# 2001
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Instrument :
BNA_M
ClientSampleId :
SSTD02015




Tgt Ion:	250	Resp:	123854
Ion	Ratio	Lower	Upper
250	100		
248	63.5	49.0	73.6
252	65.4	49.3	73.9







Abundance Scan 2481 (17.580 min): BM027544.D\data.ms (-2475) (-)

Carbazole
Concen: 18.44 ng/uL
RT: 17.580 min Scan# 2481
Delta R.T. 0.000 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45



Tgt Ion:	167	Resp:	246887
Ion	Ratio	Lower	Upper
167	100		
166	22.0	16.8	25.2
139	10.3	10.1	15.1









Scan 2579 (18.157 min): BM027544.D\data.ms (-2579) (-)
 Di-n-butylphthalate
 Concen: 18.07 ng/ul
 RT: 18.151 min Scan# 2578
 Delta R.T. -0.012 min
 Lab File: BM027561.D
 Acq: 28 Sep 2020 12:45




Instrument :
 BNA_M
 ClientSampleID :
 SSTD02015


Tgt Ion:	149	Resp:	261567
Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.5	11.3
104	6.6	5.1	7.7



Scan 2762 (19.233 min): BM027544.D\data.ms (-2762) (-)
 Fluoranthene
 Concen: 19.83 ng/ul
 RT: 19.233 min Scan# 2762
 Delta R.T. -0.006 min
 Lab File: BM027561.D
 Acq: 28 Sep 2020 12:45

Tgt Ion:	202	Resp:	419718
Ion	Ratio	Lower	Upper
202	100		
101	2.6	6.2	9.2#
100	1.9	5.0	7.6#









Pyrene
Concen: 18.27 ng/ul
RT: 19.598 min Scan# 2824
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45




Instrument :
BNA_M
ClientSampleId :
SSTD02015

Tgt Ion:	202	Resp:	422441
Ion	Ratio	Lower	Upper
202	100		
101	2.7	7.1	10.7#
100	2.2	6.1	9.1#




Butylbenzylphthalate
Concen: 19.06 ng/ul
RT: 20.504 min Scan# 2978
Delta R.T. -0.012 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:	149	Resp:	102911
Ion	Ratio	Lower	Upper
149	100		
91	58.0	59.4	89.2#
206	33.7	18.1	27.1#



Abundance Scan 3109 (21.274 min): BM027544.D\data.ms (-3162) (-)





3,3'-Dichlorobenzidine
Concen: 21.36 ng/ul
RT: 21.280 min Scan# 3110
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45


Instrument :
BNA_M
ClientSampleId :
SSTD02015

Tgt Ion:252 Resp: 155418
Ion Ratio Lower Upper
252 100
254 64.3 52.6 78.8
126 4.1 6.9 10.3#


Abundance Scan 3110 (21.280 min): BM027561.D\data.ms



Abundance Scan 3110 (21.280 min): BM027561.D\data.ms (-3077) (-)





Abundance Scan 3120 (21.339 min): BM027544.D\data.ms (-3184) (-)




Benzo(a)anthracene
Concen: 19.58 ng/ul
RT: 21.345 min Scan# 3121
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:228 Resp: 442500
Ion Ratio Lower Upper
228 100
229 19.2 15.8 23.6
226 27.0 21.5 32.3

Abundance Scan 3121 (21.345 min): BM027561.D\data.ms




Abundance Scan 3121 (21.345 min): BM027561.D\data.ms (-3088) (-)





Abundance Scan 3111 (21.286 min): BM027544.D\data.ms (-3109) (-)

Bis(2-ethylhexyl)phthalate
Concen: 18.30 ng/ul
RT: 21.286 min Scan# 3111
Delta R.T. -0.012 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45


Instrument :
BNA_M
ClientSampleId :
SSTD02015



Tgt Ion:	149	Resp:	151918
Ion Ratio	Lower	Upper	
149	100		
167	31.8	21.7	32.5
279	7.7	3.9	5.9#




Abundance Scan 3111 (21.286 min): BM027561.D\data.ms (-3079) (-)





Abundance Scan 3129 (21.392 min): BM027544.D\data.ms (-3125) (-)


Chrysene
Concen: 19.40 ng/ul
RT: 21.398 min Scan# 3130
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45




Tgt Ion:	228	Resp:	424055
Ion Ratio	Lower	Upper	
228	100		
226	29.2	23.4	35.2
229	18.9	15.8	23.8



Abundance Scan 3130 (21.398 min): BM027561.D\data.ms (-3097) (-)



Abundance Scan 3262 (22.174 min): BM027544.D\data.ms (-3262) (-)





Di-n-octyl phthalate
Concen: 15.86 ng/ul
RT: 22.180 min Scan# 3263
Delta R.T. -0.012 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45


Tgt Ion:149 Resp: 247112

Instrument :
BNA_M
ClientSampleId :
SSTD02015


Abundance Scan 3263 (22.180 min): BM027561.D\data.ms



Abundance Scan 3263 (22.180 min): BM027561.D\data.ms (-3231) (-)





Abundance Scan 3392 (22.939 min): BM027544.D\data.ms (-3392) (-)




Benzo(b)fluoranthene
Concen: 18.26 ng/ul
RT: 22.945 min Scan# 3393
Delta R.T. -0.012 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:252 Resp: 430750
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 2.2 5.1 7.7#

Abundance Scan 3393 (22.945 min): BM027561.D\data.ms




Abundance Scan 3393 (22.945 min): BM027561.D\data.ms (-3361) (-)






Abundance Scan 3400 (22.986 min): BM027544.D\data.ms (-3369) (-)

Benzo(k)fluoranthene
Concen: 18.88 ng/ul
RT: 22.992 min Scan# 3401
Delta R.T. -0.012 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Instrument :
BNA_M
ClientSampleID :
SSTD02015







Tgt Ion:	252	Resp:	444584
Ion Ratio	Lower	Upper	
252	100		
253	21.8	17.4	26.0
125	2.2	5.6	8.4#




Abundance Scan 3491 (23.521 min): BM027544.D\data.ms (-3491) (-)

Benzo(a)pyrene
Concen: 18.11 ng/ul
RT: 23.533 min Scan# 3493
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:	252	Resp:	394838
Ion Ratio	Lower	Upper	
252	100		
253	21.5	17.5	26.3
125	2.8	6.6	10.0#






Abundance Scan 3898 (25.915 min): BM027544.D\data.ms (-3898) (-)




Indeno(1,2,3-cd)pyrene
Concen: 18.80 ng/ul
RT: 25.927 min Scan# 3900
Delta R.T. -0.006 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Instrument :
BNA_M
ClientSampleId :
SSTD02015

Tgt Ion:	276	Resp:	514777
Ion	Ratio	Lower	Upper
276	100		
138	6.9	14.0	21.0#
277	24.7	20.1	30.1







Abundance Scan 3900 (25.927 min): BM027544.D\data.ms (-3898) (-)



Dibenzo(a,h)anthracene
Concen: 18.81 ng/ul
RT: 25.939 min Scan# 3902
Delta R.T. -0.012 min
Lab File: BM027561.D
Acq: 28 Sep 2020 12:45

Tgt Ion:	278	Resp:	433447
Ion	Ratio	Lower	Upper
278	100		
139	5.2	9.2	13.8#
279	23.7	19.0	28.6





Benzo(g,h,i)perylene
 Concen: 18.70 ng/ul
 RT: 26.621 min Scan# 4018
 Delta R.T. -0.012 min
 Lab File: BM027561.D
 Acq: 28 Sep 2020 12:45

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02015

Tgt Ion:	276	Resp:	423829
Ion	Ratio	Lower	Upper
276	100		
138	5.9	12.2	18.2#
277	23.6	18.6	28.0

