

Data Path : Z:\HPCHEM1\BNA_M\Data\BM092915\
 Data File : BM002778.D
 Acq On : 30 Sep 2015 10:49
 Operator : UM/IZ
 Sample : G3838-03
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E5MW1

Quant Time: Sep 30 13:04:29 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM092915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 30 06:48:30 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.70	152	105722	20.00	ng/ul	0.00
18) Naphthalene-d8	11.56	136	436004	20.00	ng/ul	0.00
36) Acenaphthene-d10	15.29	164	207269	20.00	ng/ul	0.00
62) Phenanthrene-d10	18.00	188	404435	20.00	ng/ul	0.00
78) Chrysene-d12	22.09	240	411022	20.00	ng/ul	0.00
86) Perylene-d12	24.68	264	419259	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.77	96	12263	5.39	ng/uL	0.00
5) Phenol-d5	7.81	99	238388	29.76	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.99	67	140484	31.18	ng/ul	0.00
9) 2-Chlorophenol-d4	8.21	132	224900	30.16	ng/ul	0.00
13) 4-Methylphenol-d8	9.39	113	199857	30.26	ng/ul	0.00
19) Nitrobenzene-d5	9.89	128	104388	31.40	ng/ul	0.00
22) 2-Nitrophenol-d4	10.62	143	108691	31.28	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	11.16	165	177770	28.99	ng/ul	0.00
29) 4-Chloroaniline-d4	11.68	131	272993	36.98	ng/ul	0.00
44) Dimethylphthalate-d6	14.67	166	492038	31.55	ng/ul	0.00
47) Acenaphthylene-d8	15.00	160	654928	31.47	ng/ul	0.00
52) 4-Nitrophenol-d4	15.46	143	73963	26.71	ng/ul	0.00
58) Fluorene-d10	16.27	176	425425	30.29	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	16.39	200	33633	17.05	ng/ul	0.00
71) Anthracene-d10	18.10	188	604858	31.34	ng/ul	0.00
79) Pyrene-d10	20.33	212	612503	31.93	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.51	264	687753	32.07	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
45) Dimethylphthalate	14.72	163	253917	16.02	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\Data\BM092915\
Data File : BM002778.D
Acq On : 30 Sep 2015 10:49
Operator : UM/IZ
Sample : G3838-03
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E5MW1

Quant Time: Sep 30 13:04:29 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM092915.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Sep 30 06:48:30 2015
Response via : Initial Calibration





#45

Dimethylphthalate

Concen: 16.02 ng/ul

RT: 14.72 min Scan# 1978

Delta R.T. -0.00 min

Lab File: BM002778.D


Acq: 30 Sep 2015 10:49

Instrument :

BNA_M

ClientSampleId :

E5MW1



Tgt Ion:	163	Resp:	253917
Ion	Ratio	Lower	Upper
163	100		
194	4.0	3.2	4.8
164	10.3	8.1	12.1

