

Data Path : Z:\HPCHEM1\BNA M\DATA\BM092915\
 Data File : BM002767.D
 Acq On : 30 Sep 2015 04:06
 Operator : UM/IZ
 Sample : G3798-08
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 D9G62

Quant Time: Sep 30 05:57:54 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM092915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 30 02:06:21 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.70	152	116173	20.00	ng/ul	0.00
18) Naphthalene-d8	11.56	136	492306	20.00	ng/ul	0.00
36) Acenaphthene-d10	15.29	164	237387	20.00	ng/ul	0.00
62) Phenanthrene-d10	18.00	188	456666	20.00	ng/ul	0.00
78) Chrysene-d12	22.09	240	445513	20.00	ng/ul	0.00
86) Perylene-d12	24.67	264	436855	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.76	96	9259	3.70	ng/uL	0.00
5) Phenol-d5	7.82	99	184983	21.02	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.99	67	108546	21.92	ng/ul	0.00
9) 2-Chlorophenol-d4	8.21	132	170004	20.75	ng/ul	0.00
13) 4-Methylphenol-d8	9.40	113	97130	13.38	ng/ul	0.00
19) Nitrobenzene-d5	9.89	128	78889	21.01	ng/ul	0.00
22) 2-Nitrophenol-d4	10.62	143	84139	21.44	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	11.16	165	135560	19.58	ng/ul	0.00
29) 4-Chloroaniline-d4	11.68	131	141357	16.96	ng/ul	0.00
44) Dimethylphthalate-d6	14.67	166	376340	21.07	ng/ul	0.00
47) Acenaphthylene-d8	15.00	160	295407	12.39	ng/ul	0.00
52) 4-Nitrophenol-d4	15.46	143	66400	20.94	ng/ul	0.00
58) Fluorene-d10	16.27	176	128119	7.96	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	16.39	200	21763	9.77	ng/ul	0.00
71) Anthracene-d10	18.10	188	31036	1.42	ng/ul	0.00
79) Pyrene-d10	20.33	212	91923	4.42	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.51	264	25403	1.14	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
45) Dimethylphthalate	14.72	163	158057	8.71	ng/ul	100




(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM092915\
Data File : BM002767.D
Acq On : 30 Sep 2015 04:06
Operator : UM/IZ
Sample : G3798-08
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
D9G62

Quant Time: Sep 30 05:57:54 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM092915.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Sep 30 02:06:21 2015
Response via : Initial Calibration





#45

Dimethylphthalate

Concen: 8.71 ng/ul

RT: 14.72 min Scan# 1977

Delta R.T. -0.01 min

Lab File: BM002767.D

Acq: 30 Sep 2015 04:06

Instrument :

BNA_M

ClientSampleId :

D9G62

Tot Ion:163 Resp: 158057

Ion Ratio Lower Upper

163 100

194 4.0 3.2 4.8

164 10.3 8.1 12.1

