

Data Path : Z:\HPCHEM1\BNA M\DATA\BM092915\
 Data File : BM002768.D
 Acq On : 30 Sep 2015 04:43
 Operator : UM/IZ
 Sample : G3798-09
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 D9G63

Quant Time: Sep 30 05:58:00 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM092915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 30 02:06:21 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.70	152	94460	20.00	ng/ul	0.00
18) Naphthalene-d8	11.56	136	395582	20.00	ng/ul	0.00
36) Acenaphthene-d10	15.29	164	195075	20.00	ng/ul	0.00
62) Phenanthrene-d10	18.00	188	392494	20.00	ng/ul	0.00
78) Chrysene-d12	22.09	240	403471	20.00	ng/ul	0.00
86) Perylene-d12	24.67	264	402944	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.77	96	6739	3.31	ng/uL	0.00
5) Phenol-d5	7.81	99	139569	19.50	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.99	67	76530	19.01	ng/ul	0.00
9) 2-Chlorophenol-d4	8.21	132	129408	19.42	ng/ul	0.00
13) 4-Methylphenol-d8	9.39	113	111641	18.92	ng/ul	0.00
19) Nitrobenzene-d5	9.89	128	54480	18.06	ng/ul	0.00
22) 2-Nitrophenol-d4	10.62	143	58733	18.63	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	11.16	165	101913	18.32	ng/ul	0.00
29) 4-Chloroaniline-d4	11.68	131	156669	23.39	ng/ul	0.00
44) Dimethylphthalate-d6	14.67	166	282669	19.26	ng/ul	0.00
47) Acenaphthylene-d8	15.00	160	202669	10.35	ng/ul	0.00
52) 4-Nitrophenol-d4	15.46	143	38744	14.87	ng/ul	0.00
58) Fluorene-d10	16.27	176	84452	6.39	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	16.39	200	7506	3.92	ng/ul	0.00
71) Anthracene-d10	18.10	188	23484	1.25	ng/ul	0.00
79) Pyrene-d10	20.33	212	60991	3.24	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.50	264	17319	0.84	ng/ul	-0.01


Target Compounds					Ovalue
45) Dimethylphthalate	14.72	163	121323	8.13	ng/ul 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM092915\
Data File : BM002768.D
Acq On : 30 Sep 2015 04:43
Operator : UM/IZ
Sample : G3798-09
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
D9G63

Quant Time: Sep 30 05:58:00 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM092915.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Sep 30 02:06:21 2015
Response via : Initial Calibration





#45

Dimethylphthalate

Concen: 8.13 ng/ul

RT: 14.72 min Scan# 1977

Delta R.T. -0.01 min

Lab File: BM002768.D

Acq: 30 Sep 2015 04:43

Instrument :

BNA_M

ClientSampleId :

D9G63


Tot Ion:163 Resp: 121323

Ion Ratio Lower Upper

163 100

194 3.9 3.2 4.8

164 9.8 8.1 12.1



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E

