

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM092916\
 Data File : BM007643.D
 Acq On : 30 Sep 2016 00:46
 Operator : UM/SJ
 Sample : H4955-18
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A4VM6

Manual Integrations
 APPROVED

sohil
 9/30/2016 4:02:53 PM

Quant Time: Sep 30 11:52:56 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 30 01:33:49 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	1507	0.40	ng/ul	-0.02
2) Naphthalene-d8	10.56	136	6241	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.40	164	4587	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.14	188	9838m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.35	240	9730	0.40	ng/ul	0.00
20) Perylene-d12	23.59	264	8997m	0.40	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.16	152	2897	0.40	ng/ul	0.00
14) Fluoranthene-d10	19.18	212	9440	0.41	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM092916\
Data File : BM007643.D
Acq On : 30 Sep 2016 00:46
Operator : UM/SJ
Sample : H4955-18
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A4VM6

Manual Integrations
APPROVED

sohil
9/30/2016 4:02:53 PM

Quant Time: Sep 30 11:52:56 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 30 01:33:49 2016
Response via : Initial Calibration

