


Data Path : Z:\HPCHEM1\BNA_M\Data\BM092916\
Data File : BM007644.D
Acq On : 30 Sep 2016 01:22
Operator : UM/SJ
Sample : H4955-19
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A4VM7

Manual Integrations
APPROVED

sohil
9/30/2016 4:02:51 PM

Quant Time: Sep 30 11:55:34 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 30 01:33:49 2016
Response via : Initial Calibration




Data Path : Z:\HPCHEM1\BNA_M\Data\BM092916\
Data File : BM007644.D
Acq On : 30 Sep 2016 01:22
Operator : UM/SJ
Sample : H4955-19
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A4VM7

Manual Integrations
APPROVED

sohil
9/30/2016 4:02:51 PM

Quant Time: Sep 30 11:53:38 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 30 01:33:49 2016
Response via : Initial Calibration



(10) Phenanthrene-d10 (I)

17.245min (+0.086) 0.40ng/ul

response 778034

Ion	Exp%	Act%
188.00	100	100
94.00	15.90	8.45#
80.00	18.30	6.94#
0.00	0.00	0.00




Data Path : Z:\HPCHEM1\BNA_M\Data\BM092916\
Data File : BM007644.D
Acq On : 30 Sep 2016 01:22
Operator : UM/SJ
Sample : H4955-19
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A4VM7

Manual Integrations
APPROVED

sohil
9/30/2016 4:02:51 PM

Quant Time: Sep 30 11:53:38 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 30 01:33:49 2016
Response via : Initial Calibration



TIC: BM007644.D

(10) Phenanthrene-d10 (I)

17.142min (-0.017) 0.40ng/ul m

SJ 9/30/16

response 9679

Ion	Exp%	Act%
188.00	100	100
94.00	15.90	11.28#
80.00	18.30	11.94#
0.00	0.00	0.00




Data Path : Z:\HPCHEM1\BNA_M\Data\BM092916\
Data File : BM007644.D
Acq On : 30 Sep 2016 01:22
Operator : UM/SJ
Sample : H4955-19
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A4VM7

Manual Integrations
APPROVED

sohil
9/30/2016 4:02:51 PM

Quant Time: Sep 30 11:54:09 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 30 01:33:49 2016
Response via : Initial Calibration



TIC: BM007644.D

(20) Perylene-d12 (I)

23.451min (-0.149) 0.40ng/ul

response 618511

Ion	Exp%	Act%
264.00	100	100
260.00	27.10	22.26
265.00	45.80	21.74#
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\HPCHEM1\BNA_M\Data\BM092916\
 Data File : BM007644.D
 Acq On : 30 Sep 2016 01:22
 Operator : UM/SJ
 Sample : H4955-19
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A4VM7

Manual Integrations
 APPROVED

sohil
 9/30/2016 4:02:51 PM

Quant Time: Sep 30 11:54:09 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 30 01:33:49 2016
 Response via : Initial Calibration



TIC: BM007644.D

(20) Perylene-d12 (I)

23.587min (-0.014) 0.40ng/ul m SJ 9/30/16

response 9153

Ion	Exp%	Act%
264.00	100	100
260.00	27.10	27.27
265.00	45.80	48.72
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM092916\
Data File : BM007644.D
Acq On : 30 Sep 2016 01:22
Operator : UM/SJ
Sample : H4955-19
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A4VM7

Manual Integrations
APPROVED

sohil
9/30/2016 4:02:51 PM

Quant Time: Sep 30 11:55:34 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 30 01:33:49 2016
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	1508	0.40	ng/ul	-0.02
2) Naphthalene-d8	10.56	136	6333	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.40	164	4662	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.14	188	9679m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.34	240	9798	0.40	ng/ul	0.00
20) Perylene-d12	23.59	264	9153m	0.40	ng/ul	-0.01

} SJ 9/30/16

System Monitoring Compounds

4) 2-Methylnaphthalene-d10	12.16	152	3295	0.45	ng/ul	0.00
14) Fluoranthene-d10	19.18	212	10360	0.46	ng/ul	0.00

Target Compounds

					Qvalue
13) Anthracene	17.28	178	3875	0.16 ng/ul#	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed