

Data Path : Z:\HPCHEM1\BNA M\DATA\BM092917\
 Data File : BM011787.D
 Acq On : 30 Sep 2017 03:15
 Operator : SJ/JU
 Sample : I5354-13
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Sep 30 06:38:44 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM092717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 30 06:38:23 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.93	152	318631	20.00	ng/ul	0.00
18) Naphthalene-d8	10.74	136	1444907	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.57	164	821110	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.32	188	1786129	20.00	ng/ul	0.00
75) Chrysene-d12	21.48	240	1408126	20.00	ng/ul	0.00
83) Perylene-d12	23.84	264	1167045	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.36	96	32615	4.56	ng/uL	0.00
5) Phenol-d5	7.09	99	807348	25.76	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.26	67	604278	26.59	ng/ul	0.00
9) 2-Chlorophenol-d4	7.46	132	670930	28.78	ng/ul	0.00
13) 4-Methylphenol-d8	8.63	113	682402	27.80	ng/ul	0.00
19) Nitrobenzene-d5	9.10	128	337273	29.92	ng/ul	0.00
22) 2-Nitrophenol-d4	9.82	143	383724	31.92	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.36	165	664871	28.98	ng/ul	0.00
29) 4-Chloroaniline-d4	10.87	131	925251	36.48	ng/ul	0.00
43) Dimethylphthalate-d6	13.97	166	2158496	30.65	ng/ul	0.00
46) Acenaphthylene-d8	14.27	160	2657158	31.20	ng/ul	0.00
51) 4-Nitrophenol-d4	14.76	143	296129	24.60	ng/ul	0.00
57) Fluorene-d10	15.56	176	1832409	30.22	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.68	200	209259	17.30	ng/ul	0.00
70) Anthracene-d10	17.42	188	2860944	31.48	ng/ul	0.00
76) Pyrene-d10	19.70	212	3002348	44.68	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.69	264	1772749	31.73	ng/ul	0.00


Target Compounds					Ovalue	
44) Dimethylphthalate	14.02	163	683677	10.150	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM092917\
Data File : BM011787.D
Acq On : 30 Sep 2017 03:15
Operator : SJ/JU
Sample : I5354-13
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Sep 30 06:38:44 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM092717.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 30 06:38:23 2017
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 10.150 ng/ul

RT: 14.02 min Scan# 1859

Delta R.T. -0.00 min

Lab File: BM011787.D

Acq: 30 Sep 2017 03:15

Instrument :

BNA_M

ClientSampleId :



Tot Ion:163 Resp: 683677

Ion Ratio Lower Upper

163 100

194 3.9 3.5 5.3

164 9.8 8.2 12.4



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E

