

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM092920\
 Data File : BM027565.D
 Acq On : 29 Sep 2020 12:25
 Operator : JU/CG
 Sample : SSTDO.403
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDO.4033

Quant Time: Sep 29 13:21:35 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM092920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 29 13:19:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.81	152	493	0.40	ng/ul	0.00
2) Naphthalene-d8	10.59	136	1278	0.40	ng/ul	0.00
7) Acenaphthene-d10	14.43	164	965	0.40	ng/ul	0.00
11) Phenanthrene-d10	17.17	188	2538	0.40	ng/ul	0.00
17) Chrysene-d12	21.35	240	2691	0.40	ng/ul	0.00
21) Perylene-d12	23.62	264	2745	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.18	152	1050	0.45	ng/ul	0.00
15) Fluoranthene-d10	19.20	212	3458	0.50	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.64	128	1459	0.390	ng/ul	100
5) 2-Methylnaphthalene	12.25	142	1137	0.418	ng/ul	100
6) 1-Methylnaphthalene	12.47	142	1295	1.388	ng/ul	100
8) Acenaphthylene	14.15	152	1793	0.364	ng/ul	100
9) Acenaphthene	14.50	153	1339	0.382	ng/ul	100
10) Fluorene	15.48	166	1762	0.406	ng/ul	100
12) Pentachlorophenol	16.82	266	390	0.565	ng/ul	100
13) Phenanthrene	17.22	178	3082	0.394	ng/ul	100
14) Anthracene	17.31	178	2937	0.383	ng/ul	100
16) Fluoranthene	19.23	202	4300	0.456	ng/ul	100
18) Pyrene	19.59	202	4231	0.345	ng/ul	100
19) Benzo(a)anthracene	21.34	228	4341	0.411	ng/ul	100
20) Chrysene	21.39	228	4483	0.410	ng/ul	100
22) Benzo(b)fluoranthene	22.94	252	4603	0.416	ng/ul	100
23) Benzo(k)fluoranthene	22.98	252	4672	0.406	ng/ul	100
24) Benzo(a)pyrene	23.52	252	4093	0.413	ng/ul	100
25) Indeno(1,2,3-cd)pyrene	25.91	276	5559	0.436	ng/ul#	100
26) Dibenzo(a,h)anthracene	25.92	278	4606	0.438	ng/ul	100
27) Benzo(a,h,i)perylene	26.60	276	4740	0.440	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM092920\
Data File : BM027565.D
Acq On : 29 Sep 2020 12:25
Operator : JU/CG
Sample : SST0.403
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4033

Quant Time: Sep 29 13:21:35 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM092920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 29 13:19:48 2020
Response via : Initial Calibration

