


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM092920\
 Data File : BM027562.D
 Acq On : 29 Sep 2020 09:59
 Operator : JU/CG
 Sample : DFTPP831
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP831

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM092920.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Sep 29 15:04:50 2020



AutoFind: Scans 2488, 2489, 2490; Background Corrected with Scan 2482

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	18.7	9148	PASS
68	69	0.00	2	1.9	204	PASS
69	198	0.00	100	22.5	10994	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	28.5	13915	PASS
197	198	0.00	2	0.6	304	PASS
198	198	100	100	100.0	48848	PASS
199	198	5	9	7.6	3696	PASS
275	198	10	60	33.0	16110	PASS
365	198	1	100	6.3	3053	PASS
441	198	0.01	100	13.6	6653	PASS
442	198	50	100	89.4	43674	PASS
443	442	15	24	19.2	8403	PASS

Scan 2482 (17.586 min): BM027562.D
DFTPP831

m/z	Abundance
207.10	254.0
281.00	257.0

Instrument :
BNA_M
ClientSampleId :
DFTPP831

m/z	Abundance
38.00	240.0
39.00	836.0
50.10	3975.0
51.10	11309.0
52.00	816.0
56.00	381.0
57.00	791.0
60.90	212.0
62.10	364.0
63.00	545.0
64.70	181.0
68.00	395.0
69.00	13643.0
73.10	225.0
74.10	2172.0
75.00	3533.0
76.00	901.0
77.10	15766.0
78.10	1073.0
79.00	1957.0
80.00	1051.0
81.00	1655.0
82.10	386.0
85.00	185.0
86.00	661.0
90.90	307.0
92.00	367.0
93.00	3583.0
94.00	216.0
98.00	2906.0
99.00	2296.0
101.00	545.0
102.90	235.0
104.00	595.0
104.90	558.0
106.00	277.0
107.00	6239.0
108.00	670.0
108.90	157.0
110.00	8984.0
110.90	1525.0
112.00	287.0
116.00	414.0
117.00	10726.0
117.90	746.0
122.10	441.0
123.00	551.0
123.90	401.0
125.00	374.0
127.00	17232.0
128.00	1406.0
129.00	8410.0
130.00	977.0
133.90	219.0
135.00	685.0
135.90	212.0
136.90	542.0
141.00	1691.0
141.90	525.0
143.10	239.0
145.90	291.0
147.10	881.0
148.00	2496.0
149.20	533.0
149.90	225.0
151.10	287.0
152.80	380.0
153.10	381.0
153.90	521.0
154.90	793.0
156.00	1061.0
157.00	209.0
157.90	196.0
158.90	318.0
160.00	521.0
160.90	863.0
161.90	206.0
165.00	747.0
165.90	533.0
167.00	4551.0
168.00	2636.0
168.90	280.0
169.90	208.0
171.10	189.0
172.00	259.0
173.00	410.0
174.00	918.0
175.00	1300.0
176.00	246.0
176.90	777.0
178.00	224.0
179.00	2588.0

Instrument :
BNA_M
ClientSampleId :
DFTPP831

179.90	1877.0
181.00	820.0
184.10	262.0
185.00	1305.0
186.00	9128.0
187.00	2656.0
188.00	262.0
188.90	709.0
190.80	385.0
192.00	942.0
193.00	903.0
194.10	272.0
195.90	1485.0
197.10	419.0
198.00	54792.0
199.00	4334.0
199.90	388.0
201.40	152.0
203.10	386.0
204.00	2932.0
205.00	4294.0
206.00	17408.0
207.00	2956.0
208.00	732.0
209.00	326.0
210.00	266.0
210.90	768.0
215.00	310.0
216.00	638.0
217.00	6897.0
218.00	671.0
221.00	1955.0
221.90	716.0
223.00	1455.0
224.00	10901.0
225.00	2930.0
226.10	420.0
227.00	5864.0
227.90	1010.0
228.90	1099.0
230.90	415.0
234.00	520.0
235.00	339.0
236.10	304.0
237.00	448.0
238.90	312.0
241.00	210.0
242.10	753.0
243.00	539.0
244.00	7345.0
245.00	1157.0
246.00	2348.0
246.90	431.0
248.80	217.0
253.10	244.0
253.90	239.0
255.00	40136.0
256.00	6118.0
257.00	351.0
258.00	3394.0
259.00	664.0
265.00	1131.0
266.00	240.0
270.90	220.0
273.00	1450.0
274.00	3868.0
275.00	16592.0
276.00	2130.0
277.00	2101.0
278.10	413.0
280.90	151.0
283.10	272.0
284.00	173.0
285.00	366.0
292.90	362.0
296.00	7589.0
297.00	813.0
301.80	174.0
303.00	590.0
304.10	265.0
313.90	348.0
315.00	942.0
316.00	368.0
321.00	181.0
322.90	1525.0
324.00	236.0
327.00	259.0
328.00	190.0
333.00	196.0
334.00	1053.0
334.90	195.0
341.00	170.0
345.90	505.0
351.90	293.0
353.10	267.0
354.10	435.0

Instrument :
BNA_M
ClientSampleId :
DFTPP831

365.00	3014.0
365.90	361.0
371.00	164.0
372.00	876.0
373.10	250.0
383.10	206.0
402.00	231.0
403.00	408.0
420.80	223.0
422.00	336.0
423.10	2161.0
423.90	388.0
441.10	5501.0
442.10	38752.0
443.00	7461.0
444.10	755.0

Instrument :
BNA_M
ClientSampleId :
DFTPP831

m/z	Abundance
38.10	220.0
39.00	518.0
50.10	3806.0
51.00	10701.0
51.90	646.0
55.80	255.0
57.10	610.0
62.10	241.0
62.90	378.0
68.00	217.0
69.00	13244.0
74.10	2089.0
74.90	2685.0
76.00	1075.0
77.10	14104.0
78.00	1040.0
79.00	1927.0
79.90	1314.0
81.10	1249.0
82.00	286.0
82.90	290.0
85.00	151.0
85.90	508.0
87.00	245.0
91.00	513.0
91.90	331.0
93.00	3451.0
93.90	190.0
98.00	3009.0
98.90	1815.0
100.00	173.0
100.90	903.0
102.90	293.0
103.90	446.0
104.90	481.0
107.00	6160.0
108.00	639.0
109.10	168.0
110.00	8016.0
111.00	1281.0
111.90	204.0
115.80	262.0
117.00	11058.0
117.90	639.0
122.10	447.0
123.00	501.0
123.90	238.0
124.90	352.0
127.00	16169.0
128.00	1137.0
129.00	8814.0
129.90	547.0
131.00	180.0
133.90	255.0
134.90	807.0
136.00	396.0
137.00	398.0
141.00	1463.0
141.90	382.0
143.10	331.0
145.90	333.0
146.90	749.0
148.00	2384.0
149.10	339.0
152.90	736.0
154.00	318.0
155.10	737.0
156.00	1285.0
157.20	193.0
157.90	243.0
159.00	250.0
160.00	514.0
160.90	786.0
162.00	188.0
164.90	676.0
166.10	537.0
167.00	3947.0
168.00	2125.0
169.00	389.0
170.00	169.0
170.90	243.0
171.90	330.0
172.90	491.0
173.90	755.0
175.00	2010.0
175.80	290.0
177.00	605.0
178.00	346.0
179.00	2872.0
180.00	1495.0
180.90	704.0
182.00	164.0

Instrument :
BNA_M
ClientSampleId :
DFTPP831

184.00	190.0
185.00	1080.0
186.00	9334.0
187.00	2728.0
187.90	302.0
188.90	765.0
191.00	444.0
192.00	1112.0
193.00	1152.0
194.00	296.0
196.00	1457.0
197.00	493.0
197.90	58968.0
199.00	4592.0
200.10	386.0
201.00	181.0
202.90	572.0
204.00	3455.0
205.00	5153.0
206.10	18256.0
207.00	3002.0
207.90	1033.0
208.90	330.0
210.20	468.0
211.00	842.0
214.80	414.0
215.90	638.0
217.00	6348.0
217.90	743.0
221.00	1838.0
221.90	967.0
223.00	1324.0
224.00	11208.0
225.00	3373.0
226.00	335.0
227.00	6152.0
227.90	1075.0
229.00	1312.0
230.00	184.0
231.00	485.0
233.90	512.0
235.00	576.0
235.90	378.0
237.00	380.0
238.90	276.0
240.00	244.0
240.90	444.0
242.00	596.0
243.00	597.0
244.00	9214.0
245.00	1165.0
246.00	2625.0
247.10	473.0
248.80	369.0
253.00	302.0
254.10	245.0
255.00	45880.0
256.00	7243.0
257.00	668.0
258.00	3739.0
259.00	704.0
265.00	1666.0
265.90	222.0
272.90	1244.0
274.00	3755.0
275.00	20192.0
276.00	2628.0
277.00	2311.0
277.90	346.0
280.80	198.0
282.00	168.0
282.90	296.0
284.00	217.0
284.90	331.0
293.00	610.0
296.00	8127.0
297.00	1278.0
303.10	867.0
303.90	193.0
313.90	379.0
314.90	1054.0
316.00	401.0
321.00	196.0
323.10	2132.0
324.00	267.0
326.90	531.0
327.80	169.0
333.10	468.0
334.00	1418.0
335.00	357.0
341.10	362.0
346.00	583.0
352.00	690.0
353.00	342.0
354.00	685.0
365.00	3593.0

Instrument :
BNA_M
ClientSampleId :
DFTPP831

365.90	604.0
372.00	875.0
373.00	300.0
383.10	207.0
390.00	173.0
402.00	334.0
403.00	446.0
420.90	488.0
421.90	298.0
423.00	2352.0
424.00	680.0
441.10	8284.0
442.10	52792.0
443.10	9760.0
444.00	913.0

Instrument :
BNA_M
ClientSampleId :
DFTPP831

m/z	Abundance
39.00	468.0
50.10	1961.0
51.00	5434.0
52.10	297.0
55.90	162.0
57.00	400.0
62.90	160.0
69.00	6097.0
72.90	219.0
74.10	996.0
75.10	1274.0
75.90	411.0
77.10	6957.0
78.00	401.0
79.00	714.0
80.00	683.0
81.10	603.0
82.10	174.0
85.80	207.0
86.10	190.0
92.00	295.0
93.00	1681.0
94.00	181.0
98.00	1718.0
98.90	992.0
101.00	306.0
103.90	237.0
105.00	430.0
107.00	2996.0
108.10	557.0
109.10	225.0
110.00	4310.0
110.90	740.0
116.10	236.0
117.00	4612.0
118.00	437.0
121.90	230.0
123.00	369.0
123.90	228.0
124.80	176.0
127.00	8344.0
128.00	814.0
129.00	4322.0
130.00	454.0
134.90	419.0
136.00	245.0
136.90	270.0
141.00	669.0
141.80	290.0
142.90	161.0
145.90	154.0
147.00	563.0
148.00	1463.0
149.00	182.0
153.00	217.0
153.80	326.0
154.90	562.0
156.10	651.0
157.90	264.0
158.90	249.0
159.90	280.0
161.00	351.0
164.80	258.0
165.80	170.0
167.00	2593.0
168.00	1185.0
168.90	164.0
170.80	163.0
171.10	180.0
171.90	231.0
172.90	249.0
174.00	516.0
175.00	743.0
175.90	151.0
176.20	180.0
176.70	166.0
179.00	1762.0
180.00	1111.0
180.90	460.0
184.90	580.0
186.00	5230.0
187.00	1780.0
188.20	195.0
188.90	517.0
191.00	371.0
192.00	555.0
193.00	855.0
193.90	169.0
196.00	1067.0
196.70	215.0
198.00	32784.0
198.90	2163.0

Instrument :
BNA_M
ClientSampleId :
DFTPP831

200.00	157.0
201.30	225.0
203.00	356.0
204.00	1934.0
205.10	2707.0
206.00	11322.0
207.00	1873.0
208.00	830.0
209.00	286.0
209.90	212.0
210.90	511.0
215.00	168.0
215.90	366.0
217.00	3906.0
217.90	568.0
221.00	1057.0
222.00	656.0
223.00	850.0
224.00	7375.0
225.00	1797.0
226.00	222.0
227.00	4242.0
227.90	587.0
228.90	833.0
230.90	210.0
234.00	318.0
234.90	202.0
237.00	270.0
238.80	200.0
240.90	282.0
241.80	358.0
243.00	367.0
244.00	4834.0
245.00	756.0
246.00	1291.0
246.90	359.0
248.90	182.0
254.10	210.0
255.00	28192.0
256.00	3947.0
257.00	206.0
258.00	2457.0
259.00	328.0
265.00	893.0
272.00	171.0
273.00	897.0
274.00	2344.0
275.00	11548.0
276.00	1598.0
277.00	1619.0
277.90	220.0
281.00	230.0
283.10	248.0
285.00	280.0
292.80	234.0
293.10	238.0
296.00	4735.0
296.90	481.0
303.10	660.0
303.80	161.0
314.00	307.0
315.00	331.0
316.00	331.0
323.00	1295.0
323.90	246.0
326.90	428.0
327.90	165.0
333.00	228.0
334.00	791.0
335.10	229.0
341.00	150.0
346.00	458.0
352.20	377.0
353.00	215.0
354.00	375.0
364.90	2553.0
366.00	382.0
372.10	989.0
373.00	254.0
402.10	443.0
403.10	253.0
404.10	160.0
421.00	370.0
422.10	273.0
423.10	1838.0
424.10	605.0
441.00	6174.0
442.10	39480.0
443.10	7989.0
444.00	554.0

Instrument :
BNA_M
ClientSampleId :
DFTPP831