

Data Path : Z:\HPCHEM1\BNA M\DATA\BM100117\
 Data File : BM011834.D
 Acq On : 02 Oct 2017 06:19
 Operator : SJ/JU
 Sample : I5477-13
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ET238

Quant Time: Oct 02 07:40:21 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM092717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 02 07:18:07 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.88	152	142107	20.00	ng/ul	0.00
18) Naphthalene-d8	10.68	136	645806	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.52	164	393475	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.26	188	1002546	20.00	ng/ul	0.00
75) Chrysene-d12	21.43	240	1233407	20.00	ng/ul	0.00
83) Perylene-d12	23.76	264	1106816	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.29	96	17373	5.45	ng/uL	0.00
5) Phenol-d5	7.03	99	100214	7.17	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.20	67	294786	29.08	ng/ul	0.00
9) 2-Chlorophenol-d4	7.41	132	265835	25.57	ng/ul	0.00
13) 4-Methylphenol-d8	8.58	113	187478	17.13	ng/ul	0.00
19) Nitrobenzene-d5	9.05	128	155118	30.79	ng/ul	0.00
22) 2-Nitrophenol-d4	9.76	143	172593	32.12	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.30	165	309239	30.16	ng/ul	0.00
29) 4-Chloroaniline-d4	10.83	131	23343	2.06	ng/ul	0.01
43) Dimethylphthalate-d6	13.93	166	1138441	33.73	ng/ul	0.00
46) Acenaphthylene-d8	14.22	160	1312490	32.16	ng/ul	0.00
51) 4-Nitrophenol-d4	14.72	143	30239	5.24	ng/ul	0.00
57) Fluorene-d10	15.51	176	979815	33.72	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.63	200	184381	27.15	ng/ul	0.00
70) Anthracene-d10	17.36	188	1720682	33.73	ng/ul	0.00
76) Pyrene-d10	19.65	212	2242823	38.10	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.61	264	1800687	33.98	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM100117\
Data File : BM011834.D
Acq On : 02 Oct 2017 06:19
Operator : SJ/JU
Sample : I5477-13
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ET238

Quant Time: Oct 02 07:40:21 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM092717.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 02 07:18:07 2017
Response via : Initial Calibration

