

Data Path : Z:\HPCHEM1\BNA M\DATA\BM100117\
 Data File : BM011833.D
 Acq On : 02 Oct 2017 05:44
 Operator : SJ/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02007

Quant Time: Oct 02 07:15:24 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM092717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 02 04:55:04 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	63	0.00
2	1,4-Dioxane	0.499	0.488	2.2	60	0.00
3 S	1,4-Dioxane-d8	0.449	0.420	6.5	58	0.00
4	Benzaldehyde	0.965	1.156	-19.8	65	0.00
5 S	Phenol-d5	1.967	1.868	5.0	63	0.00
6	Phenol	1.989	1.923	3.3	64	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.427	1.428	-0.1	65	0.00
8	Bis(2-Chloroethyl)ether	1.530	1.467	4.1	62	0.00
9 S	2-Chlorophenol-d4	1.463	1.485	-1.5	64	0.00
10	2-Chlorophenol	1.428	1.391	2.6	61	0.00
11	2-Methylphenol	1.461	1.408	3.6	63	0.00
12	2,2'-oxybis(1-Chloropropane	3.276	3.770	-15.1	73	0.00
13 S	4-Methylphenol-d8	1.541	1.548	-0.5	66	0.00
14	Acetophenone	2.476	2.533	-2.3	67	0.00
15 P	N-Nitroso-di-n-propylamine	1.397	1.492	-6.8	66	0.00
16	4-Methylphenol	1.606	1.589	1.1	65	0.00
17	Hexachloroethane	0.630	0.657	-4.3	66	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	68	0.00
19 S	Nitrobenzene-d5	0.156	0.151	3.2	65	0.00
20	Nitrobenzene	0.429	0.453	-5.6	73	0.00
21	Isophorone	0.805	0.801	0.5	68	0.00
22 S	2-Nitrophenol-d4	0.166	0.164	1.2	68	0.00
23 C	2-Nitrophenol	0.173	0.170	1.7	66	0.00
24	2,4-Dimethylphenol	0.394	0.407	-3.3	70	0.00
25	Bis(2-Chloroethoxy)methane	0.466	0.448	3.9	65	0.00
26 S	2,4-Dichlorophenol-d3	0.318	0.314	1.3	68	0.00
27 C	2,4-Dichlorophenol	0.302	0.305	-1.0	68	0.00
28	Naphthalene	1.045	0.993	5.0	65	0.00
29 S	4-Chloroaniline-d4	0.351	0.410	-16.8	70	0.00
30	4-Chloroaniline	0.340	0.394	-15.9	70	0.00
31 C	Hexachlorobutadiene	0.213	0.216	-1.4	71	0.00
32	Caprolactam	0.098	0.101	-3.1	74	0.00
33 C	4-Chloro-3-methylphenol	0.342	0.364	-6.4	72	0.00
34	2-Methylnaphthalene	0.741	0.733	1.1	68	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	75	0.00
36	1,2,4,5-Tetrachlorobenzene	0.676	0.648	4.1	72	0.00
37	Hexachlorocyclopentadiene	0.435	0.368	15.4	70	0.00
38 C	2,4,6-Trichlorophenol	0.434	0.427	1.6	74	0.00
39	2,4,5-Trichlorophenol	0.449	0.435	3.1	72	0.00
40	1,1'-Biphenyl	1.647	1.511	8.3	69	0.00
41	2-Chloronaphthalene	1.262	1.186	6.0	71	0.00
42	2-Nitroaniline	0.464	0.507	-9.3	81	0.00
43 S	Dimethylphthalate-d6	1.715	1.664	3.0	73	0.00
44	Dimethylphthalate	1.641	1.599	2.6	73	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM100117\
 Data File : BM011833.D
 Acq On : 02 Oct 2017 05:44
 Operator : SJ/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02007

Quant Time: Oct 02 07:15:24 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM092717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 02 04:55:04 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.302	0.306	-1.3	76	0.00
46 S	Acenaphthylene-d8	2.074	1.998	3.7	72	0.00
47	Acenaphthylene	2.081	1.980	4.9	72	0.00
48	3-Nitroaniline	0.309	0.303	1.9	76	0.00
49 C	Acenaphthene	1.417	1.347	4.9	72	0.00
50	2,4-Dinitrophenol	0.198	0.175	11.6	77	0.00
51 S	4-Nitrophenol-d4	0.293	0.289	1.4	79	0.00
52	4-Nitrophenol	0.287	0.337	-17.4	94	0.00
53	Dibenzofuran	1.946	1.891	2.8	73	0.00
54	2,4-Dinitrotoluene	0.431	0.451	-4.6	77	0.00
55	2,3,4,6-Tetrachlorophenol	0.415	0.435	-4.8	78	0.00
56	Diethylphthalate	1.725	1.753	-1.6	77	0.00
57 S	Fluorene-d10	1.477	1.459	1.2	75	0.00
58	Fluorene	1.648	1.639	0.5	76	0.00
59	4-Chlorophenyl-phenylether	0.829	0.850	-2.5	79	0.00
60	4-Nitroaniline	0.339	0.319	5.9	81	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.135	0.117	13.3	81	0.00
63	4,6-Dinitro-2-methylphenol	0.138	0.120	13.0	81	0.00
64	N-Nitrosodiphenylamine	0.591	0.531	10.2	75	0.00
65	4-Bromophenyl-phenylether	0.220	0.198	10.0	77	0.00
66	Hexachlorobenzene	0.244	0.229	6.1	79	0.00
67	Atrazine	0.229	0.213	7.0	82	0.00
68 C	Pentachlorophenol	0.161	0.141	12.4	82	0.00
69	Phenanthrene	1.119	1.034	7.6	78	0.00
70 S	Anthracene-d10	1.018	0.957	6.0	80	0.00
71	Anthracene	1.153	1.081	6.2	79	0.00
72	Carbazole	1.010	0.904	10.5	78	0.00
73	Di-n-butylphthalate	1.301	1.234	5.1	79	0.00
74 C	Fluoranthene	1.367	1.331	2.6	84	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
76 S	Pyrene-d10	0.955	0.888	7.0	87	0.00
77	Pyrene	1.192	1.084	9.1	85	0.00
78	Butylbenzylphthalate	0.498	0.448	10.0	83	0.00
79	3,3'-Dichlorobenzidine	0.380	0.328	13.7	82	0.00
80	Benzo(a)anthracene	1.119	1.139	-1.8	95	0.00
81	Bis(2-ethylhexyl)phthalate	0.744	0.680	8.6	85	0.00
82	Chrysene	1.055	1.038	1.6	91	0.00
83 I	Perylene-d12	1.000	1.000	0.0	97	0.00
84	Di-n-octyl phthalate	1.546	1.230	20.4#	88	0.00
85	Benzo(b)fluoranthene	1.180	1.151	2.5	96	0.00
86	Benzo(k)fluoranthene	1.204	1.104	8.3	99	0.00
87 S	Benzo(a)pyrene-d12	0.958	0.930	2.9	97	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM100117\
Data File : BM011833.D
Acq On : 02 Oct 2017 05:44
Operator : SJ/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02007

Quant Time: Oct 02 07:15:24 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM092717.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 02 04:55:04 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.140	1.098	3.7	97	0.00
89	Indeno(1,2,3-cd)pyrene	1.187	1.229	-3.5	99	0.00
90	Dibenzo(a,h)anthracene	0.994	1.028	-3.4	98	0.00
91	Benzo(a,h,i)perylene	0.985	1.014	-2.9	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0