

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100124\
 Data File : BM047764.D
 Acq On : 02 Oct 2024 06:12
 Operator : RC/JU
 Sample : PB163513BL
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK513

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/04/2024
 Supervised By :mohammad ahmed 10/04/2024

Quant Time: Oct 02 07:49:22 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Sep 28 02:30:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.810	152	234143	20.000	ng/u1	0.00
20) Naphthalene-d8	10.598	136	959873	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.439	164	580684	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.180	188	1166614	20.000	ng/u1	0.00
79) Chrysene-d12	21.403	240	1069806	20.000	ng/u1	0.00
88) Perylene-d12	24.403	264	1250797	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.257	96	33178	4.579	ng/uL	0.00
4) Pyridine-d5	3.669	84	469938	22.152	ng/u1	0.00
7) Phenol-d5	6.975	99	506575	22.880	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.140	67	319695	20.348	ng/u1	0.00
11) 2-Chlorophenol-d4	7.340	132	421369	26.158	ng/u1	0.00
15) 4-Methylphenol-d8	8.510	113	388338	23.591	ng/u1	0.00
21) Nitrobenzene-d5	8.963	128	195161	25.536	ng/u1	0.00
24) 2-Nitrophenol-d4	9.681	143	195568	26.703	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.222	165	365609	24.990	ng/u1	0.00
31) 4-Chloroaniline-d4	10.733	131	554362	24.509	ng/u1	0.00
46) Dimethylphthalate-d6	13.851	166	1054209	25.041	ng/u1	0.00
49) Acenaphthylene-d8	14.133	160	1251438	25.071	ng/u1	0.00
54) 4-Nitrophenol-d4	14.627	143	162634m	19.403	ng/u1	0.00
60) Fluorene-d10	15.433	176	919538	25.216	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.545	200	107925	15.809	ng/u1	0.00
73) Anthracene-d10	17.274	188	1387844	24.194	ng/u1	-0.01
81) Pyrene-d10	19.562	212	1575947	25.956	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.203	264	1579971	24.346	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100124\
Data File : BM047764.D
Acq On : 02 Oct 2024 06:12
Operator : RC/JU
Sample : PB163513BL
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK513

Quant Time: Oct 02 07:49:22 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Sep 28 02:30:36 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/04/2024
Supervised By :mohammad ahmed 10/04/2024

