

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100220\
 Data File : BM027596.D
 Acq On : 03 Oct 2020 00:12
 Operator : JU/CG
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.403

Quant Time: Oct 03 01:38:53 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 02 13:04:03 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.800	152	1222	0.40	ng/ul	0.00
2) Naphthalene-d8	10.581	136	3529	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.425	164	2418	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.169	188	6047	0.40	ng/ul	0.00
16) Chrysene-d12	21.343	240	6495	0.40	ng/ul	0.00
20) Perylene-d12	23.603	264	6804	0.40	ng/ul	#-0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.169	152	2347	0.37	ng/ul	0.00
14) Fluoranthene-d10	19.192	212	7043	0.43	ng/ul	0.00
Target Compounds						
						Qvalue
3) Naphthalene	10.631	128	3654	0.35	ng/ul	97
5) 2-Methylnaphthalene	12.245	142	2638	0.35	ng/ul	100
7) Acenaphthylene	14.144	152	3887	0.31	ng/ul	99
8) Acenaphthene	14.489	153	3028	0.34	ng/ul	96
9) Fluorene	15.475	166	3859	0.35	ng/ul	99
11) Pentachlorophenol	16.819	266	844	0.51	ng/ul	97
12) Phenanthrene	17.207	178	6742	0.36	ng/ul	99
13) Anthracene	17.301	178	6252	0.34	ng/ul	98
15) Fluoranthene	19.222	202	9065	0.40	ng/ul	98
17) Pyrene	19.584	202	9214	0.31	ng/ul#	97
18) Benzo(a)anthracene	21.326	228	9322	0.37	ng/ul	99
19) Chrysene	21.379	228	9797	0.37	ng/ul	99
21) Benzo(b)fluoranthene	22.922	252	10409	0.38	ng/ul	94
22) Benzo(k)fluoranthene	22.968	252	10909	0.38	ng/ul	95
23) Benzo(a)pyrene	23.503	252	9562	0.39	ng/ul#	90
24) Indeno(1,2,3-cd)pyrene	25.882	276	12981	0.41	ng/ul#	98
25) Dibenzo(a,h)anthracene	25.897	278	10673	0.41	ng/ul	97
26) Benzo(g,h,i)perylene	26.575	276	10849	0.41	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100220\
Data File : BM027596.D
Acq On : 03 Oct 2020 00:12
Operator : JU/CG
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.403

Quant Time: Oct 03 01:38:53 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 02 13:04:03 2020
Response via : Initial Calibration

