


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100220\
 Data File : BM027576.D
 Acq On : 02 Oct 2020 10:17
 Operator : JU/CG
 Sample : DFTPP85
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP85

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092320.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Oct 05 10:43:07 2020



AutoFind: Scans 2487, 2488, 2489; Background Corrected with Scan 2481

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	23.7	13413	PASS
68	69	0.00	2	2.8	372	FAIL*
69	198	0.00	100	23.5	13279	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	27.5	15547	PASS
197	198	0.00	2	0.7	388	PASS
198	198	100	100	100.0	56539	PASS
199	198	5	9	6.8	3842	PASS
275	198	10	60	33.3	18843	PASS
365	198	1	100	6.5	3696	PASS
441	198	0.01	100	13.4	7568	PASS
442	198	50	100	85.7	48480	PASS
443	442	15	24	19.1	9239	PASS

m/z	Abundance
207.00	331.0

Instrument :
BNA_M
ClientSampleId :
DFTPP85

m/z	Abundance
38.00	207.0
39.00	1124.0
49.10	212.0
50.00	4970.0
51.00	14753.0
52.10	743.0
55.90	318.0
56.90	744.0
61.00	228.0
62.10	505.0
63.00	599.0
65.00	276.0
67.90	407.0
69.00	13876.0
72.90	414.0
74.00	2622.0
75.00	3149.0
76.10	1301.0
77.00	17896.0
78.00	1267.0
79.00	1708.0
79.90	1247.0
81.00	1678.0
82.00	371.0
82.90	324.0
84.90	403.0
85.90	766.0
87.10	260.0
90.90	317.0
92.10	416.0
93.00	4377.0
94.00	252.0
98.00	3879.0
98.90	2091.0
100.90	811.0
102.90	267.0
104.00	665.0
104.90	575.0
106.10	254.0
107.00	7114.0
108.10	901.0
108.90	159.0
110.00	8689.0
111.00	1649.0
112.20	179.0
116.00	415.0
117.00	14039.0
117.90	1084.0
121.90	536.0
122.90	806.0
124.10	253.0
124.90	238.0
127.00	16560.0
127.90	1559.0
129.00	8110.0
130.00	865.0
131.00	217.0
133.90	556.0
135.00	779.0
135.90	312.0
136.90	665.0
139.90	155.0
141.00	1820.0
142.00	501.0
143.00	453.0
145.80	393.0
147.00	1003.0
148.00	3200.0
149.00	621.0
152.90	533.0
154.00	325.0
155.00	1100.0
156.00	1272.0
157.00	384.0
157.90	366.0
158.90	164.0
159.90	637.0
161.00	640.0
165.00	858.0
165.90	566.0
167.00	5091.0
168.00	2425.0
169.00	498.0
170.00	181.0
170.90	228.0
172.00	447.0
173.10	406.0
173.90	784.0
175.10	1407.0
175.90	233.0
177.00	758.0
178.00	308.0

Instrument :
BNA_M
ClientSampleId :
DFTPP85

179.00	2894.0
180.00	2128.0
181.00	756.0
182.20	153.0
183.80	181.0
185.00	1222.0
186.00	9447.0
187.00	3006.0
188.10	337.0
189.10	622.0
189.80	193.0
190.90	313.0
192.00	802.0
192.90	1254.0
194.20	290.0
196.00	1359.0
197.10	355.0
197.90	54560.0
198.90	4272.0
199.90	428.0
201.70	181.0
202.90	721.0
204.00	2839.0
205.00	5240.0
206.00	18656.0
207.00	2808.0
208.00	1040.0
208.90	553.0
210.00	390.0
210.90	809.0
215.00	402.0
216.00	503.0
217.00	6068.0
217.90	1106.0
221.00	1536.0
222.10	701.0
222.90	1422.0
224.00	12585.0
225.00	2981.0
226.00	467.0
227.00	6611.0
228.00	1255.0
229.00	1142.0
229.90	184.0
231.10	371.0
232.80	152.0
233.90	501.0
234.90	418.0
235.80	260.0
236.10	257.0
237.00	310.0
238.80	306.0
240.00	158.0
240.90	186.0
242.00	667.0
243.00	693.0
244.00	9089.0
245.00	1463.0
246.00	2213.0
247.00	508.0
249.00	407.0
251.90	249.0
253.00	284.0
254.10	452.0
255.00	41432.0
256.00	6187.0
256.90	596.0
258.00	3451.0
258.90	712.0
265.00	1575.0
265.90	288.0
272.90	1059.0
274.00	3306.0
275.00	16440.0
275.90	2071.0
276.90	2208.0
277.80	269.0
278.10	313.0
280.90	202.0
282.90	300.0
284.90	373.0
293.00	366.0
296.00	7467.0
296.90	1163.0
303.00	553.0
303.90	216.0
308.10	202.0
313.90	302.0
314.90	877.0
316.00	335.0
321.00	276.0
323.10	1551.0
323.90	265.0
326.90	425.0
327.90	212.0
332.20	183.0

Instrument :
BNA_M
ClientSampleId :
DFTPP85

332.90	159.0
334.00	946.0
335.00	230.0
341.10	156.0
346.00	306.0
346.80	193.0
352.00	473.0
353.30	161.0
354.10	532.0
364.90	2949.0
366.00	300.0
372.10	969.0
373.00	183.0
383.00	175.0
390.00	199.0
402.00	216.0
403.00	362.0
404.00	156.0
421.00	249.0
421.90	312.0
423.00	2082.0
424.10	481.0
441.10	5201.0
442.10	36160.0
443.10	6633.0
444.00	723.0

Instrument :
BNA_M
ClientSampleId :
DFTPP85

m/z	Abundance
37.90	447.0
39.00	875.0
48.90	304.0
50.00	5524.0
51.00	16480.0
52.10	599.0
56.00	474.0
56.90	756.0
60.90	193.0
62.00	357.0
63.10	567.0
63.90	152.0
67.90	412.0
69.00	17056.0
73.00	427.0
74.00	3266.0
75.00	3365.0
76.00	1445.0
77.10	21752.0
78.00	1627.0
79.00	2262.0
79.90	1569.0
81.00	2322.0
82.10	317.0
85.10	217.0
86.00	710.0
87.10	300.0
90.90	427.0
91.90	430.0
93.00	4910.0
93.90	286.0
95.90	262.0
98.00	4381.0
99.00	2571.0
100.10	168.0
100.90	1028.0
102.90	450.0
104.00	1013.0
105.00	791.0
105.90	378.0
107.00	9102.0
108.00	938.0
108.90	258.0
110.00	10149.0
111.00	2051.0
112.10	373.0
116.10	474.0
117.00	19032.0
118.00	1145.0
122.10	938.0
122.90	929.0
123.90	545.0
124.90	258.0
127.00	18504.0
128.00	1693.0
129.00	10641.0
130.00	940.0
130.80	155.0
131.20	172.0
133.90	529.0
135.00	978.0
136.00	459.0
137.00	481.0
140.00	409.0
141.00	2147.0
141.90	590.0
142.90	484.0
146.00	520.0
147.00	1210.0
148.00	3635.0
149.00	580.0
151.00	186.0
152.10	204.0
152.90	835.0
154.00	583.0
155.00	1092.0
156.00	1929.0
157.10	211.0
158.10	397.0
158.90	422.0
160.00	810.0
161.00	1029.0
162.00	289.0
164.90	1053.0
166.00	673.0
167.00	6448.0
168.00	3454.0
169.00	575.0
169.80	159.0
171.00	154.0
171.80	424.0
172.90	708.0

Instrument :
BNA_M
ClientSampleId :
DFTPP85

174.00	1159.0
175.00	2037.0
175.90	430.0
177.00	886.0
178.00	405.0
179.00	4310.0
180.00	2219.0
181.00	1376.0
181.90	156.0
182.90	162.0
184.00	381.0
185.00	1885.0
186.00	12350.0
187.00	3871.0
188.00	473.0
188.90	1226.0
189.90	196.0
190.80	468.0
191.90	1187.0
193.00	1437.0
193.90	201.0
195.10	186.0
196.00	1878.0
196.70	369.0
197.00	355.0
198.00	69424.0
199.00	4553.0
200.10	400.0
200.90	161.0
201.40	278.0
202.90	1192.0
204.00	4055.0
205.00	7397.0
206.00	23808.0
207.00	3914.0
208.00	1436.0
208.90	495.0
209.90	492.0
211.10	1082.0
211.70	188.0
214.90	712.0
216.00	714.0
217.00	9069.0
217.90	1054.0
221.10	2270.0
222.00	1033.0
222.90	1720.0
224.00	15126.0
225.00	3933.0
226.10	323.0
227.00	9085.0
228.00	1218.0
228.90	1870.0
230.00	294.0
230.90	722.0
232.00	175.0
232.70	201.0
233.90	584.0
235.00	514.0
236.00	467.0
236.90	655.0
238.90	400.0
239.90	376.0
241.00	365.0
241.90	828.0
243.00	1073.0
244.00	11441.0
245.00	1692.0
246.00	3197.0
247.00	741.0
249.00	562.0
252.90	314.0
253.90	374.0
255.00	57480.0
256.00	7683.0
257.00	767.0
258.00	5445.0
258.90	754.0
265.00	2402.0
265.90	349.0
270.90	182.0
273.00	1853.0
274.00	4507.0
275.00	23536.0
276.00	3237.0
277.00	3202.0
277.90	502.0
283.10	302.0
283.90	265.0
284.80	335.0
292.90	569.0
295.00	185.0
296.00	10506.0
296.90	1565.0
303.10	940.0
304.20	255.0

Instrument :
BNA_M
ClientSampleId :
DFTPP85

307.90	202.0
309.90	216.0
314.10	377.0
314.90	1271.0
315.90	507.0
323.00	2382.0
324.00	449.0
327.00	485.0
327.90	294.0
332.20	166.0
332.80	155.0
334.00	1496.0
334.90	419.0
341.10	229.0
346.00	613.0
346.90	198.0
352.00	687.0
353.00	547.0
354.00	821.0
365.00	4779.0
366.00	679.0
372.00	1283.0
372.90	189.0
382.90	245.0
402.00	421.0
403.00	454.0
403.90	200.0
420.80	450.0
422.00	379.0
423.00	2764.0
424.00	938.0
441.10	9605.0
442.10	58176.0
443.10	11373.0
444.00	1289.0

Instrument :
BNA_M
ClientSampleId :
DFTPP85

m/z	Abundance
38.10	197.0
39.10	702.0
50.00	2924.0
51.00	9007.0
52.00	605.0
55.80	200.0
57.10	459.0
62.90	550.0
64.90	206.0
67.90	298.0
69.00	8906.0
73.20	236.0
74.00	1687.0
75.00	2459.0
76.10	823.0
77.10	11739.0
78.00	691.0
79.00	1123.0
79.90	974.0
80.90	817.0
81.90	329.0
83.10	155.0
85.80	502.0
86.90	286.0
90.70	220.0
91.90	337.0
93.00	2808.0
94.10	241.0
98.00	2957.0
99.10	1441.0
100.90	642.0
102.80	184.0
103.90	508.0
105.10	542.0
106.10	183.0
107.00	5391.0
108.10	659.0
109.20	253.0
110.00	6079.0
111.00	846.0
116.10	463.0
117.00	10231.0
118.00	849.0
122.00	399.0
122.90	461.0
123.90	265.0
125.10	288.0
127.00	11578.0
128.00	1222.0
129.00	7345.0
129.90	599.0
134.00	213.0
135.00	597.0
135.90	295.0
137.00	556.0
141.00	1154.0
141.90	494.0
142.90	201.0
146.20	269.0
147.00	570.0
147.90	2377.0
149.00	506.0
150.90	241.0
152.90	411.0
154.00	339.0
154.90	702.0
156.00	1042.0
158.00	280.0
158.90	232.0
160.00	476.0
160.90	546.0
165.00	705.0
165.80	658.0
167.00	4417.0
167.90	2104.0
168.90	334.0
171.00	221.0
172.00	321.0
172.80	378.0
173.90	717.0
175.00	1110.0
176.00	328.0
176.90	408.0
179.00	2668.0
180.00	1254.0
180.90	668.0
183.90	290.0
185.00	1219.0
186.00	8060.0
187.00	2330.0
187.90	356.0
189.00	493.0

Instrument :
BNA_M
ClientSampleId :
DFTPP85

191.10	323.0
192.00	814.0
193.00	1157.0
193.90	161.0
195.90	1098.0
197.00	453.0
197.90	45632.0
199.00	2702.0
199.90	376.0
201.40	209.0
203.00	623.0
204.00	2762.0
205.00	4270.0
206.00	16688.0
207.00	2436.0
208.00	989.0
209.00	172.0
210.00	452.0
210.90	964.0
215.00	333.0
215.90	542.0
217.00	6525.0
217.90	514.0
220.90	1518.0
221.90	693.0
222.90	1092.0
224.00	10425.0
225.10	2710.0
226.00	293.0
227.00	6353.0
228.00	859.0
228.90	1238.0
229.90	166.0
230.90	315.0
234.00	431.0
235.00	584.0
236.00	430.0
237.00	463.0
238.80	249.0
240.00	184.0
240.90	251.0
241.90	636.0
243.00	662.0
244.00	7580.0
245.00	959.0
246.00	2421.0
246.80	459.0
248.80	199.0
251.90	168.0
253.10	209.0
254.10	270.0
255.00	39216.0
256.00	6357.0
257.00	641.0
258.00	3954.0
258.90	553.0
265.00	1485.0
265.90	348.0
273.00	1310.0
274.00	3550.0
275.00	16552.0
276.00	2009.0
277.00	2201.0
277.90	441.0
280.90	189.0
282.90	357.0
284.00	150.0
284.90	329.0
292.80	440.0
295.20	153.0
296.00	7678.0
296.90	1085.0
300.90	208.0
303.00	956.0
303.80	157.0
313.90	367.0
314.90	1043.0
316.10	300.0
321.00	205.0
323.00	2067.0
324.10	433.0
326.90	514.0
328.10	187.0
332.00	220.0
333.10	342.0
334.10	1481.0
335.00	411.0
341.00	240.0
345.80	347.0
346.10	371.0
346.90	274.0
352.00	770.0
353.00	420.0
353.90	517.0
365.00	3361.0
366.00	466.0

Instrument :
BNA_M
ClientSampleId :
DFTPP85

372.00	977.0
373.00	255.0
383.10	217.0
402.10	444.0
403.10	690.0
420.90	423.0
421.90	437.0
423.00	2777.0
424.00	601.0
441.00	7897.0
442.10	51104.0
443.10	9710.0
444.10	883.0

Instrument :
BNA_M
ClientSampleId :
DFTPP85