

Data Path : Z:\HPCHEM1\BNA M\DATA\BM100317\
 Data File : BM011852.D
 Acq On : 03 Oct 2017 15:48
 Operator : SJ/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV15

Quant Time: Oct 03 17:00:56 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM100317.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 03 15:47:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	1,4-Dioxane	0.430	0.409	4.9	81	0.00
3 S	1,4-Dioxane-d8	0.423	0.382	9.7	78	0.00
4	Benzaldehyde	0.857	0.883	-3.0	84	0.00
5 S	Phenol-d5	1.721	1.575	8.5	83	0.00
6	Phenol	1.703	1.552	8.9	82	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.018	0.938	7.9	80	0.00
8	Bis(2-Chloroethyl)ether	1.279	1.211	5.3	82	0.00
9 S	2-Chlorophenol-d4	1.380	1.340	2.9	83	0.00
10	2-Chlorophenol	1.349	1.312	2.7	84	0.00
11	2-Methylphenol	1.295	1.198	7.5	83	0.00
12	2,2'-oxybis(1-Chloropropane	1.635	1.506	7.9	79	0.00
13 S	4-Methylphenol-d8	1.487	1.390	6.5	84	0.00
14	Acetophenone	2.177	2.088	4.1	84	0.00
15 P	N-Nitroso-di-n-propylamine	1.114	1.095	1.7	83	0.00
16	4-Methylphenol	1.450	1.366	5.8	84	0.00
17	Hexachloroethane	0.541	0.519	4.1	83	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
19 S	Nitrobenzene-d5	0.142	0.144	-1.4	84	0.00
20	Nitrobenzene	0.344	0.345	-0.3	82	0.00
21	Isophorone	0.642	0.641	0.2	83	0.00
22 S	2-Nitrophenol-d4	0.160	0.164	-2.5	86	0.00
23 C	2-Nitrophenol	0.169	0.174	-3.0	85	0.00
24	2,4-Dimethylphenol	0.362	0.359	0.8	84	0.00
25	Bis(2-Chloroethoxy)methane	0.394	0.386	2.0	82	0.00
26 S	2,4-Dichlorophenol-d3	0.324	0.330	-1.9	85	0.00
27 C	2,4-Dichlorophenol	0.313	0.316	-1.0	85	0.00
28	Naphthalene	1.003	0.992	1.1	85	0.00
29 S	4-Chloroaniline-d4	0.354	0.408	-15.3	85	0.00
30	4-Chloroaniline	0.347	0.390	-12.4	83	0.00
31 C	Hexachlorobutadiene	0.230	0.229	0.4	84	0.00
32	Caprolactam	0.103	0.099	3.9	87	0.00
33 C	4-Chloro-3-methylphenol	0.340	0.337	0.9	83	0.00
34	2-Methylnaphthalene	0.765	0.765	0.0	85	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
36	1,2,4,5-Tetrachlorobenzene	0.678	0.669	1.3	86	0.00
37	Hexachlorocyclopentadiene	0.395	0.346	12.4	81	0.00
38 C	2,4,6-Trichlorophenol	0.431	0.431	0.0	85	0.00
39	2,4,5-Trichlorophenol	0.454	0.464	-2.2	88	0.00
40	1,1'-Biphenyl	1.599	1.570	1.8	86	0.00
41	2-Chloronaphthalene	1.251	1.235	1.3	86	0.00
42	2-Nitroaniline	0.318	0.317	0.3	82	0.00
43 S	Dimethylphthalate-d6	1.733	1.721	0.7	85	0.00
44	Dimethylphthalate	1.668	1.655	0.8	86	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM100317\
 Data File : BM011852.D
 Acq On : 03 Oct 2017 15:48
 Operator : SJ/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV15

Quant Time: Oct 03 17:00:56 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM100317.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 03 15:47:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.314	0.325	-3.5	86	0.00
46 S	Acenaphthylene-d8	2.037	2.053	-0.8	86	0.00
47	Acenaphthylene	1.975	1.983	-0.4	86	0.00
48	3-Nitroaniline	0.292	0.282	3.4	85	0.00
49 C	Acenaphthene	1.367	1.329	2.8	85	0.00
50	2,4-Dinitrophenol	0.208	0.177	14.9	88	0.00
51 S	4-Nitrophenol-d4	0.303	0.284	6.3	88	0.00
52	4-Nitrophenol	0.242	0.226	6.6	85	0.00
53	Dibenzofuran	1.984	1.976	0.4	87	0.00
54	2,4-Dinitrotoluene	0.465	0.492	-5.8	89	0.00
55	2,3,4,6-Tetrachlorophenol	0.443	0.450	-1.6	87	0.00
56	Diethylphthalate	1.693	1.702	-0.5	87	0.00
57 S	Fluorene-d10	1.530	1.523	0.5	88	0.00
58	Fluorene	1.656	1.647	0.5	87	0.00
59	4-Chlorophenyl-phenylether	0.876	0.872	0.5	87	0.00
60	4-Nitroaniline	0.340	0.301	11.5	87	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.135	0.123	8.9	90	0.00
63	4,6-Dinitro-2-methylphenol	0.137	0.128	6.6	91	0.00
64	N-Nitrosodiphenylamine	0.578	0.571	1.2	87	0.00
65	4-Bromophenyl-phenylether	0.224	0.219	2.2	88	0.00
66	Hexachlorobenzene	0.250	0.240	4.0	86	0.00
67	Atrazine	0.229	0.224	2.2	89	0.00
68 C	Pentachlorophenol	0.158	0.145	8.2	89	0.00
69	Phenanthrene	1.100	1.075	2.3	88	0.00
70 S	Anthracene-d10	0.986	0.967	1.9	87	0.00
71	Anthracene	1.119	1.120	-0.1	88	0.00
72	Carbazole	0.970	0.924	4.7	88	0.00
73	Di-n-butylphthalate	1.162	1.151	0.9	86	0.00
74 C	Fluoranthene	1.342	1.347	-0.4	89	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
76 S	Pyrene-d10	0.902	0.891	1.2	90	0.00
77	Pyrene	1.109	1.089	1.8	89	0.00
78	Butylbenzylphthalate	0.434	0.427	1.6	89	0.00
79	3,3'-Dichlorobenzidine	0.372	0.362	2.7	95	0.00
80	Benzo(a)anthracene	1.146	1.145	0.1	94	0.00
81	Bis(2-ethylhexyl)phthalate	0.646	0.632	2.2	90	0.00
82	Chrysene	1.089	1.071	1.7	92	0.00
83 I	Perylene-d12	1.000	1.000	0.0	107	0.00
84	Di-n-octyl phthalate	1.168	1.141	2.3	94	0.00
85	Benzo(b)fluoranthene	1.216	1.239	-1.9	101	0.00
86	Benzo(k)fluoranthene	1.214	1.162	4.3	96	0.00
87 S	Benzo(a)pyrene-d12	0.962	0.965	-0.3	102	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM100317\
Data File : BM011852.D
Acq On : 03 Oct 2017 15:48
Operator : SJ/JU
Sample : SSTDICV020
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SICV15

Quant Time: Oct 03 17:00:56 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM100317.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 03 15:47:54 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.168	1.168	0.0	101	0.00
89	Indeno(1,2,3-cd)pyrene	1.250	1.269	-1.5	110	0.00
90	Dibenzo(a,h)anthracene	1.027	1.060	-3.2	111	0.01
91	Benzo(a,h,i)perylene	1.034	1.039	-0.5	110	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0