


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM100416\
Data File : BM007647.D
Acq On : 04 Oct 2016 09:35
Operator : UM/SJ
Sample : DFTPP35
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP35

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM041416.M
Title : SVOA CALIBRATION
Last Update : Sat Apr 23 01:00:10 2016



AutoFind: Scans 2548, 2549, 2550; Background Corrected with Scan 2543

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.8	88773	PASS
68	69	0.00	2	1.9	1876	PASS
69	198	0.00	100	37.4	101162	PASS
70	69	0.00	2	0.6	630	PASS
127	198	10	80	44.3	119965	PASS
197	198	0.00	2	0.6	1616	PASS
198	198	100	100	100.0	270592	PASS
199	198	5	9	6.7	18103	PASS
275	198	10	60	30.3	82042	PASS
365	198	1	100	4.2	11420	PASS
441	443	0.01	100	81.9	31522	PASS
442	198	50	100	74.7	202000	PASS
443	442	15	24	19.1	38482	PASS

m/z	Abundance
38.10	751.0
39.10	4498.0
40.00	625.0
41.10	311.0
44.10	824.0
49.00	606.0
50.10	18912.0
51.10	70152.0
52.00	4630.0
54.90	424.0
56.00	1918.0
57.00	5512.0
58.00	397.0
61.00	1128.0
62.10	834.0
63.00	2872.0
64.00	630.0
65.10	1302.0
68.10	1859.0
69.00	79696.0
70.10	456.0
73.00	935.0
74.00	8786.0
75.00	13073.0
76.00	4541.0
77.10	89840.0
78.10	6974.0
79.00	6640.0
80.00	5963.0
81.00	6713.0
81.90	1840.0
83.00	1741.0
84.00	453.0
85.00	1188.0
85.90	1978.0
87.00	1103.0
87.90	363.0
89.00	336.0
91.00	1323.0
92.00	2103.0
93.00	11909.0
94.00	678.0
96.00	578.0
98.00	8625.0
99.00	9278.0
100.10	718.0
101.10	3208.0
103.00	1464.0
103.90	2037.0
105.10	2714.0
106.20	735.0
107.00	28672.0
108.00	4609.0
108.90	526.0
110.00	48696.0
111.00	8915.0
112.00	962.0
113.10	469.0
116.00	1605.0
117.00	27464.0
118.10	2020.0
122.10	2060.0
123.00	2906.0
123.90	1503.0
125.00	1258.0
126.10	496.0
127.00	89616.0
128.00	7433.0
129.00	37080.0
130.00	3393.0
131.00	792.0
132.00	480.0
132.90	322.0
134.00	989.0
135.00	2819.0
136.00	1468.0
137.00	1960.0
137.90	420.0
140.10	589.0
141.00	5728.0
142.00	1711.0
142.90	1195.0
144.00	464.0
145.00	307.0
146.00	811.0
147.00	2708.0
148.00	6619.0
149.00	1647.0
150.00	618.0
151.00	739.0
153.00	1623.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

154.00	1480.0
155.00	2855.0
156.00	3950.0
157.00	804.0
157.90	799.0
159.00	945.0
160.00	1920.0
161.00	2335.0
162.00	725.0
164.90	1803.0
166.00	1531.0
167.00	10410.0
168.00	8128.0
169.00	1274.0
171.00	613.0
172.00	1093.0
173.00	1422.0
174.10	2256.0
175.00	3877.0
176.00	989.0
177.00	1766.0
177.90	704.0
179.00	7072.0
180.00	6060.0
181.00	2799.0
181.90	427.0
184.00	593.0
185.00	3260.0
186.00	28032.0
187.00	8638.0
188.10	1536.0
189.00	1937.0
190.10	326.0
190.90	1374.0
192.00	2395.0
193.10	2995.0
193.90	570.0
196.00	4662.0
197.10	1058.0
198.00	182912.0
199.00	12607.0
199.90	1013.0
201.60	781.0
203.00	1331.0
204.00	7248.0
205.00	12343.0
206.00	51016.0
207.00	9087.0
208.10	2480.0
209.10	1089.0
210.00	771.0
211.00	2278.0
211.80	561.0
215.00	865.0
215.90	1207.0
217.00	13985.0
217.90	1793.0
219.00	342.0
221.00	6715.0
221.80	1670.0
222.90	2580.0
224.00	28272.0
225.00	8358.0
226.00	1200.0
227.00	12424.0
228.00	1924.0
229.00	2255.0
230.10	316.0
231.00	1235.0
232.80	331.0
234.00	732.0
234.90	909.0
236.00	566.0
237.00	2124.0
237.90	374.0
238.90	574.0
239.90	653.0
241.00	709.0
242.00	1227.0
243.00	1580.0
244.00	20656.0
245.00	2497.0
246.00	4852.0
247.10	678.0
247.90	337.0
248.90	795.0
250.90	309.0
252.90	501.0
254.00	758.0
255.00	111864.0
256.00	17496.0
257.10	1795.0
258.00	6289.0
259.00	1364.0
260.00	373.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

265.00	3082.0
266.00	411.0
273.00	3687.0
274.00	8784.0
275.00	51608.0
276.00	6460.0
277.00	5039.0
278.00	1010.0
281.00	689.0
283.00	978.0
284.00	330.0
285.00	838.0
291.90	311.0
292.90	1068.0
293.90	427.0
295.00	360.0
296.00	14595.0
297.00	2253.0
302.00	323.0
303.00	1657.0
304.00	516.0
309.90	347.0
314.00	811.0
315.00	2211.0
316.00	1121.0
321.00	517.0
323.10	4235.0
324.00	996.0
327.00	889.0
328.00	461.0
333.00	585.0
334.10	2780.0
334.90	662.0
340.90	537.0
345.90	870.0
352.00	1392.0
353.00	730.0
354.00	1262.0
355.10	589.0
365.00	5461.0
365.90	788.0
371.00	351.0
372.00	2078.0
373.10	660.0
383.00	434.0
401.90	872.0
403.00	1181.0
403.90	424.0
420.90	1251.0
421.90	798.0
423.00	6021.0
424.10	2471.0
425.00	327.0
441.10	14438.0
442.10	102576.0
443.00	20072.0
444.00	1785.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

m/z	Abundance
37.00	397.0
38.10	1205.0
39.10	7158.0
40.00	774.0
43.90	936.0
49.10	959.0
50.10	28112.0
51.10	108848.0
52.00	5442.0
53.00	366.0
55.00	722.0
56.00	3486.0
57.00	7271.0
58.10	410.0
61.00	1413.0
62.10	1690.0
63.00	5050.0
64.00	958.0
65.10	2195.0
68.10	1831.0
69.00	122608.0
70.00	763.0
73.00	2102.0
74.10	13608.0
75.00	20840.0
76.10	6575.0
77.10	140288.0
78.10	10221.0
79.00	10868.0
80.00	8540.0
81.00	11014.0
82.00	2440.0
83.00	2087.0
84.00	434.0
85.10	1783.0
86.00	3369.0
87.00	1407.0
88.10	805.0
91.00	2327.0
92.00	2631.0
93.00	18016.0
94.00	1401.0
95.00	319.0
96.00	733.0
97.10	467.0
98.00	14464.0
99.00	12095.0
100.10	857.0
101.00	6149.0
103.00	1743.0
104.00	4170.0
105.00	3557.0
106.00	1607.0
107.00	46696.0
108.00	6657.0
109.10	1383.0
110.00	76376.0
111.00	12864.0
112.00	2026.0
113.00	526.0
116.00	3098.0
117.00	42488.0
118.10	3278.0
118.90	449.0
120.20	478.0
122.00	3179.0
123.00	4305.0
124.00	2315.0
125.10	2004.0
126.20	798.0
127.00	143104.0
128.00	11327.0
129.00	58168.0
130.00	5309.0
131.00	934.0
131.90	620.0
132.90	325.0
133.90	1885.0
135.00	4861.0
136.00	1720.0
137.00	3079.0
138.10	603.0
139.00	376.0
139.90	890.0
141.00	7242.0
142.00	2639.0
143.00	1408.0
143.90	458.0
145.10	520.0
146.10	1392.0
147.00	4864.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

148.00	10807.0
149.00	2814.0
150.00	539.0
151.00	1223.0
151.90	917.0
153.00	2309.0
154.00	1947.0
155.00	4331.0
156.00	7140.0
157.10	1657.0
158.00	1577.0
159.00	1553.0
160.00	2758.0
160.90	3464.0
162.00	1194.0
162.90	336.0
164.00	367.0
165.00	3529.0
166.00	2548.0
167.00	16061.0
168.00	11906.0
169.00	1926.0
170.00	458.0
171.00	765.0
171.90	1777.0
173.00	1945.0
174.00	4282.0
175.00	6502.0
176.10	1570.0
177.00	3219.0
178.00	1560.0
179.00	13623.0
180.00	10176.0
181.00	4743.0
181.90	432.0
182.10	431.0
182.80	466.0
183.90	1215.0
185.00	5596.0
186.10	47048.0
187.00	14491.0
188.00	2175.0
189.00	3388.0
189.90	500.0
191.00	1877.0
192.00	4083.0
193.00	4929.0
194.00	1473.0
195.00	402.0
196.00	8622.0
196.90	1737.0
198.00	326400.0
199.00	21480.0
200.00	2008.0
201.60	1243.0
203.00	2794.0
204.00	13528.0
205.00	21328.0
206.10	95040.0
207.00	14835.0
208.00	3663.0
208.90	1435.0
209.90	1724.0
211.10	3118.0
211.90	883.0
213.10	327.0
215.00	1434.0
216.00	2243.0
217.00	26344.0
217.90	3796.0
221.00	10772.0
221.90	2809.0
223.00	5536.0
224.00	52128.0
225.00	13543.0
226.10	1891.0
227.00	25552.0
228.00	3388.0
229.00	4094.0
230.00	406.0
231.00	2159.0
232.80	424.0
234.00	1539.0
235.00	2043.0
236.00	1421.0
237.00	3259.0
238.00	521.0
239.00	766.0
240.00	1062.0
240.90	1244.0
242.00	3393.0
243.00	2960.0
244.00	36816.0
245.10	6056.0
246.00	8710.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

247.00	1900.0
248.00	348.0
249.00	1579.0
251.00	445.0
252.00	459.0
252.90	1082.0
254.10	1688.0
255.00	215808.0
256.00	32208.0
257.00	3046.0
258.00	13478.0
259.00	2255.0
261.00	317.0
263.90	432.0
265.00	5324.0
265.90	873.0
270.90	500.0
272.00	416.0
273.00	5536.0
274.00	15456.0
275.00	95024.0
276.00	13134.0
277.00	9389.0
278.00	1882.0
279.00	306.0
281.10	475.0
281.90	440.0
283.00	914.0
284.10	767.0
285.10	1596.0
289.80	428.0
293.00	2013.0
294.00	548.0
295.20	543.0
296.00	28288.0
297.00	4063.0
302.00	379.0
303.00	3109.0
304.00	736.0
307.90	456.0
314.10	1179.0
314.90	3110.0
316.00	1616.0
317.10	446.0
320.90	872.0
321.90	418.0
323.00	8823.0
324.10	1407.0
326.90	1368.0
328.00	741.0
333.00	672.0
334.00	5848.0
334.90	1528.0
340.90	949.0
346.00	1798.0
346.90	520.0
352.00	2509.0
353.10	2156.0
354.10	3072.0
355.10	628.0
365.00	12850.0
366.00	2151.0
371.00	558.0
372.00	3583.0
373.00	814.0
383.00	986.0
384.10	392.0
390.00	386.0
390.90	382.0
402.00	2141.0
403.10	2210.0
403.90	891.0
421.00	1898.0
422.10	1810.0
423.10	12754.0
424.10	4010.0
425.00	703.0
440.10	320.0
441.00	36856.0
442.00	228288.0
443.00	43216.0
444.10	4141.0
444.90	339.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

m/z	Abundance
37.00	494.0
38.00	1380.0
39.10	6301.0
40.00	666.0
41.10	340.0
44.00	1109.0
49.10	640.0
50.10	22648.0
51.10	87320.0
52.00	4026.0
53.00	337.0
55.00	377.0
56.00	2804.0
57.00	5608.0
61.00	1217.0
62.10	1541.0
63.00	3549.0
64.00	649.0
65.10	1986.0
68.00	1940.0
69.00	101184.0
70.00	672.0
73.10	1313.0
74.10	11386.0
75.10	17592.0
76.10	5900.0
77.10	111992.0
78.10	7949.0
79.10	8667.0
80.00	6425.0
81.00	8908.0
82.00	1763.0
82.90	1862.0
84.00	405.0
85.00	1594.0
86.00	2780.0
87.00	1362.0
88.00	360.0
91.00	1958.0
92.00	2562.0
93.00	15649.0
94.00	1185.0
94.90	364.0
96.00	657.0
98.00	12827.0
99.00	10375.0
99.90	964.0
101.00	4918.0
102.90	1570.0
104.00	3807.0
105.00	3013.0
106.00	1076.0
107.00	39832.0
108.00	5808.0
109.00	1470.0
110.00	66984.0
111.00	10158.0
112.00	1220.0
113.00	828.0
116.00	2662.0
117.00	37456.0
118.10	3038.0
118.90	413.0
120.00	606.0
121.00	317.0
122.00	2863.0
123.00	4126.0
123.90	1733.0
125.00	1642.0
126.20	666.0
127.00	127176.0
128.00	9480.0
129.00	51752.0
130.00	4348.0
131.00	1172.0
133.00	367.0
134.00	1551.0
135.10	4726.0
136.00	1581.0
137.10	2528.0
137.90	659.0
140.00	634.0
141.00	7662.0
142.00	2530.0
143.10	1728.0
144.00	470.0
145.10	374.0
146.10	1263.0
147.00	3867.0
148.00	9894.0
148.90	2098.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

150.00	473.0
151.00	834.0
151.90	636.0
153.00	2600.0
154.00	1882.0
155.00	4298.0
156.00	6927.0
157.00	1141.0
157.90	1372.0
158.90	1230.0
160.00	2747.0
161.00	3355.0
162.00	1010.0
162.90	513.0
164.00	517.0
165.00	2912.0
166.00	2422.0
167.00	15487.0
168.00	8966.0
169.10	1787.0
169.90	575.0
171.00	569.0
171.90	1387.0
173.00	2229.0
174.00	3091.0
175.10	6596.0
176.00	1624.0
177.00	3067.0
178.10	966.0
179.00	12059.0
180.00	8208.0
181.00	3739.0
181.90	690.0
184.00	867.0
185.00	6478.0
186.10	44400.0
187.00	13172.0
188.10	1710.0
189.00	3459.0
189.90	442.0
190.90	1453.0
192.00	4235.0
193.00	3904.0
194.00	1152.0
196.00	7401.0
197.00	2053.0
198.00	302464.0
199.00	20224.0
200.00	1522.0
201.50	1135.0
203.00	2517.0
204.10	12753.0
205.00	20216.0
206.00	90976.0
207.10	13157.0
208.00	4055.0
209.00	1442.0
210.10	1183.0
211.00	3922.0
212.00	692.0
215.00	1052.0
216.00	2171.0
217.00	25984.0
218.00	2529.0
218.90	338.0
220.00	497.0
221.00	11297.0
221.90	3487.0
223.00	5889.0
224.00	52248.0
225.00	13469.0
226.10	2057.0
227.00	22528.0
228.00	3273.0
229.00	4794.0
229.90	588.0
231.00	1891.0
232.00	305.0
233.00	432.0
234.00	1501.0
235.00	1812.0
236.00	1487.0
237.00	2989.0
237.90	449.0
239.00	1120.0
240.10	858.0
241.00	1444.0
242.00	2612.0
243.00	2500.0
244.10	37176.0
245.00	5532.0
246.00	9426.0
247.00	1819.0
247.90	523.0
249.00	1617.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

249.80	372.0
251.00	391.0
252.90	1236.0
254.10	1248.0
255.00	219136.0
256.00	33424.0
257.00	3076.0
258.00	13645.0
259.00	2106.0
260.00	346.0
261.00	515.0
265.00	6121.0
265.90	743.0
269.90	303.0
271.10	532.0
272.10	365.0
273.00	6805.0
274.00	17120.0
275.00	99496.0
276.00	13597.0
277.00	9366.0
278.00	2303.0
279.00	497.0
281.10	560.0
282.20	436.0
282.90	1144.0
283.90	792.0
285.00	1897.0
286.00	530.0
290.00	413.0
291.00	373.0
292.90	2505.0
294.00	919.0
295.00	897.0
296.00	33592.0
297.00	3667.0
298.00	471.0
300.90	341.0
302.00	586.0
303.10	3295.0
304.10	757.0
308.00	546.0
309.90	371.0
314.00	1660.0
315.00	3080.0
316.00	1522.0
317.00	444.0
320.90	607.0
322.00	321.0
323.00	9112.0
324.00	1670.0
327.00	2153.0
328.00	876.0
332.00	438.0
333.00	889.0
334.10	6442.0
335.10	1668.0
336.00	336.0
341.00	1099.0
346.00	1962.0
347.00	382.0
352.00	2737.0
353.00	2065.0
354.00	2953.0
355.00	594.0
365.00	15951.0
366.00	1999.0
371.00	779.0
372.00	4908.0
373.00	1150.0
382.90	1201.0
384.00	415.0
389.90	461.0
391.00	342.0
401.10	414.0
402.00	1826.0
403.10	2858.0
404.10	884.0
421.10	2223.0
422.00	2361.0
423.10	14230.0
424.10	5070.0
425.10	554.0
441.10	43272.0
442.10	275136.0
443.00	52160.0
444.00	5071.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

Scan 2543 (17.545 min): BM007647.D
DFTPP35

m/z	Abundance
40.00	434.0
43.90	1022.0
73.00	444.0
147.00	371.0
193.00	354.0
207.10	2311.0
207.90	443.0
221.10	310.0
281.00	1095.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35