

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM100418\
 Data File : BM017023.D
 Acq On : 05 Oct 2018 05:15
 Operator : MJ/SJ
 Sample : J5157-15
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AD7

Quant Time: Oct 12 08:21:04 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM100418MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 05 05:34:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	253754	20.00	ng/ul	0.00
18) Naphthalene-d8	10.47	136	1086095	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.34	164	580187	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.09	188	1249333	20.00	ng/ul	0.00
78) Chrysene-d12	21.27	240	1255699	20.00	ng/ul	0.00
86) Perylene-d12	23.50	264	1306489	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.25	96	25736	5.04	ng/uL	0.00
5) Phenol-d5	6.87	99	639421	30.14	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.04	67	397524	30.90	ng/ul	0.00
9) 2-Chlorophenol-d4	7.23	132	537637	30.48	ng/ul	0.00
13) 4-Methylphenol-d8	8.40	113	523688	31.51	ng/ul	0.00
19) Nitrobenzene-d5	8.85	128	259516	32.69	ng/ul	0.00
22) 2-Nitrophenol-d4	9.56	143	269163	33.33	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.10	165	509824	30.68	ng/ul	0.00
29) 4-Chloroaniline-d4	10.62	131	705520	35.34	ng/ul	0.00
44) Dimethylphthalate-d6	13.75	166	1552691	33.73	ng/ul	0.00
47) Acenaphthylene-d8	14.03	160	2010455	34.17	ng/ul	0.00
52) 4-Nitrophenol-d4	14.54	143	183960	21.15	ng/ul	0.00
58) Fluorene-d10	15.33	176	1366453	33.33	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.46	200	141621	19.22	ng/ul	0.00
71) Anthracene-d10	17.19	188	1986725	33.29	ng/ul	0.00
79) Pyrene-d10	19.48	212	2100083	37.07	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.36	264	2287533	34.20	ng/ul	0.00

Target Compounds					Qvalue
6) Phenol	6.90	94	29502	1.390 ng/ul	94
45) Dimethylphthalate	13.80	163	318036	6.977 ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM100418\
Data File : BM017023.D
Acq On : 05 Oct 2018 05:15
Operator : MJ/SJ
Sample : J5157-15
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AD7

Quant Time: Oct 12 08:21:04 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM100418MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Oct 05 05:34:09 2018
Response via : Initial Calibration

