


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100418\
Data File : BM016993.D
Acq On : 04 Oct 2018 10:28
Operator : MJ/SJ
Sample : DFTPP12
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP12

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM100418MA.M
Title : SVOA CALIBRATION
Last Update : Fri Oct 05 05:34:09 2018



AutoFind: Scans 2473, 2474, 2475; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.1	90869	PASS
68	69	0.00	2	1.5	1383	PASS
69	198	0.00	100	34.3	94386	PASS
70	69	0.00	2	0.6	535	PASS
127	198	10	80	44.9	123421	PASS
197	198	0.00	2	0.6	1740	PASS
198	198	100	100	100.0	274922	PASS
199	198	5	9	6.7	18335	PASS
275	198	10	60	29.3	80650	PASS
365	198	1	100	3.9	10805	PASS
441	198	0.01	100	14.0	38378	PASS
442	198	50	100	90.2	248064	PASS
443	442	15	24	19.1	47384	PASS

m/z	Abundance
37.00	381.0
38.10	1374.0
39.10	7659.0
39.90	1126.0
41.00	739.0
42.00	518.0
43.00	1420.0
44.00	10634.0
45.00	1296.0
49.00	1274.0
50.00	26816.0
51.00	107176.0
52.00	5448.0
53.00	514.0
55.00	854.0
56.00	3569.0
57.00	8298.0
57.90	463.0
60.00	562.0
60.90	1483.0
62.10	1577.0
63.00	4174.0
63.90	594.0
65.00	2138.0
68.00	2043.0
69.00	109424.0
70.00	606.0
73.00	1035.0
74.00	12136.0
75.00	18152.0
76.10	6054.0
77.10	122216.0
78.00	8405.0
79.00	8603.0
80.00	6412.0
81.00	8900.0
82.00	2324.0
83.00	2032.0
83.90	358.0
85.00	1551.0
86.00	2613.0
87.00	1062.0
87.80	326.0
89.00	391.0
91.00	1987.0
92.00	2212.0
93.00	14139.0
93.90	1304.0
96.10	783.0
97.20	419.0
98.00	11112.0
99.00	9884.0
100.00	992.0
101.00	5072.0
102.90	1430.0
104.00	3523.0
105.00	2960.0
106.10	922.0
107.00	37000.0
108.00	5840.0
109.10	1626.0
110.00	68552.0
111.00	11341.0
112.00	1753.0
112.90	360.0
116.10	2135.0
117.00	29168.0
118.00	2781.0
120.00	613.0
122.00	2603.0
122.90	3759.0
123.90	2113.0
125.00	1561.0
126.20	801.0
127.00	137088.0
128.00	10482.0
129.00	51168.0
130.00	4498.0
131.00	1032.0
131.90	368.0
132.90	548.0
134.00	1773.0
135.00	4282.0
136.00	1776.0
137.10	2448.0
137.90	450.0
138.80	379.0
140.00	879.0
141.00	7394.0
141.90	2161.0
143.00	1534.0

Instrument :
BNA_M
ClientSampleId :
DFTPP12

143.90	495.0
145.00	395.0
146.00	1290.0
147.00	3942.0
148.00	7697.0
149.00	2165.0
150.00	437.0
151.00	797.0
151.90	538.0
153.10	2055.0
154.00	1695.0
155.00	3761.0
156.00	5983.0
156.90	1177.0
157.90	1331.0
159.10	1108.0
160.00	2069.0
161.00	3112.0
162.00	1056.0
163.80	420.0
165.00	2580.0
166.00	2253.0
167.00	13306.0
168.00	8232.0
168.90	1584.0
170.00	467.0
170.90	759.0
172.00	1138.0
173.00	1925.0
174.00	2917.0
175.00	5716.0
176.20	1226.0
177.00	2766.0
178.00	907.0
178.90	9991.0
180.00	7138.0
181.00	3145.0
181.90	654.0
184.00	719.0
185.00	4876.0
186.00	37984.0
187.00	10956.0
188.00	1220.0
188.90	2578.0
190.00	487.0
191.00	1514.0
192.00	2829.0
193.00	3500.0
193.90	661.0
195.10	368.0
196.00	8702.0
197.10	2016.0
197.90	281152.0
199.00	19544.0
200.00	1398.0
201.40	1417.0
202.90	1856.0
204.00	10699.0
205.00	17640.0
206.00	75320.0
207.00	11605.0
208.00	2943.0
209.00	815.0
210.30	1383.0
211.10	3244.0
211.80	745.0
215.00	832.0
216.00	1602.0
217.00	21096.0
218.00	2425.0
219.80	373.0
221.00	12193.0
221.90	2294.0
223.00	4761.0
224.00	41856.0
225.00	10296.0
226.10	1398.0
227.00	17512.0
228.00	2452.0
229.00	3365.0
230.00	679.0
231.00	1843.0
232.90	380.0
233.90	865.0
235.00	1349.0
236.00	885.0
237.00	1923.0
239.00	707.0
240.00	460.0
240.90	1111.0
242.00	2710.0
243.10	2377.0
244.00	31912.0
245.00	4487.0
246.00	6234.0

Instrument :
BNA_M
ClientSampleId :
DFTPP12

247.00	1582.0
248.00	366.0
249.00	1314.0
251.00	363.0
252.00	336.0
252.90	673.0
254.10	1252.0
255.00	173824.0
256.00	25272.0
257.00	1903.0
258.00	9262.0
259.10	1793.0
260.80	338.0
265.00	3724.0
265.90	595.0
270.90	821.0
271.90	407.0
273.00	4860.0
274.00	12702.0
275.00	77072.0
276.00	10647.0
277.00	6505.0
278.00	1334.0
281.00	593.0
282.90	812.0
284.00	408.0
285.00	946.0
289.00	316.0
292.00	326.0
292.90	1697.0
294.00	639.0
295.10	437.0
296.00	21416.0
297.00	3061.0
298.00	312.0
302.00	372.0
303.00	2894.0
303.90	567.0
308.00	419.0
312.90	331.0
314.10	903.0
315.00	2337.0
316.10	1673.0
321.00	562.0
322.00	408.0
323.00	6611.0
323.90	1294.0
327.00	1358.0
328.00	776.0
332.00	442.0
332.90	469.0
334.10	4721.0
335.00	771.0
341.10	997.0
345.90	1122.0
347.20	437.0
352.10	1780.0
353.00	1411.0
354.00	2187.0
355.00	371.0
365.00	8873.0
366.00	1793.0
371.00	571.0
372.00	3120.0
373.10	849.0
383.00	1106.0
390.00	543.0
391.00	353.0
402.00	1304.0
403.00	1644.0
404.00	837.0
421.00	2091.0
422.00	1651.0
423.00	11378.0
424.00	3708.0
425.00	319.0
441.00	29248.0
442.10	196352.0
443.10	37712.0
444.10	3549.0

Instrument :
BNA_M
ClientSampleId :
DFTPP12

m/z	Abundance
37.00	462.0
38.00	1435.0
39.00	7694.0
40.00	1240.0
41.10	681.0
41.90	430.0
43.00	1317.0
44.00	10534.0
45.10	1365.0
49.10	1259.0
50.00	27056.0
51.00	107048.0
52.00	6038.0
52.90	335.0
55.10	723.0
56.00	3442.0
57.00	7763.0
58.00	368.0
60.00	564.0
61.00	1426.0
62.00	1572.0
63.00	4646.0
64.00	741.0
65.10	2271.0
68.00	2107.0
69.00	111032.0
69.90	683.0
73.00	1101.0
74.00	10740.0
75.00	18376.0
76.10	6466.0
77.10	122968.0
78.10	8335.0
79.00	9307.0
80.00	6495.0
81.00	9930.0
82.00	2234.0
83.00	2137.0
83.90	324.0
85.00	1737.0
86.00	2638.0
87.00	1283.0
87.90	465.0
89.00	323.0
90.90	2175.0
92.00	2426.0
93.00	14601.0
94.00	1048.0
95.10	349.0
96.10	811.0
96.90	302.0
98.00	11290.0
99.00	9648.0
100.00	801.0
101.00	5301.0
102.90	1992.0
104.00	3033.0
105.00	3519.0
106.10	1218.0
107.00	37832.0
108.00	5895.0
109.10	884.0
110.00	73160.0
111.00	11257.0
112.00	1323.0
113.00	455.0
116.00	2175.0
117.00	32312.0
118.00	2322.0
119.00	436.0
119.90	442.0
122.00	2384.0
123.00	4254.0
123.90	2077.0
125.00	2031.0
127.00	146048.0
128.00	10907.0
129.00	58280.0
130.00	4897.0
131.10	977.0
131.90	512.0
132.80	321.0
134.00	1757.0
135.00	4338.0
136.00	1810.0
137.00	2352.0
137.80	430.0
138.90	325.0
139.90	948.0
141.00	7863.0
142.10	2121.0

Instrument :
BNA_M
ClientSampleId :
DFTPP12

142.90	1563.0
144.00	420.0
145.00	352.0
146.00	1336.0
147.00	4088.0
148.00	8567.0
149.00	2275.0
150.00	572.0
151.00	952.0
151.90	671.0
153.00	2764.0
154.00	2024.0
155.00	3987.0
156.00	6102.0
157.10	1263.0
157.90	1473.0
158.90	1150.0
160.00	2386.0
161.00	3604.0
162.00	1065.0
164.00	629.0
165.00	3334.0
166.00	2218.0
167.00	14666.0
168.00	8987.0
169.00	1520.0
170.00	680.0
170.90	801.0
172.00	1169.0
173.00	1998.0
174.00	3578.0
175.00	5568.0
176.00	1907.0
177.00	2820.0
178.00	934.0
179.00	11925.0
180.00	8173.0
181.00	3919.0
182.00	522.0
182.90	429.0
183.90	1014.0
185.00	5866.0
186.00	44256.0
187.00	12413.0
188.00	1224.0
189.00	3180.0
189.90	486.0
191.00	1540.0
192.00	3687.0
193.00	4147.0
194.00	924.0
195.10	341.0
196.00	9928.0
196.70	2426.0
197.10	1934.0
198.00	333760.0
199.00	21800.0
199.90	1874.0
201.40	1862.0
203.00	2613.0
204.00	11470.0
205.00	21112.0
206.00	87440.0
207.00	13125.0
208.00	3624.0
209.00	1300.0
210.20	1682.0
211.00	3488.0
211.80	757.0
214.90	866.0
216.00	1762.0
217.00	23312.0
218.00	3406.0
218.90	305.0
221.00	15453.0
221.90	3021.0
223.00	5356.0
224.00	50584.0
225.00	12299.0
226.00	1246.0
227.00	20696.0
228.00	2437.0
229.00	4431.0
229.90	540.0
231.00	2091.0
232.00	350.0
232.90	456.0
233.90	1321.0
235.00	1612.0
235.90	675.0
237.00	1603.0
237.80	361.0
238.90	935.0
240.10	486.0
241.00	940.0

Instrument :
BNA_M
ClientSampleId :
DFTPP12

242.00	2663.0
243.00	3083.0
244.00	41136.0
245.00	4994.0
246.00	7814.0
247.00	1578.0
248.00	533.0
248.90	1529.0
249.90	355.0
252.00	309.0
252.90	987.0
254.20	1441.0
255.00	215488.0
256.00	30720.0
257.00	2229.0
258.00	11707.0
258.90	2167.0
259.90	346.0
260.90	454.0
264.10	381.0
265.00	5289.0
266.00	761.0
270.90	1069.0
272.00	584.0
273.00	5842.0
274.00	17664.0
275.00	99624.0
276.00	12420.0
277.00	8894.0
277.90	1243.0
281.00	536.0
282.90	1122.0
283.90	562.0
285.10	1662.0
289.10	305.0
290.00	334.0
290.90	379.0
292.10	316.0
293.00	1850.0
294.00	633.0
295.00	662.0
296.00	28936.0
297.00	4212.0
300.90	326.0
301.90	480.0
303.00	3112.0
304.00	915.0
308.00	544.0
310.10	306.0
314.00	1360.0
315.00	2932.0
316.00	2209.0
321.10	1115.0
322.10	471.0
323.00	8540.0
324.10	1745.0
327.00	1844.0
328.00	940.0
332.10	538.0
333.00	1013.0
334.10	5726.0
335.00	1470.0
336.00	362.0
341.00	982.0
341.90	385.0
346.00	2112.0
347.00	311.0
352.00	2859.0
353.10	1985.0
354.00	3419.0
355.10	599.0
365.00	13152.0
366.00	1796.0
370.00	333.0
371.00	737.0
372.10	4931.0
373.00	1339.0
383.00	1647.0
384.00	396.0
390.00	701.0
390.90	512.0
400.80	312.0
402.00	2042.0
403.00	2710.0
404.10	830.0
421.00	2339.0
422.10	2604.0
423.10	16584.0
424.00	4041.0
424.90	381.0
441.10	47352.0
442.10	301440.0
443.10	58000.0
444.00	5668.0
445.00	435.0

Instrument :
BNA_M
ClientSampleId :
DFTPP12

m/z	Abundance
38.00	828.0
39.10	4491.0
40.00	1247.0
41.00	546.0
42.00	394.0
43.00	1331.0
44.00	11033.0
45.00	1102.0
49.20	664.0
50.10	14975.0
51.00	58384.0
52.10	3187.0
55.00	539.0
55.90	1906.0
57.00	4544.0
59.90	700.0
60.90	571.0
62.00	1061.0
63.00	2965.0
64.00	365.0
65.00	1455.0
68.20	1207.0
69.00	62704.0
70.00	318.0
73.00	1038.0
74.00	6713.0
75.00	10772.0
76.00	2816.0
77.10	70352.0
78.00	4704.0
79.00	5225.0
80.00	3691.0
80.90	5040.0
82.00	1297.0
82.90	1358.0
83.90	414.0
84.90	854.0
86.10	1211.0
87.00	557.0
91.00	1100.0
92.10	1482.0
93.00	8689.0
93.90	574.0
96.00	505.0
98.00	6486.0
99.00	6110.0
99.90	751.0
101.00	3048.0
102.90	1214.0
104.00	1883.0
105.00	1793.0
105.90	584.0
107.00	23136.0
108.00	3324.0
109.10	708.0
110.00	42168.0
111.00	6154.0
111.90	879.0
113.00	396.0
116.00	1761.0
117.00	18448.0
117.90	1217.0
121.10	308.0
121.90	1942.0
123.00	2125.0
123.90	1128.0
125.00	927.0
127.00	87128.0
128.00	6438.0
129.00	33776.0
130.00	2182.0
131.00	654.0
131.80	323.0
133.00	315.0
134.00	1095.0
135.00	3081.0
135.90	998.0
137.00	1639.0
138.00	343.0
140.00	405.0
140.90	3699.0
141.90	1266.0
142.90	1151.0
144.00	305.0
146.00	793.0
147.00	2400.0
148.00	5281.0
149.00	1172.0
151.00	527.0
151.80	601.0
153.00	1476.0

Instrument :
BNA_M
ClientSampleId :
DFTPP12

153.90	922.0
155.00	2740.0
156.00	4019.0
157.10	761.0
158.10	733.0
159.00	760.0
160.00	1527.0
161.00	2208.0
161.80	474.0
164.90	1958.0
166.00	1374.0
167.00	8782.0
168.00	5581.0
169.00	1175.0
170.90	518.0
172.00	868.0
173.00	1276.0
174.00	2056.0
175.00	3764.0
176.10	1034.0
176.90	1850.0
178.10	580.0
179.00	8046.0
180.00	5532.0
181.00	2223.0
181.90	417.0
183.90	753.0
185.00	3953.0
186.00	28448.0
187.00	7910.0
188.00	690.0
188.90	2054.0
190.90	1061.0
192.00	2504.0
193.00	3025.0
194.00	768.0
194.90	458.0
195.90	5843.0
197.10	1272.0
198.00	209856.0
199.00	13662.0
199.90	864.0
201.60	1190.0
202.90	1602.0
204.00	6961.0
205.00	12754.0
206.00	56392.0
207.00	8693.0
208.00	2749.0
209.00	779.0
210.10	850.0
211.00	1899.0
213.00	370.0
214.80	549.0
216.10	1264.0
217.00	14201.0
218.00	2002.0
221.00	9464.0
221.90	2034.0
223.00	3601.0
224.10	31992.0
225.00	7785.0
226.00	843.0
227.00	13248.0
228.00	1914.0
229.00	2875.0
231.00	1239.0
233.90	1023.0
234.90	926.0
235.90	886.0
237.00	1113.0
239.00	564.0
240.90	805.0
242.10	1757.0
242.90	1797.0
244.00	25912.0
245.00	3561.0
246.00	5154.0
247.00	1078.0
248.00	317.0
249.00	1024.0
251.90	366.0
253.00	545.0
254.10	1009.0
255.00	140928.0
256.00	19408.0
257.00	1783.0
258.00	8220.0
259.00	1190.0
264.90	3320.0
265.90	513.0
270.90	547.0
271.90	506.0
273.00	4700.0
274.00	11225.0

Instrument :
BNA_M
ClientSampleId :
DFTPP12

275.00	65256.0
276.00	8834.0
277.00	5732.0
278.00	930.0
281.10	539.0
281.90	342.0
282.90	551.0
284.00	485.0
285.10	824.0
290.90	301.0
292.90	1632.0
294.00	422.0
295.20	476.0
296.00	20216.0
297.00	2455.0
301.90	354.0
303.00	2214.0
304.10	586.0
310.00	351.0
314.00	894.0
315.00	2310.0
316.00	1056.0
321.00	647.0
323.10	6592.0
324.00	1188.0
326.90	1288.0
328.00	575.0
331.90	528.0
333.10	725.0
334.00	4040.0
335.10	1088.0
341.00	793.0
346.00	1575.0
352.00	1782.0
353.00	1378.0
354.00	2471.0
355.00	402.0
365.00	10390.0
366.00	1484.0
371.00	373.0
372.00	3636.0
373.00	897.0
383.00	1304.0
384.10	418.0
390.00	600.0
391.00	358.0
402.00	1356.0
403.10	2295.0
404.00	756.0
421.00	1960.0
422.00	1814.0
423.10	13847.0
424.10	3165.0
441.10	38536.0
442.10	246400.0
443.10	46440.0
444.00	3581.0
445.10	300.0

Instrument :
BNA_M
ClientSampleId :
DFTPP12

m/z	Abundance
39.90	880.0
40.90	325.0
41.90	579.0
43.00	1440.0
44.00	9543.0
45.00	1011.0
59.90	565.0
84.00	366.0
207.00	1452.0
208.00	303.0
281.20	325.0

Instrument :
BNA_M
ClientSampleId :
DFTPP12